Gamification Enhances User Engagement and Task Performance in Prosthetic

Vision Testing

Lily M. Turkstra[1,†,*], Byron A. Johnson[1,†], Arathy Kartha[2,3], Gislin Dagnelie[2], Michael Beyeler[1,4]

[1]Department of Psychological & Brain Sciences, University of California, Santa Barbara, CA, USA

[2]Department of Ophthalmology, Johns Hopkins University School of Medicine, Baltimore, MD, USA

[3]College of Optometry, State University of New York, New York, NY, USA

[4]Department of Computer Science, University of California, Santa Barbara, CA, USA

[†]These authors contributed equally to this work.

*Corresponding author: lturkstra@ucsb.edu

Word count: 5,328

Funding information: Partially supported by the National Eye Institute of the National Institutes of Health

under Award Number R00-EY029329. The content is solely the responsibility of the authors and does not

necessarily represent the official views of the National Institutes of Health.

Author contributions: LMT, BAJ, GD, and MB conceptualized and designed the study. MB developed

the experimental software. LMT, BAJ, and AK conducted data collection. LMT, BAJ, and MB performed

data analysis. LMT, BAJ, and MB wrote the manuscript. All authors reviewed, edited, and approved the

final version of the manuscript.

Commercial relationships: None.

Abstract

Purpose: Visual function testing in retinal prosthesis users relies on repetitive psychophysical tasks that

are cognitively demanding and fatiguing. Gamification may increase engagement, but its effects on

perceptual performance in implanted users remain unclear.

Methods: Three Argus II users completed circle localization and motion direction discrimination in

clinical and gamified versions. Visual stimuli, trial structure, and response requirements were matched

within each participant; gamified versions added scoring, background music, and affectively framed end-

of-trial auditory feedback. Difficulty and response format were calibrated to individual abilities (8AFC

for two participants; 4AFC restricted to cardinal directions for one participant).

Results: Gamification improved accuracy and reduced angular error in localization but did not improve

motion discrimination. Effects were task-dependent and varied across participants, with reduced precision

in the gamified motion task for one user. Participants preferred gamified localization and reported higher

enjoyment and sustained attention; responses to gamified motion were mixed.

Conclusions: Gamification can influence measured performance and user experience in prosthetic vision

testing, but benefits are not universal and depend on task demands and cognitive load, indicating that

engagement can affect outcomes in tests often treated as objective.

Translational relevance: Personalized, engagement-aware gamified tools with adaptive difficulty may

improve the usability and scalability of prosthetic vision assessment and rehabilitation, including at-home

training.

Introduction

Visual neuroprostheses offer a promising approach to restoring functional vision for individuals with profound blindness [1,2]. Outcomes, however, have varied substantially across device generations. Early retinal systems such as Alpha-IMS and Argus II enabled some users to perform above chance on structured laboratory tasks and to demonstrate improvements on select functional vision measures (e.g., obstacle-related and object-sorting tasks) when the system was on compared with off [3–10], but real-world benefit and long-term adoption were highly heterogeneous [11–13]. In contrast, more recent retinal systems, including the subretinal photovoltaic implant PRIMA and emerging suprachoroidal approaches, have reported substantially improved spatial resolution and task performance in clinical trials [14,15], indicating that advances in device design can improve spatial resolution and performance on specific measured tasks in clinical trials.

At the same time, successful outcomes depend not only on technological performance and biotolerability, but also on user engagement, motivation, expectation management, and sustained use. Qualitative work consistently highlights usability burden, training demands, fatigue, and mismatched expectations as major barriers to sustained real-world use, even when device function can be demonstrated in controlled tasks [12,13,16–18]. These considerations underscore a persistent gap in prosthesis development: ensuring that training, habilitation, and rehabilitation protocols are engaging, accessible, and effective.

Most clinical training and assessment for prosthetic vision relies on standardized psychophysical tasks [6,19], including object localization and motion discrimination [3,7], which are used to quantify device function and guide parameter adjustments. However, these tasks can be lengthy and cognitively demanding, placing a burden on participants and limiting engagement. In practice, rehabilitation protocols also vary across sites and devices and often lack structured at-home training tools, making it difficult for users to sustain consistent practice or track progress over time [19,20]. As a result, clinical tests that are intended to provide objective measures of visual function may also be influenced by attention, motivation, and fatigue, complicating interpretation and reducing ecological validity [21–23].

One approach to addressing these challenges is to deliberately increase user motivation and engagement

during training and assessment. Gamification, defined as the integration of game elements into non-game

contexts [24], and related "serious game" approaches [25] can enhance perceptual learning and visuomotor

coordination [26]. Game-based paradigms have been used effectively in health care settings [26–28],

specifically in rehabilitation and accessibility [29–33]. Proposed mechanisms include recruitment of reward

and attention systems that may facilitate learning and retention [34–36]. At the same time, game mechanics

can introduce additional cognitive load or distraction, and user preferences vary, suggesting that

effectiveness may depend on task demands and individual differences [37–39].

Despite this potential, gamification has not been systematically evaluated in the context of retinal

prosthesis testing with implanted users. Related approaches have been successfully applied in adjacent

domains, including training and assessment for cochlear implant users [40], and in studies of simulated

prosthetic vision with sighted participants [41,42]. However, whether gamification alters perceptual

performance or user experience in individuals implanted with a retinal prosthesis remains unknown.

Here, we asked whether gamifying two standard clinical psychophysical tasks, object localization and

motion direction discrimination, influences perceptual performance and engagement in retinal implant

users. We quantified performance using accuracy and response precision, and assessed user experience

using brief surveys and open-ended probes across standard and gamified task versions. We further

explored individual differences in responsiveness to gamification and examined whether adaptive

difficulty could help sustain engagement across repeated testing. Understanding these factors is essential

for designing personalized training protocols that maximize both usability and rehabilitation outcomes in

visual prosthesis users.

Methods

Study Participants

This study involved three completely blind participants (one female and two male) with severe retinitis pigmentosa (RP), ranging from 31-90 years at the time of testing (summarized in **Table 1**). Each participant had been chronically implanted with the Argus II retinal prosthesis system prior to enrollment in the present study, following its commercial release and completion of the interventional feasibility trial (clinicaltrials.gov: NCT00407602). The present work involved behavioral task assessments only and did not include surgical intervention or changes to clinical care.

Participant	Age	Gender	Years implanted	Device usage	Prior psychophysics experience	Prior gaming experience
1	51-70	female	6-10	infrequent; primarily during research studies	extensive; frequent studies at Wilmer Eye Institute	none
2	51-70	male	6-10	rarely; occasional navigation	extensive; frequent studies at Wilmer Eye Institute	infrequent
3	31-50	male	1-5	regularly; at-home use for perceptual tasks	limited; only during follow-ups with Second Sight	extensive

Table 1. Demographic and experiential characteristics of our study participants. Because Argus II users are rare and identifiable, ages and implantation durations are reported in ranges to protect privacy.

All psychophysical experiments were conducted in the Ultra Low Vision Laboratory (ULV Lab) at the Lions Vision Research and Rehabilitation Center of the Johns Hopkins Medical Institutions in Baltimore, MD, USA. The study was approved by a centralized Institutional Review Board (WCG IRB) with local institutional oversight at Johns Hopkins University, and was conducted in accordance with the tenets of the Declaration of Helsinki. Participants provided informed consent after receiving an explanation of the study's nature and potential risks. The consent process was made accessible by reading aloud the form's text, providing a signature guide to participants, and sharing an electronic version ahead of time [43]. Medical and research staff were available to answer any participant questions. Participants were

compensated at a rate of $20 per hour, including travel time, and were allowed unlimited breaks with the

option to discontinue at any time.

Participants' experience with the Argus II varied. Participants 1 and 2 had prior involvement in

psychophysics studies at the ULV Lab. Participant 1 initially used the device for navigation but now

primarily engages with it when participating in research studies. Participant 2 previously relied on the

implant for navigation and work but has since reduced usage, citing limited benefits and functionality

without upgrades. In contrast, Participant 3, despite minimal formal training, reported frequent at-home

self-guided perceptual practice with the device.

Retinal Implant

The Argus II system comprises a 6×10 array of platinum disc electrodes, each 225 μm in diameter, spaced

575 μm apart, and subtending approximately 12×20 degrees of visual angle [44,45] (**Figure 1**). The external

head-mounted system includes a camera and wireless transmitter mounted on a pair of glasses. The

camera captures video input, which a video processing unit (VPU) converts into electrical stimulation

patterns for the retinal implant using predefined image processing algorithms.

Figure 1. Overview of the Argus II system (reused under CC-BY-ND from [46]). A) Implanted components include a hermetically

sealed electronics enclosure, a receiving antenna secured to the eye with a scleral band and sutures, and a 60-electrode array tacked

over the macula. B) External components consist of a glasses-mounted camera connected via cable to a belt- or shoulder-worn

video processing unit (VPU).

Tasks & Gamification

Two psychophysical tasks were selected for gamification: circle localization [4,7] and motion direction

discrimination [3,47] (**Figure 2**), both widely used to diagnostically assess spatial perception and motion

processing in Argus II users. While prior studies have demonstrated that prosthesis users perform above

chance in these tasks, their accuracy remains highly variable across users and significantly lower than that

of sighted individuals performing equivalent tasks under simulated prosthetic vision [5,48]. Previous work

employing these tasks with Argus II and other retinal implant users reports a marked variability in

performance between the two, with motion discrimination consistently more difficult than object

localization [48]. These tasks, which are commonly used, differ in their reported difficulties while eliciting

performance differences above chance, are therefore well-suited for evaluating the potential benefits of

gamification in prosthesis users.

Gamification was implemented by embedding task-relevant game mechanics into each paradigm while

keeping visual stimuli, response requirements, and trial structure identical to the clinical versions. The

circle localization task was reframed as a balloon-popping game, where participants had to correctly

identify the balloon's location to "pop" it by indicating its location with a game controller, accompanied

by balloon inflation sounds, a popping effect for correct responses, and deflation for incorrect responses.

Similarly, the motion direction task was transformed into a fishing game, where correct responses were

rewarded with a reeling-in sound to indicate catching a fish, while incorrect responses played a bubbling

sound to simulate the fish escaping.

Additional game elements were introduced to enhance engagement, including point-based incentives,

background music, and affectively framed auditory feedback. Importantly, feedback timing and

informational content were held constant across conditions: in both the standard and gamified tasks,

participants received auditory feedback only at the end of each trial, indicating whether the response was

correct, near-correct, or incorrect. No continuous or real-time performance feedback was provided during

stimulus presentation or response execution.

Circle Localization	Motion Discrimination

 |

 |

Figure 2. Schematic of the two experimental tasks. (Left) In the circle localization task, the target appeared in one of eight possible locations, and participants indicated its position using a game controller. (Right) In the motion discrimination task, the target's drift direction was dictated by its starting position, always traveling in a fixed direction across the screen (e.g., a target appearing in the top-left position would always move diagonally down-right). Visual stimuli were identical between clinical and gamified conditions; the gamified condition added auditory and scoring elements (e.g., themed sounds/music and point feedback), while stimulus presentation and response requirements were otherwise unchanged. Dashed circles indicate possible target locations for illustration only; they were not visible during the experiment. Target and screen sizes are not to scale (see Table 3 for actual sizes).

In the gamified condition, this end-of-trial feedback was delivered using a larger and more varied set of verbal responses (**Table 2**) designed to convey encouragement and engagement (e.g., "You're getting good at this!"), whereas the standard condition used a small, neutral set of task-oriented feedback phrases (e.g., "Correct" or "Incorrect"). Participants could request their cumulative score at any point during a trial block or session, but scores were not displayed automatically. Scores were computed from response accuracy and angular error. Background music was present only in the gamified condition and differed by task, with a carnival theme for localization and orchestral music for motion discrimination. The goal of these modifications was not to isolate the effects of individual game mechanics, but to assess whether a realistic combination of commonly used gamification elements could increase engagement and motivation when applied to standard visual function testing.

Hit	Near Miss	Miss
10 points!	Almost!	Not quite.
Ding ding ding, 10 points!	Nearly there!	Keep trying.
That was perfect!	Close, but no cigar!	Let's try that again.
You're getting good at this!	Just missed it!	No points this time.
You're unstoppable!	That was close!	Maybe next time.
Easy, isn't it?	Missed it by a hair!	Swing and a miss.
Did you secretly practice this?	You almost had it!	Well, at least you're learning.
Who knew psychophysics could be fun?	That was really close!	Nope, but you're still a winner to me.

Table 2. Examples of auditory feedback phrases used in the gamified condition, categorized by response accuracy. Hits, near misses ($\pm 45°$), and misses ($> |45°|$) triggered different verbal responses, sampled randomly from a database of approximately 25 phrases per category. Phrases were sampled with equal probability throughout the session and delivered only after the response on each trial. Feedback timing and categorical information (hit, near miss, miss) were held constant across clinical and gamified conditions; only the phrasing differed.

Experimental Setup

All testing was conducted in a darkened room in the ULV Lab to minimize ambient light interference and optimize contrast for prosthesis users. Participants were seated in front of a 69.6 cm × 39 cm television monitor (Samsung LN-S3251D) at a viewing distance ranging from 42 cm to 59.7 cm, depending on individual comfort and ability to perceive stimuli (Participant 1: 42 cm, Participant 2: 48 cm, Participant 3: 59.7 cm). Stimuli were generated using SimpleXR [49], an open-source Unity-based toolbox for virtual and augmented reality (https://github.com/simpleomnia/sxr). The monitor displayed all stimuli, and responses were recorded using a conventional video game controller. Participants (all three right-handed) used their left thumb to indicate stimulus location or direction via the controller's directional pad and confirmed their responses by pressing a separate button with their right index finger.

To provide a consistent frame of reference, tactile markers (clear tape) were placed at the center and along

the edges of the monitor, and did not interfere with stimulus perception. At the beginning of each block,

participants were instructed to reorient themselves by feeling the tactile markers. While participants were

free to move their heads to scan for stimuli, they were encouraged to avoid excessive touching of the

screen to prevent reliance on tactile cues. The monitor height was adjusted for Participant 3 to

accommodate their preferred interaction style.

All experiments were conducted with the Argus II device turned on and in standard video mode. The

video processing unit (VPU) applied the system's default contrast enhancement settings without

additional image processing.

Block Structure & Difficulty Progression

To systematically vary difficulty, the circle localization task used progressively smaller targets across

levels, while the motion direction task increased stimulus speed. Since Argus II users have highly variable

perceptual abilities, absolute stimulus parameters differed across participants. The easiest level for each

condition was calibrated individually so that participants could achieve approximately 80% accuracy,

ensuring that difficulty progression was meaningful and engaging, while keeping overall task demands

comparable.

In the circle localization task, difficulty escalated by progressively shrinking the target; for motion

direction discrimination, difficulty increased with stimulus speed. Raw values for each participant's

stimulus parameters are provided in **Table 3**.

In the gamified condition, these difficulty changes were framed as advancing through game levels,

reinforcing a sense of progression. Verbal feedback emphasized challenge and achievement, transforming

what would otherwise be passive difficulty increments into a more immersive and engaging experience.

Participant	Stimulus Size (Circle Localization)			Stimulus Speed (Motion Discrimination)		
	Level 1	**Level 2**	**Level 3**	**Level 1**	**Level 2**	**Level 3**
1	8.91°	5.35°	1.78°	0.7°/s	0.8°/s	0.9°/s
2	17.6°	12.6°	7.60°	0.9°/s	1.1°/s	1.3°/s
3	13.3°	8.86°	4.44°	1.2°/s	1.5°/s	1.8°/s

Table 3. Participant-specific stimulus parameters across three difficulty levels for circle localization and motion direction

discrimination (degrees visual angle, dva). In localization, target size decreased with increasing level. In motion discrimination,

stimulus speed increased with level; target size was held constant within each participant across motion levels (set from that

participant's Level 1 localization size) to emphasize motion discrimination rather than localization difficulty. Parameters were

individualized to reduce the likelihood that condition effects were driven by floor or ceiling performance or mismatched task

demands, rather than by task framing.

Training & Testing Procedure

The full procedure timeline is outlined in **Figure 3**. Before testing, participants completed a structured

training phase to ensure they understood the task and could reliably use the game controller for responses.

After familiarizing participants with the desktop setup, training began with a simplified version of each

task in which only two response options were available (e.g., left vs. right for localization, upward vs.

downward motion for discrimination). This allowed participants to acclimate to the task structure and

response mapping before progressing to the full task.

Initially, responses were provided verbally, and once participants demonstrated consistent understanding

of the task requirements, they transitioned to using the game controller for all responses. Training

continued until both the participant and the experimenter were confident that the task was understood and

that responses were reliable.

Figure 3. Experimental procedure. Participants first completed a training phase to familiarize themselves with the tasks and response system. Testing alternated between clinical and gamified conditions, with localization and motion discrimination tasks presented in separate blocks. Each task comprised three difficulty levels per condition, with eight trials per block. The full experiment was repeated three times across two sessions (288 trials per participant). The response format was calibrated to individual perceptual abilities (8AFC for Participants 1–2; 4AFC, cardinal directions only, for Participant 3), while trial counts and stimulus presentations were matched between clinical and gamified conditions. Surveys were administered after each condition to assess engagement and experience.

Following successful completion of the simplified training, participants advanced to the full experimental version of each task. To accommodate individual differences in perceptual abilities, response format was calibrated on a per-participant basis. Participants 1 and 2 completed an eight-alternative forced-choice (8AFC) version of the task, whereas Participant 3 completed a four-alternative forced-choice (4AFC) version restricted to the cardinal directions. In the 4AFC condition, only the four corresponding target locations or motion directions were presented, and only those labels were available as valid responses; oblique directions were neither displayed nor analyzed.

Importantly, the number of trials, stimulus presentations, block structure, and difficulty progression were held constant across participants and were matched between the clinical and gamified conditions, ensuring that differences in performance reflected task framing rather than stimulus exposure or trial count.

Statistical Analyses

All analyses were performed in R (v4.5.2). Given the small sample size ($N = 3$) and repeated-measures

structure, we modeled outcomes at the trial level using mixed-effects models and emphasized effect sizes

and uncertainty intervals.

Trial-level accuracy was analyzed using a Bayesian generalized linear mixed model (Bernoulli likelihood;

logit link; `brms`). Fixed effects included Condition (Clinical vs. Gamified), Task (Localization vs.

Motion), their interaction, and Session, with a random intercept for Participant. To account for the

differing number of alternatives across participants (4AFC vs. 8AFC), we included an offset term equal to

logit(1/m), where *m* denotes the number of alternatives on each trial. This adjusts the baseline intercept

such that a linear predictor of zero corresponds to chance-level performance (1/*m*), allowing the model

estimates to represent deviations from chance in log-odds units. We report posterior estimates with 95%

credible intervals and considered effects credible when the 95% credible interval excluded zero.

Angular error magnitude (absolute signed error, in degrees) was analyzed using a generalized linear

mixed model with a Gamma distribution and log link (`glmmTMB`), which is appropriate for non-negative,

right-skewed outcomes. The fixed- and random-effects structure matched the accuracy model.

To verify that group-level effects were not driven by any single participant and to provide distribution-

free within-subject confirmation, we additionally performed participant-level permutation tests (10,000

label permutations) comparing Gamified versus Clinical performance within each task.

Results

Individual Differences in Gamification Effects

Given the substantial variability in prosthetic vision outcomes, we first examined the impact of gamification at the individual level using non-parametric permutation tests (10,000 resamples). The effectiveness of gamification varied considerably across participants, revealing distinct response profiles in both task accuracy (**Figure 4A-C**) and perceptual precision (**Figure 4D-F**).

Participant 1 demonstrated the broadest benefit from the gamified interface, showing consistent improvements across both metrics. In terms of accuracy, they showed a significant improvement in the motion discrimination task ($p = .041$) and a marginally significant improvement in the localization task ($p = .056$). Crucially, this user also exhibited significant gains in perceptual precision. Permutation tests on mean absolute angular error revealed that gamification significantly reduced error magnitudes in both localization ($p = .016$) and motion discrimination ($p = .028$), reducing mean error by approximately 17° and 21°, respectively. This suggests that for high-performing users, gamification can enhance both the reliability of target selection and the spatial precision of their responses.

Participant 3 exhibited a task-specific benefit limited to static localization. Gamification drove a significant improvement in localization accuracy ($p = .029$) and a trend toward improved precision ($p = .085$), with a mean error reduction of 19°. However, these benefits did not extend to the dynamic motion task, where both accuracy ($p = .59$) and precision ($p = .72$) remained comparable to clinical baseline performance.

In contrast, Participant 2 illustrated the potential cognitive costs of gamification. While their accuracy remained stable across conditions ($p > .1$), their precision in the motion task significantly worsened in the gamified condition ($p = .014$), with mean error increasing by approximately 24° compared to the clinical baseline. This significant decline suggests that for some users, the additional auditory and cognitive load of the gamified motion task (i.e., the "fishing" metaphor) may have acted as a distracter rather than a motivator, interfering with the processing of the visual motion signal.

Figure 4. Individual participant performance across clinical and gamified conditions. Panels A–C show trial accuracy (%) for

Participants 1–3, and panels D–F show absolute angular error (degrees) for the same participants. Each panel compares the four

within-subject conditions: localization (clinical), localization (gamified), motion discrimination (clinical), and motion

discrimination (gamified). Error bars indicate 95% confidence intervals. Asterisks denote participant-level permutation test

results for the gamified vs. clinical contrast within each task ($p < .05$). Overall, gamification was associated with higher

localization accuracy and lower localization error in most participants, whereas effects on motion discrimination were smaller

and more heterogeneous across individuals.

Gamification Improves Accuracy in the Localization Task

To assess whether these individual benefits translated to a consistent group-level effect, we employed a

Bayesian Generalized Linear Mixed Model (GLMM). Consistent with the individual findings, the model

revealed a significant main effect of gamification on accuracy in the localization task ($\beta = 0.51$, 95% CI

[0.14, 0.89]). Post-hoc contrasts indicated that gamification increased the odds of a correct response by

approximately 1.68 times compared to the clinical condition.

However, this benefit was task-dependent at the group level. A significant Condition x Task interaction

(β = -0.52, 95% CI [-1.06, 0.02]) indicated that the accuracy gains observed in localization did not

consistently extend to the motion discrimination task. For motion discrimination, group-level

performance remained comparable to the clinical baseline (OR $\approx$ 1.01, 95% CI [0.64, 1.46]). This

suggests that while gamification reliably boosts performance in spatial localization, its ability to

overcome the sensory challenges of motion processing is highly user-dependent, as evidenced by the

contrasting outcomes of Participant 1 and Participant 2.

Precision and Error Magnitude

We further analyzed perceptual precision at the group level using a Gamma GLMM on absolute angular

error. Mirroring the accuracy results, we observed a trend toward improved precision in the gamified

localization task (p = .077). On average, angular errors in the clinical condition were 1.34 times larger

than in the gamified condition. No significant difference in precision was observed for the motion task at

the group level (p = .83), reflecting the high inter-subject variability where one participant improved

significantly while another worsened.

Absence of Learning Effects

To ensure that the observed improvements were due to the gamified interface rather than task familiarity

or practice, both the accuracy (Bayesian) and precision (Gamma) models controlled for session effects.

No significant learning effects were observed for accuracy (Session 2: β = -0.16; Session 3: β $\approx$ 0.00) or

precision (Session 2: p = 0.51; Session 3: p = 0.89). This confirms that performance changes (both

positive and negative) were driven by the task structure itself rather than sequential learning over the

course of the experiment.

Effect of Gamification on Enjoyment and Perceived Difficulty

To complement the performance analysis, participants rated their enjoyment (1-10) and ease of use, with higher values indicating greater enjoyment and greater ease (**Table 4**). Additional survey items assessed sustained attention and motivation (Yes/No/Neutral), as well as overall task preference.

Participant	Task	Δ Enjoyment ↑ (Gamified - Clinical)	Δ Difficulty ↓ (Gamified - Clinical)
1	Localization	+2	-2
1	Motion	+5	-4
2	Localization	+1	0
2	Motion	-2	+1
3	Localization	+4	-2
3	Motion	-1	0

Table 4: Changes in enjoyment (↑higher is better) and perceived difficulty (↓lower is better) between gamified and clinical conditions. Ratings were originally provided on a 1–10 scale, with higher values indicating greater enjoyment or greater difficulty. Reported values reflect the difference (Gamified – Clinical) for each participant and task, where positive values indicate increased enjoyment and difficulty in the gamified condition.

On average, gamified tasks were rated 1.5 points higher in enjoyment and 1.17 points lower in difficulty compared to their clinical counterparts, indicating that gamification not only increased engagement but also made tasks feel easier. All participants preferred the gamified localization task, while opinions on the motion task were mixed, suggesting that the impact of gamification may depend on task type. As Participant 3 remarked:

"I was having fun because I was doing better - or I was doing better because I was having fun?"

Survey responses provided insight into engagement and motivation across conditions. All participants reported being motivated in both the clinical and gamified tasks, except for Participant 1, who rated motivation as neutral in clinical localization. Similarly, all participants indicated that the tasks sustained

their attention, except for Participant 1, who reported that clinical localization failed to do so. While

gamification did not universally increase motivation, it helped maintain engagement, particularly in tasks

that might otherwise feel monotonous.

Reactions to the motion task were mixed. When asked whether gamification would encourage practice,

Participant 2 responded:

> *"Yeah, probably … not so much the fishing game, the other game was fine. The fishing game is*
>
> *kind of like teasing you… 'Better luck next time' or 'You missed it.' One way or the other, a beep*
>
> *or bang or a pop makes it a little more interesting, like a pinball machine."*

Notably, both tasks used the same auditory feedback system, suggesting differences in perception rather

than content.

Competitiveness also shaped participant experiences. Participant 3 expressed frustration:

> *"You would think I would know how to do this by now. Come on, I've had the chip for five*
>
> *years!"*

Yet, he also saw the challenge as motivating:

> *"I've always been competitive, so the better I do, the happier I am playing it. It frustrates me*
>
> *when I don't get the high score I should. It can drive you nuts to not catch the fish, but at the*
>
> *same time, just by doing the right head movements, I do find benefit from catching fish and*
>
> *popping balloons."*

In sum, gamification provided both encouragement and, at times, frustration, especially when

expectations of improvement were unmet.

Discussion

Vision rehabilitation for prosthesis users remains a significant challenge, and standard clinical assessments may not fully capture real-world functional abilities. Here, we tested whether gamification can influence perceptual performance and user experience during prosthetic vision testing. Gamification improved accuracy and reduced angular error in the localization task, consistent with the possibility that engagement influenced measured precision. In contrast, gamification did not improve, and may have hindered, motion direction discrimination, where errors remained higher. Together, these results indicate that gamification can enhance performance in some contexts but is not universally beneficial. No learning effects were observed across trials or sessions, supporting the interpretation that performance differences reflected task framing rather than short-term practice.

Gamification Enhances Performance, But Its Effects Are Task-Dependent

All participants preferred the gamified localization task over its clinical counterpart, reporting higher enjoyment and same or lower perceived difficulty (**Table 4**). However, the gamified motion task remained challenging, consistent with prior findings that motion discrimination is inherently more difficult and requires more training [50]. This suggests that while gamification can increase engagement, its impact on performance depends on task complexity.

Crucially, gamification did not just improve subjective experience, but enhanced objective performance in localization (**Figures 4–5**). Therefore, traditional acuity and function tests may not fully capture attentional and cognitive factors influencing prosthesis use. Prior research suggests that engaging tasks activate attentional networks and promote neuroplasticity [34,36], which may explain the observed improvements in localization. However, the absence of gains in the motion task suggests that gamification alone may not overcome task-intrinsic difficulty and, in some cases, may impose additional cognitive load. This pattern was also evident at the individual level, reinforcing that mechanics that support engagement for one user may distract another.

Tailoring Gamified Training to Individual Needs

Gamification's effectiveness varied across participants (**Figure 4**), reinforcing that one-size-fits-all

approaches may not be optimal for clinical applications. Individual differences (e.g., prior training, head-

scanning techniques, and baseline perceptual abilities) play a critical role in task performance [3].

Participants 1 and 2 had extensive experience with head scanning, while Participant 3 had little prior

training and required additional guidance, which may have influenced their performance. Future gamified

tasks should adapt to individual perceptual strategies, ensuring that engagement-enhancing elements do

not introduce unnecessary distractions [33]. In practice, this suggests that future at-home tools should

combine personalization with safeguards that prevent gamification from becoming a barrier to task

completion.

Limitations and Future Directions

While this study highlights the potential of gamification for perceptual testing and training, several

limitations should be considered. First, the small sample size ($N = 3$) limits generalizability, reflecting the

rarity of Argus II users available for research. Although a case-study approach provides valuable insight

into individual variability and the need for customization, larger cohorts will be necessary to

systematically examine factors such as age, prior gaming experience, and training history. Future work

should also include users of other retinal and cortical prostheses to assess whether these findings extend

beyond the Argus II system.

Second, the study was not designed to assess longer-term learning or neuroplasticity. Testing occurred

across two sessions with a limited training dose, which constrains the ability to detect practice effects or

durable changes in performance, fatigue tolerance, or attention. Accordingly, the absence of learning

effects should not be interpreted as evidence that learning does not occur in retinal implant users. Longer-

term studies incorporating extended training, retention tests, and transfer to more functional tasks will be

required to determine whether gamification supports sustained improvements beyond immediate task

performance.

Finally, gamification elements were implemented in combination to reflect realistic clinical use, rather

than to isolate individual mechanics. As a result, the contribution of specific elements such as scoring,

music, or affective feedback cannot be disentangled. While the gamified localization task was

consistently preferred, the motion discrimination task received mixed feedback, suggesting that some

design choices may introduce distraction or cognitive load in already demanding tasks. Future studies

should therefore systematically evaluate individual game mechanics and explicitly measure cognitive load

and fatigue alongside performance.

The Potential for Personalized, At-Home Gamified Training

Despite these limitations, gamified perceptual tasks could provide a scalable option for at-home vision

training, particularly for users who lack access to in-clinic rehabilitation. However, training design must

carefully manage how users experience success and failure. Prior qualitative work suggests that failures

are often internalized by users, whereas successes are frequently attributed to the device or system rather

than to the user's own skill [12,13]. This asymmetry underscores the importance of adaptive difficulty and

feedback design that preserves a high likelihood of success while still providing meaningful challenge.

Training paradigms that emphasize frequent, achievable successes may help sustain engagement and

reinforce user confidence, while reducing discouragement that can arise from repeated failure in

demanding tasks.

A personalized gamified system could support rehabilitation by dynamically adjusting difficulty based on

a user's scanning behavior, skill level, and perceptual strategies, thereby increasing task demands

gradually without becoming overwhelming. It could also enable self-paced practice that preserves user

agency while remaining clinically grounded through structured benchmarks defined collaboratively with

clinicians and users, helping align expectations for progress. Finally, it could increase accessibility by

providing consvistent training opportunities for users who cannot regularly attend specialized

rehabilitation clinics or who benefit from customized motivational supports to sustain adherence.

Beyond engagement, gamification may serve as a practical vehicle for delivering sufficient training dose,

which is a prerequisite for perceptual learning in many visual tasks. More broadly, models of attention

and reward suggest that reinforcement mechanisms could be leveraged to support sustained participation

and learning [34–36], although these longer-term effects remain to be tested directly in implanted users [22].

Future work should therefore evaluate extended at-home deployment with retention and transfer measures

to determine whether improved engagement translates into durable functional gains.

Taken together, our findings suggest that gamification can meaningfully alter both user experience and

measured performance in standard prosthetic vision tasks. Incorporating engagement-aware, adaptive task

designs into clinical testing and rehabilitation may improve the reliability of visual function assessment

and help more users access effective, sustainable training.

References

1. Weiland JD, Walston ST, Humayun MS. Electrical Stimulation of the Retina to Produce Artificial Vision. Annu Rev Vis Sci. 2016;2(1):273–94.

2. Fernandez E. Development of visual Neuroprostheses: trends and challenges. Bioelectron Med. 2018 Aug 13;4(1):12.

3. Dorn JD, Ahuja AK, Caspi A, da Cruz L, Dagnelie G, Sahel JA, et al. The Detection of Motion by Blind Subjects With the Epiretinal 60-Electrode (Argus II) Retinal Prosthesis. Jama Ophthalmol. 2013 Feb;131(2):183–9.

4. Ahuja AK, Dorn JD, Caspi A, McMahon MJ, Dagnelie G, Dacruz L, et al. Blind subjects implanted with the Argus II retinal prosthesis are able to improve performance in a spatial-motor task. Br J Ophthalmol. 2011 Apr;95(4):539–43.

5. Schaffrath K, Schellhase H, Walter P, Augustin A, Chizzolini M, Kirchhof B, et al. One-Year Safety and Performance Assessment of the Argus II Retinal Prosthesis: A Postapproval Study. JAMA Ophthalmol. 2019 Aug 1;137(8):896–902.

6. Geruschat DR, Richards TP, Arditi A, da Cruz L, Dagnelie G, Dorn JD, et al. An analysis of observer-rated functional vision in patients implanted with the Argus II Retinal Prosthesis System at three years. Clin Exp Optom. 2016 May;99(3):227–32.

7. McMahon MJ, Dorn JD, Ahuja AK, Caspi A, Filley E, Dagnelie G, et al. The Argus II Retinal Prosthesis Enables Blind Subjects to Localize Objects. Invest Ophthalmol Vis Sci. 2009 Apr 28;50(13):4589.

8. Stingl K, Bartz-Schmidt KU, Besch D, Braun A, Bruckmann A, Gekeler F, et al. Artificial vision with wirelessly powered subretinal electronic implant alpha-IMS. Proc R Soc B Biol Sci. 2013 Apr 22;280(1757):20130077.

9. Zrenner E, Bartz-Schmidt KU, Chee C, Gekeler F, Jackson TL, MacLaren R, et al. Visual outcome in 26 blind retinitis pigmentosa patients after implantation of subretinal Alpha-IMS devices. Invest Ophthalmol Vis Sci. 2014 Apr 30;55(13):5967–5967.

10. Beyeler M, Nanduri D, Weiland JD, Rokem A, Boynton GM, Fine I. A model of ganglion axon pathways accounts for percepts elicited by retinal implants. Sci Rep. 2019 June 24;9(1):1–16.

11. Ghodasra DH, Chen A, Arevalo JF, Birch DG, Branham K, Coley B, et al. Worldwide Argus II implantation: recommendations to optimize patient outcomes. BMC Ophthalmol. 2016 May 6;16(1):52.

12. Nadolskis L, Turkstra LM, Larnyo E, Beyeler M. Aligning Visual Prosthetic Development With Implantee Needs. Transl Vis Sci Technol. 2024 Nov 21;13(11):28.

13. Erickson-Davis C, Korzybska H. What do blind people "see" with retinal prostheses? Observations and qualitative reports of epiretinal implant users. PLOS ONE. 2021 Feb 10;16(2):e0229189.

14. Karapanos L, Abbott CJ, Ayton LN, Kolic M, McGuinness MB, Baglin EK, et al. Functional Vision in the Real-World Environment With a Second-Generation (44-Channel) Suprachoroidal Retinal

Prosthesis. Transl Vis Sci Technol. 2021 Aug 12;10(10):7–7.

15. Holz FG, Mer YL, Muqit MMK, Hattenbach LO, Cusumano A, Grisanti S, et al. Subretinal
Photovoltaic Implant to Restore Vision in Geographic Atrophy Due to AMD. N Engl J Med
[Internet]. 2025 Oct 20 [cited 2025 Oct 20];0(0). Available from:
https://www.nejm.org/doi/full/10.1056/NEJMoa2501396

16. Espinosa M, Nathan-Roberts D. Understanding Prosthetic Abandonment. Proc Hum Factors Ergon
Soc Annu Meet. 2019 Nov 1;63(1):1644–8.

17. Fernández E, Alfaro A, González-López P. Toward Long-Term Communication With the Brain in
the Blind by Intracortical Stimulation: Challenges and Future Prospects. Front Neurosci [Internet].
2020 Aug 11 [cited 2025 Mar 13];14. Available from:
https://www.frontiersin.org/journals/neuroscience/articles/10.3389/fnins.2020.00681/full

18. Turkstra LM, Bhatia T, Van Os A, Beyeler M. Assistive technology use in domestic activities by
people who are blind. Sci Rep. 2025 Mar 3;15(1):7486.

19. Ayton LN, Rizzo JF, Bailey IL, Colenbrander A, Dagnelie G, Geruschat DR, et al. Harmonization
of Outcomes and Vision Endpoints in Vision Restoration Trials: Recommendations from the
International HOVER Taskforce. Transl Vis Sci Technol. 2020 July 1;9(8):25–25.

20. Dagnelie G. Retinal Prostheses: Functional Outcomes and Visual Rehabilitation. In: Humayun MS,
Olmos de Koo LC, editors. Retinal Prosthesis: A Clinical Guide to Successful Implementation
[Internet]. Cham: Springer International Publishing; 2018 [cited 2020 May 15]. p. 91–104.
(Essentials in Ophthalmology). Available from: https://doi.org/10.1007/978-3-319-67260-1_6

21. Peli E. Testing Vision Is Not Testing For Vision. Transl Vis Sci Technol. 2020 Dec 1;9(13):32–32.

22. Beyeler M, Rokem A, Boynton GM, Fine I. Learning to see again: biological constraints on cortical
plasticity and the implications for sight restoration technologies. J Neural Eng. 2017 June
14;14(5):051003.

23. Meier M, Martarelli CS, Wolff W. Is boredom a source of noise and/or a confound in behavioral
science research? Humanit Soc Sci Commun. 2024 Mar 6;11(1):368.

24. Deterding S, Dixon D, Khaled R, Nacke L. From game design elements to gamefulness: defining
"gamification." In: Proceedings of the 15th International Academic MindTrek Conference:
Envisioning Future Media Environments [Internet]. Tampere Finland: ACM; 2011 [cited 2023 Nov
1]. p. 9–15. Available from: https://dl.acm.org/doi/10.1145/2181037.2181040

25. Abt CC. Serious Games. University Press of America; 1987. 200 p.

26. Fleming T, Sutcliffe K, Lucassen M, Pine R, Donkin L. Serious Games and Gamification in Clinical
Psychology. In 2020.

27. Warsinsky S, Schmidt-Kraepelin M, Rank S, Thiebes S, Sunyaev A. Conceptual Ambiguity
Surrounding Gamification and Serious Games in Health Care: Literature Review and Development
of Game-Based Intervention Reporting Guidelines (GAMING). J Med Internet Res. 2021 Sept
10;23(9):e30390.

28. Damaševičius R, Maskeliūnas R, Blažauskas T. Serious Games and Gamification in Healthcare: A
Meta-Review. Information. 2023 Feb;14(2):105.

29. Maidenbaum S, Abboud S, Amedi A. Sensory substitution: Closing the gap between basic research
and widespread practical visual rehabilitation. Neurosci Biobehav Rev. 2014 Apr 1;41:3–15.

30. Castro JB, Ramanathan A, Chennubhotla CS. Categorical dimensions of human odor descriptor
space revealed by non-negative matrix factorization. PLoS One. 2013;8(9):e73289.

31. Tepoxtecatl A, Yang C, Sehaumpai M, Seiple WH, Zhu Z. MAC-You-Vision: A Progressive
Training Application for Patients with Age-Related Macular Degeneration. In: 2023 IEEE MIT
Undergraduate Research Technology Conference (URTC) [Internet]. 2023 [cited 2025 May 13]. p.
1–5. Available from: https://ieeexplore.ieee.org/document/10534931

32. E. Contreras V, Gómez G, A. Navarro-Newball A. Towards the Gamification of Assistive
Technology for Professionals with Severe Impairments. In: 2019 International Conference on
Virtual Reality and Visualization (ICVRV) [Internet]. 2019 [cited 2025 Mar 11]. p. 176–9.
Available from: https://ieeexplore.ieee.org/document/9212977

33. Stiegler A, Zimmermann G. Gamification and Accessibility. In: Zhou J, Salvendy G, editors.
Human Aspects of IT for the Aged Population Design for Aging. Cham: Springer International
Publishing; 2015. p. 145–54.

34. Koepp MJ, Gunn RN, Lawrence AD, Cunningham VJ, Dagher A, Jones T, et al. Evidence for
striatal dopamine release during a video game. Nature. 1998 May;393(6682):266–8.

35. Bavelier D, Levi DM, Li RW, Dan Y, Hensch TK. Removing brakes on adult brain plasticity: from
molecular to behavioral interventions. J Neurosci Off J Soc Neurosci. 2010 Nov 10;30(45):14964–
71.

36. Rokem A, Silver MA. The benefits of cholinergic enhancement during perceptual learning are long-
lasting. Front Comput Neurosci. 2013 May 29;7:66.

37. Cugelman B. Gamification: What It Is and Why It Matters to Digital Health Behavior Change
Developers. JMIR Serious Games. 2013 Dec 12;1(1):e3139.

38. Furini M, Mirri S, Montangero M. Gamification and Accessibility. In: 2019 16th IEEE Annual
Consumer Communications & Networking Conference (CCNC) [Internet]. 2019 [cited 2025 Mar
11]. p. 1–5. Available from: https://ieeexplore.ieee.org/abstract/document/8651750

39. Katz B, Jaeggi S, Buschkuehl M, Stegman A, Shah P. Differential effect of motivational features on
training improvements in school-based cognitive training. Front Hum Neurosci [Internet]. 2014
[cited 2023 July 21];8. Available from:
https://www.frontiersin.org/articles/10.3389/fnhum.2014.00242

40. Lasickas T, Siim Andersen J, Serafin S, Vatti M. Cochlea: Gamifying Ear Training for Cochlear
Implant Users [Internet]. Zenodo; 2021 Nov 24 [cited 2025 Mar 11]. Available from:
https://zenodo.org/records/5724458

41. Esquenazi R, Meier K, Beyeler M, Boynton G, Fine I. Video game training improves learning of
abnormal on- off- cell population responses in sighted individuals. J Vis. 2022 Dec 5;22(14):3638.

42. Esquenazi RB, Meier K, Beyeler M, Wright D, Boynton GM, Fine I. Perceptual learning of
prosthetic vision using video game training. J Vis. 2025 Oct 7;25(12):12.

43. Wittich W, Boie NR, Jaiswal A. Methodological Approaches to Obtaining Informed Consent when
Conducting Research With Individuals With Deafblindness. Int J Qual Methods. 2023 Oct
1;22:16094069231205176.

44. Luo YH, da Cruz L. The Argus((R)) II Retinal Prosthesis System. Prog Retin Eye Res. 2016
Jan;50:89–107.

45. Yue L, Wuyyuru V, Gonzalez-Calle A, Dorn JD, Humayun MS. Retina–electrode interface
properties and vision restoration by two generations of retinal prostheses in one patient—one in
each eye. J Neural Eng. 2020 Apr;17(2):026020.

46. da Cruz L, Dorn JD, Humayun MS, Dagnelie G, Handa J, Barale PO, et al. Five-Year Safety and
Performance Results from the Argus II Retinal Prosthesis System Clinical Trial. Ophthalmology.
2016 Oct 1;123(10):2248–54.

47. Ahuja AK, Dorn JD, Caspi A, Filley E, Dagnelie G, Greenberg RJ, et al. The Argus II Retinal
Prosthesis Enables Blind Subjects to Identify the Direction of Motion. Invest Ophthalmol Vis Sci.
2009 Apr 28;50(13):4590.

48. Stronks HC, Dagnelie G. The functional performance of the Argus II retinal prosthesis. Expert Rev
Med Devices. 2014 Jan 1;11(1):23–30.

49. Kasowski J, Beyeler M. SimpleXR: An open-source Unity toolbox for simplifying vision research
using augmented and virtual reality. J Vis. 2023 Aug 1;23(9):5518.

50. Humayun MS, Dorn JD, da Cruz L, Dagnelie G, Sahel JA, Stanga PE, et al. Interim results from the
international trial of Second Sight's visual prosthesis. Ophthalmology. 2012 Apr;119(4):779–88.