A meta-analysis of bone conduction 80 Hz auditory steady state response thresholds for adults and infants with normal hearing

Emanuele Perugia[1,*] & Constantina Georga[2]

[1]Manchester Centre for Audiology and Deafness, Division of Psychology, Communication and

Human Neuroscience, School of Health Sciences, Faculty of Biology, Medicine and Health, The

University of Manchester, Manchester, UK

[2]Royal Berkshire NHS Foundation Trust, Reading, UK

*Corresponding author

E-mail: emanuele.perugia@manchester.ac.uk

ORCiD

Emanuele Perugia: 0000-0002-4392-6933

Abstract

Background: Auditory steady-state responses (ASSRs) provide an objective method for estimating hearing thresholds in individuals unable to provide behavioural responses. Bone conduction (BC) testing is required to differentiate conductive from sensorineural hearing loss. Accurate BC ASSR threshold estimation relies on "correction" factors, which are not yet well established. This meta-analysis evaluated the reliability of BC ASSR thresholds to estimate hearing thresholds at 500, 1000, 2000 and 4000 Hz.

Methods: A systematic search of PubMed, the Cochrane Library, and Embase was conducted to identify studies involving normal-hearing (NH) and hearing-impaired (HI) participants of all ages. Outcomes were (1) the difference between ASSR behavioural and ASSR thresholds, and (2) ASSR thresholds. The risk of bias was evaluated using the Newcastle-Ottawa Scale. The mean and 95% confidence intervals (CI) were calculated for the thresholds at the four frequencies. The certainty of the evidence was assessed using GRADE approach.

Results: Of records identified, 11 records met the inclusion criteria, yielding a total of 27 studies. Sample sizes ranged from 60 to 249 participants across frequencies and age groups. The quality of records ranged from low to high. Data were synthesised using random-effects models due to heterogeneity. In NH adults, the mean differences (±95% CI) between BC ASSR thresholds and behavioural thresholds were 17.0 (±4.8), 15.5 (±6.0), 13.4 (±3.3), and 12.1 (±4.1) dB at 500, 1000, 2000, and 4000 Hz, respectively. In NH infants, mean (±95% CI) BC ASSR thresholds were 17.2 (±2.2), 10.5 (±3.6), 26.4 (±2.7), and 19.9 (±4.0) dB HL at the same frequencies. The certainty of the evidence was very low.

Conclusions: BC ASSR can be a reliable method for estimating BC thresholds. However, age and frequency significantly impact BC ASSR thresholds, highlighting the need to develop of "correction" factors to accurately predict BC behavioural thresholds.

Registration: PROSPERO CRD42023422150.

Abbreviations/Glossary

AC: air-conduction threshold; AMEI: Active middle ear implant; AM: amplitude modulation; ASSR: auditory steady-state response; BC: bone-conduction threshold; CT²: circular T² test; HI: hearing-impaired participant; DPOAE: distortion-product otoacoustic emissions; HT²: Hotelling T²; NH: normal-hearing participant; MSC: magnitude squared coherence; NOS: Newcastle-Ottawa Scale; PC²: phase coherence squared; PCA: post-conceptional age; PICO: Population, Interventions, Comparators and Outcomes; SNHL: sensorineural hearing loss; TEOAE: transient evoked otoacoustic emission.

Introduction

Auditory Steady-State Responses (ASSRs) are auditory evoked potentials that are used in clinical audiology to estimate hearing thresholds [1–6]. Clinically, the technique is applicable for populations that are not able or not willing to provide behavioural thresholds, such as infants, children, and adults with intellectual difficulties or those with non-organic functional hearing loss.

The main advantage of the ASSR is that it allows objective simultaneous threshold measurements at several frequencies, reducing the testing time required [7]. A second advantage is that interpretation is primarily objective, as it relies on statistical analysis of the signal in the frequency domain [8–10], whereas interpretation of Auditory Brainstem Response (ABR) is dependent primarily on subjective interpretation of the waveform in the time domain [6,7]. The ASSR method is limited, in that, as the response is analysed in the frequency domain, it does not provide information regarding neural integrity, and so cannot diagnose neuropathy [11]. For this reason, the UK clinical ASSR protocol recommends ABR testing as the starting point for neonatal assessments [12]. For a comprehensive discussion of the strengths and weaknesses of ASSR, please refer to [6,13,14].

Bone conduction (BC) testing is important to differentiate between conductive and sensorineural hearing losses which sequentially informs the decisions around the urgency and

type of intervention. However, some information about conductive hearing loss can also be

obtained through the click-ABR [15], as well as tympanometry and acoustic immittance [16,17].

In mixed hearing losses, BC levels inform the gain set on the hearing aids and thus ensure there

is no under- or over-amplification [18]. This is particularly important for infants, where early

intervention is crucial and obtaining outcome measures on hearing aid fitting is most difficult.

Although there has been a substantial body of research on air conduction (AC) ASSRs [14], a

relatively smaller number of studies have focused on BC ASSR testing, especially when

considering normative data for threshold estimation. In general, considerable heterogeneity

exists in methods among ASSR studies [14,19] and no universal testing protocol has been

established [12,20,21].

The objective of the meta-analysis was to assesses whether hearing thresholds estimated by BC

ASSR are reliable at the four classic audiometric frequencies (i.e., 500, 1000, 2000 and 4000

Hz). The present study sought to synthesise the BC ASSR thresholds from studies conducted to

date. Research in clinical auditory electrophysiology generally consists of observational studies

Our starting intention was to collect data from all ages. Data from infants are very limited with

many studies demonstrating methodological issues, such as different procedures to record BC

ASSR [22,23], lack of behavioural BC thresholds for comparison and frequency-specific hearing

screening [24].

Methods

The study was pre-registered in PROSPERO (CRD42023422150) and followed the PRISMA

2020 statement [25,26].

Literature search, data collection, and risk of bias

Eligibility criteria

Eligibility criteria for study characteristics were based on the Population, Interventions,

Comparators and Outcomes (PICO) framework [27]. Population: both normal-hearing (NH) and

hearing-impaired (HI) participants of all ages. Interventions: BC ASSR thresholds. Comparators: BC ASSR thresholds vs. BC behavioural thresholds in adults and infants, and in NH and HI participants. Outcomes: the main outcomes were the BC ASSR thresholds in NH participants, as well as the difference between BC behavioural and ASSR thresholds. Additional (co-)factors were participants' age and hearing threshold, testing levels and frequencies.

Only studies that reported BC ASSR thresholds were included. Studies were excluded if they met any of the following criteria: (1) not published in English, (2) pre-prints (i.e., not peer-reviewed), (3) reviews, (4) animal studies, (5) BC ASSR data collected from participants with a genetic syndrome or condition, (6) BC ASSR thresholds not reported, (7) BC ASSRs recorded in cochlear implant or active middle ear implant users, (8) BC ASSRs evoked by a 40 Hz modulation rate, or (9) unavailable as a full-text version. BC ASSRs evoked by a 40 Hz modulation rate were excluded because ASSRs evoked by an 80 Hz modulation rate are recommended for assessing hearing thresholds [12], and have different neural generators and amplitude than those evoked by a 40 Hz modulation rate [13,28,29]. No restrictions were applied regarding the date of publication.

Information sources and search strategy

The MEDLINE, Embase, and Cochrane Central Register of Controlled Trials (CENTRAL) databases were searched by the authors from May 2023 to August 2023 (see Supplementary Table S1 for the full search strategies). Using the same strategy, a further search was conducted in December 2025 to update (and verify) the initial search. Additionally, backward and forward citation searchings were performed using the *Citationchaser* Shiny app [30]. Specifically, all references of the included reports (backward search) and all studies that cited the most recent included report (forward search) were scanned.

Selection process

The authors independently screened titles and abstracts of potentially eligible studies. Then, they independently reviewed full-text articles for inclusion. In both stages, disagreements between authors were solved by discussion.

Data collection process

EP extrapolated data items and entered them into a spreadsheet. CG checked data accuracy. Disagreement was solved by re-checking articles. Data collection was performed by entering values from tables in the source articles, by synthesis of values from figures in the source articles via WebPlotDigitizer [31] (i.e., Fig 3 of [24] and Fig 3 of [32]), or by a table from the related dissertation (i.e., data of [33] were taken from [34]).

Data items

Data and information collected from the studies were the following: study (authors, year, and country), participants (sample size, age, and hearing threshold and group), equipment (device, bone conductor type and position, electrode montage), protocol (stimuli and their calibration, artefact rejection, average type, detection algorithm, noise criterion, and testing frequencies and levels). Missing data are identified and reported.

Study risk of bias assessment

The Newcastle-Ottawa Scale (NOS) for cohort studies [35] was used to assess the risk of bias in the included studies. The assessment covered three categories (Selection, Comparability, and Outcome) and collectively identified over eight items. Items in the Selection and Outcome categories could be awarded a maximum of one star each, while the item in the Comparability category could receive up to two stars. Studies with 0-3 stars were considered high risk, those with 4-6 stars intermediate risk, and those with 7-9 stars low risk [36]. As stated in the pre-registration, a minimum of three papers with at least three stars on the NOS tool was required

143 for inclusion in the synthesis. No sensitivity analyses were planned. The two authors

144 independently assessed the studies and resolved any discrepancies through discussion.

Effect measures, synthesis methods, and certainty assessment

Effect measures

147 Analyses were performed on the BC ASSR thresholds for NH participants, and on the difference

148 between their behavioural and ASSR BC thresholds. The thresholds were analysed as a function

149 of participants' age and testing frequency. Mean, SD, effect size and 95% confidence intervals

150 (CIs) were calculated. Due to the insufficiency of reported results for HI participants, analyses of

151 the BC ASSR thresholds for these participants could not be performed.

Synthesis methods

The scope of the synthesis was twofold. First, it aimed to compare BC behavioural and ASSR

thresholds for both NH and HI participants, with the goal of establishing correction factors for

accurately predicting BC behavioural thresholds from BC ASSR thresholds. Second, it involved

pooling BC ASSR thresholds for NH participants to establish typical BC ASSR thresholds with

the aim of using these as clinical discharge criteria.

However, only three studies (two in adults and one in infants) reported results for HI

participants. In general, most articles reported only ASSR BC thresholds, i.e., behavioural

thresholds were missing. Given the thresholds variability associated with age and frequency,

data subgroups were made based on participants' age (i.e., adults or infants) and on frequency

(i.e., 500, 1000, 2000, and 4000 Hz).

To assess statistical heterogeneity of the data, the I^2 index [37] was performed within each

subgroup. Specifically, the I^2 index measures the proportion of total variability due to between-

study heterogeneity [38]. Either fixed-effect modelling (in case of $I^2 < 50\%$; i.e., low

heterogeneity) or random-effect modelling (otherwise) are used to synthesise the data. The I^2

index is based on the Chi-squared test, with a statistically significant (i.e., between-study

heterogeneity) p-value threshold of < 0.10 [37]. In the fixed-effect model, the heterogeneity is

neglected and the difference among studies is assumed as being due to chance. Instead, in the

random-effect model, the heterogeneity is incorporated.

The objective of the meta-analysis was to assess whether hearing thresholds estimated by BC

ASSR are reliable at the four classic audiometric frequencies (i.e., 500, 1000, 2000 and 4000

Hz) and for both adults and infants. To be considered a reliable threshold, the 95% confidence

intervals around the mean of the synthesised data should be within 10 dB, as BC ASSR stimuli

are generally presented in 10 dB steps [22].

All analyses were run in R (version 4.3.1; [39]) using the packages *PRISMA2020* [40], *meta*

[41], and the *Tidyverse* family [42].

Certainty assessment

The GRADE approach was used to assess the certainty of the evidence. As the reports included

in the meta-analysis were non-randomised studies, by definition, the body of evidence initially

began with a low-certainty rating. The following criteria were considered for downgrading the

certainty of evidence, where appropriate: unexplained heterogeneity or inconsistency of results,

indirectness of evidence, imprecision of results, and a high probability of publication bias. The

certainty of evidence could be upgraded in the presence of a large effect [43]. The GRADEpro

GDT software was used to prepare the Summary of Findings tables.

Results

Study selection

The database searches identified 98 records. After removal of duplicates, 59 records were

screened. Twenty-six records were excluded because they were not in English, were preprints or

conference abstracts, were reviews, were not relevant, involved genetic conditions, or involved

participants with cochlear implants or active middle ear implants. Four records were not

available. Consequently, 29 records underwent full-text review. Of these, 20 records were

excluded because they did not performed BC ASSR thresholds, did not report BC ASSR

thresholds, reported dB values that could not be converted, or were conducted at 40 Hz. Citation

searching of the included records identified 265 additional records, which were scanned, and

two met the final selection criteria. In total, 11 records were remained, which included 27

individual studies [22–24,32,33,44–49]. A summary of the selection procedure is presented in

Fig. 1.

The number of studies was larger than the number of reports, as several reports tested different

groups of participants. Specifically, Small & Stapells (2006) [45] measured thresholds in both

preterm and post-term infants. Small et al. (2007) [23] reported three different experiments

involving three different groups of infants. Small & Stapells (2008a) [24] had three different age

groups. Small & Stapells (2008b) [46] and Casey & Small [33] recorded thresholds in infants

and adults. Swanepoel et al. (2008) [47] measured thresholds in NH infants and in three groups

of HI infants (conductive HL, mild-to-moderate and severe-to-profound sensorineural hearing

loss [SNHL]). Ishida et al (2011) [48] tested two groups of NH adults and a group of HI adults.

Small & Hu (2011) [32] reported thresholds obtained in young infants, older infants, and adults.

Ismaila et al., [49] measured thresholds in NH infants and in two groups of HI infants (mild-to-

moderate and severe-to-profound SNHL).

Figure 1: PRISMA 2020 flow diagram of the study selection procedure.

Study characteristics

The characteristics of the 11 included studies are presented in Tables 1 and 2. The Correspondent author on eight out of 11 studies were based at the University of British Columbia (Vancouver, Canada) as the group led by David R. Stapells and Susan A. Small have produced many contributions in the field [22–24,32,33,45,46]. The other studies were based at the University of Toronto (Canada) [44], University of Pretoria (South Africa) [47], and Al-Azhr University (Cairo, Egypt) [49]. All studies tested participants within different age groups (e.g., adults and infants) and/or different hearing groups (e.g., NH and HI).

To generate and record the responses, the Rotman MASTER system [50] was used in eight out of 11 studies, and the GSI AUDERA (Grason-Stadler, Eden Prairie, Minnesota, USA) used twice [47,49]. The B71 bone transducer (Radioear, New Eagle, Pennsylvania, USA) was used in all studies. It was positioned on the temporal bone in eight studies, on the mastoid in two studies [47,49], and on the forehead in a study [44].

Different types of stimuli were used across studies. 100% amplitude modulated (AM) tones were used in two studies [22,44], with mixed modulated tones (100% AM and 10% to 25% FM) in another eight studies [23,24,32,45–49]. Exponentially AM tones (100% AM and 5% FM) were used in Casey & Small (2014) [33]. Carrier frequencies tested were 500, 1000, 2000, and 4000 Hz in all studies, with 250 Hz also tested by Swanepoel et al (2008) [47]. Online artefact rejection was set by all but two studies [47,49], with threshold for rejection of epochs between ±40 and 80 µV. Weighted [51], rather than linear averaging of sweeps was used in six studies [23,24,32,33,46,48].

Table 1: Study Characteristics of the 11 Included Studies. AABR: automatic auditory brain stem response; AC: air-conduction thresholds; BC: bone-conduction thresholds; DPOAE: distortion-product otoacoustic emissions; HI: hearing-impaired participants; NA: not available; NH: normal-hearing participants; PCA: post-conceptional age; SNHL: sensorineural hearing loss; TEOAE: transient evoked otoacoustic emission.

Study	Country	Sample	Age (Mean)	Hearing status	Hearing Estimation
Lins et al (1996) [44]	Canada	8 adults	NA	NH	Behavioural AC & BC @ 500 to 4000 Hz
Small & Stapells (2005) [22]	Canada	10 adults	Range : 18-48 years	NH	Behavioural AC & BC ≤ 20 dB HL @ 250 to 8000 Hz
Small & Stapells (2006) [45]	Canada	29 preterm infants	34.5 week PCA	NH	AABR at 35 dB nHL
		14 postterm infants	17 week	NH	DPOAE
Small et al (2007) [23]	Canada	10 postterm infants	17 week	NH	AABR at 35 dB nHL or DPOAE
		15 preterm infants (9 were included in the analysis of ASSR threshold)	34.5 week PCA	NH	AABR at 35 dB nHL
		13 postterm infants	15 week	NH	DPOAE
Small & Stapells (2008a) [24]	Canada	35 young infants (but 9 from [23]; 14 from [45]; 12 from [a])	16 week	NH	
		13 older infants	18.2 months	NH	AABR at 35 dB nHL or DPOAE
		18 adults (but 10 from [22])	22.9 years	NH	Behavioural AC & BC thresholds ≤ 25 dB HL @ 500-4000 Hz
Small & Stapells (2008b) [46]	Canada	14 infants (with 1 participated also in [23])	21 week	NH	DPOAE
		11 adults	23 years	NH	Behavioural AC thresholds ≤ 25 dB HL @ 500-4000 Hz
Swanepoel et al (2008) [47]	South Africa	13 infants (21 ears)	3.6 years	NH	Tympanograms, DPOAE, AC click-evoked ABR thresholds, AC ASSR thresholds
		18 infants (35 ears)	2.7 years	Conductive HL	
		7 infants (13 ears)	2.5 years	Mild-to-Moderate SNHL	

a   The manuscript referred to "Small & Stapells (2005b)", but this was not in the reference of the report. We think that the correct reference is Small & Stapells (2008b) [46] because the number of infants tested using a two-channel EEG.

		13 infants (23 ears)	2.2 years	Severe-to-profound SNHL	
Ishida et al (2011) [48]	Canada	16 adults	26.8 years	NH	Behavioural AC thresholds ≤ 20 dB HL @ 500-4000 Hz
		10 adults	24.6 years	NH	
		40 adults	46.0 years	SNHL	Behavioural AC thresholds > 20 & ≤ 55 dB HL @ 500-4000 Hz; & BC thresholds
Small & Hu (2011) [32]	Canada	22 young infants (but for 12 from [23])	2.1 months	NH	TEOAE
		10 older infants	15.3 months	NH	
		20 adults	26.5 years	NH	Behavioural AC & BC thresholds ≤ 25 dB HL @ 500-4000 Hz
Casey & Small (2014) [33]	Canada	23 infants	10.5 months	NH	TEOAE or tympanometry
		12 adults	29.9 years	NH	Behavioural AC & BC thresholds ≤ 25 dB HL @ 500-4000 Hz & air–bone gaps ≤10 dB
Ismaila et al (2016) [49]	Egypt	20 infants (35 ears)	4.5 years	NH	AC play audiometry mean = 15 dB HL, AC click-evoked ABR thresholds mean = 30 dBnHL, tympanograms
		20 infants (36 ears)		Mild-to-moderate SNHL	AC & BC play audiometry mean = 30-50 dB HL, AC click-evoked ABR thresholds mean = 40-60 dBnHL, tympanograms
		20 infants (35 ears)		Conductive HL	AC play audiometry mean = 25-50 dB HL, Normal BC play audiometry, AC click-evoked ABR elevated, tympanograms

Table 2: Hearing Test Information of the 11 Included Studies AM: amplitude modulation; CR: clear response; CF: carrier frequency; FM: frequency modulation; MF: modulation frequency; HL: hearing loss; HT2: Hotelling T2; PC2: phase coherence squared; RA: response absent; SNHL: sensorineural hearing loss.

Study	Equipment	Placement	Stimulus	Sweep	Electrode montage	Artefact rejection	Average	Detection algorithm (CR)	Noise criterion (RA)	Testing levels
Lins et al (1996) [44]	NA	Forehead	Modulated tones with 100% AM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 512 data points and lasted a total of 12.06 seconds.	Cz and at the nape f the neck	Epoch > ±40 uV in amplitude rejected.	Normal	F ratio & HT² with $p <$ 0.05	64 sweeps were not SIG	Started at 50 dB SPL, decreased in 10 dB step until all resp were absent
Small & Stapells (2005) [22]	Rotman MASTER research system	Temporal Bone	Modulated tones with 100% AM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 1024 data points and lasted a total of 13.107 seconds.	Noninverting on Cz. Inverting at the nape of the neck, just below the hairline. Ground on high forehead.	Epoch > ±40 uV in amplitude rejected.	Normal	F ratio with $p < 0.05$	noise < 11 nV & $p >$ 0.05 OR noise < 10 nV & $p \geq$ 0.30	10 dB steps at: (1) 50 to 10 dB HL for noninverted stimuli, (2) 50 to 30 dB HL for inverted stimuli, and (3) 50 to 20 dB HL for alternated stimuli.
Small & Stapells (2006) [45]	Rotman MASTER research system	Temporal Bone	Modulated tones with 100% AM & 25% FM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 1024 data points and lasted a total of 13.107 seconds. A minimum of 7 sweeps per condition.	Noninverting on midline at the high forehead. Inverting at the nape of the neck, just below the hairline. Ground at	Epoch > ±40 uV in amplitude rejected.	Normal	F ratio with $p < 0.05$ for at least two consecutive sweeps.	noise < 11 nV & $p >$ 0.05 OR noise < 10 nV & $p \geq$ 0.30	

					the low forehead.					
Small et al (2007) [23]	Rotman MASTER research system	Temporal Bone	Modulated tones with 100% AM & 25% FM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 1024 data points and lasted a total of 13.107 seconds. A minimum of 7 sweeps per condition.	Noninverting on forehead. Inverting at the nape of the neck, just below the hairline. Ground on high forehead.	Epoch > ±40 uV in amplitude rejected.	Weighted	F ratio with $p < 0.05$ for at least two consecutive sweeps.	noise < 11 nV & $p >$ 0.05 OR noise < 10 nV & $p ≥$ 0.30	10-dB steps from 50 to -10 dB HL
Small & Stapells (2008a) [24]	Rotman MASTER research system	Temporal Bone	Modulated tones with 100% AM & 25% FM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 1024 data points and lasted a total of 13.107 seconds. A minimum of 7 sweeps per condition.	For 23 of the 35 young infants, all older infants and adults: Noninverting on forehead. Inverting at the nape of the neck. Ground on high forehead. For 12 of the 35 young infants, 2 2 inverting on the L & R mastoids	Epoch > ±40 uV in amplitude rejected.	Weighted	F ratio with $p < 0.05$ for at least two consecutive sweeps.	noise < 11 nV & $p >$ 0.05 OR noise < 10 nV & $p ≥$ 0.30	10-dB steps
Small & Stapells (2008b) [46]	Rotman MASTER research system	Temporal Bone	Modulated tones with 100% AM & 25% FM Multiple CF:	16 epochs of 1024 data points and lasted a total of 13.107	Noninverting on high forehead. Inverting on mastoids.	Epoch > ±40 uV in amplitude rejected.	Weighted	F ratio with $p < 0.05$ for at least two consecutive sweeps.	noise < 11 nV & $p >$ 0.05 OR noise < 10 nV & $p ≥$	10-dB steps from 50 to -10 dB HL

			0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	seconds. A minimum of 7 sweeps per condition.	Ground on low forehead.				0.30	
Swanepoel et al (2008) [47]	Gsi AUDERA	Mastoid	Modulated tones with 100% AM & 10% FM Multiple CF: 0.25, 0.5, 1, 2, & 4 kHz MF: 67, 74, 81, 88, & 95 Hz	NA	Noninverting on high forehead (Fz). Inverting on ipsilateral mastoid. Ground on contralateral mastoid.			PC^2 with p <0.03	1uV	10- & 5- dB steps
Ishida et al (2011)[48]	Rotman MultiMASTER	Temporal Bone	Modulated tones with 100% AM & 25% FM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 1024 data points and lasted a total of 13.107 seconds. A min of 12 sweeps per condition.	Noninverting on Cz. Inverting on midline at the nape. Ground on high forehead.	Epoch > ±57 uV in amplitude rejected.	Weighted	F ratio with $p < 0.05$ for at least 7 consecutive sweeps.	noise1 < 11 nV & $p >$ 0.05 OR ASSR amp < 10 nV & $p ≥$ 0.30	Started at 20 or 40 dB HL, increased or decreased in 10-dB steps
Small & Hu (2011) [32]	Rotman MultiMASTER	Temporal Bone	Modulated tones with 100% AM & 25% FM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 77, 85, 93, & 101 Hz	16 epochs of 1024 data points and lasted a total of 13.107 seconds. A min of 10 sweeps per condition.	Noninverting on high forehead. Inverting on ipsilateral mastoids. Ground on forehead.	Epoch > ±80 uV in amplitude rejected.	Weighted	F ratio with $p < 0.05$ for at least 2 consecutive sweeps.	noise < 11 nV & $p >$ 0.05 OR ASSR amp < 10 nV & $p ≥$ 0.30	Random start between 10 to 30 dB HL, increased or decreased in 10-dB steps
Casey & Small (2014) [33]	Rotman MultiMASTER	Temporal Bone	Exponential modulated tones with 100% AM& 5% FM	16 epochs of 1024 data points and lasted a total of 13.107	Noninverting on Fz. Inverting on ipsilateral mastoids.	Epoch > ±80 uV in amplitude rejected.	Weighted	F ratio with $p < 0.05$ for at least 3 consecutive sweeps.	noise < 11 nV & $p >$ 0.05 OR ASSR amp < 10 nV & $p ≥$	Random start between 10 to 30 dB HL, increased or

			Multiple CF: 0.5, 1, 2, & 4 kHz MF: 78, 85, 93, & 101 Hz	seconds. A min of 10 consecutive sweeps	Ground on low forehead.				0.30	decreased in 10-dB steps
Ismaila et al (2016) [49]	Gsi AUDERA	Mastoid	Modulated tones with 100% AM & 10% FM Multiple CF: 0.5, 1, 2, & 4 kHz MF: 74, 81, 88, & 95 Hz	NA	Noninverting on high forehead (Fz). Inverting on ipsilateral mastoid. Ground on contralateral mastoid.			PC² with p <0.03	1mV	10-dB steps

Detection algorithm and noise criterion

Multiple detection algorithms have been employed for ASSR analysis: F ratio, Hotelling T^2 (HT^2), phase coherence squared (PC^2), as well as combinations of the above [9,52]. The F ratio for hidden periodicity compares the power of the signal (i.e., the modulation frequency) to the mean power of N adjacent frequencies. The F ratio follows an F-distribution with [2 and N] degrees of freedom [7,9]. PC^2 and HT^2 are considered repeated measurements of the variance of the response [7], as they are based on individual epochs rather than an average epoch [9]. HT^2 is the two-dimensional analogue of univariate t-tests, where the real and imaginary components of the epoch spectra form a two-dimensional vector. The HT^2 test assesses whether the means of the real and imaginary components are significantly different from zero [7,9]. The PC^2, in contrast, considers only the phase of the response and is based on the Rayleigh test [7,52].

An ASSR system requires an algorithm for determining the likelihood of the absence or presence of a response. For the former, usually a Signal to Noise Ratio (SNR) analysis and/or noise criterion threshold are involved and, for the latter, a Residual Noise (RN) analysis is required. The efficiency with which this is carried out will determine the accuracy in threshold estimation. It has been shown that increased number of sweeps will increase SNR and reduce RN resulting in increased threshold estimation accuracy [53]. It is therefore important to consider the effects of detection technique, the effects of number of sweeps, and noise criterion for stopping averaging in the current analysis.

The detection algorithms employed depended on the device. The Rotman MASTER decision is based on F ratio while the GSI AUDERA is based on PC^2 (see [10] for a comparison). In eight studies, the noise criterion for considering a response absent was defined as either noise < 11 nV and $p > 0.05$, or noise < 10 nV and $p \geq 0.30$ [22–24,32,33,45,46,48]. Other studies used a non-significant detection of ASSR for 64 sweeps [44], a noise criterion threshold of 1 µV [47] or 1 mV [49].

Effect measures

Four reports, yielding five studies, compared BC behavioural and BC ASSR thresholds in NH adults [44,45,48,32]. Ishida et al (2011) [48] compared BC behavioural and BC ASSR thresholds in two different experiments with NH and HI participants. The data from these studies were used to carry out a meta-analysis.

BC ASSR thresholds in NH adults were estimated in seven reports, yielding eight studies [22,24,32,33,44,46,48]. The data from these studies were used to carry out a second meta-analysis.

 BC ASSR thresholds in NH infants were estimated in eight reports, yielding 13 studies [23,24,32,33,45–47,49]. The data from these studies were used to carry out a third meta-analysis.

Thresholds in HI adults were compared only by Ishida et al. (2011) [48]. Only two studies [33,49] compared BC behavioural and BC ASSR thresholds in NH infants. Similar, two studies [47,49] compared BC behavioural and BC ASSR thresholds in HI infants. These data were not used for meta-analyses.

In total, three meta-analyses were performed: (1) BC ASSR thresholds vs BC behavioural in NH adults, (2) BC ASSR thresholds in NH adults, and (3) BC ASSR thresholds in NH infants.

Hearing threshold

For adult participants, behavioural AC and, in five studies, BC thresholds were obtained, so their hearing thresholds were clinically ascertained.

For infant participants, Casey & Small (2014) [33] performed visual reinforcement audiometry to estimate hearing threshold at different frequencies [54]. Ismaila et al. (2016) [49] performed play audiometry but reported only BC thresholds.

Hearing was screened via automatic auditory brain stem response (AABR) and/or distortion-product otoacoustic emissions (DPOAE) by five studies [23,24,45–47] with also tympanograms

by Swanepoel et al (2008) [47]. Hearing screening in Small & Hu (2011) [31]. was based only on

transient evoked otoacoustic emission (TEOAE). However, AABR and OAE do not provide

frequency-specific estimates of auditory function. This is their main drawback for hearing

screening [6,55].

Same participants in multiple experiments or studies

Small et al. (2007) [23] compared the effects of BC coupling methods (i.e., elastic-band vs hand-

held) and oscillator placements (temporal, mastoid and forehead) on ASSR threshold. The two

coupling methods were recorded in the same infants; here we included the elastic-band

thresholds because their standard deviation was smaller than for the hand-held thresholds. The

three oscillator placements were recorded in the same infants; here we included temporal bone

thresholds as it was the placement most commonly used in the other articles.

Two studies [24,32] combined new participants' thresholds with those previously reported, or

combined thresholds from different articles. These thresholds are included here because their

mean and standard deviation were different.

Risk of bias in studies

A summary of these assessments is provided in Table 3. Six studies were classified as high risk

[23,24,32,45–47] because the evaluation of hearing thresholds in infants was not frequency-

specific (see above). Four studies were classified as intermediate risk [22,33,44,49], and one

study as low risk [48]. Since none of the studies included follow-up, only one star was assigned

in the Outcome category. Cumulatively, all studies received at least three stars and were

therefore included in the synthesis.

Table 3: Risk of bias for each of the included study.

Study	Selection	Comparability	Outcome	Total
Lins et al (1996) [44]	2	1	1	4
Small & Stapells (2005) [22]	3	0	1	4
Small & Stapells (2006) [45]	1	1	1	3

Small et al (2007) [23]	1	1	1	3
Small & Stapells (2008a) [24]	1	1	1	3
Small & Stapells (2008b) [46]	1	1	1	3
Swanepoel et al (2008) [47]	1	1	1	3
Ishida et al (2011) [48]	4	1	1	6
Small & Hu (2011) [32]	1	1	1	3
Casey & Small (2014) [33]	2	1	1	4
Ismaila et al (2016) [49]	2	1	1	4

Results of individual studies and syntheses

BC ASSR vs behavioural thresholds in NH adults

Fig 2 shows the forest plot for BC ASSR vs behavioural thresholds in NH adults as a function of the frequency (see also Supplementary Table S2). The number of participants per frequency was between 60 and 62. Heterogeneity was less than 50% only at 2000 Hz (I^2 = 43%, τ^2 = 6.1, p = 0.14). It was high at 500 (I^2 = 52%, τ^2 = 15.4, p = 0.08), 1000 (I^2 = 68%, τ^2 = 33.0, p = 0.01), and 4000 Hz (I^2 = 58%, τ^2 = 11.9, p = 0.06). For consistency, a random effects model was carried out for each frequency, using the inverse variance method, the restricted maximum-likelihood estimator for τ^2, and untransformed (raw) means.

Means (and 95% confident intervals) of the random effects model for each frequency were 17.0 (12.3/21.8) dB at 500 Hz, 15.5 (9.5/21.4) dB at 1000 Hz, 13.4 (10.1/16.7) dB at 2000 Hz, and 12.1 (8.0/16.2) dB at 4000 Hz.

Figure 2: Forest plot for BC ASSR vs behavioural thresholds in NH adults as a function of the frequency. Behav: behaviour; SD: standard deviation; MD: mean difference; CI = confident intervals.

BC ASSR thresholds in NH adults

Fig 3 shows the forest plot for BC ASSR thresholds in NH adults as a function of the frequency (see also Supplementary Table S3). Seven studies reported BC ASSR thresholds in NH adults, with Ishida et al (2011) [48] reporting the thresholds on two different cohorts. The number of participants per frequency was between 97 and 99. High heterogeneity was present at each frequency as $I^2 = 62\%$ ($\tau^2 = 12.7$, $p < 0.01$) at 500 Hz; $I^2 = 65\%$ ($\tau^2 = 17.9$, $p < 0.01$) at 1000 Hz; I^2

= 80% (τ^2 = 26.3, p < 0.01) at 2000 Hz; and I^2 = 89% (τ^2 = 41.2, p < 0.01) at 4000 Hz.

Therefore, a random effects model was carried out, using the inverse variance method, a

restricted maximum-likelihood estimator for τ^2 and untransformed (raw) means.

Means (and 95% confident intervals) of the random effects model for each frequency were 24.6

(21.4/27.8) dB HL at 500 Hz, 19.2 (15.4/23.0) dB HL at 1000 Hz, 20.00 (16.1/23.9) dB HL at

2000 Hz, and 16.3 (11.4/21.1) dB HL at 4000 Hz.

Figure 3: Forest plot for BC ASSR thresholds in NH adults as a function of the frequency. SD: standard deviation; MRAW: mean raw, i.e., untransformed means; CI = confident intervals.

BC ASSR thresholds in NH infants

Fig 4 shows the forest plot for BC ASSR thresholds in NH infants as a function of the frequency (see also Supplementary Table S4). Six studies reported BC ASSR thresholds in NH infants, with additional four studies reporting the thresholds on different cohorts [23,24,32,45]. The number of participants per frequency was between 247 and 249. High heterogeneity was present at each frequency as $I^2 = 54\%$ ($\tau^2 = 8.5$, $p < 0.01$) at 500 Hz; $I^2 = 93\%$ ($\tau^2 = 39.6$, $p < 0.01$) at 1000 Hz; $I^2 = 82\%$ ($\tau^2 = 18.0$, $p < 0.01$) at 2000 Hz; and $I^2 = 90\%$ ($\tau^2 = 48.3$, $p < 0.01$) at 4000 Hz. Therefore, a random effects model was carried out, using the inverse variance method, a restricted maximum-likelihood estimator for τ^2 and untransformed (raw) means.

Means (and 95% confident intervals) of the random effects model for each frequency were 17.2 (15.0/19.4) dB HL at 500 Hz, 10.5 (6.9/14.1) dB HL at 1000 Hz, 26.4 (23.7/29.1) dB HL at 2000 Hz, and 19.9 (15.9/23.9) dB HL at 4000 Hz.

Figure 4: Forest plot for BC ASSR thresholds in NH infants as a function of the frequency. SD: standard deviation; MRAW: mean raw, i.e., untransformed means; CI = confident intervals.

Spurious responses

Of particular interest are the minimum intensity levels at which spurious (artefactual) results were observed. Table 4 summarises the spurious results reported in the included studies. The level at which spurious responses occur increased as a function of frequency. Similarly, the percentage of participants with spurious responses increased with frequency. Therefore, the BC ASSR threshold at low frequencies, such as 500 Hz, and moderate intensities, such as 40 dB HL, may not be reliable. Swanepoel and colleagues [47] questioned the clinical value of BC ASSR thresholds measured at or below 500 Hz. Current UK guidance recommends that BC ASSR should not be performed at 500 Hz and beyond 40 dB nHL for frequencies between 1000 and 4000 Hz [12].

BC stimuli presented at moderate and high intensities elicited spurious responses, primarily at 500 Hz and 1000 Hz. At these frequencies and range of intensities, BC ASSR amplitudes were abnormally larger when compared either to those at low intensities or to AC ASSR amplitudes [56]. The reasons for these findings, along with potential solutions, are discussed elsewhere [22,56,57].

Table 4: Spurious (non-auditory) responses reported in the included papers. SNHL: sensorineural hearing loss.

Study	Note
Small & Stapells (2005) [22]	The percentage of participants who had thresholds ≤20 dB HL were 60, 40, 90, and 60% at 500, 1000, 2000, and 4000 Hz, respectively. The percentage of participants who had thresholds ≤30 dB HL were 90, 70, 100, and 100% at 500, 1000, 2000, and 4000 Hz, respectively.
Small & Stapells (2006) [45]	For the post-term infants, 90% or more of the subjects had responses present at minimum intensities of 30, 10, 40, and 30 dB HL for 500, 1000, 2000, and 4000 Hz, respectively. For the preterm infants, 90% or more of the subjects had responses present at minimum intensities of 40, 30, 50, and 50 dB HL for 500, 1000, 2000, and 4000 Hz, respectively.
Swanepoel et al (2008) [47]	Spurious BC responses for children with severe-to-profound SNHL. See their Table 4 for % spurious for children with mild-to-moderate SNHL. Min levels where spurious results were seen: 40 dB HL @ 500 Hz; 60 dB HL @ 1000, 2000, 4000.
Ismaila et al (2016) [49]	Min levels where spurious results were seen: 52 dB HL @ 500 Hz; 66.5 dB HL @ 1000 Hz; 69 dB HL @ 2000 Hz; 64 dB HL @ 4000 Hz.

Certainty assessment

The GRADE Summary of Findings tables for BC ASSR thresholds versus behavioural thresholds in NH adults are presented in Table 5, and those for BC ASSR thresholds in NH adults and infants are presented in Table 6. As the included studies were cohort studies, the body of evidence began with a low-certainty rating. The certainty was subsequently downgraded to very low due to serious inconsistency, reflected by high heterogeneity (in both adults and infants), and serious indirectness (in infants), as the evaluation of hearing thresholds was not frequency-specific.

Table 5: Summary of Findings tables for BC ASSR thresholds versus behavioural thresholds in NH adults.

BC ASSR thresholds compared to Behavioural BC thresholds in NH adults			
Patient or population: NH adults **Setting**: Audiology departments **Intervention**: BC ASSR thresholds **Comparison**: BC behavioural thresholds			
Outcomes	**№ of participants (studies)**	**Certainty of the evidence (GRADE)**	**Anticipated absolute effects**
			BC ASSR thresholds vs Behav BC thresholds
500 Hz	60 (5 non-randomised studies)	⊕○○○ Very low[a]	MD 17.0 **dB HL higher** (12.3 higher to 21.8 higher)
1000 Hz	61 (5 non-randomised studies)	⊕○○○ Very low[a]	MD 15.5 **dB HL higher** (9.5 higher to 21.4 higher)
2000 Hz	61 (5 non-randomised studies)	⊕○○○ Very low[a]	MD 13.4 **dB HL higher** (10.1 higher to 16.7 higher)
4000 Hz	62 (5 non-randomised studies)	⊕○○○ Very low[a]	MD 12.1 **dB HL higher** (8.0 higher to 16.2 higher)
MD: mean difference; **Behav**: Behavioural			
GRADE Working Group grades of evidence **High certainty**: we are very confident that the true effect lies close to that of the estimate of the effect. **Moderate certainty**: we are moderately confident in the effect estimate: the true effect is likely to be close to the estimate of the effect, but there is a possibility that it is substantially different. **Low certainty**: our confidence in the effect estimate is limited: the true effect may be substantially different from the estimate of the effect. **Very low certainty**: we have very little confidence in the effect estimate: the true effect is likely to be substantially different from the estimate of effect.			

Explanations
a. High heterogeneity

Discussion

The present meta-analysis included 11 papers spanning 20 years of research in BC ASSR. No

other available meta-analyses have specifically considered BC ASSR or BC ABR.

An accurate BC ASSR threshold estimation relies on precise and reliable "correction" values that

allow the prediction of behavioural thresholds from the measured ASSRs values. A possible

definition of a reliable BC ASSR threshold is that its 95% confidence interval is within 10 dB of

the mean of each model. The width of the confidence interval ranged from 2.2 to 6.0 dB HL,

indicating that it remained within 10 dB.

Our main finding is the elevation of BC ASSR thresholds compared to behavioural thresholds, as

a function of both age group and frequency. For NH adults, the difference between behavioural

and ASSR thresholds ranged from 12.1 to 17.0 dB, while ASSR thresholds ranged from 16.3 to

24.6 dB HL across the frequency range of 500-4000 Hz. For NH infants, ASSR thresholds

ranged from 10.5 to 26.4 dB HL within the same frequency range.

The BSA Early Assessment Guidance [[58], Appendix H] provides provisional corrections for

ABR thresholds as a function of frequency and age. For NH adults, the provisional corrections

range from 11.8 to 20.4 dB, which are similar to our findings. However, for NH infants, the

provisional corrections range from 13.6 to 19.6 dB, which differ from our findings. Therefore,

new correction values should account for the elevations reported here.

To explore a less-well established technique such as BC ASSRs, thresholds should be compared

with a gold standard technique such as behavioural thresholds. This is achievable in adults, but

not possible in neonates as behavioural assessment only becomes possible around seven months

and even then, a lengthy period may be required to determine accurate behavioural thresholds.

A compromise would be to compare the BC ASSR thresholds to those obtained using BC ABR.

Our original interest was to run meta-analyses on BC ASSR thresholds in HI adults and infants.

However, only three studies reported BC ASSR thresholds for these groups: one study for HI

adults and two for HI infants. Therefore, only NH thresholds were analysed.

Table 6: Summary of Findings tables for BC ASSR thresholds in NH adults and infants

BC ASSR thresholds in NH adults and infants			
Patient or population: NH adults and infants **Setting**: Audiology departments **Intervention**: BC ASSR thresholds			
Outcomes	**Nº of participants (studies)**	**Certainty of the evidence (GRADE)**	**Anticipated absolute effects** **BC ASSR thresholds**
Adults			
500 Hz	97 (8 non-randomised studies)	⊕◯◯◯ Very low[a]	mean **24.6 dB HL higher** (21.4 higher to 27.8 higher)
1000 Hz	98 (8 non-randomised studies)	⊕◯◯◯ Very low[a]	mean **19.2 dB HL higher** (15.5 higher to 23.0 higher)
2000 Hz	98 (8 non-randomised studies	⊕◯◯◯ Very low[a]	Mean **20.0 dB HL higher** (16.1 higher to 23.9 higher)
4000 Hz	99 (8 non-randomised studies)	⊕◯◯◯ Very low[a]	mean **16.3 dB HL higher** (11.4 higher to 21.1 higher)
Infants			
500 Hz	248 (13 non-randomised studies)	⊕◯◯◯ Very low[a,b]	mean **17.2 dB HL higher** (15.0 higher to 19.4 higher)
1000 Hz	249 (13 non-randomised studies)	⊕◯◯◯ Very low[a,b]	mean **10.5 dB HL higher** (6.9 higher to 14.1 higher)
2000 Hz	247 (13 non-randomised studies)	⊕◯◯◯ Very low[a,b]	mean **26.4 dB HL higher** (23.7 higher to 29.1 higher)
4000 Hz	249 (13 non-randomised studies)	⊕◯◯◯ Very low[a,b]	mean **19.9 dB HL higher** (15.9 higher to 23.9 higher)
GRADE Working Group grades of evidence **High certainty**: we are very confident that the true effect lies close to that of the estimate of the effect. **Moderate certainty**: we are moderately confident in the effect estimate: the true effect is likely to be close to the estimate of the effect, but there is a possibility that it is substantially different. **Low certainty**: our confidence in the effect estimate is limited: the true effect may be substantially different from the estimate of the effect. **Very low certainty**: we have very little confidence in the effect estimate: the true effect is likely to be substantially different from the estimate of effect.			

Heterogeneity

Heterogeneity was strongly apparent across the studies. This was due to multiple factors, such

as stimuli, statistical indices, stopping criteria, and threshold definitions used. Furthermore,

several studies reported a large standard deviation (up to 16.9 dB HL) of BC ASSR thresholds.

Nevertheless, the average BC ASSR thresholds were consistently positive, meaning they were

always higher than the behavioural thresholds.

Because of the small number of studies available, it was not possible to assess the effect of

recording parameters on the BC ASSR thresholds. A meta-analysis on AC ASSR thresholds

revealed that the effect of recording parameters, such as type of modulation (i.e., amplitude or

mixed modulated) and maximum number of sweeps, varied for different frequencies and

hearing status [19].

Infant thresholds

The measure of BC ASSR thresholds for infants and children should be interpreted with extra

caution because of their age. Specifically, the reported age varied from 34.5 weeks PCA to 4.5

years. Including such a large age range might be a confound for BC ASSR thresholds as, at least

for AC ASSR thresholds, thresholds tend to decrease in NH infants as a function of age [59,60].

Our rationale for including the full age range was that the heterogeneity and variability of the

studies were greater than any potential effects of age. For instance, ASSR threshold (and

standard deviation) at 2000 Hz was nearly independent of Age: 20.0 (12.8) dB HL in 21-week-

old infants [46], 20.50(10.5) dB HL in 10.5-month infants [33], 23.6 (6.5) dB HL in 3.6-year-old

children [47], and 20.0 (6.5) dB HL in 4.5-year-old children [49]. However, to rule out the

possible confound due to Age, a further meta-analysis was conducted excluding the studies in

children [47,49]. Means (and 95% confident intervals) of the random effects model for each

frequency were 16.2 (14.0/18.4) dB HL at 500 Hz, 8.9 (5.6/12.1) dB HL at 1000 Hz, 27.5

(24.8/30.2) dB HL at 2000 Hz, and 18.9 (14.3/23.4) dB HL at 4000 Hz. Thus, the overall mean

difference was 0.64 dB HL between the model including children and the one excluding them.

BC ASSR thresholds for adults and infants

Among others, Casey & Small (2014) [33] compared BC ASSR thresholds in adults and infants.

Fig 5 shows a summary of the meta-analysis results for BC ASSR thresholds in adults (blue

squares) and infants (pink circles) in the same conditions, i.e., only ASSR.

Adult thresholds are higher at low frequencies (500 and 1000 Hz) and lower at high frequencies

(2000 and 4000 Hz) when compared to those of infants. The adult threshold at 2000 Hz was

about the same of those at 1000 and 4000 Hz. Instead, the infant threshold at 2000 Hz was

clearly higher than those at other frequencies. Since BC ASSR thresholds change as a function of

age and frequency, correction factors are needed for different population and frequency.

Figure 5: Summary of the meta-analysis results for BC ASSR thresholds in adults (beige square) and infants (purple circle). The ASSR threshold dB (attenuator) as a function of the frequency.

Variability of BC ASSR thresholds

In the current meta-analyses, the standard deviations for BC ASSR thresholds across frequency was between 16.1 (at 500 Hz) and 24.8 (at 4000 Hz) dB HL for adults, and between 17.7 (at 500 Hz) and 32.4 (at 4000 Hz) dB HL for infants. Comparing these with AC ASSR thresholds, the standard deviation was reported between 9.9 and 12.1 dB HL for adults [19], and between 7 and 15 dB SL for infants (Table 9-2 in [13]). Comparing BC ASSR with BC tone burst ABR, the standard deviation was between 6.2 and 10.5 dB HL for adults, and between 7.2 and 10 for infants [61,62]. Therefore, the inter-subject variability across frequency in BC ASSR threshold was larger than those in AC ASSR and BC ABR thresholds.

It has been suggested that the substantial variability in BC ASSR thresholds may be due to factors such as variations in transducer placement, the condition of occluded versus non-occluded ears, differences in stimulus calibration methods [22]. Other possible contributing factors include test duration, artefact rejection thresholds, average type, detection algorithms, and noise criteria.

The aforementioned standard deviation values refer to inter-subject variability. Since BC ASSR thresholds can be influenced by several factors, intra-subject variability should also be evaluated. Future studies should assess the test-retest reliability of BC ASSR threshold [63].

Possible limitations

Beside the range of age of the infants discussed earlier, another limitation was that most of the included studies were conducted by Stapells and Small. Contribution from other laboratories or countries should help to improve the quality of research such as, for instance, removing risk of bias. A related issue was that most of the thresholds were collected using the Rotman MASTER system, a first-generation system that is no longer available. The lack of thresholds obtained using a second generation system [20] should be addressed in future research.

458 There were also limitations in the review processes used. For instance, the exclusion of non-

459 English articles and pre-prints, which could have improved the quality of the research by

460 diversifying the data. Nevertheless, we are confident in our outcome and the overall conclusions.

Conclusions

Findings from this meta-analysis suggest that ASSR can be a reliable technique to estimate BC

thresholds in NH populations. In the frequency range of 500-4000 Hz, the ASSR threshold

ranged from 16 to 24 dB HL in NH adults, and from 10 to 26 dB HL in NH infants. Future work

is needed to expand the measure of BC ASSR thresholds in NH infants, as well as to start

investigating both HI adults and infants. Only then can we reliably define normative BC ASSR

thresholds for different populations.

Declarations

Registration

470 This systematic review has been registered in the international prospective register of systematic

471 reviews (PROSPERO) under the registration number: CRD42023422150.

Author contributions

473 EP: Conceptualisation, Methodology, Software, Validation, Formal analysis, Investigation,

474 Resources, Data Curation, Writing - Original Draft, Writing - Review & Editing, Visualization,

475 Supervision, Project administration. CG: Conceptualisation, Methodology, Validation,

476 Investigation, Resources, Data Curation, Writing - Original Draft, Writing - Review & Editing,

477 Supervision, Project administration.

Availability of data and material

All data generated or analyses during this study are included in this published article and its Supplementary Table S5.

Competing interests

The authors declare that they have no competing interests.

Funding

E.P. is supported by the NIHR Manchester Biomedical Research Centre (NIHR203308).

Acknowledgments

We thank Timothy S. Wilding and Garreth Prendergast for helpful comments on an earlier version of this paper.

Supporting information

S1 Table. Search strategy

S2 Table. BC ASSR vs behavioural thresholds in NH adults as a function of the frequency.

S3 Table. BC ASSR thresholds in NH adults as a function of the frequency.

S4 Table. BC ASSR thresholds in NH infants as a function of the frequency.

S5 Table. Data.

494 **References**

1. Galambos R, Makeig S, Talmachoff PJ. A 40-Hz auditory potential recorded from the human scalp. Proc Natl Acad Sci USA. 1981;78: 2643–2647.

2. Lins OG, Picton TW. Auditory steady-state responses to multiple simultaneous stimuli. Electroencephalogr Clin Neurophysiol. 1995;96: 420–432.

3. John MS, Lins OG, Boucher BL, Picton TW. Multiple auditory steady-state responses (MASTER): stimulus and recording parameters. Audiology. 1998;37: 59–82.

4. John MS, Dimitrijevic A, van Roon P, Picton TW. Multiple auditory steady-state responses to AM and FM stimuli. Audiol Neurootol. 2001;6: 12–27. doi:46805

5. Herdman AT, Stapells DR. Thresholds determined using the monotic and dichotic multiple auditory steady-state response technique in normal-hearing subjects. Scand Audiol. 2001;30: 41–49.

6. Hall JW. eHandbook of Auditory Evoked Responses: Principles, Procedures & Protocols. 2015.

7. Picton TW, John MS, Dimitrijevic A, Purcell D. Human auditory steady-state responses. Int J Audiol. 2003;42: 177–219. doi:10.3109/14992020309101316

8. Rance G, Rickards FW, Cohen LT, De Vidi S, Clark GM. The automated prediction of hearing thresholds in sleeping subjects using auditory steady-state evoked potentials. Ear Hear. 1995;16: 499–507.

9. Valdes JL, Perez-Abalo MC, Martin V, Savio G, Sierra C, Rodriguez E, et al. Comparison of statistical indicators for the automatic detection of 80 Hz auditory steady state responses. Ear Hear. 1997;18: 420–429. doi:10.1097/00003446-199710000-00007

10. Luts H, Wouters J. Comparison of MASTER and AUDERA for measurement of auditory steady-state responses. Int J Audiol. 2005;44: 244–253.

11. Emara AA, Gabr TA. Auditory steady state response in auditory neuropathy. J Laryngol Otol. 2010;124: 950–956. doi:10.1017/S0022215110000630

12. British Society of Audiology. Practice Guidance: Auditory Steady State Response (ASSR) Testing. 2023. Available: https://www.thebsa.org.uk/wp-content/uploads/2022/04/OD104-115-Practice-Guidance-for-ASSR-Testing-1.pdf

13. Rance G. The Auditory Steady-State Response: Generation, Recording, and Clinical Application. Plural Publishing; 2008.

14. Korczak P, Smart J, Delgado R, Strobel TM, Bradford C. Auditory steady-state responses. J Am Acad Audiol. 2012;23: 146–170. doi:10.3766/jaaa.23.3.3

15. Gorga MP, Reiland JK, Beauchaine KA. Auditory brainstem responses in a case of high-frequency conductive hearing loss. J Speech Hear Disord. 1985;50: 346–350. doi:10.1044/jshd.5004.346

16. Prieve BA, Feeney MP, Stenfelt S, Shahnaz N. Prediction of conductive hearing loss using wideband acoustic immittance. Ear Hear. 2013;34 Suppl 1: 54S-59S. doi:10.1097/AUD.0b013e31829c9670

17. Prieve BA, Vander Werff KR, Preston JL, Georgantas L. Identification of conductive hearing loss in young infants using tympanometry and wideband reflectance. Ear Hear. 2013;34: 168–178. doi:10.1097/AUD.0b013e31826fe611

18. Dillon H. Hearing aids. 2nd ed. Sydney: Boomerang Press; 2012.

19. Tlumak AI, Rubinstein E, Durrant JD. Meta-analysis of variables that affect accuracy of threshold estimation via measurement of the auditory steady-state response (ASSR). Int J Audiol. 2007;46: 692–710. doi:10.1080/14992020701482480

20. Sininger YS, Hunter LL, Hayes D, Roush PA, Uhler KM. Evaluation of Speed and Accuracy of Next-Generation Auditory Steady State Response and Auditory Brainstem Response Audiometry in Children With Normal Hearing and Hearing Loss. Ear Hear. 2018;39: 1207–1223. doi:10.1097/AUD.0000000000000580

21. Michel F, Jørgensen KF. Comparison of threshold estimation in infants with hearing loss or normal hearing using auditory steady-state response evoked by narrow band CE-chirps and auditory brainstem response evoked by tone pips. Int J Audiol. 2017;56: 99–105. doi:10.1080/14992027.2016.1234719

22. Small SA, Stapells DR. Multiple auditory steady-state responses to bone-conduction stimuli in adults with normal hearing. J Am Acad Audiol. 2005;16: 172–183. doi:10.3766/jaaa.16.3.5

23. Small SA, Hatton JL, Stapells DR. Effects of bone oscillator coupling method, placement location, and occlusion on bone-conduction auditory steady-state responses in infants. Ear Hear. 2007;28: 83–98. doi:10.1097/01.aud.0000249787.97957.5b

24. Small SA, Stapells DR. Maturation of bone conduction multiple auditory steady-state responses. Int J Audiol. 2008;47: 476–488. doi:10.1080/14992020802055284

25. Page MJ, McKenzie JE, Bossuyt PM, Boutron I, Hoffmann TC, Mulrow CD, et al. The PRISMA 2020 statement: an updated guideline for reporting systematic reviews. BMJ. 2021;372: n71. doi:10.1136/bmj.n71

26. Page MJ, Moher D, Bossuyt PM, Boutron I, Hoffmann TC, Mulrow CD, et al. PRISMA 2020 explanation and elaboration: updated guidance and exemplars for reporting systematic reviews. BMJ. 2021;372: n160. doi:10.1136/bmj.n160

27. McKenzie J, Brennan S, Ryan R, Thomson H, Johnston R, Thomas J. Chapter 3: Defining the criteria for including studies and how they will be grouped for the synthesis [last updated August 2023]. In: Higgins J, Thomas J, Chandler J, Cumpston M, Li T, Page M, et al., editors. Cochrane Handbook for Systematic Reviews of Interventions version 65 Cochrane. Cochrane; 2024. Available: https://www.cochrane.org/authors/handbooks-and-manuals/handbook/current/chapter-03

28. Luke R, De Vos A, Wouters J. Source analysis of auditory steady-state responses in acoustic and electric hearing. Neuroimage. 2017;147: 568–576. doi:10.1016/j.neuroimage.2016.11.023

29. Weisz N, Lithari C. Amplitude modulation rate dependent topographic organization of the auditory steady-state response in human auditory cortex. Hear Res. 2017;354: 102–108. doi:10.1016/j.heares.2017.09.003

30. Haddaway NR, Grainger MJ, Gray CT. Citationchaser: A tool for transparent and efficient forward and backward citation chasing in systematic searching. Res Synth Methods. 2022;13: 533–545. doi:10.1002/jrsm.1563

31. Rohatgi A. Webplotdigitizer: Version 4.6. 2022. Available: https://automeris.io/WebPlotDigitizer/

32. Small SA, Hu N. Maturation of the occlusion effect: a bone conduction auditory steady state response study in infants and adults with normal hearing. Ear Hear. 2011;32: 708–719. doi:10.1097/AUD.0b013e31821de1b0

33. Casey K-A, Small SA. Comparisons of auditory steady state response and behavioral air conduction and bone conduction thresholds for infants and adults with normal hearing. Ear Hear. 2014;35: 423–439. doi:10.1097/AUD.0000000000000021

34. Casey K-A. Comparisons of auditory steady-state response and behavioural air- and bone-conduction thresholds in infants and adults with normal hearing. Master of Science, The University of British Columbia. 2012.

35. Wells G, Shea B, O'Connell D, Peterson J, Welch V, Losos M, et al. The Newcastle-Ottawa Scale (NOS) for assessing the quality of non-randomised studies in meta-analysis. 2018 [cited 1 Jul 2020]. Available: http://www.ohri.ca/programs/clinical_epidemiology/oxford.asp

36. Lo CK-L, Mertz D, Loeb M. Newcastle-Ottawa Scale: comparing reviewers' to authors' assessments. BMC Med Res Methodol. 2014;14: 45. doi:10.1186/1471-2288-14-45

37. Deeks JJ, Higgins JP, Altman DG. Chapter 10: Chapter 10: Analysing data and undertaking meta-analyses [last updated November 2024]. In: Higgins JP, Thomas J, Chandler J, Cumpston M, Li T, Page M, et al., editors. Cochrane Handbook for Systematic Reviews of Interventions version 65 Cochrane. Cochrane; 2024. Available: https://www.cochrane.org/authors/handbooks-and-manuals/handbook/current/chapter-10

38. Rücker G, Schwarzer G, Carpenter JR, Schumacher M. Undue reliance on I(2) in assessing heterogeneity may mislead. BMC Med Res Methodol. 2008;8: 79. doi:10.1186/1471-2288-8-79

39. R Core Team. R: A Language and Environment for Statistical Computing. R Foundation for Statistical Computing, Vienna, Austria. 2024. Available: https://www.R-project.org/

40. Haddaway NR, Page MJ, Pritchard CC, McGuinness LA. PRISMA2020: An R package and Shiny app for producing PRISMA 2020-compliant flow diagrams, with interactivity for optimised digital transparency and Open Synthesis. Campbell Systematic Reviews. 2022;18: e1230. doi:10.1002/cl2.1230

41. Balduzzi S, Rücker G, Schwarzer G. How to perform a meta-analysis with R: a practical tutorial. Evid Based Ment Health. 2019;22: 153–160. doi:10.1136/ebmental-2019-300117

42. Wickham H, Averick M, Bryan J, Chang W, McGowan LD, François R, et al. Welcome to the Tidyverse. Journal of Open Source Software. 2019;4: 1686. doi:10.21105/joss.01686

43. Schünemann H, Higgins J, Vist G, Glasziou P, Akl E, Skoetz N, et al. Chapter 14: Completing 'Summary of findings' tables and grading the certainty of the evidence [last updated August 2023]. In: Higgins J, Thomas J, Chandler J, Cumpston M, Li T, Page M, et al., editors. Cochrane Handbook for Systematic Reviews of Interventions version 65 Cochrane. Cochrane; 2024. Available: https://www.cochrane.org/authors/handbooks-and-manuals/handbook/current/chapter-14

44. Lins OG, Picton TW, Boucher BL, Durieux-Smith A, Champagne SC, Moran LM, et al. Frequency-specific audiometry using steady-state responses. Ear Hear. 1996;17: 81–96. doi:10.1097/00003446-199604000-00001

45. Small SA, Stapells DR. Multiple auditory steady-state response thresholds to bone-conduction stimuli in young infants with normal hearing. Ear Hear. 2006;27: 219–228. doi:10.1097/01.aud.0000215974.74293.b9

46. Small SA, Stapells DR. Normal ipsilateral/contralateral asymmetries in infant multiple auditory steady-state responses to air- and bone-conduction stimuli. Ear Hear. 2008;29: 185–198.

47. Swanepoel DW, Ebrahim S, Friedland P, Swanepoel A, Pottas L. Auditory steady-state responses to bone conduction stimuli in children with hearing loss. Int J Pediatr Otorhinolaryngol. 2008;72: 1861–1871. doi:10.1016/j.ijporl.2008.09.017

48. Ishida IM, Cuthbert BP, Stapells DR. Multiple auditory steady state response thresholds to bone conduction stimuli in adults with normal and elevated thresholds. Ear Hear. 2011;32: 373–381. doi:10.1097/AUD.0b013e318201c1e5

49. Ismaila N, El-Saiid E, El Sebaii A, Fadel H. Reliability of auditory steady-state response to bone conduction stimuli in assessing hearing loss in children. Egypt J Otolaryngol. 2016;32: 196–201. doi:10.4103/1012-5574.186532

50. John MS, Picton TW. MASTER: a Windows program for recording multiple auditory steady-state responses. Comput Methods Programs Biomed. 2000;61: 125–150.

51. John MS, Dimitrijevic A, Picton TW. Weighted averaging of steady-state responses. Clin Neurophysiol. 2001;112: 555–562. doi:10.1016/s1388-2457(01)00456-4

52. Cebulla M, Stürzebecher E, Wernecke KD. Objective detection of the amplitude modulation following response (AMFR). Audiology. 2001;40: 245–252. doi:10.3109/00206090109073118

53. Luts H, Wouters J. Hearing assessment by recording multiple auditory steady-state responses: the influence of test duration. Int J Audiol. 2004;43: 471–478. doi:10.1080/14992020400050060

54. British Society of Audiology. Recommended Procedure: Visual Reinforcement Audiometry. 2014. Available: https://www.thebsa.org.uk/wp-content/uploads/2014/06/OD104-37-Recommended-Procedure-Visual-Reinforcement-Audiometry-2014-1.pdf

55. Dhar S. Otoacoustic emissions: principles, procedures, and protocols. San Diego, California ; Plural Publishing, Inc.; 2012.

56. Picton TW, John MS. Avoiding electromagnetic artifacts when recording auditory steady-state responses. J Am Acad Audiol. 2004;15: 541–554.

57. Small SA, Stapells DR. Artifactual responses when recording auditory steady-state responses. Ear Hear. 2004;25: 611–623.

58. British Society of Audiology. Guidelines for the Early Audiological Assessment and Management of Babies Referred from the Newborn Hearing Screening Programme. 2021. Available: https://www.thebsa.org.uk/wp-content/uploads/2022/10/OD104-98-BSA-Practice-Guidance-Early-Assessment-Management.pdf

59. Rance G, Tomlin D. Maturation of auditory steady-state responses in normal babies. Ear Hear. 2006;27: 20–29. doi:10.1097/01.aud.0000191944.03797.5a

60. Sousa AC, Didoné DD, Sleifer P. Longitudinal Comparison of Auditory Steady-State Evoked Potentials in Preterm and Term Infants: The Maturation Process. Int Arch Otorhinolaryngol. 2017;21: 200–205. doi:10.1055/s-0036-1584888

61. Foxe JJ, Stapells DR. Normal infant and adult auditory brainstem responses to bone-conducted tones. Audiology. 1993;32: 95–109. doi:10.3109/00206099309071860

62. Vander Werff KR, Prieve BA, Georgantas LM. Infant air and bone conduction tone burst auditory brain stem responses for classification of hearing loss and the relationship to behavioral thresholds. Ear Hear. 2009;30: 350–368. doi:10.1097/AUD.0b013e31819f3145

63. Israelsson K-E, Bogo R, Berninger E. Reliability in Hearing Threshold Prediction in Normal-Hearing and Hearing-Impaired Participants Using Mixed Multiple ASSR. J Am Acad Audiol. 2015;26: 299–310. doi:10.3766/jaaa.26.3.9