1   **Original Research**

2   **Title:**

3   **Comparative Study of Point-of-Care Hemoglobin Meters Compared with an Automated**

4   **Hematology Analyzer in Kumasi, Ghana**

5   **Running Title:** Accuracy of Point-of-Care Hemoglobin Meters

6

7   **Authors:**

8   Amankwaah Lordina[1], Acheampong George A.[2]

9   [1] Department of Laboratory Technology, Kumasi Technical University

10   [2]Department of Clinical Microbiology, Kwame Nkrumah University of Science and

11   Technology

12

13   ***Corresponding Author**

14   **Name:** Lordina Amankwaah

15   **Address:** Kumasi Technical University, Ghana

16   **Email:** amankwaahlaudina505@gmail.com

17

18

19

20   Source(s) of support: The authors received no funding for the study.

21   Conflicting Interest: There are no conflicting interests.

22

Abstract

Background:

Anemia is one of the most prevalent global health challenges, particularly among women of reproductive age and children. Accurate hemoglobin estimation is essential for diagnosing anemia, particularly in resource-limited settings where point-of-care devices are widely used.

Aim:

To evaluate the diagnostic accuracy of two point-of-care hemoglobin meters compared with a fully automated hematology analyzer.

Materials and Methods:

A comparative cross-sectional study was conducted among 100 participants at Aniniwaa Medical Centre, Kumasi. Hemoglobin levels were measured using a fully automated analyzer (reference method), Urit and Mission Hb meters. Agreement was assessed using Bland–Altman analysis, and diagnostic performance was evaluated using sensitivity and specificity.

Results:

The prevalence of anemia was 28% using the analyzer, compared to 60% and 64% using the Urit and Mission meters, respectively. Both meters demonstrated 100% sensitivity but lower specificities (Urit 55.6%; Mission 50.0%). Bland–Altman analysis indicated negative biases (Urit = –1.665 g/dL; Mission = –1.55 g/dL) indicating underestimation of hemoglobin levels.

Conclusion:

Hemoglobin (Hb) meters offer convenience and portability for field screening but overestimate anemia prevalence. The fully automated analyzer remains more accurate and reliable for diagnosing anemia in clinical settings.

Keywords: Hemoglobin, Anemia, Point-of-care testing, Hemoglobin Meter, Diagnostic Accuracy

47 **Introduction**

48 Hemoglobin is a protein in red blood cells that carries oxygen and is composed of a protein

49 called heme, which binds oxygen [1, 2]. Hemoglobin is formed in the red blood cells which is

50 produced in the bone marrow and is made up four polypeptide chains which are two alpha

51 and two beta chains in a normal healthy individual [3]. Hemoglobin estimation indirectly

52 measures the amount of iron in the blood to produce efficient red blood cells [4]. Methods

53 such as the hematocrit method, the drabkin's solution method, the fully automated method,

54 and point of care testing hemoglobin meters can be used in hemoglobin estimation.

55 Anemia is a widespread public health issue affecting people across all age groups and is often

56 described as low levels of blood, mainly red blood cells [5]. According to the World Health

57 Organization (WHO), anemia is defined as hemoglobin levels below 13.0 g/dL in adult males

58 and 12.0 g/dL in non-pregnant females and 11.0 g/dL in pregnant women [6]. Since the

59 clinical management of anemia depends on identifying both the presence and severity of

60 hemoglobin reduction, hemoglobin analysis is often the first-line investigation in suspected

61 cases.

62 In resource-limited settings, hemoglobin meters are mostly used as point-of-care testing for

63 rapid screening of anemia or hemoglobin levels in community and outreach settings because

64 of their portability. However, questions remain about their accuracy and reliability in

65 detecting anemia compared to the fully automated hematology analyzers, which are

66 considered the gold standard in clinical laboratories [7]. According to the National Institutes of

67 Health, hemoglobin (Hb) meters use light absorbance principles to measure Hb concentration

68 in blood, directly or after a chemical reaction that alters the hemoglobin's light-absorbing

69 properties, and light absorbed is proportional to the hemoglobin concentration [4].

70 The fully automated method uses hemoglobin analyzer to estimate hemoglobin levels by

71 using the principle of measuring changes in electrical resistance as cells pass through a

sensing area and analysing the scattering of light by cells [4]. Given the widespread use of

hemoglobin meters in primary healthcare and outreach programs, it is crucial to assess how

their results compare to those from fully automated analyzers.

In remote and low-resource areas where electricity access and laboratory infrastructure is

limited, health facilities rely on Hb meters as an alternative to fully automated analyzers [8]. It

is essential to assess the agreement between hemoglobin meters and the fully automated

analyzer to determine whether Hb meters are reliable methods of Hb estimation. This study

will provide valuable information about the performance of Hb meters potentially aiding in

better decision making for anemia screening programs in remote and low-resource settings.

Materials and Methods

Study Design and Setting

A comparative cross-sectional study was conducted at Aniniwaa Medical Centre, Kumasi,

Ghana to evaluate the agreement and diagnostic performance of point-of-care hemoglobin

meters against a fully automated hematology analyzer. Aniniwaa Medical Centre is a facility

equipped with both point-of-care and automated laboratory systems.

Study Population

A total of 100 participants referred for routine complete blood count (CBC) testing at

Aniniwaa Medical Centre were consecutively recruited.

Inclusion Criteria: Consenting patients undergoing CBC testing at Aniniwaa Medical

Centre.

Exclusion Criteria: Patients with visibly haemolysed or clotted samples.

Sample Collection and Analysis

Approximately 2 mL of blood was collected into EDTA tube and analyzed within 30 minutes

using the fully automated analyzer (Mindray hematology analyzer) and two portable

hemoglobin meters (URIT-12 and Mission Hb Meter), using venous aliquots according to

manufacturer guidelines. The hemoglobin values from all three methods were recorded in a

structured data collection sheet for further analysis.

Statistical Analysis

Data were analyzed using descriptive statistics. Diagnostic performance was assessed using

sensitivity and specificity. Agreement between methods was evaluated using Bland–Altman

analysis. A p-value < 0.05 was considered statistically significant.

Ethical Clearance

The Institute of Research, Innovation and Development in Kumasi Technical University

(KsTU) approved the study (Ref No.: IRID/EC2025/HS0041). Verbal informed consent was

obtained from all participants and no identifiable features were used in the study.

Results

Demographics of Participants

The study recruited a total of 100 participants comprising of 71% females and 29% males.

The median age was 26.0 (21.0 – 37.0) years. A summary of the demographic characteristics

is shown in **Table 1**.

Characteristics	Range (Cat)	Frequency	%
	0–18 (1)	16	16.0
	19–30 (2)	46	46.0
Age	31–50 (3)	27	27.0
	51–70 (4)	7	7.0
	71–90 (5)	4	4.0
Gender	Males	29	29.0

| | Females | 71 | 71.0 |

Table 1: Demographic characteristics of the study participants.

%= percentage. Cat = category

Prevalence of Anemia as classified by the three estimation methods

Participants with Hb levels less than 11.0 g/dL were classified as anemic. The fully

automated analyzer recorded the least prevalence of anemia with a rate of 28% followed by

the Urit method classification, 60% and the Mission Hb meter method recorded the highest

prevalence of 64% as shown in **Table 2**.

Table 2: Prevalence of anemia as classified by the methods of estimation

Methods	Anemia		Normal	
	Frequency	**Percentage (%)**	**Frequency**	**Percentage (%)**
Urit	60	60	40	40
Mission	64	64	36	36
FAM	28	28	72	72

121 *FAM= fully automated analyzer*

122 **Prevalence of anemia by age and gender using the three estimation methods**

123 The urit meter recorded a high prevalence of anemia in females (73.2%) than males (27.6%).

124 Anemia was prevalent among females in the age range 19–30 (40%) using the Urit meter.

125 **Figure 1** shows the Urit meter prevalence of anemia by age and gender.

126
127 **Figure 1: Prevalence of anemia by the Urit Hb meter**

128 The mission hemoglobin meter recorded a high prevalence of anemia in females (76.1%) and

129 anemia incidence was higher in the 31-50 age group (85.2%) as represented in **Figure 2**.

130
131 **Figure 2: Prevalence of anemia by the Mission Hb meter**

132

133 The fully automated analyzer (reference) found an anemia prevalence of 36.6% in females

134 and 6.9% in males, shown in **Figure 3.**

Figure 3: Prevalence of anemia by the fully automated method.

Assessing the Accuracy of the Hemoglobin meters to the Reference Analyzer Method

Tables 3 and 4 below demonstrates the accuracy of the Urit and Mission Hb meters to the

fully automated analyzer in classifying participants into normal and anemic. The Urit

overestimated anemia with a false positive of 32, sensitivity of 100% and a specificity of

55.6%.

Table 3: Comparison of anemia status by the Urit and the reference analyzer

		Fully automated analyzer (Reference)			
		Anemia	Normal	Total	p-value
Urit estimation	**Anemia**	28	32	60	**<0.001**
	Normal	0	40	40	
	Total	28	72	100	

145   *Compared using the Chi-square test, p<0.05 was considered statistically significant.*

147 The Mission Hb meter recorded a false positive of 36, a 100% sensitivity and showed a

148 specificity of 50% against the standard fully automated analyzer.

149 **Table 4:**

		Fully automated analyzer (Reference)			
		Anemia	Normal	Total	p-value
Mission estimation	Anemia	28	36	64	<0.001
	Normal	0	36	36	
	Total	28	72	100	

Comparison of anemia status by the Mission and the reference analyzer

Compared using Chi-Square test of independence, p<0.05 considered statistically significant.

Assessment of the Level of Agreement between the Hemoglobin meters and the

Reference Analyzer Method

The Bland-Altman plot was used to determine the level of agreement between the two testing

methods (Urit and Mission) and the reference method. The mean bias for Urit meter

hemoglobin readings was -01.665 g/dL, with limits of agreement ranging from -4.443 to

1.123 g/dL as shown in **Figure 4.** The Mission hemoglobin meter recorded similar negative

bias, with a mean bias of -1.55 g/dL. The lower and the upper limits of agreement were -4.50

and 1.40 g/dL respectively (**Figure 5**).

Figure 4: Mean (Urit vs FAM) and Difference (Urit-FAM)

Bias= -1.665

Lower limit of agreement (LLOA) = -4.433

Upper limit of agreement (ULOA) = 1.123

Figure 5: Mean (Mission vs FAM) and Difference (Mission-FAM)

Bias= -1.55

LLOA= -4.50

ULOA= 1.40

Discussion

This study compared the hemoglobin concentrations obtained from the Urit Hb meter,

Mission Hb meter, and the Mindray Fully Automated analyzer from 100 participants. The

automated analyzer classified fewer participants as anemic compared to the point-of-care

(POC) Hb meters. This suggests that the Hb meters may be overestimating the prevalence of

anemia, potentially due to device-specific calibration differences or sample type.

Similar findings by [8] in Ghana observed that POC devices tend to report lower hemoglobin

values compared to automated hematology analyzers in malaria-endemic populations, which

led to the overestimation of anemia prevalence. In India, [9] found discrepancies between Hb

meters and laboratory analyzers, noting that Bland–Altman plots revealed systematic biases.

More recently, [10] confirmed that while Hb meters remain useful for screening in field

surveys, they require strict validation against venous samples analyzed on automated systems

to avoid inflated prevalence estimates.

From a public health perspective, these discrepancies are important. If POC devices

consistently overestimate anemia prevalence, program planners may allocate resources

inefficiently or misinterpret the severity of anemia in a population. However, their portability

makes them indispensable tools in rural settings where automated analyzers are unavailable.

Prevalence of Anemia by gender and age differences

Anemia was classified based on a cut value of <11.0 g/dL. Across methods, females recorded

higher anemia prevalence than males with Urit, Mission, and automated analyzer classifying

73.2%, 76.1% and 36.6% of females as anemic respectively. These gender disparities are

biologically plausible. In women of reproductive age, iron requirements are elevated due to

menstrual blood losses, pregnancies, and lactation [11]. Pregnancy in particular imposes

increased demands for iron to support fetal growth and expansion of maternal blood volume,

predisposing to anemia when dietary intake or supplementation is inadequate [12]. In Ghana,

studies have repeatedly highlighted the burden of anemia in women with [13] reporting that

anemia prevalence among Ghanaian women of reproductive age exceeded 40%, consistent

with WHO's classification of a severe public health problem.

Using the Urit meter, anemia was most prevalent among younger adults (19–30 years),

accounting for 40% of all anemic cases. The Pearson Chi-Square test revealed a statistically

significant difference with (X^2 = 17.881, p<0.001). By contrast, the Mission Hb meter

suggested that the highest prevalence occurred among participants aged 31–50 years, with

85.2% classified as anemic in this group (X^2= 15.446, p<0.001). The fully automated

analyzer reported lower overall prevalence (28%) but similarly identified higher rates among

reproductive-aged women (57.1%) (X^2= 9.023, p=0.003).

These findings can be understood in relation to life-stage iron requirements and

comorbidities. In younger women, menstruation and pregnancy are primary drivers of anemia

[14]. This aligns with the Urit meter's results, where nearly three-quarters of females were

classified as anemic. In middle-aged adults (31–50 years), chronic inflammation, liver

disease, or parasitic infections (such as hookworm or schistosomiasis) may contribute to

anemia in these age groups [15] as suggested by the Mission Hb meter. Globally, WHO

estimates that the burden of anemia peaks in men and women of reproductive age, with

prevalence decreasing slightly after menopause in women and remaining consistently low in

adult men [11]. The current findings mirror these trends but also suggest device-specific

variability in classifying anemia by age group.

Accuracy of the Hemoglobin meters to the Reference Analyzer Method

The Urit overestimated anemia with a false positive of 32, sensitivity of 100% and a

specificity of 55.6%. This difference was statistically significant as shown by the Chi-Square

223 test, ($X^2 = 25.926$, $p < 0.05$). Similarly, the Mission Hb meter recorded a false positive of 36, a

224 100% sensitivity and showed a specificity of 50% against the standard fully automated

225 analyzer. The Pearson Chi-Square test showed a statistically significant difference, ($X^2 =$

226 21.875, $p < 0.001$).

227 The assessment of accuracy revealed that both the Urit and Mission Hb meters demonstrated

228 perfect sensitivity (100%) compared with the fully automated analyzer. In practical terms,

229 none of the true anemic cases were missed, which is advantageous in clinical and field

230 screening because it minimizes the risk of failing to detect individuals requiring intervention.

231 However, the Hb meters showed poor specificity suggesting that while the meters are highly

232 reliable in detecting anemia when present, they tend to overestimate its prevalence by

233 misclassifying non-anemic individuals as anemic. Such overestimation was further supported

234 by the statistically significant Chi-square results ($p < 0.05$), confirming that the discrepancy

235 between the Hb meters and the reference analyzer was not due to chance.

236 This pattern is consistent with earlier studies. [9] reported that point-of-care Hb meters often

237 over diagnose anemia compared with automated analyzers. In Ghana, [8] similarly observed

238 that portable Hb meters tended to overestimate anemia prevalence in malaria-endemic areas,

239 cautioning against their use as stand-alone diagnostic tools in epidemiological surveys.

240 Recent evaluations by [10], also highlighted same limitation, emphasizing the importance of

241 validating point-of-care devices against venous samples using Bland–Altman methods.

242 The high sensitivity of Hb meters makes them suitable for screening programs where the

243 priority is to detect all potential anemic individuals. However, the low specificity can inflate

244 prevalence estimates and potentially lead to unnecessary supplementation or misallocation of

245 resources. This limitation underscores the importance of using Hb meters as preliminary

246 screening tools rather than definitive diagnostic methods, particularly in research or policy

247 contexts where accurate prevalence data are required.

248   **Level of Agreement between Urit, Mission, and the Reference Automated Analyzer**

249   The Bland–Altman analysis provided further insight into the level of agreement between the

250   portable Hb meters (Urit and Mission) and the fully automated analyzer. The Hb meters

251   demonstrated a negative mean bias, with the Urit meter showing a bias of –1.665 g/dL and

252   the Mission meter –1.55 g/dL. This consistent underestimation indicates that, on average, the

253   Hb meters reported lower hemoglobin values than the reference analyzer.

254   The limits of agreement (LOA) for the Urit meter (–4.433 to 1.123 g/dL) and the Mission

255   meter (–4.50 to 1.40 g/dL) were relatively wide indicating that individual readings between

256   the meters and the reference analyzer could differ by as much as ±4 g/dL. Such wide limits

257   raise concerns about reliability at the individual diagnostic level.

258   These findings align with prior studies that reported systematic underestimation by point-of-

259   care devices. [10] found that both HemoCue and Mission Hb meters demonstrated mean

260   negative bias when compared with automated hematology analyzers, with similar wide

261   LOAs. In Ghana, [8] also noted poor agreement between Hb meters and laboratory analyzers,

262   cautioning that while useful for screening, they were less appropriate for precise diagnostics.

263   The negative bias and wide LOA imply that Hb meters overestimate the prevalence of

264   anemia. An individual classified as anemic by a portable meter may not always be confirmed

265   as such by the reference analyzer, highlighting the risk of false positives. Nevertheless, the

266   relatively consistent bias across both Urit and Mission meters shows that systematic

267   calibration adjustments might improve their performance in local contexts.

268   The Bland–Altman analysis demonstrates that although Hb meters agree reasonably well with

269   the reference method in terms of trend, their underestimation and wide LOA reduce their

270   accuracy for individual-level diagnosis. Confirmatory testing with automated analyzers

271   remains essential, especially in clinical decision-making and epidemiological reporting.

272

273 **Conclusion**

274 The Hb meters overestimate anemia prevalence confirmed by their high sensitivity but low

275 specificity, The Bland–Altman analysis further showed that both Hb meters consistently

276 underestimated Hb values compared to the reference analyzer indicating that, while useful for

277 screening, the Hb meters may not provide sufficient accuracy for individual diagnostic

278 purposes without confirmatory testing. While Hb meters provide portable and cost-effective

279 results, their limited accuracy compared to automated analyzers can affect clinical decision-

280 making and public health reporting.

281

282

283

284

285

286 **References**

287 1. Ahmed MH, Ghatge MS, Safo MK. Hemoglobin: Structure, function and allostery.

288 Subcell Biochem. 2020;94:345-382. doi:10.1007/978-3-030-41769-7_14

289 2. Farid Y, Bowman NS, Lecat P. Biochemistry, hemoglobin synthesis. In: StatPearls

290 [Internet]. Treasure Island (FL): StatPearls Publishing; 2025. Available from:

291 https://www.ncbi.nlm.nih.gov/books/NBK536912/

292 3. Marengo-Rowe AJ. Structure-function relations of human hemoglobins. Baylor Univ

293 Med Cent Proc. 2006;19(3):239-245.

294 4. Whitehead RD Jr, Mei Z, Mapango C, Jefferds MED. Methods and analyzers for

295 haemoglobin measurement in clinical laboratories and field settings. Ann N Y Acad

296 Sci. 2019;1450(1):147-171. doi:10.1111/nyas.14124

297 5. Gallagher PG. Anemia in the pediatric patient. Blood. 2022;140(6):571-593.

6. World Health Organization. Hemoglobin concentrations for the diagnosis of anemia
and assessment of severity. Geneva: WHO; 2011. Available from:
https://www.who.int/vmnis/indicators/hemoglobin.pdf

7. Lewis SM, Bain BJ, Bates I. Dacie and Lewis practical haematology. 12th ed.
London: Elsevier; 2016.

8. Nkrumah B, Nguah SB, Sarpong N, Dekker D, Idriss A, May J, et al. Hemoglobin
estimation by the HemoCue® portable hemoglobin photometer in a resource-poor
setting. BMC Med. 2011;9:16. doi:10.1186/1741-7015-9-16

9. Ranjan V, Rana R, Khillan K, Chauhan K. A comparative quality evaluation of point-
of-care methodology for testing hemoglobin in blood donors by two different
technologies. Curr Med Res Pract. 2020;10(3):90-92.

10. Saengnipanthkul S, Tawiangnan C, Sirikarn P, Supasai S, Sitthikarnkha P,
Techasatian L, et al. Accuracy of noninvasive total hemoglobin measurement for
detecting anemia in infants: A cross-sectional study. BMJ Open.
2024;14(12):e083356.

11. World Health Organization. Hemoglobin concentrations for the diagnosis of anemia
and assessment of severity. Geneva: WHO; 2023. Available from:
https://www.who.int/publications/i/item/WHO-NMH-NHD-MNM-11.1

12. Smith ER, Hoodbhoy Z, Hotwani A, Jehan F, Khan A, Nisar I, et al. Protocol for the
redefining maternal anemia in pregnancy and postpartum (ReMAPP) study. PLoS
One. 2021;16(7):e0321943.

13. Ewusie JE, Ahiadeke C, Beyene J, Hamid JS. Prevalence of anemia among under-5
children in Ghana. BMC Public Health. 2014;14:626.

14. World Health Organization. Anaemia in women and children. Geneva: WHO; 2021.
Available from: https://www.who.int/data/nutrition/nlis/info/anaemia

15. Abizari AR, Ali Z. Dietary patterns and associated factors of schooling Ghanaian

adolescents. J Health Popul Nutr. 2019;38(1):5. doi:10.1186/s41043-019-0162-8