TITLE PAGE

Full Title: Machine Learning for Pattern Detection in Cochlear Implant FDA

Adverse Event Reports

Short running title: Cochlear Implant Adverse Event Patterns

Authors:

Matthew G. Crowson MD FRCSC^{1,2}

Amr Hamour MD MBT¹

Vincent Lin MD FRCSC¹

Joseph M. Chen MD FRCSC¹

Timothy C. Y. Chan, PhD²

1 Department of Otolaryngology-HNS, Sunnybrook Health Sciences Center, University of Toronto, Toronto, Ontario

2 Department of Mechanical & Industrial Engineering, University of Toronto, Toronto, Ontario

Key words: cochlear implants; machine learning; adverse events

Meeting Submission: This paper was accepted for presentation at the ACI 'Ci2020' Annual Meeting, Orlando, Florida, March 18-22.

Manuscript Submission:

- (1) Each of the authors indicated above have contributed to, read and approved this manuscript.
- (2) **FINANCIAL DISCLOSURE:** no authors have disclosures related to this manuscript.
- (3) **CONFLICT DISCLOSURE:** no authors have conflicts related to this manuscript.
- (4) In consideration of the journal reviewing and editing my submission, the authors undersigned transfers, assigns and otherwise conveys all copyright ownership in the event that such work is published.

Corresponding Author:

Matthew G. Crowson MD MPA FRCSC
Sunnybrook Health Sciences Center
University of Toronto
matt.crowson@mail.utoronto.ca
Toronto, Ontario, Canada

ABSTRACT

Importance: The United States Food & Drug Administration (FDA) passively monitors medical device performance and safety through submitted medical device reports (MDRs) in the Manufacturer and User Facility Device Experience (MAUDE) database. These databases can be analyzed for patterns and novel opportunities for improving patient safety and/or device design.

Objectives: The objective of this analysis was to use supervised machine learning to explore patterns in reported adverse events involving cochlear implants.

Design: The MDRs for the top three CI manufacturers by volume from January 1st 2009 to August 30th 2019 were retained for the analysis. Natural language processing was used to measure the importance of specific words. Four supervised machine learning algorithms were used to predict which adverse event narrative description pattern corresponded with a specific cochlear implant manufacturer and adverse event type - injury, malfunction, or death.

Setting: U.S. government public database.

Participants: Adult and pediatric cochlear patients.

Exposure: Surgical placement of a cochlear implant.

Main Outcome Measure: Machine learning model classification prediction

accuracy (% correct predictions).

Results: 27,511 adverse events related to cochlear implant devices were submitted to the MAUDE database during the study period. Most adverse events involved patient injury (n = 16,736), followed by device malfunction (n = 10,760), and death (n = 16). Submissions to the database were dominated by Cochlear Corporation (n = 13,897), followed by MedEL (n = 7,125), and Advanced Bionics (n = 6,489). The random forest, linear SVC, naïve Bayes and logistic algorithms were able to predict the specific CI manufacturer based on the adverse event narrative with an average accuracy of 74.8%, 86.0%, 88.5% and 88.6%, respectively.

Conclusions & Relevance: Using supervised machine learning algorithms, our classification models were able to predict the CI manufacturer and event type with high accuracy based on patterns in adverse event text descriptions.

Key words: cochlear implants; machine learning; adverse events

Level of evidence: 3

INTRODUCTION

The United States Food & Drug Administration (FDA) passively monitors medical device performance and safety through submitted medical device reports (MDRs) in the Manufacturer and User Facility Device Experience (MAUDE) database.^{1,2} The MDRs contain the type of adverse event (i.e., injury, device malfunction or death), the device, manufacturer and a free text narrative of the event. A theoretical advantage of using a centralized device database over a single-institution's experience is a wider breadth and depth of incidents including all manufacturers of approved medical devices in all fields of medicine and surgery. Frequently cited analyses of cardiovascular devices (e.g., catheters and stents), obstetrical devices, and robots have produced insights into the nature and frequency of device-related adverse events using the MAUDE database.^{3–9}

Machine learning algorithms are a subdomain of artificial intelligence and are used to 'learn' patterns in large, heterogeneous data to generate predictions. These predictions can augment human decision making and can uncover novel relationships in large datasets that are not observable using traditional statistical techniques. Natural language processing (NLP) – a subset of artificial intelligence – is concerned with the autonomous deduction of meaning and intent from human language using algorithms. Freeform narrative text is ubiquitous in clinical and research settings, and significant recent interest has been established in applying NLP coupled with machine learning to unlock embedded patterns and insights. Successful efforts using NLP with narrative clinical data have been applied to structure narratives within electronic health records,^{10–20} mining drug

adverse effect incidence from social media platforms,²¹ pharmaceutical and vaccine safety surveillance,^{22–30} automatic identification of complications during procedures,^{31–33} and patient safety event detection.^{34–39}

The MAUDE MDR database has been analyzed to assess cochlear implant (CI) electrode migration,⁴⁰ and temporal trends in CI device failure, surgical site infections, and other adverse events related to CI.⁴¹ However, as of the writing of this manuscript, no prior analyses used machine learning and natural language processing techniques to analyze the narrative text in CI adverse event reports. Considering the different methodologic approach inherent to machine learning, we hypothesized that we may be able provide additional insights in the study of CI adverse event reports. Specifically, the primary objective of this analysis was to use supervised machine learning to predict the manufacturer of a CI device based on the adverse event narrative text. A secondary objective was to explore patterns in the narrative text by the type of adverse event. By examining patterns in adverse events related to specific manufacturers and types of adverse events, we hoped to mine for novel opportunities for device improvement and patient safety.

METHODS

This study involved publicly available data, so a review through the Sunnybrook Health Sciences Center institutional review board (IRB) was not undertaken.

Manufacturer and User Facility Device Experience (MAUDE) Database. We mined the MAUDE database for all available CI-related MDRs submitted to the FDA over the period from January 1st 2009 to August 30th 2019. The MAUDE database contains event type, date, device name, event narrative description and manufacturer in each MDR. The narrative description in the MDRs for the top three CI manufacturers by volume of adverse events were retained for the analysis.

Data Processing. Free-text narrative data needs to be processed before it can be used as input for a machine learning algorithm in natural language processing tasks. The narrative text from individual adverse event reports was cleaned by removing blank rows, standardizing text case, removing common ‘stop-words’ (e.g., "a," "and," "but," "how," "or," "the," "who," and "what") and non-numeric characters and punctuation. We lemmatized individual words to remove inflectional endings and retain the ‘base’ dictionary form of the word. After the text data were cleaned, we used term frequency-inverse document frequency (TF-IDF) natural language processing to measure the importance of specific words in each adverse event proportional to its frequency within the entry and the entire database. The importance value for words are converted into a specific

numerical vector values (i.e., vectorization) that allow the comparison between specific words.

Machine Learning Algorithms. After the text data were prepared, we deployed a five-fold cross validation. Cross-validation is a resampling procedure used to evaluate machine learning models. In five-fold cross validation, the dataset is randomly sampled to generate five different folds whereby each fold is used a testing fold during an iterative training-testing process. In each fold, we used four subsets (80% of the total dataset) to train the model and the remaining subset (20% of the total dataset) to test model performance on data unseen by the model during training. This process was repeated four additional times, and the final accuracy was the aggregate accuracy over all five of the test sets.

The analytic goal of the machine learning algorithms was to classify the narrative text from individual adverse events as i) belonging to one of the three manufacturers, and ii) corresponding to the type of adverse event report – injury, malfunction, or death – in separate analyses (**Figure 1**). Based on their successful use in other applications of natural language processing text classification, we deployed a random forest classifier (200 estimators with a maximum depth of 3), linear support vector classifier (SVC), multinomial Naïve Bayes (mNB) classifier, and multinomial logistic regression classifier (Python *scikit-learn* 0.21.3 package).^{42,43} Several metrics were calculated to evaluate the individual classification performance for each model including precision, recall, and the F_1 score. The precision is the ratio of true positives (TP) over the sum of the true positives (TP) and false positives (FP): $TP/(TP+FP)$. The model recall is

the ratio of TP over the sum of TP and false negatives (FN): $TP/(TP+FN)$. The F_1 score is the weighted average of the precision and recall: $2 \times (\text{precision} \times \text{recall}) / (\text{precision} + \text{recall})$.

Model Interpretation. To illustrate the utility of the machine learning algorithms in performing classification of adverse event narrative text, we asked the best performing algorithm to predict the manufacturer from novel narrative text generated by the authors as examples of common device-related adverse events. The models were interpreted using the interpretable model-agnostic explanations (LIME) package (*version 0.1.1.36*).⁴⁴ The LIME approach provides a qualitative understanding between the input variables and the output generated by the classifier algorithm. We deployed LIME to provide a probability (%) of the text sample belonging to a given manufacturer along with the most important text features that account for each prediction. The ELI5 package (*version 0.9.0*) was used to compute the weights of the most common positive and negative predictors for each manufacturer. The numerical weights represent how much each word contributed to the final manufacturer category prediction.

Hardware & Software Environment. The model training and analyses were completed on a custom-built desktop computer with an AMD Ryzen Threadripper 1950X processor, 32GB of read-only memory (RAM), and a GeForce RTX 2070 8GB graphics card. Data processing and algorithm computation was completed using the Python version 3.6 environment running in an Ubuntu Linux operating system.

RESULTS

27,512 cochlear implant adverse events from the top three most common manufacturers were submitted to the MAUDE database from January 1st, 2009 to August 30th, 2019. Submissions to the database were led by Cochlear Corporation (n = 13,897; 50.5%), followed by Med-EL (n = 7,125; 25.9%), and Advanced Bionics (n = 6,489; 23.6%). Most adverse events involved patient injury (n = 16,736; 60.8%), followed by device malfunction (n = 10,760; 39.1%), and death (n = 16; 0.1%).

Machine learning algorithm classification accuracy performance. Our final adverse event report narrative database contained 1,275,668 words with a vocabulary size of 7,858 unique words, and a maximum adverse event report narrative length of 477 words. Based on the free text narrative data alone, the random forest, linear SVC, naïve Bayes and logistic algorithms were able to predict the specific CI manufacturer based on the adverse event narrative with an average accuracy of 74.8%, 86.0%, 88.5% and 88.6%, respectively, after five-fold cross validation (**Figure 2**). For the adverse event type classification task, the random forest, linear SVC, naïve Bayes and logistic algorithms produced an average accuracy of 69.7%, 84.0%, 85.4% and 85.5%, respectively (**Figure 3**).

Computed classification metrics demonstrated the logistic regression model had the highest precision for MedEL and the highest recall for Cochlear Corporation (**Table 1**). In predicting the adverse event type, the linear SVC model achieved the highest precision for 'injury', whereas the logistic regression achieved the

highest precision for ‘malfunction’. Several positive and negative predictors for each manufacturer and adverse event type were identified (**Table 2**). The logistic regression model was used to generate predictions on novel text examples for specific CI manufacturers (**Table 3**). The most influential positive predictors appear to represent the style in how the manufacturers structured their report. For instance, MDRs submitted by Advanced Bionics were predicted using words such as “*recipient*” and “*pt*” (assumed to be short form for ‘patient’). Whereas “*clinic*” and “*audiologist*” were more predictive of Cochlear corporation MDRs. While these findings are not useful in identifying manufacturer-specific trends in adverse events, it is interesting that the algorithm identified an etymological style of reporting unique to each manufacturer.

DISCUSSION

The objectives of this analysis were to apply natural language processing to the MAUDE MDR database to predict specific CI manufacturers and adverse event types based on the narrative text explanation of the adverse event. All the machine learning algorithms deployed against the MAUDE CI dataset achieved relatively high classification accuracy. While the model predictions themselves may hold little clinical utility, the insights derived from the model interpretations are useful in exploring the characteristics and quality of the narratives in the adverse reports. Interestingly, the most predictive words for each of the manufacturers resembled more of the writing style in the reporting adverse events, and less of patterns in specific clinically relevant adverse events. The narrative words predictive of specific adverse event types were intuitive and provided face validity support to our approach.

A recurring criticism of machine learning is that the algorithms are a ‘black box’ that lack practical interpretability. One of the benefits of using non-neural network algorithms, as we have chosen, is that analytic methods to interpret how the algorithm arrived at a specific prediction are available. We interpreted the logistic regression model’s predictions by tabulating the most predictive words calculated using a weighted measure of importance for each manufacturer and adverse event type. On our analysis of the most important narrative words, we believe the algorithm picked up on style of narrative writing rather than specific clinical concepts. Moreover, considering accuracy was high, we hypothesize that the

manufacturers MDR narratives were internally consistent through the use of templates, report writing training, and/or the same personnel.

To further interrogate our algorithms' ability to mine clinically useful trends in adverse event reports, we predicted the manufacturer base on our author-generated short adverse event narratives. In our examples, MedEL was predicted to be associated with the narratives describing CI dislodgement and circuit failure. For our example depicting electrode migration, the classification probability was approximately equal between MedEL and Advanced Bionics. Further analysis of these specific manufacturer adverse event relationships could be validated by cross-referencing specific model outcomes in independently published reports if they were to be made available. Because of the lack of standardization and weighting in reporting in the MAUDE database, we must strongly emphasize that we cannot make inferences or generalize these or other adverse events to a specific manufacturer.

Our study has important limitations that need mentioning. The most pressing limitation of our analysis is rooted in the circumstance that the MAUDE database is composed of non-standardized manufacturer-supplied adverse event reports. We cannot independently verify the accuracy of adverse events reports, nor assume that the adverse events are reported in a consistent manner. We assume that that the classification of the adverse event type is the appropriate label for each adverse event report. We cannot generalize on the frequency of specific adverse event types attributed to manufacturers and their devices. In particular, there are relatively few adverse event reports describing 'death' and

thus our algorithms may fail to generalize to mortality reports when provided with new data. CI investigators have previously called for improvements to the adverse event reporting for CIs in the MAUDE database as well as guidelines for more specific detail in 'who' and 'what' should be contained within the reports.⁴⁵ Anecdotal reports from manufacturers' adverse events reporting compliance staff suggest that a standardized reporting system is under development with all three major CI manufacturers so that the adverse event report data will be more uniform in format and quality. We welcome this initiative that will enhance the quality and utility of the MAUDE database for future analyses.

Using supervised machine learning algorithms, our classification model was able to accurately predict the CI manufacturer based on adverse event text descriptions. From the interpretation of the model outputs, we believe the algorithms were able to accurately determine the adverse event based on the reporting style of the narrative. While we did not observe novel clinical concepts with specific association to manufacturers or adverse event types, we believe this shortcoming is less a function of our methodologic approach. Our attempt to derive clinically significant patterns in adverse events from this database was limited, in part, by the quality of the data inherent to the MAUDE database. Moreover, it is plausible that the pooled manufacturers adverse event data does not have sufficient granularity for which to tease out small differences between manufacturers. Compared to single institution cohort analyses, this 'big data' approach to an independent, national database may allow for further analysis of previously undiscovered or unreported trends in adverse events with specific manufacturers' devices. Further work is underway to refine our approach to

explore patterns at the device model level. We emphatically encourage efforts to standardize the CI database reporting methods to increase the quality of subsequent analyses.

ACKNOWLEDGEMENTS

This work is supported in part by a 2018 Connaught Global Challenge Award from the University of Toronto.

REFERENCES

- 1 Ensign LG, Cohen KB. A Primer to the Structure, Content and Linkage of the FDA's Manufacturer and User Facility Device Experience (MAUDE) Files. *EGEMS (Wash DC)*. 2017;5(1):12. doi:10.5334/egems.221.
- 2 Gurtcheff SE. Introduction to the MAUDE database. *Clin Obstet Gynecol*. 2008;51(1):120-123. doi:10.1097/GRF.0b013e318161e657.
- 3 Andonian S, Okeke Z, Okeke DA, et al. Device failures associated with patient injuries during robot-assisted laparoscopic surgeries: a comprehensive review of FDA MAUDE database. *Can J Urol*. 2008;15(1):3912-3916.
- 4 Haber K, Hawkins E, Levie M, Chudnoff S. Hysteroscopic morcellation: review of the manufacturer and user facility device experience (MAUDE) database. *J Minim Invasive Gynecol*. 2015;22(1):110-114. doi:10.1016/j.jmig.2014.08.008.
- 5 Delaney JW, Li JS, Rhodes JF. Major complications associated with transcatheter atrial septal occluder implantation: a review of the medical literature and the manufacturer and user facility device experience (MAUDE) database. *Congenit Heart Dis*. 2007;2(4):256-264. doi:10.1111/j.1747-0803.2007.00107.x.
- 6 Friedman DCW, Lendvay TS, Hannaford B. Instrument Failures for the da Vinci Surgical System: a Food and Drug Administration MAUDE Database Study. *Surg Endosc*. 2013;27(5):1503-1508. doi:10.1007/s00464-012-2659-8.
- 7 Gurtcheff SE, Sharp HT. Complications associated with global endometrial ablation: the utility of the MAUDE database. *Obstet Gynecol*. 2003;102(6):1278-1282. doi:10.1016/j.obstetgynecol.2003.07.007.
- 8 Hafiz AM, Kalra A, Ramadan R, et al. Clinical or Symptomatic Leaflet Thrombosis Following Transcatheter Aortic Valve Replacement: Insights from the U.S. FDA MAUDE Database. *Structural Heart*. 2017;1(5-6):256-264. doi:10.1080/24748706.2017.1366086.
- 9 Everett KD, Conway C, Desany GJ, et al. Structural Mechanics Predictions Relating to Clinical Coronary Stent Fracture in a 5 Year Period in FDA MAUDE Database. *Ann Biomed Eng*. 2016;44(2):391-403. doi:10.1007/s10439-015-1476-3.
- 10 Hripcsak G, Friedman C, Alderson PO, DuMouchel W, Johnson SB, Clayton PD. Unlocking clinical data from narrative reports: a study of natural language processing. *Ann Intern Med*. 1995;122(9):681-688. doi:10.7326/0003-4819-122-9-199505010-00007.
- 11 Hyun S, Johnson SB, Bakken S. Exploring the ability of natural language processing to extract data from nursing narratives. *Comput Inform Nurs*. 2009;27(4):215-23; quiz 224-5. doi:10.1097/NCN.0b013e3181a91b58.
- 12 Kreimeyer K, Foster M, Pandey A, et al. Natural language processing systems for capturing and standardizing unstructured clinical information: A systematic review. *Journal of Biomedical Informatics*. 2017;73:14-29. doi:10.1016/j.jbi.2017.07.012.

- 13 Liao KP, Cai T, Savova GK, et al. Development of phenotype algorithms using electronic medical records and incorporating natural language processing. *BMJ*. 2015;350:h1885. doi:10.1136/bmj.h1885.
- 14 Marafino BJ, Park M, Davies JM, et al. Validation of Prediction Models for Critical Care Outcomes Using Natural Language Processing of Electronic Health Record Data. *JAMA Netw Open*. 2018;1(8):e185097. doi:10.1001/jamanetworkopen.2018.5097.
- 15 Ohno-Machado L. Realizing the full potential of electronic health records: the role of natural language processing. *J Am Med Inform Assoc*. 2011;18(5):539. doi:10.1136/amiajnl-2011-000501.
- 16 Ohno-Machado L, Nadkarni P, Johnson K. Natural language processing: algorithms and tools to extract computable information from EHRs and from the biomedical literature. *J Am Med Inform Assoc*. 2013;20(5):805. doi:10.1136/amiajnl-2013-002214.
- 17 Savova GK, Masanz JJ, Ogren PV, et al. Mayo clinical Text Analysis and Knowledge Extraction System (cTAKES): architecture, component evaluation and applications. *J Am Med Inform Assoc*. 2010;17(5):507-513. doi:10.1136/jamia.2009.001560.
- 18 Sheikhalishahi S, Miotto R, Dudley JT, Lavelli A, Rinaldi F, Osmani V. Natural Language Processing of Clinical Notes on Chronic Diseases: Systematic Review. *JMIR Med Inform*. 2019;7(2):e12239. doi:10.2196/12239.
- 19 Walker AM, Zhou X, Ananthakrishnan AN, et al. Computer-assisted expert case definition in electronic health records. *Int J Med Inform*. 2016;86:62-70. doi:10.1016/j.ijmedinf.2015.10.005.
- 20 Weng W-H, Waghlikar KB, McCray AT, Szolovits P, Chueh HC. Medical subdomain classification of clinical notes using a machine learning-based natural language processing approach. *BMC Med Inform Decis Mak*. 2017;17(1):155. doi:10.1186/s12911-017-0556-8.
- 21 Bian J, Topaloglu U, Yu F. Towards Large-scale Twitter Mining for Drug-related Adverse Events. *SHB12 (2012)*. 2012;2012:25-32. doi:10.1145/2389707.2389713.
- 22 Baer B, Nguyen M, Woo EJ, et al. Can Natural Language Processing Improve the Efficiency of Vaccine Adverse Event Report Review? *Methods Inf Med*. 2016;55(2):144-150. doi:10.3414/ME14-01-0066.
- 23 Carrell DS, Cronkite D, Palmer RE, et al. Using natural language processing to identify problem usage of prescription opioids. *Int J Med Inform*. 2015;84(12):1057-1064. doi:10.1016/j.ijmedinf.2015.09.002.
- 24 Gurulingappa H, Mateen-Rajput A, Toldo L. Extraction of potential adverse drug events from medical case reports. *J Biomed Semantics*. 2012;3(1):15. doi:10.1186/2041-1480-3-15.
- 25 Haerian K, Varn D, Vaidya S, Ena L, Chase HS, Friedman C. Detection of pharmacovigilance-related adverse events using electronic health records and automated methods. *Clin Pharmacol Ther*. 2012;92(2):228-234. doi:10.1038/clpt.2012.54.
- 26 Hazlehurst B, Naleway A, Mullooly J. Detecting possible vaccine adverse events in clinical notes of the electronic medical record. *Vaccine*. 2009;27(14):2077-2083. doi:10.1016/j.vaccine.2009.01.105.

- 27 Lependu P, Iyer SV, Fairon C, Shah NH. Annotation Analysis for Testing Drug Safety Signals using Unstructured Clinical Notes. *J Biomed Semantics*. 2012;3 Suppl 1:S5. doi:10.1186/2041-1480-3-S1-S5.
- 28 Li Y, Salmasian H, Harpaz R, Chase H, Friedman C. Determining the reasons for medication prescriptions in the EHR using knowledge and natural language processing. *AMIA Annu Symp Proc*. 2011;2011:768-776.
- 29 Munkhdalai T, Liu F, Yu H. Clinical Relation Extraction Toward Drug Safety Surveillance Using Electronic Health Record Narratives: Classical Learning Versus Deep Learning. *JMIR Public Health Surveill*. 2018;4(2):e29. doi:10.2196/publichealth.9361.
- 30 Wang X, Hripcsak G, Markatou M, Friedman C. Active computerized pharmacovigilance using natural language processing, statistics, and electronic health records: a feasibility study. *J Am Med Inform Assoc*. 2009;16(3):328-337. doi:10.1197/jamia.M3028.
- 31 Murff HJ, FitzHenry F, Matheny ME, et al. Automated identification of postoperative complications within an electronic medical record using natural language processing. *JAMA*. 2011;306(8):848-855. doi:10.1001/jama.2011.1204.
- 32 Penz JFE, Wilcox AB, Hurdle JF. Automated identification of adverse events related to central venous catheters. *Journal of Biomedical Informatics*. 2007;40(2):174-182. doi:10.1016/j.jbi.2006.06.003.
- 33 Wang H, Zhang W, Zeng Q, Li Z, Feng K, Liu L. Extracting important information from Chinese Operation Notes with natural language processing methods. *Journal of Biomedical Informatics*. 2014;48:130-136. doi:10.1016/j.jbi.2013.12.017.
- 34 Hripcsak G, Bakken S, Stetson PD, Patel VL. Mining complex clinical data for patient safety research: a framework for event discovery. *Journal of Biomedical Informatics*. 2003;36(1-2):120-130. doi:10.1016/j.jbi.2003.08.001.
- 35 Melton GB, Hripcsak G. Automated detection of adverse events using natural language processing of discharge summaries. *J Am Med Inform Assoc*. 2005;12(4):448-457. doi:10.1197/jamia.M1794.
- 36 Murff HJ, Patel VL, Hripcsak G, Bates DW. Detecting adverse events for patient safety research: a review of current methodologies. *Journal of Biomedical Informatics*. 2003;36(1-2):131-143. doi:10.1016/j.jbi.2003.08.003.
- 37 Stockwell DC, Kane-Gill SL. Developing a patient safety surveillance system to identify adverse events in the intensive care unit. *Crit Care Med*. 2010;38(6 Suppl):S117-25. doi:10.1097/CCM.0b013e3181dde2d9.
- 38 Szekendi MK, Sullivan C, Bobb A, et al. Active surveillance using electronic triggers to detect adverse events in hospitalized patients. *Qual Saf Health Care*. 2006;15(3):184-190. doi:10.1136/qshc.2005.014589.
- 39 Taggart M, Chapman WW, Steinberg BA, et al. Comparison of 2 Natural Language Processing Methods for Identification of Bleeding Among Critically Ill Patients. *JAMA Netw Open*. 2018;1(6):e183451. doi:10.1001/jamanetworkopen.2018.3451.
- 40 Connell SS, Balkany TJ, Hodges AV, Telischi FF, Angeli SI, Eshraghi AA. Electrode migration after cochlear implantation. *Otol Neurotol*. 2008;29(2):156-159. doi:10.1097/MAO.0b013e318157f80b.

- 41 Tambyraja RR, Gutman MA, Megerian CA. Cochlear implant complications: utility of federal database in systematic analysis. *Arch Otolaryngol Head Neck Surg.* 2005;131(3):245-250. doi:10.1001/archotol.131.3.245.
- 42 Nadkarni PM, Ohno-Machado L, Chapman WW. Natural language processing: an introduction. *J Am Med Inform Assoc.* 2011;18(5):544-551. doi:10.1136/amiajnl-2011-000464.
- 43 Pedregosa F, Varoquaux G, Gramfort A, Michel V, Thirion B, Grisel O. Scikit-learn: Machine Learning in Python. *Journal of Machine Learning Research.* 2011;12:2825-2830. <http://www.jmlr.org/papers/volume12/pedregosa11a/pedregosa11a.pdf>.
- 44 Ribeiro MT, Singh S, Guestrin C. "Why Should I Trust You?". In: Krishnapuram B, Shah M, Smola A, Aggarwal C, Shen D, Rastogi R, eds. Proceedings of the 22nd ACM SIGKDD International Conference on Knowledge Discovery and Data Mining - KDD '16. New York, New York, USA: ACM Press; 2016:1135-1144.
- 45 Raz Y. The utility of the MAUDE database in researching cochlear implantation complications. *Arch Otolaryngol Head Neck Surg.* 2005;131(3):251. doi:10.1001/archotol.131.3.251.

TABLES

Table 1. Machine learning model classification performance metrics. ***Weighted average:** averaging the support-weighted mean per label.*

	<i>Logistic Regression</i>			<i>Random Forest Classifier</i>			<i>Linear SVC</i>			<i>Multinomial Naïve Bayes</i>		
Manufacturer Category	Precision	Recall	F₁ Score	Precision	Recall	F₁ Score	Precision	Recall	F₁ Score	Precision	Recall	F₁ Score
Cochlear Corporation	0.90	0.93	0.92	0.70	0.98	0.82	0.90	0.93	0.91	0.92	0.92	0.92
MedEL	0.91	0.90	0.90	0.92	0.85	0.88	0.91	0.89	0.90	0.90	0.91	0.90
Advanced Bionics	0.82	0.78	0.80	0.96	0.21	0.35	0.82	0.77	0.79	0.81	0.79	0.80
<i>Weighted average</i>	0.89	0.89	0.89	0.82	0.77	0.73	0.88	0.88	0.88	0.89	0.89	0.89
Adverse Event Type	Precision	Recall	F₁ Score	Precision	Recall	F₁ Score	Precision	Recall	F₁ Score	Precision	Recall	F₁ Score
Injury	0.90	0.93	0.92	0.70	0.98	0.82	0.91	0.93	0.92	0.89	0.92	0.90
Malfunction	0.90	0.85	0.87	0.93	0.37	0.53	0.89	0.86	0.87	0.88	0.83	0.85
Death	0.00	0.00	0.00	0.00	0.00	0.00	1.00	0.50	0.67	0.00	0.00	0.00
<i>Weighted average</i>	0.90	0.90	0.90	0.79	0.73	0.70	0.90	0.90	0.90	0.88	0.88	0.88

Table 2. Top 15 most influential positive and negative predictive words for each manufacturer and adverse event type. Generated using the *ELI5* package.

Manufacturer					
Advanced Bionics		Cochlear Corp		MedEL	
Weight	Feature	Weight	Feature	Weight	Feature
5.10	recipient	5.23	per	6.89	description
3.57	reportedly	3.97	subsequently	6.18	user
1.69	pt	3.83	clinic	5.29	failure
1.66	inspection	2.17	submit	4.87	active
1.59	considers	2.10	date	4.74	evaluation
1.52	obtain	1.72	audiologist	4.17	explanted
1.50	supplemental	1.67	report	3.74	conclusion
1.48	test	1.46	decision	3.63	damage
1.45	schedule	1.41	due	3.49	hear
1.40	process	1.36	associate	3.40	returned
1.36	inform	1.30	july	3.11	housing
1.35	reportable	1.28	experience	2.95	return
1.33	pass	1.27	remains	2.93	base
1.24	lock	1.25	subsequent	2.89	situ
1.17	company	1.16	request	2.73	measurement
-1.27	audiologist	-1.66	sound	-2.04	decision
-1.42	conclusion	-1.76	description	-2.08	reportable
-1.44	situ	-1.83	final	-2.09	daughter
-1.50	housing	-1.87	explanted	-2.09	submit
-1.50	cause	-1.88	housing	-2.23	associate
-1.69	active	-1.89	channel	-2.31	august
-1.79	hear	-1.91	user	-2.32	allegation
-1.86	description	-2.06	investigation	-2.35	inspection
-1.90	damage	-2.23	follow	-2.46	date
-2.09	user	-2.26	damage	-2.46	considers
-2.27	clinic	-2.33	returned	-2.49	experience
-2.34	report	-2.75	reportedly	-3.09	recipient
-3.04	subsequently	-3.03	evaluation	-3.14	clinic
-3.16	submit	-3.36	failure	-4.72	subsequently
-3.39	per	-4.66	recipient	-5.00	per

Adverse Event Type					
Death		Injury		Malfunction	
Weight	Feature	Weight	Feature	Weight	Feature
21.67	expire	7.84	match	7.92	lock

9.82	away	7.24	reimplanted	6.80	explanted
6.66	stroke	7.24	infection	6.19	another
6.48	brain	7.07	final	5.84	failure
6.08	pcv	7.07	investigation	5.21	new
6.03	series	6.43	conclusion	5.06	equipment
5.90	anesthesia	5.90	returned	5.04	intermittency
5.82	meningitis	5.58	evaluate	5.01	evaluation
5.75	pass	4.85	extrusion	4.34	explant
5.64	acute	4.80	prescribe	4.10	consistent
5.47	arrest	4.11	array	3.99	malfunction
5.39	ekg	4.00	migration	3.72	return
5.22	readmitted	3.90	extrude	3.63	accord
5.13	flu	3.64	skin	3.62	test
5.11	cardiac	3.34	antibiotic	3.59	remain
-2.12	failure	-3.54	implement	-3.33	subsequent
-2.16	reimplantation	-3.61	remain	-3.61	skin
-2.18	site	-3.71	return	-3.85	extrude
-2.22	contributed	-3.75	accord	-4.01	migration
-2.28	test	-4.02	malfunction	-4.10	array
-2.30	post	-4.15	consistent	-4.75	prescribe
-2.32	submit	-4.34	explant	-4.85	extrusion
-2.48	implanted	-5.03	intermittency	-5.54	evaluate
-2.68	experience	-5.04	evaluation	-5.83	returned
-2.77	report	-5.06	equipment	-6.48	conclusion
-2.81	unknown	-5.17	new	-7.07	final
-3.00	explanted	-5.78	failure	-7.11	investigation
-3.03	electrode	-6.17	another	-7.20	reimplanted
-3.08	recipient	-6.72	explanted	-7.26	infection
-4.67	device	-7.88	lock	-7.84	match

Table 3. Manufacturer predictions using the trained logistic regression model on novel adverse event examples.

<p>A) Example – Implant Dislodged</p> <p>Manufacturer Prediction Probability (%):</p> <table><tr><td>Cochlear Corp</td><td>0.32</td></tr><tr><td>MedEL</td><td>0.52</td></tr><tr><td>Advanced Bionics</td><td>0.16</td></tr></table>	Cochlear Corp	0.32	MedEL	0.52	Advanced Bionics	0.16	<p>“The surgeon reported that the implant was dislodged”</p> <p>Influential Text Features:</p> <table><tr><td>NOT MedEL</td><td>MedEL</td></tr><tr><td>surgeon</td><td>dislodged</td></tr><tr><td>0.05</td><td>0.04</td></tr><tr><td></td><td>implant</td></tr><tr><td></td><td>0.02</td></tr><tr><td>reported</td><td></td></tr><tr><td>0.00</td><td></td></tr><tr><td></td><td>was</td></tr><tr><td></td><td>0.00</td></tr></table> <p>Text with highlighted features:</p> <p>“The surgeon reported that the implant was dislodged.”</p>		NOT MedEL	MedEL	surgeon	dislodged	0.05	0.04		implant		0.02	reported		0.00			was		0.00				
Cochlear Corp	0.32																													
MedEL	0.52																													
Advanced Bionics	0.16																													
NOT MedEL	MedEL																													
surgeon	dislodged																													
0.05	0.04																													
	implant																													
	0.02																													
reported																														
0.00																														
	was																													
	0.00																													
<p>B) Example - Electrode Migration</p> <p>Manufacturer Prediction Probability (%):</p> <table><tr><td>Cochlear Corp</td><td>0.21</td></tr><tr><td>MedEL</td><td>0.39</td></tr><tr><td>Advanced Bionics</td><td>0.40</td></tr></table>	Cochlear Corp	0.21	MedEL	0.39	Advanced Bionics	0.40	<p>“The patient had performed well for 3 months after activation. However, they had a sudden loss in performance and an x-ray determined that the electrode had migrated/extruded out of the cochlea”</p> <p>Influential Text Features:</p> <table><tr><td>NOT MedEL</td><td>MedEL</td></tr><tr><td></td><td>migrated/extruded</td></tr><tr><td></td><td>0.06</td></tr><tr><td>well</td><td></td></tr><tr><td>0.04</td><td></td></tr><tr><td>sudden</td><td></td></tr><tr><td>0.03</td><td></td></tr><tr><td>performed</td><td></td></tr><tr><td>0.03</td><td></td></tr><tr><td>however,</td><td></td></tr><tr><td>0.03</td><td></td></tr></table> <p>Text with highlighted features:</p> <p>“The patient had performed well for 3 months after activation. however, they had a sudden loss in performance and an x-ray determined that the electrode had migrated/extruded out of the cochlea.”</p>		NOT MedEL	MedEL		migrated/extruded		0.06	well		0.04		sudden		0.03		performed		0.03		however,		0.03	
Cochlear Corp	0.21																													
MedEL	0.39																													
Advanced Bionics	0.40																													
NOT MedEL	MedEL																													
	migrated/extruded																													
	0.06																													
well																														
0.04																														
sudden																														
0.03																														
performed																														
0.03																														
however,																														
0.03																														
<p>C) Example – Circuit Failure</p> <p>Manufacturer Prediction Probability (%):</p> <table><tr><td>Cochlear Corp</td><td>0.32</td></tr><tr><td>MedEL</td><td>0.46</td></tr><tr><td>Advanced Bionics</td><td>0.23</td></tr></table>	Cochlear Corp	0.32	MedEL	0.46	Advanced Bionics	0.23	<p>“The patient reports being struck in the head near his implant site. He noted an immediate drop in performance. The implant was removed, and circuit failure was determined after analysis by the manufacturer”</p> <p>Influential Text Features:</p> <table><tr><td>NOT MedEL</td><td>MedEL</td></tr><tr><td>failure</td><td></td></tr><tr><td>0.10</td><td></td></tr><tr><td></td><td>drop</td></tr><tr><td></td><td>0.05</td></tr><tr><td></td><td>immediate</td></tr><tr><td></td><td>0.05</td></tr><tr><td>analysis</td><td></td></tr><tr><td>0.04</td><td></td></tr><tr><td>near</td><td></td></tr><tr><td>0.03</td><td></td></tr></table> <p>Text with highlighted features:</p> <p>“The patient reports being struck in the head near his implant site. he noted an immediate drop in performance. the implant was removed, and circuit failure was determined after analysis by the manufacturer”</p>		NOT MedEL	MedEL	failure		0.10			drop		0.05		immediate		0.05	analysis		0.04		near		0.03	
Cochlear Corp	0.32																													
MedEL	0.46																													
Advanced Bionics	0.23																													
NOT MedEL	MedEL																													
failure																														
0.10																														
	drop																													
	0.05																													
	immediate																													
	0.05																													
analysis																														
0.04																														
near																														
0.03																														

FIGURE LEGENDS

Figure 1. Analytic workflow for preparing and analyzing MAUDE-derived cochlear implant medical device reports.

Figure 2. Box and whisker plot demonstrating the test-set classification accuracy (%; y-axis) for predicting the specific cochlear implant manufacturer after five-fold cross validation for each of the four machine algorithms. The lines within the boxes denote the mean classification accuracy.

Figure 3. Box and whisker plot demonstrating the test-set classification accuracy (%; y-axis) for predicting adverse event type after five-fold cross validation for each of the four machine algorithms. The lines within the boxes denote the mean classification accuracy.

Cochlear Implant MDR Record

per the clinic reported, a small opening of around 0.5mm in diameter at recipient's implant site. revision surgery was completed on -- 2019 by surgeon to cover the opening. device remains in-situ.

- Weigh down most frequent text terms and sequences & scale up the rare/less frequent (*i.e. generate TF-IDF statistic*)
- Convert weightings to numerical vector values (vectorization)

80%

20%

Training Set

Test Set

Naïve Bayes

Support Vector Classifier

Random Forest

Logistic Regression

Model Outcome Classification Accuracy (%)

A.

**Narrative Data
Preprocessing**

B.

**Feature
Extraction**

C.

**Model
Training**
Five-fold
Cross-Validation

D.

**Model
Interpretation**