Limited specificity of SARS-CoV-2 antigen-detecting rapid diagnostic tests at low temperatures

Verena Haage¹, Andres Moreira-Soto¹, Jilian A. Sacks², Victor Corman^{1,3}, Christian Drosten^{1,3}, Felix Drexler^{1,3}

¹Institute of Virology, Charité-Universitätsmedizin Berlin, Corporate Member of Freie Universität Berlin, Humboldt-Universität zu Berlin, and Berlin Institute of Health, Berlin, Germany

² Foundation for Innovative New Diagnostics (FIND), Geneva, Switzerland

³ German Centre for Infection Research (DZIF), Associated Partner Charité-Universitätsmedizin Berlin, Berlin, Germany

*Correspondence: Jan Felix Drexler, Helmut-Ruska-Haus, Institute of Virology, Campus Charité Mitte, Charitéplatz 1, 10098 Berlin, Germany; Email: felix.drexler@charite.de

Keywords: SARS-CoV-2, rapid antigen tests, specificity, temperature stability, diagnostics, cold settings

Abstract

SARS-CoV-2 antigen-detecting rapid diagnostic tests (Ag-RDTs) are available within and outside of health care settings to enable increased access to COVID-19 diagnosis. These environments include provisional testing facilities, lacking temperature control; as outside temperatures fall, recommended testing temperatures cannot be guaranteed. We report impaired specificity in two out of six Ag-RDTs when used at 2-4°C, indicating that testing in cold settings might cause false-positive results potentially entailing unwarranted quarantine assignments and incorrect incidence estimates.

Introduction

SARS-CoV-2 antigen-detecting tests were introduced into the global market during late 2020 and are now widely used, in both the global north and south. Many regions located in the northern hemisphere such as the United States or Europe are currently severely affected by the second wave of the COVID-19 pandemic (1, 2). To manage testing demand many physicians, healthcare and public health systems have opened external testing stations such as ‘diagnostic streets’ or drive-through facilities in urban settings (3). In Berlin, a German metropolis with a population of about 3.5 million inhabitants, several outside testing stations in tents or pavilions are in use, produced by manufacturers who offer special corona test cabins (4). These facilities are often of provisional nature, for example in the form of unheated tents. In the winter months, temperatures in Europe or the U.S. can range from -10°C to 10°C (5, 6), well below the recommended operating temperatures of most Ag-RDTs. Most manufacturers of SARS-CoV-2 Ag-RDTs specify storage conditions between 2-30°C, but stipulate that tests be equilibrated to room temperature (15-30°C) at the time of use to guarantee their performance. With temperatures around freezing point during the winter months, unheated testing facilities cannot always comply with these conditions. Moreover, as

the national COVID-19 reference laboratory in Germany, we were contacted by physicians from two German metropolis both reporting an unusual high number of positive SARS-CoV-2 Ag-RDTs from outside testing facilities. To test whether operation of SARS-CoV-2 Ag-RDTs at low outside temperatures impacts their performance, we selected six (**Table 1**) out of the 139 currently commercially available SARS-CoV-2 Ag-RDTs (7) based on clinical performance data (8) and market availability, for validation of specificity when stored and/or operated at cold outside temperatures of 2-4°C.

Materials and Methods

Tested conditions included: (a) storage and testing of Ag-RDTs at recommended conditions (15-30°C); (b) storage of tests at recommended conditions, pre-incubation of tests for 30mins to cold-temperatures (2-4°C) and testing at cold temperatures (2-4°C); and (c) and storage of tests at 2-4°C followed by testing at 2-4°C (**Figure 1A**). Analytical specificity was examined by testing for cross-reactivity with common cell culture-derived respiratory viruses including the ubiquitous human coronaviruses (HCoV) HCoV-229E (2.9×10^7 copies/ml), HCoV-OC43 (1.0×10^6 copies/ml), influenza virus A H1N1 (7.8×10^6 copies/ml) and rhinovirus A (2.2×10^6 copies/ml). Concentrations were selected according to the guidelines on analytical specificity testing for SARS-CoV-2 Ag-RDTs published by the German Federal institute for vaccines and biomedicines (9). 20µl of viral cell culture supernatant were added to proprietary lysis buffer or as an internal control, and 20µl of lysis buffer were directly applied to test cassettes for validation experiments. Tests were performed in duplicates, according to the test-specific supplier instructions for use and results were scored independently by two persons. We additionally tested ten healthy laboratory members who previously volunteered for a SARS-CoV-2 Ag-RDT validation study (8). Healthy volunteers were

without symptoms of respiratory tract infection and tested negative for SARS-CoV-2 by RT-qPCR (10). All subjects received instructions on self-sampling, swabs were immediately dissolved in 1ml PBS and 20µl of PBS were added to proprietary buffer for testing. In a recent study, self-sampling was shown to be a reliable alternative to professional nasopharyngeal swabs for Ag-RDTs (11).

Results

Two of the six SARS-CoV-2 Ag-RDTs showed impaired specificity (**Figure 1B**) when stored at room temperature, but exposed to 2-4°C for 30 minutes prior to testing at 2-4°C as cross-reactivity with common respiratory viruses, and false-positive results occurred in healthy volunteers in the form of weak, but clearly visible bands (**Figure 1C**). In one test (test I), unspecific reactivity was only observed upon short-term incubation at 2-4°C followed by test operation at 2-4°C, but not after long-term storage at 2-4°C. In contrast, the other test yielding unspecific results (test II) yielded almost identically unspecific results after both, short- and long-term storage at 2-4°C and operation at 2-4°C (**Figure 1B**). On the one hand, those data highlight differences between test devices. On the other hand, our results may hint at effects of relatively rapid temperature changes on some tests for unknown reasons, potentially including environmental factors such as condensation. Results were reproducible and functionality of tests was confirmed by determining their level of detection using serial dilutions of SARS-CoV-2 nucleoprotein (SARS-CoV-2-N) at recommended conditions as previously described (8).

Discussion

Our study highlights that specificity of SARS-CoV-2 Ag-RDTs may be impaired when operating tests at temperatures that differ from recommended conditions (15-30°C),

leading to false-positive results. These results were observed for only certain test brands including one of the Ag-RDT currently listed for emergency use by the World Health Organization (12, 13), highlighting that each test may need to be considered specifically and broader validation of temperature robustness of SARS-CoV-2 RDTs should be performed. Moreover, impaired sensitivity of SARS-CoV-2 Ag-RDTs at elevated temperatures has been reported previously (14), highlighting the importance of test operation at recommended temperatures. Of note, all tests studied here were shown to be highly specific when operated at recommended conditions in prior studies (8, 14), underlining that impaired specificity is not a test-intrinsic problem but owed to test operation under conditions beyond those defined by the manufacturer.

Our data imply that caution must be taken when offering SARS-CoV-2 Ag-RDT diagnostics in settings lacking temperature control, including diagnostic streets or drive-through testing stations in car parks (3). Especially under provisional conditions when minimally-trained or unexperienced staff perform diagnostic tests, the establishment of laboratory workflows and compliance with the conditions recommended by the manufacturer are vital to ensure accurate testing (15). Until January 14th 2021, severity of regional lockdown measures in Germany is increased upon trespassing 7-day incidences of 200 per 100.000 inhabitants (16), highlighting the potential consequences of false-positive test results beyond potentially unwarranted individual quarantine assignments, if those results were reported to public health authorities without confirmation by a gold standard test such as RT-PCR (10).

Funding

The study was partially supported by the Foundation for Innovative New Diagnostics (FIND), including procurement of some test kits. The findings and conclusions contained within are those of the authors and do not necessarily reflect positions or policies of FIND.

Conflict of interest: The authors declare no conflict of interest.

Table 1. Overview of SARS-CoV-2 rapid antigen tests included in the study

ID	Test	Manufacturer	Lot No.
I	Panbio™ COVID-19 Ag Rapid Test	Abbott Laboratories	41ADF012A
II	ActivXpress + COVID-19 Antigen Complete Testing Kit	Edinburgh Genetics	AG20200905
III	COVID-19 Ag	Genedia	643X2005
IV	ichroma - COVID-19 Ag	Boditech Med	SRQHA27
V	COVID-19 Antigen Rapid Test Kit	JOYSBIO (Tianjin) Biotechnology Co., Ltd.	2020092409
VI	SARS-CoV-2 Rapid Antigen Test	Roche Diagnostics*	QCO3020083 QCO390003I/Sub:I-2 QCO390011A/Sub:A-2

* equals STANDARD Q COVID-19 Test by SD Biosensor, Inc.

Figure 1. Experimental setup and results for specificity testing of SARS-CoV-2

Ag-RDTs. A. (a): storage and operation at recommended conditions; (b): recommended storage, pre-incubation at 2-4°C for 30mins and operation at 2-4°C; (c): storage and operation at 2-4°C. **B.** Specificity of SARS-CoV-2 Ag-RDTs decreases at low temperatures. (a): rt storage- rt testing; (b): rt storage – 2-4°C 30 min pre-incubation – 2-4°C testing; (c): 2-4°C storage – 2-4°C testing. red: positive; white: negative; salmon: weak band, result unclear. rt: room temperature. I: Abbott; II: ActivXpress; III: Genedia; IV: ichroma; V: JOYSBIO; VI: Roche. **C.** Example for observed cross-reactivity of the ActivXpress test with Influenza virus A H1N1 and SARS-CoV-2 nucleoprotein as positive control (SARS-CoV-2-N; 5µg/ml) when tested under condition (b): recommended storage, pre-incubation at 2-4°C for 30mins and operation at 2-4°C.

References

1. <https://coronavirus.jhu.edu/us-map>. Accessed 14.01.2021.
2. <https://www.ecdc.europa.eu/en/cases-2019-ncov-eueea>. Accessed 14.01.2021.
3. Shah A, Challener D, Tande AJ, Mahmood M, O'Horo JC, Berbari E, Crane SJ. 2020. Drive-Through Testing: A Unique, Efficient Method of Collecting Large Volume of Specimens During the SARS-CoV-2 (COVID-19) Pandemic. *Mayo Clinic Proceedings* 95:1420-1425.
4. https://www.dwd.de/DE/wetter/vorhersage_aktuell/berlin_brandenburg/vhs_bbb_node.html. Accessed 11.01.2021.
5. <https://climate.copernicus.eu/european-temperature>. Accessed 14.01.2021.
6. <https://www.ncdc.noaa.gov/climate-information/statistical-weather-and-climate-information>. Accessed 14.01.2021.
7. <https://www.finddx.org/covid-19/pipeline>. Accessed 13.01.2021.
8. Corman VM, Haage VC, Bleicker T, Schmidt ML, Mühlemann B, Zuchowski M, Jó Lei WK, Tscheak P, Möncke-Buchner E, Müller MA, Krumbholz A, Drexler JF, Drosten C. 2020. Comparison of seven commercial SARS-CoV-2 rapid Point-of-Care Antigen tests. *medRxiv* doi:10.1101/2020.11.12.20230292:2020.11.12.20230292.
9. Sinne ABMfS-C-Ai, Antigenschnelltests. vAST.
10. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DK, Bleicker T, Brünink S, Schneider J, Schmidt ML, Mulders DG, Haagmans BL, van der Veer B, van den Brink S, Wijsman L, Goderski G, Romette J-L, Ellis J, Zambon M, Peiris M, Goossens H, Reusken C, Koopmans MP, Drosten C. 2020. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Eurosurveillance* 25:2000045.
11. Lindner AK, Nikolai O, Kausch F, Wintel M, Hommes F, Gertler M, Krüger LJ, Gaeddert M, Tobian F, Lainati F, Köppel L, Seybold J, Corman VM, Drosten C, Hofmann J, Sacks JA, Mockenhaupt FP, Denking CM. 2020. Head-to-head comparison of SARS-CoV-2 antigen-detecting rapid test with self-collected anterior nasal swab versus professional-collected nasopharyngeal swab. *medRxiv* doi:10.1101/2020.10.26.20219600:2020.10.26.20219600.
12. https://www.who.int/diagnostics_laboratory/eual/listing/en/. Accessed 14.01.2021.
13. https://www.who.int/diagnostics_laboratory/200922_eul_sars_cov2_product_list.pdf?ua=1. Accessed 14.01.2021.
14. Haage V, de Oliveira-Filho EF, Moreira-Soto A, Kühne A, Fischer C, Sacks J, Corman VM, Müller MA, Drosten C, Drexler JF. 2021. Impaired performance of SARS-CoV-2 antigen-detecting rapid tests at elevated temperatures. *medRxiv* doi:10.1101/2021.01.06.21249314:2021.01.06.21249314.
15. Fischer C, Mögling R, Melidou A, Kühne A, Oliveira-Filho EF, Wolff T, Reiche J, Broberg E, Meijer A, Leitmeyer K, Drexler JF, Reusken C. 2020. Variable sensitivity in molecular detection of SARS-CoV-2 in European Expert Laboratories: External Quality Assessment, June - July 2020. *J Clin Microbiol* doi:10.1128/jcm.02676-20.
16. <https://www.bundesregierung.de/breg-de/themen/coronavirus/corona-massnahmen-1734724>. Accessed 14.01.2021.