TITLE PAGE

Full-length title:

Implementation of an in-house real-time reverse transcription-PCR assay to detect the emerging SARS-CoV-2 N501Y variants

Short title (for the running head): qPCR for SARS-CoV-2 N501Y variants

Author list: Marielle BEDOTTO¹, Pierre-Edouard FOURNIER^{1,2}, Linda HOUHAMDI^{1,2}, Philippe COLSON^{1,2}, Didier RAOULT^{1,2}*

Affiliations: ¹ IHU Méditerranée Infection, 19-21 boulevard Jean Moulin, 13005 Marseille, France; ² Aix-Marseille Univ., Institut de Recherche pour le Développement (IRD), Assistance Publique - Hôpitaux de Marseille (AP-HM), MEPHI, 27 boulevard Jean Moulin, 13005 Marseille, France

* Corresponding author: Didier RAOULT, IHU - Méditerranée Infection, 19-21 boulevard Jean Moulin, 13005 Marseille, France. Tel.: +33 413 732 401, Fax: +33 413 732 402; email: didier.raoult@gmail.com

Key words (5): SARS-CoV-2; Covid-19; N501Y variants; qPCR; diagnosis

Word counts: abstract: 176; text: 497

Figure: 0; Table: 1; References: 6.

ABSTRACT

The SARS-CoV-2 pandemic has been associated with the emergence of several variants with a mutated spike glycoprotein that are of substantial concern regarding their transmissibility and ability to evade immune responses. This warrants implementing strategies for their detection and surveillance. We have set up an in-house one-step real-time reverse transcription-PCR (qPCR) assay that specifically detects SARS-CoV-2 spike N501Y variants. Our assay was positive for all 6 patients found spike N501Y-positive by genome sequencing. Ten cDNA samples for each of the 10 Marseille variants identified by genome sequencing and three nasopharyngeal samples of a spike N501Y-negative variant (Marseille-4) that predominates locally tested negative. All negative controls among which 5 SARS-CoV-2-negative nasopharyngeal samples tested negative. First use in the setting of diagnosis on 51 nasopharyngeal samples from SARS-CoV-2-positive but Marseille-4-negative patients showed positivity in 5 cases further confirmed by sequencing as from spike N501Y variant-infected patients. Thus, our in-house qPCR system was found reliable for the detection of the N501Y substitution and allowed estimating preliminarily that spike N501Y variant prevalence was 4% among SARS-CoV-2 diagnoses since January 2020.

TEXT

The SARS-CoV-2 pandemic has been associated with the occurrence of several viral variants with a mutated spike glycoprotein (S). Those currently of greatest concern carry N501Y substitution within the spike (S) receptor binding domain (RBD). Indeed, they have become predominant in England (20I/501Y.V1) [1] and were detected in South Africa (20H/501Y.V2) [2] and Brazil [3]. The 20I/501Y.V1 variant has started to spread worldwide including in France [4]. It has been reported as 50-74% more transmissible than preexisting strains, suspected to evade anti-spike antibodies [1], and it caused a reinfection [5]. Its real-time detection is critical to manage patients appropriately, monitor and assess its epidemiological and clinical features, and survey cases of immune escape post-infection or vaccination. An alternative comprehensive detection strategy is warranted considering the very large number of SARS-CoV-2 cases.

We have implemented an in-house one-step real-time reverse transcription-PCR (qPCR) assay that specifically detects SARS-CoV-2 N501Y variants by targeting nucleotide position 23,063 within S gene where A>U leads to N501Y. SARS-CoV-2 genomes from the GISAID database with or without N501Y were used to design primers and a hydrolysis probe. The sequences of these latter and the qPCR conditions are shown in Table 1. Our N501Y-specific assay was positive for all 6 patients identified as 20I/501Y.V1-positive by genome sequencing and for all 3 additional contact patients for whom sequences were not obtained due to low viral load (Ct of qPCR diagnosis test= 32.0-34.0) leading to no amplification by two conventional PCR. Ten cDNA samples, one for each of the 10 Marseille variants [6] identified by genome sequencing, and three additional nasopharyngeal (NP) samples of the Marseille-4 variant that predominates locally ($\approx$ two-thirds of diagnoses) tested negative. All negative controls (5 SARS-CoV-2-negative NP samples and 20 RNA extract-free qPCR

68 mixes) tested negative. The first use for diagnosis purpose conducted on 51 NP samples from
69 SARS-CoV-2 positive (Ct= 10.2-30.8) but Marseille-4-negative patients showed positivity for
70 5 samples, which were confirmed by sequencing as from N501Y variant-infected patients.

71 Our in-house qPCR system was found reliable to detect specifically the N501Y
72 substitution and preliminarily allowed estimating 20I/501Y.V1 variant prevalence to 4%
73 among our current SARS-CoV-2 diagnoses since January. This assay showed better
74 sensitivity than conventional PCR and should be able to detect different N501Y variants. A
75 commercialized RT-PCR diagnosis assay (TaqPath RT-QPCR test) allows the indirect
76 identification of the 20I/501Y.V1 variant by detecting its ORF1a and N genes but not its S
77 gene due to a deletion at positions 21,766-21,772 [4]. However this deletion is also present in
78 strains devoid of the N501Y substitution and was reported in 0.6% of recent SARS-CoV-2
79 diagnoses in France [4]. Moreover, it is absent from the South African and Brazilian N501Y
80 variants, which prevents their identification. Finally, our in-house qPCR test can be widely
81 and easily deployed in laboratories as it runs on any open qPCR microplate platform, does not
82 require technical workers' training, and is as cheap as other in-house qPCR assays. Such
83 approach should allow adapting continuously diagnosis strategies for new SARS-CoV-2
84 variants.

85

86

87 **Author contributions**

88 Conceived and designed the study: PC, DR. Contributed materials/analysis tools: MB, PEF,
89 LH, PC. Analyzed the data: all authors. Wrote the paper: PC, DR.

90

91 **Acknowledgments -Funding**

This work was supported by the French Government under the “Investments for the Future” program managed by the National Agency for Research (ANR), Méditerranée-Infection 10-IAHU-03 and was also supported by Région Provence Alpes Côte d’Azur and European funding FEDER PRIMMI (Fonds Européen de Développement Régional-Plateformes de Recherche et d’Innovation Mutualisées Méditerranée Infection), FEDER PA 0000320 PRIMMI.

Conflicts of interest

The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper. Funding sources had no role in the design and conduct of the study; collection, management, analysis, and interpretation of the data; and preparation, review, or approval of the manuscript.

Ethics

This study has been approved by the ethics committee of the University Hospital Institute Méditerranée Infection, Marseille, France, with the registration number 2020-029.

References

- [1] K. Leung, M.H.Shum, G.M.Leung, T.T.Lam, J.T.Wu. Early transmissibility assessment of the N501Y mutant strains of SARS-CoV-2 in the United Kingdom, October to November 2020, Euro. Surveill 26 (2021) .
- [2] H. Tegally, E. Wilkinson, M. Giovanetti, A. Iranzadeh, V. Fonseca, J. Giandhari, et al. Emergence and rapid spread of a new severe acute respiratory syndrome-related coronavirus 2 (SARS-CoV-2) lineage with multiple spike mutations in South Africa. medRxiv preprint (2020) doi: <https://doi.org/10.1101/2020.12.21.20248640>.

- [3] N.R. Faria, I.M. Claro, D. Candido, L.A. Moyses Franco, P.S. Andrade, T.M. Coletti, et al. Genomic characterisation of an emergent SARS-CoV-2 lineage in Manaus: preliminary findings, Virological.org (2021) <https://virological.org/t/genomic-characterisation-of-an-emergent-sars-cov-2-lineage-in-manaus-preliminary-findings/586>.
- [4] A. Bal, G. Destras, A. Gaymard, K. Stefic, J. Marlet, S.A. Eymieux, et al. Two-step strategy for the identification of SARS-CoV-2 variant of concern 202012/01 and other variants with spike deletion H69-V70, France, August to December 2020, medRxiv preprint. doi: <https://doi.org/10.1101/2020.11.10.20228528>.
- [5] D. Harrington, B. Kele, S. Pereira, X. Couto-Parada, A. Riddell, S. Forbes, et al. Confirmed reinfection with SARS-CoV-2 variant VOC-202012/01, Clin. Infect. Dis. (2021) ciab014. doi: 10.1093/cid/ciab014. Online ahead of print.
- [6] P. Colson, A. Levasseur, J. Delerce, H. Chaudet, V. Bossi, M. Ben Kheder, et al. Dramatic increase in the SARS-CoV-2 mutation rate and low mortality rate during the second epidemic in summer in Marseille. IHU preprint (2020) doi: <https://doi.org/10.35088/68c3-ew82>.

TABLE

Table 1. Primers, probe and qPCR conditions

Name	Sequence (5'-3')	Positions *
<i>Primers:</i>		
Pri_IHU_N501Y_F1	ATCAGGCCCGGTAGCACAC	22,980-22,997
Pri_IHU_N501Y_R1	AAACAGTTGCTGGTGCATGT	23,135-23,116
<i>Probe (6FAM-labelled):</i>		
Pro_IHU_C_GB_1_MBP	CCACTTATGGTGTGGTTACCAA	23,058-23,080

* in reference to SARS-CoV-2 genome GenBank Accession no. NC_045512.2 (Wuhan-Hu-1 isolate). The nucleotide carrying the mutation is covered by the probe and underlined.

The qPCR was performed by adding 5 µL of extracted viral RNA to 15 µL of reaction mixture containing 5 µL of 4X TaqMan Fast Virus 1-Step Master Mix (Thermo Fisher Scientific, Grand Island, NY, USA), 0.5 µL of forward primer (10 pmol/µL), 0.5 µL of reverse primer (10 pmol/µL), 0.4 µL of probe (10 pmol/µL), and 8.6 µL of water. PCR conditions are as follows: reverse transcription at 50°C for 10 min, then a hold at 95°C for 20 sec followed by 40 cycles comprising a step at 95°C for 15 sec and a step at 60°C for 60 sec.

This qPCR was run on a LC480 thermocycler (Roche Diagnostics, Mannheim, Germany).