Correction of measurement error in a commercial multiple-breath washout device

Florian Wyler^{*1} (<https://orcid.org/0000-0002-1232-1392>)

Marc-Alexander Oestreich^{*1,2} (<https://orcid.org/0000-0001-9641-3691>)

Bettina S. Frauchiger^{1,3} (<https://orcid.org/0000-0002-9519-9328>)

Kathryn Ramsey^{#1} (<https://orcid.org/0000-0003-4574-6917>)

Philipp Latzin^{#1} (<https://orcid.org/0000-0002-5239-1571>)

* Authors contributed equally to this work as first author

# Authors contributed equally to this work as last author

¹ Pediatric Respiratory Medicine, Inselspital, University Children's Hospital of Bern, University of Bern, Switzerland.

² Graduate School for Health Sciences, University of Bern, Switzerland.

³ Graduate School of Cellular and Biomedical Sciences, University of Bern, Switzerland

Abstract

Rationale:

Nitrogen multiple-breath washout (N₂MBW) is an established technique to assess functional residual capacity (FRC) and ventilation inhomogeneity in the lung. Accurate measurement of gas concentrations is essential for the appropriate calculation of clinical outcomes.

Objectives:

We investigated the accuracy of oxygen and carbon dioxide measurements used for the indirect calculation of nitrogen concentration in a commercial MBW device (Exhalyzer D, Eco Medics AG, Duernten, Switzerland) and its impact on FRC and lung clearance index (LCI).

Methods:

High precision calibration gas mixtures and mass spectrometry were used to evaluate sensor

NOTE: This preprint reports new research that has not been certified by peer review and should not be used to guide clinical practice.

output. We assessed the impact of corrected signal processing on FRC and LCI in a dataset of healthy children and children with cystic fibrosis using custom analysis software.

Results:

We found inadequate correction for the cross sensitivity of the oxygen and carbon dioxide sensors in the Exhalyzer D device. This results in an overestimation of expired nitrogen concentration, and consequently FRC and LCI outcomes. Breath-by-breath correction of this error reduced mean (SD) FRC by 8.9 (2.2)% and LCI by 11.9 (4.0)%. It also resulted in almost complete disappearance of the tissue nitrogen signal at the end of measurements.

Conclusions:

Inadequate correction for cross sensitivity between the oxygen and carbon dioxide gas sensors of the Exhalyzer D device leads to an overestimation of FRC and LCI. Correction of this error is possible and could be applied by re-analysing the measurements breath-by-breath in an updated software version.

Corresponding Author

Prof. Philipp Latzin, MD, PhD
Inselspital, University Children's Hospital of Bern
Freiburgstrasse 8, 3010 Bern, Switzerland
E-mail: philipp.latzin@insel.ch
Phone: +41 31 632 9353

Author contributions

F.W. and M.O. designed performed and analysed experiments. M.O., B.F and K.R created and organized the databases for retrospective analysis. F.W. wrote the manuscript, with major contributions from M.O, K.R. and P.L. All authors checked the final version of the manuscript.

53   **Grants, Gifts, Equipment, Drugs**

54   Eco Medics AG (Duernten, Switzerland) provided a research version of their commercial
55   software Spiroware 3.2.1 including insight on signal processing algorithms and helped with
56   the acquisition of mass spectrometry measurements. This project was funded by the Swiss
57   National Science Foundation, Grant Nr. 182719 (P. Latzin) and 168173 (K. Ramsey)

Introduction

Nitrogen multiple-breath washout (N_2 MBW) is an established technique to assess ventilation inhomogeneity and lung volumes during relaxed tidal breathing^{1–3}. The test starts at the end of a relaxed expiration, with the lung at functional residual capacity (FRC). As subjects inhale pure oxygen (100% O_2), resident nitrogen (N_2) is gradually washed out. The main outcome, the lung clearance index (LCI), represents the expired volume required to reach 2.5% of the starting N_2 concentration expressed in multiples of the FRC.

The N_2 MBW technique is increasingly being used as a functional measure capable of monitoring early disease progression more sensitive than spirometry and to capture treatment effects in patients with conditions such as cystic fibrosis (CF) and primary ciliary dyskinesia (PCD)^{4–8}. The endorsement of the North American Cystic Fibrosis Foundation and the European Cystic Fibrosis Society Clinical Trials Network consolidated the LCI as an endpoint in clinical trials in children and adults with CF^{3,9–12}. As recommended by different societies, most data on the efficacy of disease modifying therapies in CF^{13,14} were obtained with the Exhalyzer D / Spiroware setup (Eco Medics AG, Duernten, Switzerland) and this is the device of choice in multiple ongoing studies with the LCI as a primary endpoint (www.clinicaltrials.gov: NCT04138589, NCT03320382, NCT04026360, NCT02657837). Consequently, the correct measurement of N_2 MBW is of critical importance, notably in studies that serve to guide drug approval. In this device N_2 concentration is not measured directly, but relies on indirect determination of nitrogen, through accurate measurement of oxygen (O_2) and carbon dioxide (CO_2)¹⁵.

In this study we investigated the accuracy of indirect measurement of N_2 during MBW tests using the Exhalyzer D device by i) assessing the sensor accuracy in the Exhalyzer D's O_2 and CO_2 sensors, ii) establishing a correction for any observed sensor error and iii) assessing the effect size of sensor error on clinical outcomes (FRC, LCI) as well as on tissue nitrogen.

Methods

Study design

This was a combined prospective, experimental and retrospective, observational study to assess sensor accuracy as well as the impact of sensor inaccuracy using existing N₂MBW data. We included experimental data from gas mixture measurements and mass spectrometry (Eco Medics AG, Duernten, Switzerland). We characterized the impact of the developed correction functions on MBW outcomes with data from healthy children¹⁶ and children diagnosed with CF (Swiss Cystic Fibrosis Infant Lung Development (SCILD) cohort)^{6,17}. The Ethics Committee of the Canton of Bern, Switzerland approved the study protocol (PB_2017-02139).

i) Sensor accuracy

To measure sensor accuracy over the wide range of concentrations encountered in a N₂MBW measurement, we tested the Exhalyzer D sensors with gas mixtures containing known reference concentrations, and compared concentrations of CO₂ and O₂ measured by the Exhalyzer D to known reference concentrations (supplemental Table 1). Specifications from manufacturers of technical gas mixtures (Carbagas AG, Muri bei Bern, Switzerland) and mass spectrometry measurements (Eco Medics AG, Duernten, Switzerland) served as reference.

Equipment

Applying Dalton's law of partial pressures, the Exhalyzer D (Eco Medics AG, Duernten, Switzerland) computes an indirect nitrogen concentration that is based on the measurement of oxygen, carbon dioxide and an assumed fixed ratio of nitrogen and argon (supplemental Equation 1)¹⁸. While dry atmospheric air contains nitrogen (N₂), oxygen (O₂), argon, carbon dioxide (CO₂) and traces of neon, helium, methane and krypton, it is assumed that all gases except oxygen and carbon dioxide (which are affected by tidal breathing) remain in proportion to each other.

Sensor characteristics

The O₂ sensor used in the Exhalyzer D setup (X3004 OEM sensor, Oxigraf Inc., Sunnyvale, CA, USA) is located within the main structure of the Exhalyzer D. Gas from the patients breathing stream is sampled at a rate of 200 mL/min via a Nafion tube¹⁸. The CO₂ sensor (Capnostat 5, Respironics Inc., Wallingford, CT, USA) is located within the breathing stream. The Exhalyzer D measures both O₂ and CO₂ using laser absorption spectroscopy, a technique where the absorption of light is measured at specific frequencies characteristic to each gas¹⁹. The absorption spectrum of both gases is affected by a variety of factors, including pressure, temperature, and the presence of other gases (manufacturer's manual). The presence of O₂ in the gas mixture affects the absorption spectrum of CO₂ and vice versa.

Software for gas sensor measurements

The manufacturer provided a research version of the Exhalyzer D software (Spiroware 3.3.0 Research, Eco Medics AG, Duernten, Switzerland) which enabled technical gas measurements with direct recording of raw, uncorrected sensor outputs. This software version is, for the purpose of MBW tests, identical with the commercial release (Spiroware 3.2.1) on the level of signal processing and analysis.

Technical gases

Sensor accuracy was tested over the complete range of concentrations present in N₂MBW measurements. Twelve technical gas mixtures (Carbagas AG, Muri bei Bern, Switzerland), each containing different combinations of CO₂ (2.5%, 5%, 7.5%) and O₂ (30%, 60%, 90%) as well as three mixtures containing only CO₂ and O₂, were used (supplemental Table 1). Technical gas mixtures were manufactured at 2% mixture precision and 1% measurement precision. Additionally, a series of mass spectrometry measurements was carried out, where N₂ was kept at 2% to mimic the MBW end of test condition, while CO₂ varied from 0-6% with the rest of the mixture being O₂.

Measurement conditions for technical gases and mass spectrometry

Sensor accuracy of the Exhalyzer D was assessed under two conditions: i) at body temperature ($37.5 \pm 1.5^\circ \text{C}$), humidity saturated (MR850 Heated Humidifier, Fisher & Paykel, Auckland, New Zealand), with technical gas mixture specifications as reference, and ii) at 35°C in a dry, climate-controlled chamber with mass spectrometry (AMIS 2000, Innovision ApS, Odense, Denmark and Red-y Smart Controller, Voegtlin Instruments GmbH, Muttensz, Switzerland) as reference. Gas concentrations were measured in triplicates, on two Exhalyzer D devices (calibrated according the manufacturers manual) for technical gas measurements and one mass spectrometry device. Additional calibration measurements were carried out before and after each set of measurements.

Signal processing for gas sensors

The CO_2 signal measured was subsequently corrected with the relevant signal processing of a standard Exhalyzer D measurement to achieve final sensor readings that were equivalent to current standard processing during a MBW measurement: For condition i) a temperature/humidity-dependant “ATP correction” factor was applied to the CO_2 -signal to correct for humidity and transform the measured fraction into atmospheric temperature and pressure (ATP) conditions (Eq. 2 in OLS). For both condition i) and ii), the CO_2 signal was corrected with an O_2 -dependant cross sensitivity correction factor (Eq. 3 in OLS).

ii) Correction function fitting

We combined the data of the technical gas mixture and mass spectrometry measurements to construct a correction function for the O_2 and CO_2 sensors. Sensor error was defined as the absolute difference between measured concentrations of O_2 and CO_2 by the Exhalyzer D and reference values provided by mass spectrometry and the technical gas mixture manufacturer. We then fitted a 2nd-degree two-parameter polynomial through the error for each sensor, as a function of measured O_2 and CO_2 (Eq. 4 in OLS): Fitting was performed using MATLAB 2017b (Mathworks, Natick, Massachusetts, USA). This characterization of the measurement error as a function of measured gas concentrations could then be directly used as a correction function

for the analysis of MBW measurements, by adding the fitted error for each combination of O₂ and CO₂ to the respective measured concentrations.

iii) Effect size of sensor correction

Retrospective analysis

We characterized the impact of measurement error on MBW outcomes using a custom Python script developed by our group, designed to replicate standard Spiroware 3.2.1 signal processing and outcome calculation of MBW data. The impact of sensor error was assessed by comparing the outcomes of standard analysis (Exhalyzer D and Spiroware 3.2.1, Eco Medics AG, Duernten, Switzerland) with the outcomes of a corrected algorithm modified to include the O₂-CO₂ cross sensitivity correction functions outlined above.

For this, we re-analysed 357 MBW measurements of 85 healthy controls (HC, mean age: 11.1y, SD: 3.8y, range: 6-17.8y) and 62 subjects diagnosed with cystic fibrosis (CF, mean age: 9.5y, SD: 3.5y, range: 4-17.3y) from previous studies^{6,16} (OLS Table 2 for details). We included study visits with at least two acceptable, quality controlled N₂-MBW measurements, in accordance with recent consensus guidelines²⁰.

Statistical testing

For statistical testing of significance we used STATA 16 (StataCorp LLC, College Station, Texas, USA). Unpaired t-tests were used for comparisons between healthy controls and children with cystic fibrosis. Paired t-tests were performed for differences between standard and corrected processing of MBW files.

Results

i) Sensor accuracy

O₂ sensor accuracy

We found that the O₂ sensor of the Exhalyzer D device has a non-linear O₂- and CO₂- dependant measurement error (Figure 1). The degree of error in the O₂ sensor is higher with increasing concentrations of CO₂ between the range of 0 – 7.5%. At conditions that reflect the end of the current standard MBW end of test (N₂ around 2%, expired CO₂ around 5%, rest O₂), the O₂ sensor error leads to an underestimation of absolute expiratory O₂-concentrations of 0.8%. This in turn leads to an overestimation of calculated concentrations of N₂, and makes up the majority (86%) of a total sensor error which in the example above leads to measured concentrations of N₂ of 2% instead of the corrected 1.03% (Table 1).

CO₂ sensor accuracy

Additionally we found a residual error in the CO₂ sensor, currently not corrected for. The CO₂ sensor reading is already corrected by a factor that depends on the concentration of O₂ (see methods). However, the CO₂ sensor seems to display a different cross-sensitivity measurement error than the current signal processing takes into account (Figure 2). This residual error appears to be mainly dependant on CO₂ concentration (Figure 2A) and only mildly dependant on O₂ concentration (Figure 2B). Overall, the impact of this residual error in CO₂ is lower than the error of the O₂ sensor. It also leads to an overestimation of N₂, making up 14% of the total sensor error in the standard end of test condition (Table 1).

The combined error of both sensors is given in Figure 3. As illustrated, the resulting error in N₂ is higher in very low N₂ concentrations and at higher CO₂ concentrations. The gas concentrations that exist at a typical end of N₂MBW measurements (corrected N₂ around 2% and CO₂ around 5%) unfortunately result in a relatively large absolute N₂-error of 0.94% and even larger relative N₂-error of 47.2% (Table 1). The lowered N₂ concentrations throughout the entire MBW measurement (Figure 4A) lead to an end of test that is systematically reached earlier (Figure 4B).

iii) Effect size of sensor correction

Sensor correction impact on MBW outcomes

We re-analysed 357 MBW measurements from healthy controls (HC) and children with cystic fibrosis (CF) using the sensor correction function outlined above in a custom software. Applying the O₂ and CO₂ sensor correction function had a significant impact on all MBW outcomes (Table 1). Following the sensor correction, mean (SD) FRC and LCI decreased by 8.9 (2.2) % and 11.9 (4.0) %, respectively. The cumulative expired volume (breathing required by the patient, CEV) decreased by 19.6 (5.0) %. The reduced CEV is explained by lower end-expiratory concentrations of N₂, which lead to an earlier end of test. Decreased FRC is explained by slightly lower concentrations of N₂ throughout the measurement. The decrease in CEV is more pronounced than for FRC, and with LCI being the ratio of those two outcomes (LCI = CEV/FRC), this leads to an overall decrease in LCI.

Outcome differences due to the sensor correction can vary greatly for individual measurements (Table 2), and without re-analysis on a signal processing level it is difficult to predict how much outcomes of one measurement will change due to the sensor correction. However, outcomes on the level of a larger number of measurements before and after the correction correlate strongly. Linear fits of corrected outcomes vs standard outcomes have R² values of 0.99 for FRC, and 0.96 for LCI (Figure 5).

The significance of differences in LCI and FRC [L] observed between healthy controls and children with CF in standard processing were preserved after sensor correction (Table 3). On the other hand, the difference in FRC [mL/kg] between groups only became significant after correction.

The change in outcomes following correction is dependent on the magnitude of the outcomes themselves, for both FRC and LCI (Figure 6).

Sensor correction impact on tissue nitrogen

Due to the fact that N₂ is currently overestimated in the presence of CO₂, there is a non-zero N₂ concentration calculated by the Exhalyzer D, even without any N₂ present. This minimum measurable N₂ concentration in current standard processing can be estimated by summing up

the O₂ and CO₂ sensor correction functions (Figure 7B). In conditions reflecting the end of expirations in conditions where N₂ is absent after correction (CO₂ around 5%, rest O₂), the measured concentration of N₂ by the Exhalyzer D in standard processing is 0.94% (Table 1, Figure 7B, intersection of green line with x-axis). For this given concentration of CO₂, the measured N₂ concentration using an Exhalyzer D could theoretically never reach lower end-expiratory values than this. If the Exhalyzer D is therefore used to quantify the impact of tissue nitrogen in long MBW measurements, a significant part of what appears to be a diffusion equilibrium of tissue nitrogen disappears after the sensor correction is applied (Figure 7A). The higher the end-expiratory concentrations of CO₂, the greater the measured (artificial) concentration of N₂ when real N₂ is at 0%. (Figure 7B).

Discussion

Summary

We report a significant measurement error in the Eco Medics Exhalyzer D N₂MBW device. At high concentrations of O₂, and natural end-expiratory concentrations of CO₂, the Exhalyzer D sensors underestimate O₂ and CO₂ gas concentrations and therefore overestimate end-expiratory concentrations of N₂. This strongly affects the end of test criterion, and causes significant error in the assessment of FRC and LCI. It also creates a significant artificial overestimation of measured tissue nitrogen at the end of MBW tests.

i) Sensor accuracy

We are the first group to identify this sensor cross sensitivity error. Previous validation studies as well as the internal testing of Ecomedics did not make special mention of the end of test, end-expiratory conditions examined here¹⁵. It is worth noting that this error remains relatively constrained for the readings of the individual sensors themselves (ca.1% relative error of a sensor reading) even in the most extreme case (low N₂, high CO₂). However, due to the design of the Exhalyzer D MBW device, and due to the fact that high O₂ and high CO₂ concentrations occur at the end of the MBW measurement, this sensor error translates into a relative N₂ error in the range of 47% at the corrected end of test condition. This measurement error greatly exceeds the recommendations for manufacturers outlined in the ATS/ERS consensus statement of measuring within relative 5% of tracer gas concentration²⁰.

ii) Correction function

The sensor error we observed in this study appears systematic and reproducible across different Exhalyzer D systems, with multiple different ground truth references confirming our findings. The correction function required to correct for the sensor error is simple and can be implemented in future signal processing of Spiroware, as well as applied retrospectively to existing data.

iii) Effect size of sensor correction

Sensor correction impact on MBW outcomes

The current erroneous calculated N₂ reading in the end of test end-expiratory condition leads to substantially lowered MBW outcomes. This result provides a new perspective on previously described differences between N₂ and SF6 MBW measurements²¹. It also offers a potential explanation for the differences observed between N₂MBW outcomes measured using the Exhalyzer D and devices by other manufacturers such as ndd Medizintechnik AG (Zürich, Switzerland)²². In both studies, it was observed that primary MBW outcomes from the Exhalyzer D were higher than in the other systems, an observation that may be explained by the systematic overestimation of N₂ by the Exhalyzer D reported in this study. The direction of the change after correction suggests that differences between devices will be smaller. In order to validate this in detail, original data need to be reloaded using the sensor correction described here. Fortunately, the N₂ error has been an overestimation rather than an underestimation, as measurements can now be re-analysed without the worry that the trials might not have recorded data long enough to reach the end of test in the corrected measurement.

A major concern that arises with the publication of this study is that it calls into question previously published results obtained using the Exhalyzer D. It is to be expected that effect sizes, confidence intervals and significance values of MBW outcomes in such studies will change. However, while the impact of the sensor error has effects which are difficult to predict on the level of individual measurements, the impact on MBW outcomes on a large enough number of files appears more systematic. It is therefore reasonable to be optimistic that publications which observed significant differences in MBW outcomes between two groups or as a treatment effect may continue to see a difference after the sensor correction re-analysis – even though the changes observed here suggest that effect sizes will be smaller. In our opinion, re-analysis of MBW measurements should become a priority for those studies where e.g. drug approval was or is based on affected N₂MBW data.

Sensor correction impact on tissue nitrogen

One previously-described feature of N₂ MBW tests is the fact that towards the end of a N₂MBW test the concentration of N₂ in the lungs drops so low that a noticeable amount of N₂ diffuses from the body into the lungs^{21,23}. The results of this study would suggest that the impact of tissue N₂ is significantly lower than previously measured with Exhalyzer D devices. The error measured in the end-expiratory, high-O₂ condition seems to suggest that the Exhalyzer D could not have measured any end-expiratory (CO₂ around 5%) concentrations of N₂ lower than 0.94%, which would significantly perturb estimates of tissue nitrogen. The sensor correction functions introduced in this paper would therefore eliminate a substantial part of the observed tissue nitrogen in measurements performed with the Exhalyzer D (approximately 1% N₂ at equilibrium).

Strenghts and limitations

The main strength of this study is the fact that we were able, through detailed understanding of the underlying signal processing of the Exhalyzer D, to characterize the precise impact of an observed error in gas sensors on the clinical outcomes LCI and FRC. The findings from the technical gases were confirmed by measurements using a Mass Spectrometer, both giving similar results. We developed a correction function and were thus able to estimate the measurement error precisely and correct for it.

The main limitation of this study is the fact that we only had a finite number of gas samples with finite precisions to test the sensors with. We chose a selection of gas concentrations from our range of interest which would exhibit cross-sensitivity effects, but could ultimately not cover the entire range of concentration combinations in MBW measurements using technical gases. However, the phase of the measurement where sensor accuracy is the most relevant for accurate MBW outcomes is the end of test, where thanks to mass spectrometry measurements we can describe the sensor error with high certainty.

Due to the minor contribution of argon (1.2% of N₂) and the negligible contribution of neon, helium, methane, krypton etc.) these gases were omitted in the gas mixtures. In this study, the portion of gases other than oxygen and carbon dioxide is therefore referred to as nitrogen.

This does not change anything about the relative errors observed, although in examples closer to the reality of an MBW test, 1.2% of the fraction attributed here to N₂ would be considered Argon.

Outlook

In the process of conducting the research for this paper, we contacted the manufacturer for information regarding their sensor configurations and questions regarding sensor settings and signal processing. They are working on a solution to both re-analyse past measurements and to improve the signal processing for future measurements in an upcoming version of Spiroware.

Conclusion

An error in the cross sensitivity correction between the oxygen and carbon dioxide gas sensors of the Exhalyzer D device leads to an overestimation of FRC and LCI. Correction of this error is possible but needs to be applied breath-by-breath by re-analysing the measurements in an updated software version.

References

1. Usemann J, Yammine S, Singer F, Latzin P. Inert gas washout: background and application in various lung diseases. *Swiss Med Wkly.* 2017;147:w14483. doi:10.4414/smw.2017.14483.
2. Robinson PD, Goldman MD, Gustafsson PM. Inert gas washout: theoretical background and clinical utility in respiratory disease. *Respiration.* 2009;78(3):339–355. doi:10.1159/000225373.
3. Kent L, Reix P, Innes JA, Zielen S, Le Bourgeois M, Braggion C, Lever S, Arets HGM, Brownlee K, Bradley JM, et al. Lung clearance index: evidence for use in clinical trials in cystic fibrosis. *J Cyst Fibros.* 2014;13(2):123–138. doi:10.1016/j.jcf.2013.09.005.
4. Perrem L, Stanojevic S, Shaw M, Jensen R, McDonald N, Isaac SM, Davis M, Clem C, Guido J, Jara S, et al. Lung Clearance Index to Track Acute Respiratory Events in School-age Children with Cystic Fibrosis. *Am J Respir Crit Care Med.* 2020. doi:10.1164/rccm.202006-2433OC.
5. Stanojevic S, Davis SD, Retsch-Bogart G, Webster H, Davis M, Johnson RC, Jensen R, Pizarro ME, Kane M, Clem CC, et al. Progression of Lung Disease in Preschool Patients with Cystic Fibrosis. *Am J Respir Crit Care Med.* 2017;195(9):1216–1225. doi:10.1164/rccm.201610-2158OC.
6. Frauchiger BS, Binggeli S, Yammine S, Spycher B, Krüger L, Ramsey KA, Latzin P. Longitudinal Course of Clinical Lung Clearance Index in Children with Cystic Fibrosis. *Eur Respir J.* 2020. doi:10.1183/13993003.02686-2020.
7. Nyilas S, Bauman G, Pusterla O, Sommer G, Singer F, Stranzinger E, Heyer C, Ramsey K, Schlegtendal A, Benzrath S, et al. Structural and Functional Lung Impairment in Primary Ciliary Dyskinesia. Assessment with Magnetic Resonance Imaging and Multiple-breath Washout in Comparison to Spirometry. *Ann Am Thorac Soc.* 2018;15(12):1434–1442. doi:10.1513/AnnalsATS.201712-967OC.
8. Kinghorn B, McNamara S, Genatossio A, Sullivan E, Siegel M, Bauer I, Clem C, Johnson RC, Davis M, Griffiths A, et al. Comparison of Multiple-breath Washout and Spirometry in

Children with Primary Ciliary Dyskinesia and Cystic Fibrosis and Healthy Controls. *Ann Am Thorac Soc.* 2020. doi:10.1513/AnnalsATS.201905-375OC.

9. Saunders C, Jensen R, Robinson PD, Stanojevic S, Klingel M, Short C, Davies JC, Ratjen F. Integrating the multiple-breath washout test into international multicentre trials. *J Cyst Fibros.* 2020;19(4):602–607. doi:10.1016/j.jcf.2019.11.006.

10. Subbarao P, Milla C, Aurora P, Davies JC, Davis SD, Hall GL, Heltshe S, Latzin P, Lindblad A, Pittman JE, et al. Multiple-Breath Washout as a Lung Function Test in Cystic Fibrosis. A Cystic Fibrosis Foundation Workshop Report. *Ann Am Thorac Soc.* 2015;12(6):932–939. doi:10.1513/AnnalsATS.201501-021FR.

11. Milla CE, Ratjen F, Marigowda G, Liu F, Waltz D, Rosenfeld M. Lumacaftor/ivacaftor in Patients Aged 6-11 Years with Cystic Fibrosis and Homozygous for F508del-CFTR. *Am J Respir Crit Care Med.* 2017;195(7):912–920. doi:10.1164/rccm.201608-1754OC.

12. Davies J, Sheridan H, Bell N, Cunningham S, Davis SD, Elborn JS, Milla CE, Starner TD, Weiner DJ, Lee P-S, et al. Assessment of clinical response to ivacaftor with lung clearance index in cystic fibrosis patients with a G551D-CFTR mutation and preserved spirometry: a randomised controlled trial. *Lancet Respir Med.* 2013;1(8):630–638. doi:10.1016/S2213-2600(13)70182-6.

13. Davies JC, Sermet-Gaudelus I, Naehrlich L, Harris RS, Campbell D, Ahluwalia N, Short C, Haseltine E, Panorchan P, Saunders C, et al. A phase 3, double-blind, parallel-group study to evaluate the efficacy and safety of tezacaftor in combination with ivacaftor in participants 6 through 11 years of age with cystic fibrosis homozygous for F508del or heterozygous for the F508del-CFTR mutation and a residual function mutation. *J Cyst Fibros.* 2020. doi:10.1016/j.jcf.2020.07.023.

14. Ratjen F, Hug C, Marigowda G, Tian S, Huang X, Stanojevic S, Milla CE, Robinson PD, Waltz D, Davies JC, et al. Efficacy and safety of lumacaftor and ivacaftor in patients aged 6–11 years with cystic fibrosis homozygous for F508del-CFTR: a randomised, placebo-controlled phase 3 trial. *The Lancet Respiratory Medicine.* 2017;5(7):557–567. doi:10.1016/S2213-2600(17)30215-1.

15. Singer F, Houltz B, Latzin P, Robinson P, Gustafsson P. A realistic validation study of a new nitrogen multiple-breath washout system. *PLoS ONE*. 2012;7(4):e36083. doi:10.1371/journal.pone.0036083.
16. Anagnostopoulou P, Latzin P, Jensen R, Stahl M, Harper A, Yammine S, Usemann J, Foong RE, Spycher B, Hall GL, et al. Normative data for multiple-breath washout outcomes in school-aged Caucasian children. *Eur Respir J*. 2020;55(4). doi:10.1183/13993003.01302-2019.
17. Korten I, Kieninger E, Yammine S, Regamey N, Nyilas S, Ramsey K, Casaulta C, Latzin P, For TSSG. The Swiss Cystic Fibrosis Infant Lung Development (SCILD) cohort. *Swiss Med Wkly*. 2018;148:w14618. doi:10.4414/smw.2018.14618.
18. Gustafsson PM, Robinson PD, Lindblad A, Oberli D. Novel methodology to perform sulfur hexafluoride (SF₆)-based multiple-breath wash-in and washout in infants using current commercially available equipment. *J Appl Physiol (1985)*. 2016;121(5):1087–1097. doi:10.1152/japplphysiol.00115.2016.
19. Cummings B, Hamilton ML, Ciaffoni L, Pragnell TR, Peverall R, Ritchie GAD, Hancock G, Robbins PA. Laser-based absorption spectroscopy as a technique for rapid in-line analysis of respired gas concentrations of O₂ and CO₂. *J Appl Physiol (1985)*. 2011;111(1):303–307. doi:10.1152/japplphysiol.00119.2011.
20. Robinson PD, Latzin P, Verbanck S, Hall GL, Horsley A, Gappa M, Thamrin C, Arets HGM, Aurora P, Fuchs SI, et al. Consensus statement for inert gas washout measurement using multiple- and single- breath tests. *Eur Respir J*. 2013;41(3):507–522. doi:10.1183/09031936.00069712.
21. Jensen R, Stanojevic S, Gibney K, Salazar JG, Gustafsson P, Subbarao P, Ratjen F. Multiple-breath nitrogen washout: a feasible alternative to mass spectrometry. *PLoS ONE*. 2013;8(2):e56868. doi:10.1371/journal.pone.0056868.
22. Poncin W, Singer F, Aubriot A-S, Lebecque P. Agreement between multiple-breath nitrogen washout systems in children and adults. *J Cyst Fibros*. 2017;16(2):258–266. doi:10.1016/j.jcf.2016.11.004.

- 422 23. Nielsen N, Nielsen JG, Horsley AR. Evaluation of the impact of alveolar nitrogen
423 excretion on indices derived from multiple-breath nitrogen washout. PLoS ONE.
424 2013;8(9):e73335. doi:10.1371/journal.pone.0073335.

Figure 1

Figure 1: Observed absolute error between known reference and measured O₂ concentrations. (dots: mean of error for one gas mixture, error bars: +/- SD of error). Curves represent a two parameter quadratic polynomial fitted through the error values (see OLS for details), represented here as curves for given CO₂ or O₂ concentrations. Dots represent measurements performed with 12 technical gas mixtures as reference, triangles represent mass spectrometry reference measurements of 6 mixtures at the end of test condition (2% N₂). **A)** Absolute O₂ error as a function of CO₂ concentration, **B)** Absolute O₂ error as a function of O₂ concentration. Dashed curves all represent the same fit $f(O_2, CO_2)$.

Figure 2

Figure 2: Observed absolute error between known and measured CO₂ concentrations. (dots: mean of error of one gas mixture, error bars: +/- SD of error). Curves represent a two parameter quadratic polynomial fitted through the error values (see OLS for details), represented here as curves for given CO₂ or O₂ concentrations. Dots represent measurements performed with 12 technical gas mixtures as reference, triangles represent mass spectrometry reference measurements on 6 mixtures at the end of test condition (2% N₂). **A)** Absolute CO₂ error as a function of CO₂ concentration, **B)** Absolute CO₂ error as a function of O₂ concentration. Dashed curves all represent the same fit $f(O_2, CO_2)$.

443 **Figure 3**

444
445 *Figure 3: N₂ error as a function of measured N₂. Illustration of the impact of the O₂ and CO₂ correction functions*
446 *on final measured N₂. **A)** Summed up correction functions of Figure 1 and Figure 2, and therefore absolute error in*
447 *N₂, represented as curve for selected concentrations of CO₂. **B)** Relative error in measurement of N₂ ((standard-*
448 *corrected)/corrected) as a function of N₂.*

Figure 4

Figure 4: Illustration of the effect of the sensor correction on the N_2 signal and consequently on the end of test of an example MBW measurement. Traced in blue is the signal output of the standard signal processing, corrected signal is shown in red. Vertical dashed lines represent the end of test for the current standard and corrected measurement respectively. The black line corresponds to $1/40^{\text{th}}$ of the initial N_2 concentration (end of test criterium). **A)** Time course of N_2 throughout a standard MBW measurement. **B)** Zoom into the critical period of end of test determination. In this example the test ends 5 breaths earlier in the corrected measurement compared to standard.

Figure 5

Figure 5: MBW outcomes (A) LCI (B) FRC) after sensor correction (corrected) vs current standard (standard), in healthy controls (HC) and patients with cystic fibrosis (CF).

Figure 6

Figure 6: Bland-Altman plot of the absolute difference (corrected – standard) in MBW outcomes (**A**) LCI **B**) FRC) due to sensor correction, plotted as a function of the outcomes themselves (mean of corrected and standard).

Figure 7

Figure 7: Illustration of the effect of the sensor correction function on nitrogen measurement in the late phase of MBW tests. **A)** Example of the equilibrium N_2 reached in a very long continued MBW measurement using the Exhalyzer D, displaying a greatly decreased N_2 -back-diffusion equilibrium (tissue nitrogen). **B)** Corrected N_2 plotted against standard N_2 in conditions around the end of test condition (2% N_2).

Table 1

Condition	Standard			Corrected			N ₂ Error		Contribution	
	[N ₂]	[CO ₂]	[O ₂]	[N ₂]	[CO ₂]	[O ₂]	abs	rel	[CO ₂]	[O ₂]
Standard test-end [%]	2	5	93	1.03	5.14	93.8	0.97	94.4	14.2	85.8
Corrected test-end [%]	2.94	4.87	92.2	2	5	93	0.94	47.2	13.9	86.1
No nitrogen [%]	0.94	4.87	94.2	0	5	95	0.94			

Table 1: Specific examples of sensor error impact on measurement of N₂ in three conditions of interest. Standard concentrations denote concentrations measured in standard Spiroware 3.2.1 processing. Corrected concentrations correspond to concentrations after sensor correction is applied. N₂ error summarizes the absolute difference between N₂ in standard vs. corrected, as well as the relative error (standard-corrected)/corrected. The relative contribution of each sensor to the total error is listed under "Contribution". For simplicity, the presence of Argon is omitted here.

479 **Table 2**

	n	Standard	Corrected	Difference					p-value
				mean	mean [%]	SD [%]	95% CI [%]		
FRC [mL/kg]	147	43.9	40.0	3.9	-8.9	2.2	-9.3 - -8.6		<0.0001
CEV [L]	147	14.9	11.9	3.0	-19.6	5.0	-20.4 - -18.8		<0.0001
LCI [TO]	147	8.3	7.3	1.0	-11.9	4.0	-12.5 - -11.2		<0.0001

480 Bold print indicates statistical significance.

481 *Table 2: Effects of sensor correction on main MBW outcomes. Relative difference is calculated as 100*(corrected*
482 *- standard)/standard.*

483 **Table 3**

		n	Standard			Corrected			Difference				
			mean	SD	p-value [#]	mean	SD	p-value [#]	mean	%	95% CI		p-value*
LCI [TO]	Healthy controls	85	7.12	0.51		6.30	0.40		0.82	11.5	0.76	- 0.88	<0.0001
	Cystic fibrosis	62	9.99	2.21		8.69	1.77		1.30	13.1	1.14	- 1.47	<0.0001
	Difference		-2.87		<0.0001	-2.38		<0.0001					
FRC [L]	Healthy controls	85	1.87	0.95		1.73	0.89		0.14	7.7	0.07	- 0.13	<0.0001
	Cystic fibrosis	62	1.31	0.61		1.17	0.53		0.14	10.7	0.12	- 0.16	<0.0001
	Difference		0.56		<0.0001	0.56		<0.0001					
FRC [mL/kg]	Healthy controls	85	45.1	10.5		41.6	9.8		3.51	7.8	3.33	- 3.70	<0.0001
	Cystic fibrosis	62	42.3	8.0		37.9	7.2		4.40	10.4	4.03	- 4.76	<0.0001
	Difference		2.86		0.0631	3.75		0.0086					

*paired t test; [#]unpaired t test

Bold print indicates statistical significance.

485 *Table 3: Summary of the differences in Lung Clearance Index (LCI) and functional residual capacity (FRC) between healthy controls (HC) and and patients with cystic fibrosis (CF)*
486 *in the retrospective dataset before (standard) and after (corrected) the application of the sensor correction function.*