

Supplementary Table 2. Ag-RDT results with a professional-collected NMT swab and AN swab in RT-PCR positive patients, together with RT-PCR cycle threshold (Ct) values, viral load (in descending order), and duration of symptoms. The positive percent agreement between NMT and AN samples on Ag-RDT, and the respective sensitivities compared to RT-PCR are shown.

No.	Ag-RDT		RT-PCR (OP/NP)		Days of Symptoms
	Prof. NMT	Prof. AN	Ct value	Viral load ¹	
1	pos. (+++)	pos. (+++)	12.65	10.22 ²	1
2	pos. (+++)	pos. (+++)	14.63	9.63 ²	2
3	pos. (+++)	pos. (+++)	15.19	9.46 ²	1
4	pos. (+++)	pos. (+++)	15.25	9.44 ²	2
5	pos. (+++)	pos. (+++)	15.42	9.39 ²	3
6	pos. (+++)	pos. (+++)	15.66	9.32 ²	4
7	pos. (+++)	pos. (+++)	15.71	9.31 ²	1
8	pos. (+++)	pos. (+++)	15.74	9.30 ²	3
9	pos. (+++)	pos. (+++)	15.89	9.25 ²	1
10	pos. (+++)	pos. (+++)	16.58	9.05 ²	2
11	pos. (+++)	pos. (+++)	16.73	9.00 ²	2
12	pos. (+++)	pos. (+++)	16.80	8.98 ²	2
13	pos. (+++)	pos. (+++)	17.08	8.90 ²	4
14	pos. (+++)	pos. (+++)	17.43	8.80 ²	2
15	pos. (+++)	pos. (+++)	17.56	8.76 ²	4
16	pos. (+++)	pos. (+++)	17.63	8.74 ²	2
17	pos. (+++)	pos. (+++)	20.73	8.58 ³	2
18	pos. (+++)	pos. (++)	18.46	8.49 ²	3
19	pos. (+++)	pos. (+++)	18.68	8.42 ²	2
20	pos. (+++)	pos. (+++)	21.30	8.41 ³	1
21	pos. (+++)	pos. (+++)	21.33	8.40 ³	2
22	pos. (+++)	pos. (+++)	18.81	8.39 ²	7
23	pos. (+++)	pos. (+++)	21.75	8.28 ³	4
24	pos. (+++)	pos. (+++)	19.47	8.19 ²	2
25	pos. (+++)	pos. (+++)	22.27	8.12 ³	7
26	pos. (+++)	pos. (+++)	20.33	7.93 ²	4
27	neg.	neg.	20.59	7.86 ²	1
28	pos. (+++)	pos. (+++)	20.88	7.77 ²	1
29	pos. (++)	pos. (+)	20.95	7.75 ²	6
30	pos. (+++)	pos. (+++)	26.81	6.78 ³	1
31	neg.	neg.	25.15	6.50 ²	3
32	neg.	neg.	29.28	6.05 ³	1
33	pos. (++)	pos. (++)	27.18	5.90 ²	5
34	neg.	neg.	27.28	5.87 ²	1
35	pos. (++)	pos. (++)	29.74	5.14 ²	5
36	neg.	neg.	31.38	4.65 ²	5
Sensitivity 31/36 (86.1%)		Sensitivity 31/36 (86.1%)	¹ log ₁₀ SARS-CoV2 RNA copies/ml ² TibMolbiol assay, E-gene target ³ Roche Cobas SARS-CoV-2 assay (E-gene, T2 target)		
Positive percent agreement 100.0%					

Prof., professional sampling; self, self-sampling; AN, anterior nasal; NMT, nasal mid-turbinate; Ag-RDT, Antigen-based rapid diagnostic tool; RT-PCR, Real-time polymerase chain reaction; OP, oropharyngeal; NP, nasopharyngeal; Ct, cycle threshold; No, Number; pos. (+), weak positive; pos. (++) , positive; pos. (+++), strong positive; neg., negative