

Supplementary Table 4. The relevant steps of self NMT-sampling with numbers of participants who precisely followed the instructions, according to observations of study physicians.

Individual steps of self NMT-sampling N=96		Participants following precisely instructions
- removal of the swab from packing	n (%)	80 (83.3%)
- swab only touched at handle		94 (97.9%)
- reinclination of head		83 (86.5%)
- insertion about 2 cm		92 (95.8%)
- angle of the sampling direction		83 (86.5%)
- rotation for minimum 15 seconds		68 (70.8%)
- rotation minimum 4 times		84 (87.5%)
- rubbing against the nasal walls N=95		59 (62.1%)
- bilateral sampling	93 (96.9%)	