Indexing Cerebrovascular Health Using Near-infrared Spectroscopy

Rashid Afkhami^{1*} Frederick Rohan Walker² Saadallah Ramadan³
 Rachel Wong² Sarah Johnson¹

¹ School of Electrical Engineering & Computing, The University of Newcastle

² School of Biomedical Sciences & Pharmacy, The University of Newcastle

³ School of Health Science, The University of Newcastle

Abstract

Near-infrared spectroscopy (NiRS) is a relatively new technology of brain imaging with its potential in the assessment of cerebrovascular health only recently discovered. Encouraging early results suggest that NiRS can be used as an inexpensive and portable cerebrovascular health tracking device using a recently proposed pulse relaxation function (PReFx). In this paper, we propose a new NiRS timing index, TI_{NiRS} , of cerebrovascular health. TI_{NiRS} is a novel use of the NiRS technology. TI_{NiRS} is motivated by the previously proved relationship of the timing of the reflected wave with vascular resistance and compliance in the context of pressure waveforms. We correlated both TI_{NiRS} and PReFx against age, a non-exercise cardiorespiratory fitness (CRF) index, and two existing indices of cerebrovascular health, namely Transcranial Doppler (TCD) augmentation index, AI_{TCD} , and magnetic resonance imaging (MRI) blood flow pulsatility index, PI_{MRI} . The TI_{NiRS} correlations with Age, CRF, PI_{MRI} and AI_{TCD} all are significant, i.e., $r = 0.53$ ($p = 0.002$), $r = -0.44$ ($p = 0.011$), $r = 0.45$ ($p = 0.012$) and $r = 0.46$ ($p = 0.010$), respectively. PReFx, however, did not have significant correlations with any of the vascular health factors. The proposed timing index is a reliable indicator of cerebrovascular aging factors in the NiRS waveform.

¹rashid.afkhami@newcastle.edu.au

Introduction

Vascular stiffening is an important health indicator and influences health in a manner that is independent of blood pressure [1]. The American Heart Association recommends that vascular stiffening be used as a predictor of cardiovascular disease risk [1]. Vascular stiffness is inversely proportional to compliance and reflects arterial resistance to stretching and recoiling in response to blood ejection from the heart. In a young or otherwise healthy vascular system, compliant arteries help smooth out the vascular pulsatility induced by blood pressure. As compliance decreases, blood pressure causes structural changes within the arterial walls that adversely affect their function [2, 3].

Another change that occurs with reduced vascular compliance is increased pulse wave velocity (PWV) [4]. PWV is the speed of the pressure wave traveling from the heart towards the periphery. Simultaneous measurement of arterial pulse in a proximal and a distal arterial sites e.g. common carotid and femoral gives the travel time of the pressure waveform. The distance between the measurement sites can be measured over the body surface. Hence, PWV is calculated as distance divided by travel time between two measurement sites [5]. PWV is commonly measured by applanation tonometry, however, other non-invasive technologies such as MRI and Doppler ultrasound can also be used [5]. A photoplethysmograph (PPG) which uses infrared light emitters and detectors to measure blood volume changes in periphery (e.g. fingertip, ankle, wrist or earlobe), is similarly used to estimate PWV by dividing subject height to the travel time [6, 7, 8].

When a pressure wave reaches a vascular bifurcation point, a portion of the wave is reflected back towards the heart. The reflected wave eventually augments an incident wave coming from the heart, thereby increasing the pulse pressure. Due to the increased PWV in a stiff arterial system, the reflected wave travels faster, thereby reaching the incident wave earlier in systole and increasing the local pulse pressure even further. Both forward and reflected waves leave a peak or an inflection point on the pressure waveform. The time difference between these peaks is believed to be an estimation of the travel time and used in calculation of PWV and PPG SI [9, 6, 10, 11].

Increased arterial stiffness can damage the arterial endothelium in small arteries and can, over time, produce a pathology known as small vessel disease (SVD) [2]. The brain, in particular, is highly vulnerable to SVD, as the blood flow to the sub-surface brain structures is compromised [12, 13, 14]. Based

on a comprehensive review, cerebral SVD is believed to be one of the main causes of vascular cognitive impairments and accounts for a quarter of all acute ischaemic strokes [15].

Cerebrovascular compliance is commonly assessed non-invasively using indices derived from blood pressure and flow readings. Transcranial Doppler (TCD) ultrasound and magnetic resonance imaging (MRI) are the modalities most widely used to acquire cerebrovascular compliance and cerebrovascular health indices. TCD measures blood flow velocity; i.e., the speed at which the blood cells move inside a vessel. While it is possible to estimate compliance using TCD blood flow velocity [16], the method relies on estimates of vascular cross-sectional area, arterial inflow and venous outflow, which are not readily measurable. The TCD indices such as the pulsatility index, PI_{TCD} , and augmentation index, AI_{TCD} , are the commonly reported TCD cerebrovascular health indicators [17, 18, 19]. Of these two, PI_{TCD} is the most commonly used TCD index due to its ease of calculation. However, AI_{TCD} has been shown to have a stronger correlation with cerebrovascular health indicators such as age and cardiorespiratory fitness (CRF) [20]. The MRI pulsatility index (PI_{MRI}) is the commonly adopted MRI-based cerebrovascular health measure [14]. While MRI has a high spatial resolution for localized measurements, it is less commonly used as it is both costly and time consuming, making it unsuitable for simple and routine monitoring of vascular health in the general population.

Near-infrared spectroscopy (NiRS) is a relatively new technology that measures regional blood volume changes based on absorption of near-infrared light, mainly by oxygenated blood inside cerebral arteries [21, 22, 23]. NiRS offers high temporal resolution and is sufficiently sensitive to detect changes in blood volume during the cardiac cycle. Therefore, NiRS allows local studies of cerebrovascular behaviour and possible assessment of cerebrovascular health [21]. The recently proposed pulse relaxation function (PReFx) is the first NiRS based approach to the indexing of cerebrovascular compliance [21]. PReFx measures the deformation of the blood volume waveform during the systolic relaxation phase due to the presence of reflected waves. It has been demonstrated to correlate with CRF and inversely correlate with age [21, 24, 25]. PReFx is an area measurement in the averaged NiRS signal. The PReFx algorithm assumes that the first positive peak after the diastolic minimum is the systolic peak, which may not be the case when the systolic maximum is an inflection point rather than a peak. Therefore, we are motivated to use an index that can be easily applied to NiRS signals. In our

previous works we have proposed a new timing index (TI) which has been shown to significantly correlate with vascular health factors such as age and CRF in a mathematical model of the pressure wave [11] and in an experimental study using TCD waveforms [20]. TI is defined as the inverse of the reflection time, i.e. the time difference between the systolic and reflected peaks, and using a tube-based model of the arterial tree, the timing index has been proven to be directly controlled by vascular aging factors including compliance, resistance and PWV [11]. In fact, it has been shown that the reflection time is more than the time it takes for the blood waveform to travel the distance between the measurement site and the reflection site and the waveform faces an additional delay before turning back from the reflection site which is a function of resistive and capacitive properties of the vascular system beyond the reflection site [11]. Therefore, the reflection time is a sum of two time delays, one is a line delay controlled by PWV and physical distance between the measurement and reflection sites and the other is a load delay forced by the vascular properties beyond the reflection site. All these parameters are known to change with ageing and vascular health and thus making TI a powerful vascular stiffness index [11]. As such, we propose a new NiRS timing index (TI_{NiRS}) to directly measure the timing of reflected waves. Here, the NiRS timing index is defined as the inverse of the time between the systolic and reflected waveform peaks in the NiRS signal. Then, we will examine the correlation of the proposed index with other health indices including age, CRF and the most currently used vascular health measures derived from other imaging modalities, namely TCD augmentation index and MRI pulsatility index.

Methods

Protocol and Data Collection

Thirty eight adults (23 female, 15 male, age range = 24-67 years, mean age = 41.7 years) were recruited from the Newcastle, Australia. They provided informed consent prior to assessment. The study protocol was approved by the University of Newcastle Human Research Ethics Committee and is registered in the Australian and New Zealand Clinical Trials Registry (ACTRN12619000144112). All experiments were performed in accordance with relevant guidelines and regulations. TCD recordings from these participants

were previously reported in [20]. Each participant attended the Hunter Medical Research Institute and the University of Newcastle Callaghan campus over two consecutive days and participated in three scanning sessions. Participants were asked to refrain from consuming caffeine before their scans. Height, weight, age, sex and resting heart rate were recorded for each participant and a physical activity questionnaire was completed.

TCD ultrasound (DopplerBox X; Compumedics DWL, Singen, Germany) was used to record resting-state cerebral blood flow velocity from the right and left middle cerebral arteries. Participants were also scanned on a 3T MRI scanner (Magnetom Prisma, Siemens Healthineers, Erlangen, Germany), equipped with 64-channel receive only head coil, while a standard built-in dual channel body coil was used for RF transmission. Blood flow was quantified using a phase contrast flow quantification sequence ($TR = 26.5\text{ms}$, $TE = 6.9\text{ms}$, slice thickness = 5mm, matrix 256×256) of a single excitation with a velocity encoding value of 120cm/s on middle cerebral arteries. Finally, participants were instructed to relax and sit still for 300 seconds of resting-state NiRS recording. A frequency-domain NiRS (ISS Imagent, Champaign, Illinois, USA) was used at a 110MHz modulation frequency and sampling rate of 39.0625Hz. The setup montage was designed to cover the majority of the frontal lobe and consisted of four detectors, each crossed with 16 time-multiplexed sources (eight of which operated at 690nm and the other eight at 830nm), making a total of 64 channels. Participants were asked to wear an Equivital EQ02 LifeMonitor sensor belt during this session, which recorded ECG and NiRS signals simultaneously.

During the TCD imaging session, resting-state heart rate was measured using an heart rate monitoring device (Omron HEM-7320). Participants assumed a sitting position for approximately five minutes during headpiece setup before recording commenced. During TCD recording, the heart rate was measured once per minute three times. The mean of these three heart rate measurements was used as the resting-state heart rate.

Calculating the Indices

MRI Pulsatility Index

MRI phase contrast images were processed using the scanner software (Siemens Syngo) by manually placing a region of interest around the middle cerebral artery and quantifying the flow. The MRI pulsatility index, PI_{MRI} , was then

obtained by subtracting the minimum flow from the maximum flow and dividing it by the mean flow over the cardiac cycle [14]; i.e.,

$$PI_{MRI} = \frac{F_{\max} - F_{\min}}{\bar{F}}, \quad (1)$$

where F indicates the flow and the bar sign indicates the mean value. Three subjects did not have an MRI scan and among the remaining 35 subjects, PI_{MRI} was calculated either for both middle cerebral arteries ($n = 21$) or for a single middle cerebral artery (MCA) ($n = 14$) based on data quality. The indices from right and left MCAs were then averaged for each subject.

TCD Augmentation Index

Calculation of the augmentation index, AI_{TCD} , requires identification of three characteristic points on an averaged Doppler waveform. Peak systole (V_{sys}), peak diastole (V_{dia}) and the peak of the reflected wave (V_{refl}) are the parameters required to calculate the augmentation index [18]:

$$AI_{TCD} = \frac{V_{refl} - V_{dia}}{V_{sys} - V_{dia}}. \quad (2)$$

Thus, the timing of the diastolic peaks was first determined automatically in MATLAB using a peak detector function. Then, each signal was averaged with respect to its diastolic peaks, resulting in an averaged signal covering the duration of a single cardiac cycle. The systolic and reflected peaks were set manually on the averaged signal. The TCD data for two subjects were excluded from the analysis due to noise. The AI_{TCD} index was calculated using data from both MCAs (and averaged) for 34 subjects and from a single MCA for two subjects, depending on data quality.

Evaluating Cardiorespiratory Fitness

The CRF was determined using the non-exercise test model proposed in [26]. It involves a self-reported physical activity score, age, sex, body mass index and resting heart rate. See the *supplementary materials* for the physical activity questionnaire and cardiorespiratory fitness formula.

NiRS Indices

Optical light intensity NiRS signals were used to derive two different indices: the pulse relaxation function (PReFx) and the proposed timing index, TI_{NiRS} . The pre-processing of the recordings was done similarly for both methods and was adopted from [21] using the “*p_pod*” package [27] which is a MATLAB-based software module developed by the same group. Specifically, channels with source detector distances of 20-60 mm were selected and went through a quality-control process passing 1283 channels out of the original $38 \times 64 = 2432$ channels. These channels were then normalized by dividing by their mean values and then bandpass-filtered to 0.5-5.0 Hz to capture the dynamics of the pulse signal and remove any low-frequency elements and moving artefacts. The resulting signal was averaged with respect to the ECG R peaks, and then rotated around the x-axis (see Fig. 1). The result is called the arterial pulsation signal, s . The rotation was performed so that the systolic peak would appear as a positive peak and the shape of the signal would look similar to that of a blood pressure, flow or volume signal. This facilitates understanding of the waveform. The arterial pulsation signals were visually inspected and channels that did not have a pulse shape were removed (i.e., low blood volume at the beginning and end representing diastole, and an increase in between representing systole). Some 964 channels passed the pre-processing stage for further analysis. Channels with source detector distances between 25 and 50 mm proved to be most likely to generate usable waveforms.

PReFx: The pulse relaxation function is calculated based on the methods explained in [21]. In brief, s_{D1} and s_{D2} peaks occur as minima in the s signal in the first and the second cardiac cycles and s_S occurs as the first local maximum after s_{D1} (see Fig. 2). $\{s_{D1}, s_{D2}\}$ and s_S are intended to represent the diastolic and systolic peaks, respectively. However, s_S is not necessarily a systolic peak, see Fig.2b where s_S is the reflected peak and the systolic peak indicated by an arrow. Then, the area enclosed by the s signal between s_S and s_{D2} is calculated as A and inserted into

$$PReFx = \frac{A}{B} - 0.5. \quad (3)$$

Where B is the area of a rectangle formed by s_S and s_{D2} at its diagonal ends (i.e. the rectangle formed by blue and red lines in Fig. 2). Note that if the drop in volume between peaks was linear, PReFx would equal

Figure 1: (a) Ten-second-long raw NiRS signal (blue) and simultaneously recorded ECG lead I (red). (b) Averaged and flipped NiRS s signal.

zero. After communicating with the authors of [21], channels with PReFx value outside the interval of $[-0.1, 0.4]$ were removed and other channels were manually excluded from the calculations where reflected peaks were selected by the algorithm (see Fig. 2b for an example). A total of 245 additional channels were removed in this process. Finally, for each subject with more than ten PReFx values, an averaged PReFx value was calculated as the index of vascular compliance. We were able to calculate PReFx for 29 of the 38 subjects.

TI_{NiRS}: We define the timing index as the inverse of the time between the forward and reflected waveform peaks. In order to find the timing of these peaks, named t_{sys} and t_{refl} for peak systole and peak reflection, respectively, for a given signal, s , a completely automated algorithm is proposed (Sample data and codes available at <https://github.com/Rashid-Afkhami>). The algorithm uses the second derivative to locate key points in the signal, this is based on similar approaches that have been proposed for locating inflection point in the pressure waveform [28]. Firstly, the second derivative, s_2 , of s is calculated. Next, the following time points are defined on the s and the s_2 signals: $\{t_{p1}, t_{p2}, \dots\}$, the time of the peaks on the s signal and $\{t_{zc1}, t_{zc2}, \dots\}$ the time of zero-crossings on the s_2 signal. If the number of s_2 zero-crossings is less than 4, the channel is excluded. Otherwise,

$$t_{\text{sys}} = \min(t_{p1}, t_{zc2}). \quad (4)$$

Then, to define t_{refl} the algorithm checks only the time period between t_{zc3}

Figure 2: Examples of calculating PReFx for two different s signals. The signal s is shown in black, the s_S , s_{D1} and s_{D2} peaks are marked with black circles, and the blue and red rectangles are formed with s_S and s_{D2} at their diagonals. The shaded area, A , is the area enclosed by s between s_S and s_{D2} and the red sides of the rectangle. The calculated PReFx values for (a) and (b) are 0.136 and 0.081, respectively.

and t_{zc4} . If s has a positive peak, t_{refl} is set as the time point corresponding to this peak, t_{pi} , otherwise, it is set as the time when s_2 reaches its minimum, t_{min} (all in the specified time period between t_{zc3} and t_{zc4}) i.e.,

$$t_{refl} = \begin{cases} t_{pi} & \text{if } s \text{ contains a positive peak} \\ t_{min} & \text{else} \end{cases} \quad (5)$$

After all channels are processed for a subject, channels with t_{sys} and t_{refl} outside their mean ± 1.5 std (i.e., standard deviation) or outside a predefined threshold were also excluded from further calculations. The predefined threshold for t_{sys} was any value < 125 ms and for t_{refl} it was any value > 500 ms. Then, if ten or more channels remained, the t_{sys} and t_{refl} values for each channel were used to calculate the timing index as:

$$TI_{NIRS} = \frac{1}{t_{refl} - t_{sys}} \quad (6)$$

The timing index values of different channels were then averaged for each participant. Of the original 38 subjects, we were able to calculate a TI for 32

of them. Examples of the proposed algorithm are shown in Fig. 3 when the algorithm is applied to in vivo and simulated signals. Three main types of the arterial pulsation signals, s , are shown in this figure where the detected t_{sys} and t_{refl} correspond to either a peak point or an inflection point on the recorded signal. The simulated signals are formed by adding three waveforms, an incident forward-traveling waveform, a reflected waveform and a re-reflected waveform (which is also traveling towards the periphery). The purpose of the simulated waveforms is to better visualize where the underlying waveforms reach their peak and how the algorithm performs in terms of detecting the peak times. For each simulated case the underlying waveforms are arranged in a form that the resulted observed signal (black line) resembles an in vivo case. Using a NiRS device only the composite signal s is detected and the wave components shown in Fig. 3b, d and f are for illustrative purposes only.

Statistical Analysis

Statistical analysis was carried out in MATLAB using the inbuilt *corr* function. Lilliefors test of normality with a 5% significance level was used to assess the data distribution. Age and PI_{MRI} were found to be non-normally distributed ($p = 0.0072$ and 0.0020 , respectively). All other distributions were normal. Pearson or Spearman correlation coefficients were calculated for normally-distributed and non-normally-distributed data, respectively. False discovery rate correction was used to adjust significance levels with an initial $\alpha = 0.05$.

Results

The correlations of the proposed NiRS timing index and existing NiRS PReFx index with the indices derived from other imaging modalities, age and CRF are reported in Table 1. Corresponding scatter-plots are shown in Fig. 4. Least-squares linear models have not been provided for Fig 4.G and Fig 4.H due to small slopes. Significant correlations after correcting for multiple comparisons are marked with an asterisk. Note that PReFx is an index of vascular compliance whereas TI_{NiRS} is an index of vascular stiffness that is inversely proportional to compliance. Therefore, the two indices show opposite signs when correlated with the same factors. The proposed systolic

Figure 3: Three main types of in vivo arterial pulsation signals are shown on the left column (a), (c) and (e) and an equivalent simulated signal is shown on the right column (b), (d) and (f), respectively. Black lines are the observed recorded blood volume signal, s and the red lines are the second derivatives, s_2 . The gray lines represent the underlying forward, reflected and re-reflected waveforms. The x-axis represents time in seconds.

Table 1: Correlation matrix with correlation coefficients (and p -values). Asterisks indicate statistically significant correlations after correction for multiple comparisons using a false discovery rate

	CRF	Age	AI _{TCD}	PI _{MRI}
TI _{NIRS}	-0.44* (0.011)	0.53* (0.002)	0.46* (0.010)	0.45* (0.012)
PReFx	0.24 (0.203)	-0.40 (0.030)	-0.26 (0.173)	-0.01 (0.978)

peak and reflected wave peak detection algorithm for TI_{NIRS} detected 228 channels (out of 964 input channels) with a systolic peak corresponding to an inflection point (mean participant age for these channels = 42.5 years). Both TI_{NIRS} (the peak detection algorithm) and PReFx indices are affected by the sampling frequency of 39.0625Hz (i.e. ≈ 25 ms sampling interval) used in this study; a faster sampling rate would allow a higher resolution and more accurate results.

Discussion

The NiRS indices are designed to measure vascular stiffness, which is well correlated with age [1] and cardiorespiratory fitness [29]. Thus, a common approach with any vascular index is to compare its relationship with age and CRF [30, 21] and here we have done the same. The PReFx correlation coefficient reported here for age is consistent with existing studies; however, the correlation with CRF is lower than previously reported. PReFx correlations with age and CRF in the adult population have so far been studied by the same research group using two different datasets [21, 24]. Reported correlations with age are $r = -0.39$ (mean age = 69.87 years) [21] and $r = -0.43$ (mean age = 47.8 years) [24]; a correlation of $r = -0.60$, mean age = 58.8 years was reported in [25] for a combination of the two datasets. Our reported correlation, $r = -0.40$ ($p = 0.030$), although consistent with the previous values, is not significant. Reported correlations with CRF are $r = 0.416$ [21] and $r = 0.32$ [24], however, our correlation with CRF of $r = 0.24$ ($p = 0.203$) was not significant. Given that the mean age for our sample is 41.7 years, we suspect that the PReFx measure has a lower correlation in younger subjects

Figure 4: Scatter plots of the correlations with least-squares fit line. (a), (b), (c) and (d) represent TI_{NiRS} scatter plots with CRF, age, AI_{TCD} and PI_{MRI} , respectively. (e), (f), (g) and (h) represent $PReFx$ scatter plots with CRF, age, AI_{TCD} and PI_{MRI} , respectively. The red bars in (a) and (e) show TI_{NiRS} and $PReFx$ standard deviations, respectively.

than older ones. Based on the performance of the PReFx algorithm on the current data, we have identified two potential sources that can distort the PReFx calculation and potentially cause lower correlations. Firstly, there are cases where the PReFx algorithm detects reflected peaks instead of systolic peaks as the s_S points. The PReFx algorithm assumes the first local maximum after the minimum peak, i.e. s_S , to be the systolic peak, which is not accurate when the true systolic peak is in the form of an inflection point (e.g., see Fig. 2b and 3b). Based on our analysis these cases account for almost one-fourth of all channels (228 of the 964 studied channels) and are more likely to occur in older participants in which the reflected wave moves faster and reaches the incident wave sooner than in younger participants. The second problem occurs in channels with a notch in the signal after the systolic and reflected peaks, which is likely caused by closure of the aortic valve (similar to the dicrotic notch seen in the pressure waveform [31]). This notch can appear as a large deformation in the signal (see Fig. 2a for an example), which will change PReFx values independently of vascular health factors. Among the aforementioned two potential causes of low correlation for PReFx, the first was eliminated in the results reported in Table 1 by discarding channels where s_S was assigned to the reflected peak. However, if all the channels are included in the calculations, it will remove the significant correlation of PReFx with age; i.e., $r = -0.19$.

While no previous NiRS studies have reported a timing index, the correlations of TI_{NiRS} with age and CRF match those of the TCD reported by our group [11]. That is, correlations of $r = 0.70$ and $r = 0.50$ between TCD TI and age in two different datasets and a correlation of $r = -0.79$ between TCD TI and CRF. The proposed NiRS timing index, TI_{NiRS} , also shows significant correlations with age ($r = 0.53$, $p = 0.002$) and CRF ($r = -0.44$, $p = 0.011$). TI_{NiRS} correlated more strongly with both age and CRF than PReFx. This suggests that TI may be a more sensitive index than PReFx to age and lifestyle related cerebrovascular changes in middle-aged participants.

Based on a previously proven association between PI_{MRI} and SVD [14], we expect NiRS indices to correlate with the MRI pulsatility index which is an indicator of vascular health. PReFx showed no significant correlation with PI_{MRI} ($r = -0.01$, $p = 0.978$), whereas TI_{NiRS} correlated significantly with PI_{MRI} ($r = 0.45$, $p = 0.012$). PI_{MRI} indicates vascular health by measuring the pulsatility of the blood flow. High pulsatility means that fast traveling reflected waves join incident waves and contribute to increased maximum flow peaks. Similarly, TI_{NiRS} reaches greater values when the reflected wave

peak is closer to the systolic peak and pulsatility is high. PReFx also changes based on the location of the reflected wave; high PReFx values correspond to channels with high deformation from a perfect sinusoidal waveform, i.e., the existence of a reflected peak. However, the presence of a dicrotic notch may also affect the PReFx index by distorting the wave shape while having no effect on PI or TI. Thus, in channels with a prominent dicrotic notch the PReFx may be high while pulsatility is low, which may explain the lower correlation for PReFx.

TI_{NIRS} showed a significant correlation with the TCD augmentation index ($r = 0.46$, $p = 0.010$), while the correlation of PReFx with AI_{TCD} was not significant ($r = -0.26$, $p = 0.173$). The augmentation index is a widely accepted index used in both blood pressure and flow velocity (TCD) signals, and is a strong predictor of vascular aging and stiffening [32, 18, 19, 33]. The AI captures the augmentation of the incident wave by the reflected wave, with smaller values reflecting no augmentation when the arteries are healthy and the propagation speed is low. TI indexes this delay between the incident and reflected waves directly, which is why AI_{TCD} and TI_{NIRS} are significantly correlated. The PReFx also correlates with AI as larger PReFx values can correspond to late reflected waves leaving a peak on the systolic-to-diastolic segment of the arterial pulsation signal and thus increasing the area under the curve. Nevertheless, the correlation may be weakened when the presence of dicrotic notch increases PReFx regardless of the reflected wave.

Overall, NiRS blood volume waveforms are believed to closely resemble the changes in blood pressure [21] which is also supported by the use of infrared light in PPG and its morphological similarity to the arterial blood pressure waveform [22, 34]. Thus, we believe that the proposed NiRS timing index is affected similarly by vascular health parameters in the same way as the timing index derived from the blood pressure waveforms [11]. As blood pressure can not be readily measured in cerebral vessels, a NiRS-based timing index can provide an alternative approach.

Conclusion

In this paper, we presented a novel NiRS based cerebrovascular stiffness index, TI_{NIRS}. The TI_{NIRS} correlates significantly with age, CRF and other cerebrovascular health indices derived from TCD and MRI data, indicating it has high performance in tracking changes in the cerebrovascular system.

NIRS offers a potentially valuable means of indexing vascular health and has superior cost, portability and widespread implementation potential compared with existing techniques.

References

- [1] R. R. Townsend et al. Recommendations for improving and standardizing vascular research on arterial stiffness. *Hypertension*, 66(3):698–722, 2015.
- [2] Marielle M.F. Poels, Keren Zaccai, Germaine C. Verwoert, and et. al. Arterial stiffness and cerebral small vessel disease. *Stroke*, 43(10):2637–2642, 2012.
- [3] Julie C. Kohn, Marsha C. Lampi, and Cynthia A. Reinhart-King. Age-related vascular stiffening: Causes and consequences. *Frontiers in Genetics*, 6:112, 2015.
- [4] Gerard M. London, Bruno Pannier, and Michel E. Safar. Arterial stiffness gradient, systemic reflection coefficient, and pulsatile pressure wave transmission in essential hypertension. *Hypertension*, 74(6):1366–1372, 2019.
- [5] Jochen Steppan, Viachaslau Barodka, Dan E. Berkowitz, and Daniel Nyhan. Vascular stiffness and increased pulse pressure in the aging cardiovascular system. *Cardiology Research and Practice*, 2011(24):263585, 2011.
- [6] S. C. Millasseau, R. P. Kelly, J. M. Ritter, and P. J. Chowienczyk. Determination of age-related increases in large artery stiffness by digital pulse contour analysis. *Clinical Science*, 103(4):371–377, 08 2002.
- [7] Richard J. Woodman, Bronwyn A. Kingwell, Lawrence J. Beilin, Sarah E. Hamilton, Anthony M. Dart, and Gerald F. Watts. Assessment of central and peripheral arterial stiffness*: Studies indicating the need to use a combination of techniques. *American Journal of Hypertension*, 18(2):249–260, 02 2005.

- [8] Q. Yousef, M. B. I. Reaz, and M. A. M. Ali. The Analysis of PPG Morphology: Investigating the Effects of Aging on Arterial Compliance. *Measurement Science Review*, 12(6):266–271, December 2012.
- [9] Jun Sugawara, Koichiro Hayashi, and Hirofumi Tanaka. Distal shift of arterial pressure wave reflection sites with aging. *Hypertension*, 56(5): 920–925, 2010.
- [10] Philip J Chowienczyk et al. Photoplethysmographic assessment of pulse wave reflection. *Journal of the American College of Cardiology*, 34(7): 2007–2014, 1999.
- [11] Rashid Afkhami and Sarah Johnson. Wave reflection: More than a round trip. *Medical Engineering & Physics*, 92:40–44, 2021.
- [12] Qian Li, Yang Yang, Cesar Reis, Tao Tao, Wanwei Li, Xiaogang Li, and John H. Zhang. Cerebral small vessel disease. *Cell Transplantation*, 27(12):1711–1722, 2018.
- [13] T. Michael De Silva and Frank M. Faraci. Contributions of aging to cerebral small vessel disease. *Annual Review of Physiology*, 82(1):275–295, 2020.
- [14] J. Birnefeld et al. Cerebral arterial pulsatility is associated with features of small vessel disease in patients with acute stroke and TIA: A 4D flow MRI study. *J Neurol*, 267(3):721–730, 2020.
- [15] Leonardo Pantoni. Cerebral small vessel disease: from pathogenesis and clinical characteristics to therapeutic challenges. *The Lancet Neurology*, 9(7):689 – 701, 2010.
- [16] Emmanuel Carrera, Dong-Joo Kim, Gianluca Castellani, Christian Zweifel, Peter Smielewski, John D. Pickard, and Marek Czosnyka. Effect of hyper- and hypocapnia on cerebral arterial compliance in normal subjects. *Journal of Neuroimaging*, 21(2):121–125, 2011.
- [17] RG Gosling and DH King. Arterial assessment by Doppler-shift ultrasound. *Proceedings of The Royal Society of Medicine-London*, 67(6): 447–449, 1974.

- [18] Kozo Hirata, Toshio Yaginuma, Michael F. O'Rourke, and Masanobu Kawakami. Age-related changes in carotid artery flow and pressure pulses. *Stroke*, 37(10):2552–2556, 2006.
- [19] Gary F. Mitchell et al. Arterial stiffness, pressure and flow pulsatility and brain structure and function: The Age, Gene/Environment Susceptibility – Reykjavik Study. *Brain*, 134(11):3398–3407, 11 2011.
- [20] Rashid Afkhami, Rachel Wong, Saadallah Ramadan, Frederick Rohan Walker, and Sarah Johnson. Indexing cerebrovascular health using Transcranial Doppler ultrasound. *Ultrasound in Medicine & Biology*, 47(4): 919–927, 2021.
- [21] Monica Fabiani et al. Taking the pulse of aging: Mapping pulse pressure and elasticity in cerebral arteries with optical methods. *Psychophysiology*, 51(11):1072–1088, 2014.
- [22] Edward D. Chan, Michael M. Chan, and Mallory M. Chan. Pulse oximetry: Understanding its basic principles facilitates appreciation of its limitations. *Respiratory Medicine*, 107(6):789 – 799, 2013.
- [23] R. Afkhami, F. R. Walker, S. Ramadan, and S. J. Johnson. A dynamic model of brain hemodynamics in near-infrared spectroscopy. *IEEE Transactions on Biomedical Engineering*, 67(7):2103–2109, 2020.
- [24] Chin Hong Tan et al. Mapping cerebral pulse pressure and arterial compliance over the adult lifespan with optical imaging. *Plos One*, 12 (2), 2017.
- [25] Chin Hong Tan, Kathy A. Low, et al. Optical measures of cerebral arterial stiffness are associated with white matter signal abnormalities and cognitive performance in normal aging. *Neurobiology of Aging*, 84: 200 – 207, 2019.
- [26] Radim Jurca et al. Assessing cardiorespiratory fitness without performing exercise testing. *American Journal of Preventive Medicine*, 29(3): 185–193, 2005.
- [27] University of Illinois. University of illinois cognitive neuroimaging laboratory optical brain imaging software, 2017. URL <https://www.nitrc.org/projects/uofi-cn1-opt/>. [Last Accessed: March 23rd 2020].

- [28] Patrick Segers et al. Assessment of pressure wave reflection: Getting the timing right! *Physiological Measurement*, 28(9):1045–1056, aug 2007.
- [29] Colin A. Boreham, Isabel Ferreira, Jos W. Twisk, Alison M. Gallagher, Maurice J. Savage, and Liam J. Murray. Cardiorespiratory fitness, physical activity, and arterial stiffness. *Hypertension*, 44(5):721–726, 2004.
- [30] E. Patvardhan, K. S. Heffernan, J. Ruan, M. Hession, P. Warner, R. H. Karas, and J. T. Kuvin. Augmentation index derived from peripheral arterial tonometry correlates with cardiovascular risk factors. *Cardiology Research and Practice*, 2011, 2011.
- [31] Mahesh Nirmalan and Paul M Dark. Broader applications of arterial pressure wave form analysis. *Continuing Education in Anaesthesia Critical Care & Pain*, 14(6):285–290, 02 2014.
- [32] Katsuhiko Kohara et al. Radial augmentation index: A useful and easily obtainable parameter for vascular aging. *American Journal of Hypertension*, 18(S1):11S–14S, 01 2005.
- [33] Kurji Alyakhan, Chantel Debert, William A. Whitelaw, Jean M. Rawling, Richard Frayne, and Marc J. Poulin. Differences between middle cerebral artery blood velocity waveforms of young and postmenopausal women. *The Journal of The North American Menopause Society*, 13(2): 303–313, 2006.
- [34] M.D. Reisner, Andrew, Ph.D. Shaltis, Phillip A., Devin McCombie, and Ph.D. Asada, H Harry. Utility of the Photoplethysmogram in Circulatory Monitoring. *Anesthesiology: The Journal of the American Society of Anesthesiologists*, 108(5):950–958, 05 2008.