
Supplemental material

**Supplemental Figure 1.** Signal to cut off ratio (S/Co) by calendar month April (4) until September (9). Red lines represent thresholds. Abbott-NP (1.4) (n=54 positive) based on the manufacture's recommendations. Spike (0.190) (n=228 positive); RBD (0.186) (n=104 positive); and NP (0.396) (n=214 positive). Abbott- Architect SARS-Cov-2 IgG assay which targets the nucleocapsid antigen and three in-house IgG ELISA chemiluminescent assays recognizing distinct recombinant viral antigens: full length spike glycoprotein (Spike), spike glycoprotein receptor binding domain (RBD), and nucleocapsid (NP).



	Overall		April	May	June	July	Aug	Sept	
Province									
	British Columbia	1358	15%	228 (15%)	228 (15%)	227 (15%)	229 (15%)	228 (15%)	218 (14%)
	Alberta	1862	21%	309 (21%)	309 (21%)	309 (21%)	308 (21%)	308 (21%)	319 (21%)
	Saskatchewan	439	5%	77 (5%)	54 (4%)	77 (5%)	77 (5%)	77 (5%)	77 (5%)
	Manitoba	496	6%	79 (5%)	102 (7%)	80 (5%)	78 (5%)	78 (5%)	79 (5%)
	Ontario	3889	43%	648 (43%)	647 (43%)	648 (43%)	649 (43%)	649 (43%)	648 (43%)
	Atlantic	955	11%	159 (11%)	159 (11%)	159 (11%)	159 (11%)	160 (11%)	159 (11%)
Age group									
	17-19	185	2%	23 (2%)	19 (1%)	36 (2%)	33 (2%)	44 (3%)	30 (2%)
	20-29	1525	17%	253 (17%)	248 (17%)	240 (16%)	232 (15%)	286 (19%)	266 (18%)
	30-39	1655	18%	286 (19%)	270 (18%)	284 (19%)	255 (17%)	244 (16%)	316 (21%)
	40-49	1453	16%	249 (17%)	247 (16%)	207 (14%)	233 (16%)	271 (18%)	246 (16%)
	50-59	1895	21%	336 (22%)	319 (21%)	338 (23%)	333 (22%)	292 (19%)	277 (18%)
	60-69	1728	19%	284 (19%)	300 (20%)	282 (19%)	312 (21%)	276 (18%)	274 (18%)
	70-79	518	6%	66 (4%)	90 (6%)	103 (7%)	95 (6%)	77 (5%)	87 (6%)
	80+	40	0%	3 (0%)	6 (1%)	10 (1%)	7 (0%)	10 (1%)	4 (0%)
Sex									
	Female	4250	47%	719 (48%)	699 (47%)	669 (45%)	725 (48%)	728 (49%)	710 (47%)
	Male	4749	53%	781 (52%)	800 (53%)	831 (55%)	775 (52%)	772 (51 %)	790 (53%)

Supplemental Table 1. Baseline Characteristics of blood donor populations over time

## Supplemental Table 2. Assay Characteristics overall and over time; based on the BLCM Posterior Means (95% Creditable Interval),

## Informative

	Overall				April/May		June/July		August/September	
	Sensitivity	PPV	Specificity	NPV	Sensitivity	Specificity	Sensitivity	Specificity	Sensitivity	Specificity
Spike	93.5%	27.7%	98.1%	99.8%	95.0%	98.0%	93.6%	98.8%	93.6%	97.6%
	(88.7, 97.3)	(23.2, 33.7)	(97.9, 98.4)	(99.6, 100.0)	(90.1, 98.2)	(97.4, 98.4)	(88.0, 97.6)	(98.4, 99.2)	(88.0, 97.6)	(97.0, 98.1)
RBD	89.1%	59.8%	99.5%	99.8%	89.3%	99.5%	89.2%	99.6%	88.7%	99.3%
	(84.1, 93.5)	(50.0, 71.4)	(99.3, 99.7)	(99.6, 99.9)	(83.7, 93.8)	(99.3, 99.8)	(98.4, 99.2)	(83.6, 93.9)	(83.0, 93.4)	(99.0, 99.6)
NP	78.8%	21.0%	98.2%	99.6%	79.9%	99.5%	80.5%	97.3%	78.5%	97.6%
	(74.1, 83.2)	(17.2, 25.3)	(97.9, 98.4)	(99.3, 99.7)	(74.9, 84.5)	(99.2, 99.7)	(75.6, 84.9)	(96.7, 97.9)	(73.5, 83.3)	(97.1, 98.1)
Abbott-	58.5%	87.5%	99.8%	99.4%	77.3%	99.7%	60.2%	99.8%	64.4%	99.9%
NP	(46.3, 70.6)	(81.3, 93.8)	(99.7, 99.9)	(99.1, 99.6)	(58.7, 92.5)	(99.5, 99.9)	(41.2, 78.5)	(99.5, 99.9)	(45.6, 83.0)	(99.8, 100.0)

## Supplemental Table 3. Assay Characteristics overall and over time; based on the BLCM Posterior Means (95% Creditable Interval),

## Weakly Informative

	Overall				April/May		June/July		August/September	
	Sensitivity	PPV	Specificity	NPV	Sensitivity	Specificity	Sensitivity	Specificity	Sensitivity	Specificity
Spike	92.9%	27.1%	98.1%	99.9%	96.7%	97.9%	92.0%	98.7%	92.2%	97.5%
	(83.9, 99.4)	(24.6, 29.8)	(97.8, 98.4)	(99.9, 100.0)	(87.2, 100.0)	(97.3, 98.4)	(76.9, 99.8)	(98.3, 99.1)	(77.2, 99.8)	(96.9, 98.0)
RBD	90.6%	57.8%	99.5%	99.9%	91.3%	99.4%	93.2%	99.4%	89.7%	99.2%
	(80.5, 98.1)	(51.9, 62.5)	(99.3, 99.6)	(99.9, 100.0)	(77.0, 99.0)	(99.1, 99.7)	(77.9, 99.9)	(99.1, 99.7)	(73.5, 99.5)	(98.8, 99.5)
NP	75.8%	23.0%	98.1%	99.8%	84.8%	99.4%	89.4%	97.2%	70.5%	97.5%
	(64.6, 85.9)	(20.6, 25.7)	(97.8, 98.4)	(99.8, 99.9)	(66.6, 97.6)	(99.1, 99.6)	(74.1, 99.3)	(96.6, 97.8)	(52.1, 88.0)	(96.9, 98.1)
Abbott-	58.7%	67.2%	99.8%	99.7%	79.4%	99.6%	62.9%	99.6%	63.9%	99.8%
NP	(46.2, 71.2)	(64.8, 70.4)	(99.7, 99.9)	(99.6, 99.8)	(59.8, 94.5)	(99.4, 99.8)	(42.4, 82.1)	(99.4, 99.8)	(44.1, 83.9)	(99.6, 99.9)