

Supplemental Table 1

Comparative evaluation results of SARS-CoV-2 antigen RDT passing the sensitivity criteria  
(in alphabetical order of manufacturers)

		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17
Manufacturer		Abbott Rapid Diagnostics Jena GmbH	ACON Biotech (Hangzhou) Co., Ltd	Aesku Diagnostics GmbH	Affimedix	Amazing Biotech (Shanghai) Co., Ltd	Ameda Labordiagnostik GmbH	AmonMed (Xiamen) Biotechnology Co.,Ltd.	Anbio (Xiamen) Biotechnology Co., Ltd	Anhui Deepblue Medical Technology Co., Ltd.	ASAN PHARM.CO.,LTD.	Atlas Link Technology Co.,Ltd.	Avalun	AXIOM Gesellschaft für Diagnostica und Biochemica mbH	Azure Biotech Inc.	Becton Dickinson	Beijing Beier Bioengineering Co., Ltd.	Beijing Hotgen Biotech Co., Ltd.
Test name		Panbio™ Covid-19 Ag Rapid Test Device	Flowflex SARS-CoV-2-Antigenschnelltest (Nasopharynxtpuffer)	Aesku Rapid SARS-CoV-2 Rapid Test	TestNOW® - COVID-19 Antigen	CoroVisio Covid-19 Ag Versieglungsrollchen Teststreifen (Kolloidales Gold)	AMP Rapid Test SARS-CoV-2 Ag	COVID-19 Antigen Rapid Test Kit (Colloidal Gold)	Rapid Covid-19 Antigen Test (Colloidal Gold )	COVID-19 (SARS CoV-2) Antigen Test Kit (Colloidal Gold)	Asan Easy Test COVID-19 Ag	Nova Test SARS-CoV-2 Antigen Rapid Test Kit	Ksmart® SARS-CoV2 Antigen Rapid Test	Axiom Diagnostics COVID-19 Ag Schnelltest	Dia Sure Covid-19 Antigen Rapid Test Device (Nasopharyngeal/Oropharyngeal Swab)	BD Veritor™ System for Rapid Detection of SARS-CoV-2	Covid-19 Antigen Schnelltest	Neuartiges Coronavirus (2019-nCoV)-Antigentest (Kolloidales Gold)
Panel 1 V1		Panel 1 V2		Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V1	Panel 1 V2
Pool Nr.	CT	Pool Nr.	CT															
1	17,55	1	17,31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
6	20,20	2	19,08	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
5	20,28	3	19,62	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
3	20,38	5	20,60	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	20,54	4	20,61	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4	20,98	6	21,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
7	21,71	12	22,12	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1
12	21,82	7	22,15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
8	21,95	8	22,32	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1
9	22,14	16	22,88	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
11	22,34	9	23,13	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1
16	22,55	11	23,13	1	1	0	1	1	1	1	1	1	0	1	1	1	1	1
10	22,88	10	23,21	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1
13	23,32	15	24,38	1	1	0	0	1	1	1	1	1	1	1	0	1	1	1
17	24,00	23	24,45	1	1	0	1	0	1	1	1	1	1	1	0	1	1	1
23	24,04	17	24,81	1	0	1	1	1	1	1	1	1	1	1	0	0	0	1
15	24,14	14	24,97	1	1	1	1	1	1	1	1	1	1	1	0	0	0	1
14	24,28	25	25,07	1	0	0	1	0	1	1	1	1	1	1	0	0	0	1
27	25,14	24	25,20	1	0	0	1	1	1	1	1	1	0	1	0	0	0	1
24	25,24	13	25,29	1	1	1	1	1	1	1	1	1	1	1	1	0	0	1
31	25,27	19	25,45	0	0	1	0	0	1	1	1	1	0	1	0	0	0	1
18	25,30	21	25,95	1	0	0	1	1	0	1	1	1	0	0	0	1	0	1
26	25,47	27	26,12	1	0	0	1	0	1	1	1	1	0	1	0	1	0	1
19	25,50	31	26,24	1	0	0	1	0	1	1	1	1	0	0	0	0	0	0
21	25,54	26	26,32	1	0	0	1	0	1	0	0	1	0	1	0	0	0	0
22	25,87	32	26,64	1	0	0	1	0	1	0	0	0	0	0	0	0	0	0
20	26,27	35	26,66	1	0	0	1	0	1	1	1	1	0	1	0	0	0	1
32	26,44	36	27,05	0	0	0	0	0	1	1	1	1	0	0	1	0	0	0
35	27,04	30	27,24	1	0	0	1	0	1	1	1	0	1	0	0	0	0	1
28	27,14	29	27,34	0	0	0	0	0	1	0	1	1	0	1	0	0	0	0
29	27,15	28	27,41	1	0	0	0	0	1	0	0	0	0	1	0	0	0	0
40	27,65	22	27,42	0	0	1	0	1	1	1	1	1	0	1	0	0	0	1
34	27,89	34	27,82	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0
36	28,13	40	28,19	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0
38	28,14	18	28,33	1	0	1	0	0	1	1	1	1	1	1	0	0	0	1
42	28,43	33	28,92	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0
30	28,86	38	29,36	1	0	0	0	0	1	1	0	1	1	0	0	0	0	0
33	28,96	20	29,46	0	0	0	0	1	1	1	1	1	0	1	0	0	0	1
44	29,24	42	29,48	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0
25	29,70	44	29,51	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0
39	29,76	39	30,12	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0
45	30,10	37	30,13	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0
41	30,13	41	30,14	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
37	30,54	45	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
43	31,05	48	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
46	31,54	46	31,34	0	0	0	0	0	0	1	0	0	0	0	1	0	0	0
48	32,06	43	31,61	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0
47	35,19	47	34,55	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
49	35,22	50	35,83	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
50	36,36	49	36,04	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0

Supplemental Table 1

Comparative evaluation results of SARS-CoV-2 antigen RDT passing the sensitivity criteria  
(in alphabetical order of manufacturers)

		18	19	20	21	22	23	24	25	26	27	28	29	30	31	32	33	34
Manufacturer		Beijing Lepu Medical Technology Co., Ltd	Beijing Tigsun Diagnostics Co.,Ltd.	BIOMERICA Inc.	BIONOTE	BioRepair GmbH	BIOSYNEX SWISS SA	BTNX, Inc. (Biotrend Chemikalien GmbH)	Chil Tibbi Mal. San. Tic. Ltd. Şti	Core Technology Co., Ltd.	DNA Diagnostic A/S.	Edinburgh Genetics Limited	Eurobio Scientific	Fujirebio Inc. (Mast Diagnostica GmbH)	Genrui Biotech Inc.	Gensure Biotech Inc.	Getein Biotech, Inc.	Green Cross Medical Science Corp. (Weko Pharma GmbH)
Test name		SARS-CoV-2 Antigen Rapid Test Kit	Tigsun COVID-19 Saliva Antigen Rapid Test	COVID-19-Antigen-Schnelltest (Nasopharyngeal-Abstrich)	NowCheck® COVID 19 Ag Test	Covid 19 Antigen Schnelltest	BIOSYNEX COVID-19 Ag BSS	Rapid Response COVID-19 Rapid Test Device	COVID-19 Antigen Schnell Test (Nasopharyngeal / Oropharyngeal Tupfer - Kassette)	Canea COVID-19 Antigen Schnelltest	Covid-19 Antigen Detection Kit	Edinburgh Genetics ActivXpress+ COVID-19 Antigen Complete Testing Kit	EBS SARS-CoV-2 Ag Rapid Test	ESPLINE® SARS-CoV-2	Genrui SARS-CoV-2 Antigen Test Kit (Colloidal Gold)	GensureTM COVID 19 Antigen Rapid Test Kit	One Step Test for SARS-CoV-2 Antigen (Colloidal Gold)	Genedia W Covid-19 Ag
Panel 1 V1		Panel 1 V2		Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1
Pool Nr.	CT	Pool Nr.	CT	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1
1	17,55	1	17,31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
6	20,20	2	19,08	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
5	20,28	3	19,62	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
3	20,38	5	20,60	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	20,54	4	20,61	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4	20,98	6	21,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
7	21,71	12	22,12	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
12	21,82	7	22,15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
8	21,95	8	22,32	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
9	22,14	16	22,88	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
11	22,34	9	23,13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
16	22,55	11	23,13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
10	22,88	10	23,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
13	23,32	15	24,38	1	1	1	1	0	1	0	1	1	1	1	1	1	1	1
17	24,00	23	24,45	1	1	1	1	1	1	0	1	1	1	1	1	1	1	1
23	24,04	17	24,81	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0
15	24,14	14	24,97	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0
14	24,28	25	25,07	1	1	1	1	0	1	1	1	1	1	1	1	1	1	0
27	25,14	24	25,20	0	1	1	1	0	1	0	1	1	0	1	1	0	1	0
24	25,24	13	25,29	1	1	1	1	1	1	1	1	1	1	0	1	1	1	0
31	25,27	19	25,45	1	1	1	1	0	1	1	1	1	1	1	1	0	1	0
18	25,30	21	25,95	0	1	0	1	0	1	0	1	0	0	1	1	0	1	1
26	25,47	27	26,12	0	1	0	1	0	1	0	1	1	1	1	1	0	1	1
19	25,50	31	26,24	0	1	0	1	0	0	0	0	0	0	1	0	0	1	0
21	25,54	26	26,32	0	1	1	1	0	1	0	0	0	1	0	0	0	1	0
22	25,87	32	26,64	0	1	0	1	0	1	0	0	0	1	0	1	0	1	0
20	26,27	35	26,66	0	1	0	1	1	1	0	0	0	0	0	1	0	1	0
32	26,44	36	27,05	0	1	0	0	1	1	0	0	0	0	0	0	0	1	0
35	27,04	30	27,24	0	1	0	1	0	1	0	0	0	0	1	0	0	1	0
28	27,14	29	27,34	1	1	0	0	0	1	0	0	0	0	0	1	0	1	0
29	27,15	28	27,41	0	1	0	1	1	0	0	0	0	0	0	0	0	1	0
40	27,65	22	27,42	1	1	0	0	0	1	1	1	1	0	0	1	0	1	0
34	27,89	34	27,82	0	1	0	1	0	0	0	0	0	0	0	0	0	1	0
36	28,13	40	28,19	0	0	0	1	0	0	0	0	0	0	0	N/A	0	0	0
38	28,14	18	28,33	1	1	1	1	0	1	1	1	1	1	1	1	1	1	0
42	28,43	33	28,92	0	1	0	0	0	0	0	0	0	0	0	N/A	0	0	0
30	28,86	38	29,36	0	1	0	1	1	0	0	0	0	0	0	0	0	1	0
33	28,96	20	29,46	0	1	1	0	0	1	0	1	1	1	0	1	0	1	0
44	29,24	42	29,48	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
25	29,70	44	29,51	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0
39	29,76	39	30,12	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0
45	30,10	37	30,13	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0
41	30,13	41	30,14	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0
37	30,54	45	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
43	31,05	48	31,19	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0
46	31,54	46	31,34	0	1	0	0	0	1	0	1	0	0	0	0	0	0	0
48	32,06	43	31,61	invalid	0	0	0	0	0	0	0	0	0	0	0	0	0	0
47	35,19	47	34,55	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
49	35,22	50	35,83	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
50	36,36	49	36,04	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0

Supplemental Table 1

Comparative evaluation results of SARS-CoV-2 antigen RDT passing the sensitivity criteria  
(in alphabetical order of manufacturers)

		35	36	37	38	39	40	41	42	43	44	45	46	47	48	49	50	51	
Manufacturer		Guangdong Hecin Scientific, Inc.	Guangdong Wesail Biotech Co., Ltd.	Guangzhou Wondfo Biotech Co. Ltd	Hangzhou Clongene Biotech Co., Ltd.	Hangzhou Immuno Biotech Co., Ltd.	Hangzhou Laihe Biotech Co., Ltd. (Lissner Q) GmbH	Hangzhou Lysun Biotechnology Co., Ltd.	Hangzhou Testsea Biotechnology Co., Ltd	Humasis Co., Ltd.	IVC Pragen Healthcare	Jiangsu Diagnostics Biotechnology Co., Ltd	Jiangsu Medomics Medical Technology Co., Ltd (CIV care impuls Vertrieb)	Joinstar Biomedical Technology Co., Ltd (CIV care impuls Vertrieb)	Labnovation Technologies, Inc.	Lumigenex (Suzhou) Co., Ltd.	LumiQuick Diagnostics Inc.	LumiraDX	
Test name		2019-nCoV Antigen Test Kit (colloidal gold method)	COVID-19 Ag Test Kit	Wondfo SARS-CoV-2 Antigen Test (Lateral Flow Method)	Clungene COVID-19 Antigen Rapid Test	IMMUNOBIO SARS-CoV-2 Antigen-Schnelltest (COVID-19 Ag)	Lyher Novel Coronavirus (COVID-19) Antigen Test Kit (Colloidal Gold)	Lysun COVID-19 Antigen Rapid Test Device (Colloidal Gold)	Testsealabs® Rapid Test Kit COVID-19 Antigen Test Cassette	COVID-19 Ag Test	GenBody COVID-19 Ag	COVID-19 Antigen Rapid Test Cassette (Colloidal Gold)	SARS-CoV-2-Antigen-Testkit (LFIA)	COVID-19 Antigen Schnelltest (Colloidal Gold)	Labnovation SARS-CoV-2 Antigen Rapid Test Kit (Immunochromatography)	PocRoc SARS-CoV-2, Antigen Schnelltest Set (Kolloidales Gold)	QuickProfile Covid-19 Antigen Test Card	LumiraDx SARS-CoV-2 Ag Test	
Panel 1 V1		Panel 1 V2		Panel 1 V2		Panel 1 V1		Panel 1 V2		Panel 1 V2		Panel 1 V2		Panel 1 V1		Panel 1 V2		Panel 1 V1	
Pool Nr.	CT	Pool Nr.	CT	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1
1	17,55	1	17,31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
6	20,20	2	19,08	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
5	20,28	3	19,62	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
3	20,38	5	20,60	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	20,54	4	20,61	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4	20,98	6	21,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
7	21,71	12	22,12	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
12	21,82	7	22,15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
8	21,95	8	22,32	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
9	22,14	16	22,88	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
11	22,34	9	23,13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
16	22,55	11	23,13	0	1	1	1	1	1	0	1	1	1	1	1	1	1	1	1
10	22,88	10	23,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
13	23,32	15	24,38	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
17	24,00	23	24,45	0	1	0	1	0	1	0	1	0	1	0	1	0	1	1	1
23	24,04	17	24,81	1	1	1	0	1	0	1	0	1	1	1	1	1	1	1	1
15	24,14	14	24,97	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1
14	24,28	25	25,07	0	1	0	1	0	1	1	1	1	1	0	1	0	1	1	1
27	25,14	24	25,20	0	1	0	1	0	0	1	1	0	1	0	1	0	1	1	1
24	25,24	13	25,29	1	1	0	0	1	0	1	0	1	1	1	1	1	1	1	0
31	25,27	19	25,45	0	0	0	0	1	0	1	0	0	1	1	1	1	1	1	0
18	25,30	21	25,95	0	1	0	1	0	1	0	1	1	0	1	0	1	0	1	1
26	25,47	27	26,12	0	1	0	1	0	1	1	1	1	0	1	0	1	1	1	1
19	25,50	31	26,24	0	1	0	1	0	1	0	0	1	1	0	1	0	1	1	1
21	25,54	26	26,32	0	1	0	0	0	1	0	0	1	0	1	0	1	1	1	0
22	25,87	32	26,64	0	1	0	1	0	1	0	0	1	0	1	0	1	1	1	1
20	26,27	35	26,66	0	1	0	1	0	1	1	0	1	0	1	0	1	1	1	1
32	26,44	36	27,05	0	0	0	0	0	1	0	0	1	0	1	0	1	1	1	1
35	27,04	30	27,24	0	1	0	0	0	1	1	0	1	1	0	1	0	1	1	1
28	27,14	29	27,34	0	0	0	0	0	1	1	0	0	1	0	0	0	1	1	0
29	27,15	28	27,41	0	1	0	1	0	0	1	0	0	1	0	0	0	1	1	1
40	27,65	22	27,42	0	0	0	0	0	1	1	0	0	1	0	1	1	1	1	0
34	27,89	34	27,82	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0
36	28,13	40	28,19	0	1	0	0	0	0	0	0	0	0	1	N/A	0	0	1	1
38	28,14	18	28,33	1	0	0	0	1	1	1	0	1	1	0	1	1	1	1	0
42	28,43	33	28,92	0	0	0	0	0	0	0	0	0	0	0	N/A	0	0	0	0
30	28,86	38	29,36	0	1	0	0	0	1	0	0	1	0	1	0	1	1	1	1
33	28,96	20	29,46	0	0	0	0	0	1	1	0	0	1	1	0	0	1	1	0
44	29,24	42	29,48	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0
25	29,70	44	29,51	0	0	0	1	0	0	0	0	0	0	1	0	0	1	1	1
39	29,76	39	30,12	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0
45	30,10	37	30,13	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
41	30,13	41	30,14	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
37	30,54	45	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
43	31,05	48	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
46	31,54	46	31,34	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
48	32,06	43	31,61	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0
47	35,19	47	34,55	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0
49	35,22	50	35,83	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
50	36,36	49	36,04	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0

Supplemental Table 1

Comparative evaluation results of SARS-CoV-2 antigen RDT passing the sensitivity criteria (in alphabetical order of manufacturers)

		52	53	54	55	56	57	58	59	60	61	62	63	64	65	66	67	68
Manufacturer		MEDsan GmbH	Merlin Biomedical (Xiamen) Co., Ltd.	Mölab GmbH	MP Biomedicals Germany GmbH	nal von minden GmbH	NanoEntek Inc.	Nanjing Norman Biological Technology Co., Ltd	Nantong Diagnos Biotechnology Co., Ltd.	New Gene (Hangzhou) Bioengineering Co., Ltd.	Novatech Tibbi Cihaz Ürünleri San. Ve Tic. A.S.	Oncosem Onkolojik Sistemler San. Ve Tic. A.S.	PCL, Inc.	PerGrande BioTech Development Co., Ltd.	Precision Biosensor Inc. (Axon Lab AG)	ProGnosis Biotech	Quidel Corporation	Qingdao Hightop Biotech Co., Ltd.
Test name		MEDsan® SARS-CoV-2 Antigen Rapid Test	SARS-CoV-2 Antigen Rapid Test Cassette	mö-screen Corona Antigen Test	Rapid SARS-CoV-2 Antigen Test Card	NADAL® COVID-19 Ag Schnelltest	Frend™ COVID-19 Ag	Novel Coronavirus (2019-nCoV) Antigen Testing Kit (Colloidal Gold)	COVID-19 Antigen Saliva Test Kit (Colloidal Gold)	Covid-19-Antigen-Testkit	novatech®-Ag SARS-CoV-2 Covid-19 Antigen Rapid Test	CAT Antigen Covid Rapid Test	PCL COVID19 Ag Gold Saliva	SARS-CoV-2 Antigen Detection Kit (Colloidal Gold Immuno-chromatographic Assay)	Exdia COVID-19-Ag Test	Rapid Test Ag 2019-nCoV	Sofia SARS Antigen FIA	Hightop SARS-CoV-2 (Covid-19) Antigen Rapid Test
Panel 1 V1		Panel 1 V2		Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2
Pool Nr.	CT	Pool Nr.	CT															
1	17,55	1	17,31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
6	20,20	2	19,08	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
5	20,28	3	19,62	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
3	20,38	5	20,60	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	20,54	4	20,61	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4	20,98	6	21,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
7	21,71	12	22,12	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
12	21,82	7	22,15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
8	21,95	8	22,32	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
9	22,14	16	22,88	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1
11	22,34	9	23,13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
16	22,55	11	23,13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
10	22,88	10	23,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
13	23,32	15	24,38	1	1	0	0	1	1	1	1	0	1	1	1	1	1	1
17	24,00	23	24,45	1	1	1	1	0	1	1	0	1	1	1	1	1	1	1
23	24,04	17	24,81	1	1	1	1	0	1	1	1	1	1	1	1	1	0	1
15	24,14	14	24,97	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1
14	24,28	25	25,07	1	1	1	1	0	1	1	1	1	1	0	1	1	1	1
27	25,14	24	25,20	1	1	1	1	0	1	1	0	1	1	0	1	1	0	1
24	25,24	13	25,29	0	1	1	1	1	1	1	1	1	1	0	1	1	0	1
31	25,27	19	25,45	0	1	0	1	0	1	1	0	1	1	0	1	1	0	1
18	25,30	21	25,95	1	1	1	1	0	1	1	0	1	1	0	1	1	1	0
26	25,47	27	26,12	1	1	1	1	1	1	1	0	1	1	0	1	0	1	1
19	25,50	31	26,24	1	1	1	0	0	0	1	0	0	0	0	1	0	0	0
21	25,54	26	26,32	1	1	1	1	0	0	1	0	0	1	0	1	0	0	0
22	25,87	32	26,64	1	1	1	0	0	0	1	0	0	0	0	1	1	0	0
20	26,27	35	26,66	1	1	1	0	0	0	1	1	0	0	0	1	0	0	1
32	26,44	36	27,05	1	1	0	1	0	0	0	1	0	0	1	0	1	0	0
35	27,04	30	27,24	1	1	1	0	0	0	0	1	1	0	1	0	1	0	0
28	27,14	29	27,34	0	1	0	0	0	0	0	1	0	0	1	0	0	0	1
29	27,15	28	27,41	0	1	1	0	0	0	0	1	0	0	1	1	1	0	0
40	27,65	22	27,42	0	1	0	1	0	0	1	1	1	1	0	0	1	0	1
34	27,89	34	27,82	0	1	0	0	0	0	0	1	0	0	0	0	0	0	0
36	28,13	40	28,19	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0
38	28,14	18	28,33	1	1	0	1	0	1	1	1	1	1	1	1	1	0	1
42	28,43	33	28,92	0	0	0	0	0	0	0	0	0	0	1	0	1	0	0
30	28,86	38	29,36	0	1	0	0	0	0	1	1	0	0	0	1	0	0	0
33	28,96	20	29,46	0	1	0	0	0	0	1	1	0	0	0	0	1	0	0
44	29,24	42	29,48	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0
25	29,70	44	29,51	0	0	1	0	0	0	0	0	0	0	0	1	0	0	0
39	29,76	39	30,12	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
45	30,10	37	30,13	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0
41	30,13	41	30,14	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0
37	30,54	45	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
43	31,05	48	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
46	31,54	46	31,34	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
48	32,06	43	31,61	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0
47	35,19	47	34,55	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
49	35,22	50	35,83	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
50	36,36	49	36,04	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0

Supplemental Table 1

Comparative evaluation results of SARS-CoV-2 antigen RDT passing the sensitivity criteria  
(in alphabetical order of manufacturers)

		69	70	71	72	73	74	75	76	77	78	79	80	81	82	83	84	85	
Manufacturer		R-Biopharm AG	Safecare Biotech Hangzhou Co., Ltd.	Salofa OY	ScheBo Biotech AG	SD BIOSENSOR (Roche Diagnostics GmbH)	SD BIOSENSOR	SD BIOSENSOR	SGA Mühendislik DAN. EG. İcve DIS.Ltd.STI	Shenzhen Lvshiyuan Biotechnology Co., Ltd.	Shenzhen Microprofit Biotech Co., Ltd.	Shenzhen Watmind Medical Co.,Ltd.	Shenzhen Watmind Medical Co.,Ltd.	Shenzhen Zhenrui Biotech co.Ltd.	Siemens Healthineers	Sugentech, Inc.	Toda Pharma	Triplex International Biosciences (China) Co., Ltd.	
Test name		RIDA® QUICK SARS-CoV-2 Antigen	Safecare COVID-19 Ag Rapid Test Kit (Swab)	salocor SARS-CoV-2 Antigen Rapid Test Cassette (Nasopharyngeal swab)	ScheBo SARS-CoV-2 Quick Antigen	SARS-CoV-2 Rapid Antigen Test	STANDARD™ F COVID-19 Ag FIA	STANDARD™ Q COVID-19 Ag Test	V-Chek SARS-CoV-2 Rapid Ag Test Kit (Colloidal Gold)	Green Spring SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold)	fluorecare COVID-19 SARS-CoV-2 Spike Protein Test Kit (Colloidal Gold Chromatographic Immunoassay)	SARS-CoV-2 Ag Diagnostic Test Kit (Colloidal Gold)	SARS-CoV-2 Ag Diagnostic Test Kit (Immunofluorescence)	Zhenrui COVID-19 (SARS-COV-2) Antigen Test Kits	CLINITEST® Rapid COVID-19 Antigen Test	SGTI-flex COVID-19 Ag	Toda Coronadiag Ag	SARS-CoV-2 Antigen Rapid Test Kit	
Panel 1 V1		Panel 1 V2		Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2
Pool Nr.	CT	Pool Nr.	CT	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V1	Panel 1 V2
1	17,55	1	17,31	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
6	20,20	2	19,08	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
5	20,28	3	19,62	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
3	20,38	5	20,60	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
2	20,54	4	20,61	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
4	20,98	6	21,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
7	21,71	12	22,12	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
12	21,82	7	22,15	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
8	21,95	8	22,32	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
9	22,14	16	22,88	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
11	22,34	9	23,13	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
16	22,55	11	23,13	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1
10	22,88	10	23,21	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
13	23,32	15	24,38	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1
17	24,00	23	24,45	1	1	0	1	1	1	0	1	1	1	1	1	1	1	1	1
23	24,04	17	24,81	1	1	1	1	0	1	1	1	1	1	1	0	1	1	1	1
15	24,14	14	24,97	1	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1
14	24,28	25	25,07	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1
27	25,14	24	25,20	0	1	0	1	1	1	0	1	1	1	1	1	1	1	1	1
24	25,24	13	25,29	1	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1
31	25,27	19	25,45	0	1	0	1	0	0	1	1	1	1	1	0	1	1	1	1
18	25,30	21	25,95	1	1	0	1	1	1	N/A	1	1	1	1	0	1	0	1	1
26	25,47	27	26,12	1	1	0	1	1	1	0	1	1	1	1	0	1	1	1	1
19	25,50	31	26,24	0	0	0	1	1	1	0	1	0	1	0	0	1	1	1	1
21	25,54	26	26,32	1	1	0	1	0	0	1	0	1	1	0	1	1	1	1	1
22	25,87	32	26,64	0	0	0	1	1	1	0	1	0	1	0	0	1	0	1	1
20	26,27	35	26,66	0	1	0	1	0	0	1	0	1	1	0	1	1	1	1	1
32	26,44	36	27,05	0	1	0	1	0	0	1	1	1	1	0	1	1	1	1	1
35	27,04	30	27,24	0	1	0	1	1	1	0	1	0	1	0	1	1	1	1	1
28	27,14	29	27,34	0	1	0	1	0	0	1	0	1	0	0	1	1	1	1	1
29	27,15	28	27,41	0	0	0	1	0	0	1	0	1	0	0	1	1	1	1	1
40	27,65	22	27,42	0	1	0	1	0	0	1	1	1	1	0	0	1	1	1	1
34	27,89	34	27,82	0	0	0	1	0	0	1	1	1	1	0	1	1	1	1	1
36	28,13	40	28,19	0	0	N/A	1	1	1	N/A	1	0	1	0	0	1	0	1	0
38	28,14	18	28,33	0	1	1	1	0	1	0	1	1	1	1	1	1	1	1	1
42	28,43	33	28,92	0	0	0	0	0	0	N/A	1	0	1	0	0	1	0	1	0
30	28,86	38	29,36	0	1	0	1	0	0	0	1	0	1	1	0	1	1	1	1
33	28,96	20	29,46	0	0	0	1	0	0	0	1	1	1	0	1	1	1	1	1
44	29,24	42	29,48	0	0	0	1	0	0	0	1	0	1	0	0	0	1	1	1
25	29,70	44	29,51	0	0	0	0	1	0	0	0	0	0	0	1	0	0	0	0
39	29,76	39	30,12	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0
45	30,10	37	30,13	0	0	0	0	0	0	0	1	1	1	0	0	0	0	1	0
41	30,13	41	30,14	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	1
37	30,54	45	31,19	0	0	0	1	0	0	N/A	0	0	0	0	0	0	0	1	0
43	31,05	48	31,19	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
46	31,54	46	31,34	0	0	0	0	0	0	0	1	0	0	0	0	0	0	1	0
48	32,06	43	31,61	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1
47	35,19	47	34,55	0	0	0	0	0	0	0	1	0	0	0	0	0	0	1	0
49	35,22	50	35,83	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0
50	36,36	49	36,04	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0

Supplemental Table 1

Comparative evaluation results of SARS-CoV-2 antigen RDT passing the sensitivity criteria  
(in alphabetical order of manufacturers)

		86	87	88	89	90	91	92	93	94	95	96
Manufacturer		ulti med Products (Deutschland) GmbH	Vitrosens Biyoteknoloji Ltd. Sti	Wantai (Beijing) Wantai Biological Pharmacy Enterprise Co., Ltd.)	Wuhan EasyDiagnosis Biomedicine Co., Ltd	Wuhan Life Origin Biotech Joint Stock Co., Ltd.	Wuhan UNScience Biotechnology Co., Ltd.	Xiamen Boson Biotech Co., Ltd.	Xiamen WIZ Biotech Co., Ltd.	Zet Medikal Tekstil Dis Ticaret Ltd. STI.	Zhejiang Anji Saianfu Biotech Co.,Ltd.	Zhejiang Orient Gene Biotech Co.,Ltd
Test name		COVID-19 Antigen Speicheltest (Immunochromatographie)	RapidFor SARS-CoV-2 Rapid Antigen Test Colloidal Gold	SARS-CoV-2 Ag Rapid Test (FIA)	COVID-19 (SARS-CoV-2) Antigen Test Kit	SARS-CoV-2 Antigen Assay Kit (Immunochromatography)	SARS-CoV-2 Antigen Rapid Test Kit	SARS-CoV-2 Antigen Schnelltest	Wizbiotech SARS-CoV-2 Antigen Rapid Test	softec SARS COV-2 (Covid-19) Antigen Test Kit	reOpenTest COVID-19 Antigen Rapid Test (Colloidal Gold)	Coronavirus Ag Rapid Test Cassette (Swab)
Panel 1 V1		Panel 1 V2		Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1
Pool Nr.	CT	Pool Nr.	CT	Panel 1 V2	Panel 1 V2	Panel 1 V1	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V2	Panel 1 V1
1	17,55	1	17,31	1	1	1	1	1	1	1	1	1
6	20,20	2	19,08	1	1	1	1	1	1	1	1	1
5	20,28	3	19,62	1	1	1	1	1	1	1	1	1
3	20,38	5	20,60	1	1	1	1	1	1	1	1	1
2	20,54	4	20,61	1	1	1	1	1	1	0	1	1
4	20,98	6	21,21	1	1	1	1	1	1	1	1	1
7	21,71	12	22,12	1	1	1	1	1	1	1	1	1
12	21,82	7	22,15	1	1	1	1	1	1	1	1	1
8	21,95	8	22,32	1	1	1	1	1	1	1	1	1
9	22,14	16	22,88	1	1	1	1	1	1	1	1	1
11	22,34	9	23,13	1	1	1	1	1	1	1	1	1
16	22,55	11	23,13	1	1	1	1	1	1	1	1	1
10	22,88	10	23,21	1	1	1	1	1	1	0	1	1
13	23,32	15	24,38	1	1	1	0	1	0	1	1	1
17	24,00	23	24,45	1	1	1	0	1	0	N/A	0	1
23	24,04	17	24,81	1	1	1	1	1	1	1	1	1
15	24,14	14	24,97	1	1	1	1	1	1	1	1	1
14	24,28	25	25,07	1	1	1	1	1	0	1	1	1
27	25,14	24	25,20	1	0	1	1	0	1	0	1	1
24	25,24	13	25,29	1	1	1	1	1	1	1	1	1
31	25,27	19	25,45	1	1	1	1	1	1	0	1	1
18	25,30	21	25,95	1	0	1	1	0	1	N/A	0	1
26	25,47	27	26,12	1	1	1	1	0	1	0	1	1
19	25,50	31	26,24	1	0	1	1	0	0	invalid	0	1
21	25,54	26	26,32	1	0	1	1	0	1	0	1	1
22	25,87	32	26,64	1	0	1	1	0	0	0	0	1
20	26,27	35	26,66	1	0	1	1	0	0	0	1	1
32	26,44	36	27,05	1	0	1	1	0	1	0	0	1
35	27,04	30	27,24	1	0	1	1	0	0	0	0	1
28	27,14	29	27,34	1	0	1	1	0	0	0	0	1
29	27,15	28	27,41	1	0	1	0	0	0	0	0	1
40	27,65	22	27,42	1	1	0	1	1	1	1	1	0
34	27,89	34	27,82	1	0	1	0	0	0	0	0	1
36	28,13	40	28,19	1	0	1	0	0	0	N/A	0	1
38	28,14	18	28,33	1	1	1	1	1	1	N/A	1	1
42	28,43	33	28,92	1	0	0	0	0	0	N/A	0	1
30	28,86	38	29,36	1	0	1	1	0	0	0	0	1
33	28,96	20	29,46	1	1	0	1	1	0	0	0	1
44	29,24	42	29,48	1	0	0	0	0	0	0	0	0
25	29,70	44	29,51	0	0	1	0	0	0	0	0	1
39	29,76	39	30,12	0	0	0	0	0	0	invalid	0	0
45	30,10	37	30,13	0	0	0	0	0	0	0	0	0
41	30,13	41	30,14	1	0	0	0	0	0	0	0	0
37	30,54	45	31,19	0	0	0	0	0	0	N/A	0	0
43	31,05	48	31,19	0	0	0	0	0	0	invalid	0	0
46	31,54	46	31,34	1	0	0	0	0	0	invalid	0	0
48	32,06	43	31,61	0	0	0	0	0	0	0	0	0
47	35,19	47	34,55	0	0	0	0	0	0	0	0	0
49	35,22	50	35,83	0	0	0	0	0	0	1	0	0
50	36,36	49	36,04	0	0	0	0	0	0	invalid	0	0