TITLE

Harmony COVID-19: a ready-to-use kit, low-cost detector, and smartphone app for point-of-care SARS-CoV-2 RNA detection

ONE SENTENCE SUMMARY Harmony COVID-19: point-of-care SARS-CoV-2 RNA detection

AUTHORS

Nuttada Panpradist^{1,2}, Robert G. Atkinson¹, Michael Roller¹, Enos Kline¹, Qin Wang¹, Ian T. Hull¹, Jack H. Kotnik^{1,3}, Amy K. Oreskovic¹, Crissa Bennett¹, Daniel Leon¹, Victoria Lyon³, Peter D. Han^{4,5}, Lea M. Starita^{4,5}, Matthew J. Thompson³, Barry R. Lutz^{1*}

AFFILIATIONS

1 Department of Bioengineering,
2 Global Health for Women, Adolescents, and Children, School of Public Health,
3 Department of Family Medicine,
4 Department of Genome Sciences, University of Washington, Seattle, WA, USA
5 Brotman Baty Institute for Precision Medicine, Seattle, WA, USA

*** CORRESPONDING AUTHOR**

B. R. L. (blutz@uw.edu)

3720 15th Ave NE, Campus Box 355061, Seattle, Washington, United States 98195

ABSTRACT (149 words)

RNA amplification tests allow sensitive detection of SARS-CoV-2 infection, but their complexity and cost are prohibitive for expanding COVID-19 testing. We developed “Harmony COVID-19”, a point-of-care test using inexpensive consumables, ready-to-use reagents, and a simple device that processes up to 4 samples simultaneously. Our lyophilized reverse-transcription, loop-mediated isothermal amplification (RT-LAMP) can detect as little as 15 SARS-CoV-2 RNA copies per reaction, and it can report as early as 17 min for samples with high viral load (2×10^5 RNA copies per reaction). Analysis of RNA extracted from clinical nasal specimens ($n = 101$) showed 95% concordance with RT-PCR, including 100% specificity in specimens positive for other viruses and bacteria. Analysis of contrived samples in the nasal matrix showed detection of 92% or 100% in samples with ≥ 20 or ≥ 100 particles per reaction, respectively. Usability testing showed 95% accuracy by healthcare workers operating the test for the first time.

WORD COUNT (9,878 words)

INTRODUCTION

In 2019, an outbreak in China of SARS-CoV-2, the causative pathogen of COVID-19, rapidly became a global pandemic(1), and after a year it has infected 100 million people and killed 2 million people(2). Multiple measures have been used to contain the spread of COVID-19. Governments imposed universal "stay-home" orders to minimize transmission(3) which in turn has harmed mental health, social life, and the economy(4). It has been proposed(5) and demonstrated(6) that widespread COVID-19 testing and contact tracing could be a solution for re-openings, allowing subsets of a community to resume work and begin to restore the economy. Moreover, testing is essential for re-opening of international borders. Several countries have enforced a "fit-to-fly" policy; international travelers must test negative for COVID-19 within 72h before boarding international flights(7). Fast and sensitive point-of-care (**POC**) testing could facilitate the safe return to functioning domestic and international economies. POC testing could allow testing to expand to geographic regions that have limited access to centralized laboratory facilities, enable essential businesses to regularly test employees, and allow easy and efficient community testing by public health officials and clinics. Importantly, POC testing is critical in settings that require rapid turnaround time such as emergency urgent care facilities or points of introduction such as airports.

Here, we report the development of Harmony COVID-19 — a complete moderate-throughput sample-to-result system for sensitive POC detection of SARS-CoV-2 RNA. Harmony simplifies testing with ready-to-use assay reagents, an easy-to-use dedicated smartphone interface, and an inexpensive isothermal heater/detector device to enable POC testing. The assay includes three redundant SARS-CoV-2 targets to avoid false negatives due to viral mutation and an internal amplification control (**IAC**) in each test to avoid false negatives in case of assay failure. The heater/detector detects two-color real-time fluorescence for the SARS-CoV-2 redundant targets and IAC, and results are automatically reported on the smartphone. We evaluated the system using two panels of specimens - a panel of extracted nasal specimens from individuals with respiratory symptoms, and a blinded panel of contrived specimens from the XPRIZE Rapid Covid Testing competition – and we conducted usability testing of Harmony by healthcare workers (**HCWs**). These evaluations have demonstrated that Harmony is a complete system for sample-to-result testing that is high performance and sufficiently simple for CLIA-waived settings.

RESULTS

Workflow and operation of COVID-19 Harmony

The Harmony test kit and detector were engineered for low cost and simplicity of use to enable POC testing (**Fig.1a**). The test involves taking a nasal swab, eluting the swab in a lysis buffer, and transferring the buffer to a reaction tube containing ready-to-use

reagents. The reagent tube is inserted into the custom-designed heater/reader operated by a cell phone that provides instructions and displays the result.

The Harmony assay uses a unique variation of reverse transcription loop-mediated isothermal amplification (**RT-LAMP**) that allows fluorescence detection of multiple targets in the same reaction. The assay amplifies three genomic regions of the SARS-CoV-2 nucleocapsid phosphoprotein reported by a green fluorescence signal (FAM) and an engineered internal amplification control (**IAC**) reported by a red fluorescence signal (TEX 615, alternative for Texas Red®). For samples with no target (**Fig. 1B**, left; NTC), detection of the IAC confirms that the reaction was functional; absence of target and IAC indicates a failed reaction. For samples with SARS-CoV-2 RNA (**Fig. 1B**, middle; 200 copies/rxn), the green signal indicates a positive test, regardless of whether IAC is detected. Harmony using dry reagents detects down to 200 copies of SARS-CoV-2 RNA on a swab (**Fig. 1B**, right; ~20 copies/rxn, corresponding to 0.5 copies/ μ L). Real-time detection reports a positive sample as soon as the target signal appears, allowing detection in <30 min for samples ≥ 2000 copies per swab (**Fig. 1C**), with earlier results for higher viral load. The three SARS-CoV-2 targets are reported into the green fluorescence channel to provide a triply-redundant assay to make it robust against viral mutations (**Fig. 1D**). A positive test result is reported when any of the three SARS-CoV-2 targets are present, and a negative test result is reported if SARS-CoV-2 RNA is undetected, and the IAC is detected. If neither target nor IAC is detected, it indicates a

test failure, and the result is reported as indeterminate (**IND**). **table S1** lists the sequences of primers/probes used in this RT-LAMP assay.

Development and analytical sensitivity of ready-to-use reagents.

We have developed a lyophilized RT-LAMP mixture that contains primers, sequence-specific fluorescence probes, IAC DNA template, polymerases, supporting proteins, and co-factors (**Fig. 2A**) and is activated by adding the nasal swab eluate containing additional buffer, salts, and detergent. We initially screened excipient formulations that can preserve enzymatic activity and do not significantly interfere with the primer or probe behaviors. Mannitol did not drastically impact the core assay compared to trehalose at the same concentration (**fig. S1**) and did not require increasing the reaction temperature to maintain the assay speed. Using a commercial real-time thermal cycler, lyophilized RT-LAMP detected down to 15 RNA copies/rxn (0.38 copies/ μ L), and IAC was detected in all negative samples (**Fig. 2B**). RT-LAMP amplification of 2×10^5 RNA copies/rxn using the dry reagents generated detectable signal in 17 min (**Fig. 2C**). We subsequently tested our in-house lyophilized reagents on a larger number of technical replicates of SARS-CoV-2 RNA at 20 copies/rxn using multiple batches, and 90% (36/40, 95% confidence interval (**CI**): 76-97%) were amplified (**Fig. 2D**). Medians of time to detection for FAM and TEX 615 signals were 28.6 min (range: 28.0-29.4) and 31.7 min (range: 31.1-32.7), respectively. In the presence of simulated nasal matrix (salt, mucin, human DNA), we observed a 10-min delay in amplification of both SARS-CoV-2 RNA and IAC DNA (**Fig. 2E**), but all targets were detected correctly. Greater

interference or other harm to the reaction will further delay IAC amplification; thus, the IAC serves as an indicator of the reaction's function and can prevent reporting false negatives due to interference or damaged reagents. By setting a 1-hour reaction time, the assay can suffer some delay while still correctly identifying positive samples. Specificity testing against closely-related pathogens MERS and SARS-CoV-1 showed no cross-reactivity (**fig. S2**).

Hardware: housing and real-time fluorescence reader device

The Harmony system includes a heater/detector device controlled by a dedicated cell phone and plastic housing with a sample setup station (**Fig. 3A**). The heater/detector device (**Fig. 3B**) includes an aluminum heater block with four wells and a spring-loaded heated lid to prevent condensation. A proportional integral derivative (**PID**) feedback loop running on a microcontroller allows recovery of the temperature setpoint within 3-4 min after disturbance from opening the lid (**Fig. 3C**). Two LEDs (**Fig. 3D**) provide the excitation light by shining down through the top of each sample tube. Fluorescence emission passes through holes drilled on the side of the heater block. Each sample well has an emission filter/photodiode set on each side to detect two emission wavelengths (**Fig. 3E**). A blue LED (dominant wavelength 470nm) is used to excite FAM, and the emission photodiode sits behind a dielectric 550nm longpass filter. A yellow LED (dominant wavelength 587nm) is used to excite TEX 615, and a second emission photodiode sits behind a 630nm longpass dielectric filter.

Thermostable Inorganic Pyrophosphatase enables real-time LAMP detection in the Harmony COVID-19 device

RT-LAMP generates inorganic pyrophosphates that precipitate from the solution. The cloudiness created by precipitates has been used as a visual readout signal for LAMP(8), but here the cloudiness scattered light and caused non-specific signals in NTC samples (**Fig. 3F**, top) when tested on device. Adding inorganic pyrophosphatase in the RT-LAMP reaction eliminated built-up pyrophosphate and prevented non-specific signal in negative reactions (**Fig. 3F**, bottom). Interestingly, this effect was not observed when the reaction was tested in the commercial thermal cycler (BioRad CFX96).

Multi-purpose Android mobile app: instruction, temperature control, and real-time analysis of results

The phone software provides interactive guidance for setting up the test. It prompts the user to input a sample ID by scanning a barcode, capturing a photo, or typing a sample ID (**Fig. 4, A-D**). The software also carries out the closed-loop temperature control for the heater block and lid heater, and it displays the status of the heater block to prompt the user to insert the reaction tubes only after the heater reaches the reaction temperature (**Fig. 4E**). During the test run, the screen displays the test status (“analysis in progress”) and elapsed time, or it reports any detected errors (“failed run”) (**Fig. 4F**). Importantly, the software performs real-time data analysis that detects fluorescence intensity rise above a dynamically and automatically computed background level. Signals from SARS-CoV-2 can appear early in the test, especially for high viral loads,

and the real-time analysis allows reporting of positive tests immediately after they are detected (**Fig. 4G**). Negative results and IND results are reported at the end of the test run time (**Fig. 4H**). Software analysis is described in the Methods section.

Evaluation using extracted RNA from clinical specimens

Clinical specimens collected from individuals positive for SARS-CoV-2 ($n = 33$) or for other pathogenic respiratory infections ($n = 68$) were re-tested by RT-qPCR(9) to measure the SARS-CoV-2 viral load and provide the reference test result. Using RT-qPCR, 68 specimens were confirmed to be SARS-CoV-2 negative (**Fig. 5A**, left), 30 specimens were confirmed to be SARS-CoV-2 positive (**Fig. 5A**, right), and three previously-positive specimens were only detected by one RT-qPCR assay (either N1 or N2) and classified as inconclusive (**INC**, **Fig. 5A**, middle). Specimens were tested in duplicate on Harmony using the ready-to-use reagents (**Fig. 5A**), arranged from low to high virus concentration as determined by RT-qPCR. Out of 30 RT-qPCR-positive specimens, Harmony detected 25 specimens in both technical replicates and 4 specimens in one of the two replicates. In the three RT-qPCR INC results, one sample was detected by Harmony in one of the two replicates. Excluding the RT-qPCR INC specimens and applying a stringent requirement for both Harmony replicates to be detected to report a positive test, Harmony achieved 95% accuracy (93/98, 95%CI: 88-98%), 83% sensitivity (25/30, 95%CI: 65-94%), and 100% specificity (68/68, 95%CI: 94-100%). If replicates are treated independently, Harmony detected 54/60 positive specimens (90% sensitivity, 95%CI: 79-96%). For specimens with ≥ 10 copies/rxn,

Harmony detected 24/25 specimens (96% sensitivity, 95%CI: 80-100%), and for specimens with ≥ 40 copies/rxn, Harmony detected 18/18 specimens (100% sensitivity, 95%CI: 81-100%).

The clinical panel (**Fig. 5A**) included 68 specimens that were SARS-CoV-2 negative but contained other respiratory pathogens including influenza, rhinovirus, and seasonal coronavirus (**Fig. 5B**). Harmony correctly identified 68/68 specimens (**Fig. 5A**, left) as negative (100%, 95%CI: 94-100% specificity). **Fig. 5C** summarizes the performance of the extracted clinical specimens.

Evaluation of contrived SARS-CoV-2 viral particle specimens without extraction A

panel of 69 contrived samples from the COVID-19 XPRIZE competition were tested on the Harmony system with ready-to-use reagents (**Fig. 5, D and E**). For SARS-CoV-2 virus in PBS, Harmony detected 100% of samples containing ≥ 50 SARS-CoV-2 viral particles/rxn, and 89% of samples containing ≥ 20 viral particles/rxn in PBS (despite PBS being a non-ideal buffer). In nasal matrix, Harmony COVID-19 detected 100% of samples containing ≥ 100 viral particles/rxn, and 92% of samples containing ≥ 20 viral particles/rxn. In human saliva, Harmony detected 100% of samples with ≥ 250 viral particles/rxn and 58% with ≥ 20 viral particles/rxn. Harmony had zero false positives across all samples.

Feasibility testing: sample-to-result system tested by healthcare workers and laboratory personnel

To complete the system, we developed a test kit that included all components needed for running a test, including the swab, elution buffer, a fluid transfer device, and a ready-to-use reagent tube. Swabs were pre-loaded with SARS-CoV-2 DNA to serve as a control sample to evaluate performance by HCWs and laboratory personnel (**LPs**).

In the first phase of user testing (**Fig. 6A**), we compared the accuracy and reproducibility of two methods for transferring the eluate to the reaction tube when operated by LPs ($n = 5$) and HCWs (such as nurses, medical students, and dental students; $n = 10$). The first method (**Fig. 6B** and shown-previously in **Fig. 1**) used a unified sampler-dispenser. The second method (**Fig. 6C**) used a transfer pipette to transfer a fixed volume of fluid to the reaction tube. Regardless of the baseline skill of operators, transfer pipettes yielded more reproducible recovered volumes than the unified dispenser (F-test, $p < 0.001$). Recovered volumes of the two methods were not significantly different between HCWs and LPs (student's t-test, two-sided, $p = 0.90$; **Fig. 6D**). User feedback on the transfer methods and preferences on tube sizes are summarized in **table S2**.

Next, HCWs and LPs executed the full Harmony workflow on contrived SARS-CoV-2 DNA swab samples (**Fig. 6**). Using the transfer pipette workflow (**Fig. 6E**), the test accuracy was 100% (10/10) for LPs and 95% (19/20) for HCWs. Surprisingly, we found that the HCWs performed better than the LPs when using the unified dispenser system (**Fig. 6F**), both accuracy and IND rates. HCWs had significantly lower (Z-test, two-tailed,

p<0.05) IND results (1/20) than LPs (3/10) when using the unified dispenser system.

Excluding IND results, the HCWs had 89% accuracy, higher than LPs with 71%

accuracy (Z-test, two-tailed, $p < 0.05$), when using the unified dispenser system.

DISCUSSION

This work presents a combination of multi-disciplinary engineering (molecular, mechanical, electrical, software, human interface design) to develop a complete system for SARS-CoV-2 detection at the POC. The system is inexpensive and simple yet has features comparable to laboratory-based tests (**Table 1A**). The assay has high sensitivity and specificity in clinical and contrived specimens, and HCWs successfully performed the workflow with high accuracy.

Harmony uses ready-to-use reagents that improve speed and accuracy for assay setup.

Unlike other lyophilized RT-LAMP²⁵, our lyophilized reagents are magnesium-free, which prohibits any enzymatic activity that may occur before assay initiation. Our lyophilized RT-LAMP is activated upon the addition of sample which contains magnesium and detergent. Thus, it should be less prone to moisture degradation than magnesium-containing lyophilized RT-LAMP.

The lyophilized Harmony assay detected as low as 15 copies/rxn of synthetic SARS-CoV-2 RNA and detected all extracted RNA samples from clinical specimens at ≥ 40 copies/rxn. Here, we used only 10 μ L RNA in 40 μ L reaction to allow head-to-head sensitivity comparison to the CDC RT-PCR protocol that uses 5 μ L RNA in a 20 μ L reaction. The

lyophilized reagents can accommodate up to 40 μ L of sample (4x than used here), which could further increase Harmony sensitivity.

Harmony lyophilized RT-LAMP is robust against simulated nasal matrix containing mucin, salt, and human genomic DNA. The impact of salts and mucin was observed as a delayed reaction time for both target and IAC. To account for this delay, we adjusted the assay time to 90 min when analyzing specimens with nasal matrix. We found excellent sensitivity when testing samples with ≥ 50 viral particles in nasal matrix (estimated 1250 copies/swab for 1 mL elution buffer and 40 μ L reaction), which would be sufficient to detect SARS-CoV-2 in most nasal specimens (10^3 - 10^9 SARS-CoV-2 copies/swab) from individuals within the first 3-5 days after the symptom onset²⁶. Further, Harmony detected all samples with ≥ 100 viral particles in nasal matrix (estimated 2500 copies/swab for 1 mL elution buffer and 40 μ L reaction). While saliva was not our intended specimen type, we found a moderate sensitivity when testing the XPRIZE blinded saliva panel. Heat treatment of saliva shown to enable detection by RT-PCR²⁷ would likely improve the sensitivity of our RT-LAMP in detecting SARS-CoV-2 in saliva.

Isothermal amplification, like RT-LAMP in Harmony, is often promoted as ideal for POC testing because it can be carried out using a constant temperature heat source, such as a water bath, heat block, or even more novel sources (chemical heat source, sunlight, body temperature)(10). However, reported tests often have other limitations that may prevent their use in POC testing (**Table 1B**). Common limitations include use of reagents use of refrigerated or frozen reagent stocks that must be formulated using

laboratory pipettes, manual steps for sample processing that require laboratory skill, or detection steps that rely heavily on user interpretation or analysis by a user's cell phone. Harmony is a complete sample-to-result system that uses ready-to-use reagents for simple 1-minute setup and a dedicated device that reports results without extra post-amplification steps or reliance on user interpretation.

Some RT-LAMP tests use end-point lateral flow strips for detection²⁸, which adds a user step, but more importantly, should not be done at the POC because it exposes the testing site to amplicons that will give false positive results in subsequent tests. Others use in-tube detection by eye(11),(11) or using a cell phone(12) which increases the user burden and could introduce additional sources of error. Real-time detection used in Harmony COVID-19 removes the need for extra detection steps or user interpretation. Real-time detection also allows Harmony to report positive results as soon as the SARS-CoV-2 signal appears rather than waiting for an end-point detection method. High viral load samples could be detected in as little as 20 min, which could enable more timely infection control measures to limit viral spread. Harmony real-time detection allows testing without opening tubes after amplification, requires no extra steps or interpretation by the user, and allows reporting a positive sample earlier than end-point detection methods.

The Harmony device automates heating, multiplexed detection, and data interpretation. We integrated a simple heater with low-cost LEDs and sensors to enable real-time two-color detection of RT-LAMP fluorescence in four independent sample wells. The device costs <US\$300 in parts for prototypes, while the lowest cost commercial RT-PCR

machine on the market is >US\$5000(13). Moreover, the Harmony device is portable and energy-efficient, requiring <15 watts and operable by USB power sources. These features are attractive for use at the POC in the U.S. and resource-limited settings. Lastly, the simple workflow and analysis of Harmony is appropriate for POC testing. Test operation involves a few simple steps that take ~1 minute, and there are no further user steps after starting the test. Here, HCWs were able to operate the tests without training using only simple written instructions and watching the procedure on video. The simplicity of the Harmony procedure and completeness of the stand-alone system can be a significant advantage for onboarding to ambulatory care settings and potentially community testing locations.

In this study, we evaluated the system using RNA extracted from clinical samples stored in viral transport medium (**VTM**) and contrived specimens in unknown nasal and saliva matrix (XPRIZE), which may not reflect performance in fresh clinical samples. These samples were used due to the difficulty of obtaining fresh dry swabs intended for this POC test, and RNA was extracted from clinical samples because VTM is irrelevant for immediate testing at the POC. Nevertheless, this testing shows excellent analytical reactivity to clinical specimens, lack of cross-reactivity with other respiratory pathogens, analytical sensitivity comparable to the CDC RT-PCR, and resilience in the presence of nasal matrix.

While our assay is relatively fast and sensitive, there is potential to improve performance. We ran RT-LAMP at 64°C, which is 9°C above the optimal temperature for the RT enzyme. We hypothesize that cooling of the heat block when the lid is open,

followed by a few minutes to ramp back up to 64°C, provided an opportunity for the RT enzyme to operate. In contrast, the reaction speed of the temperature-stable DNA polymerase in this assay is optimal at higher temperature. Thus, operating the device with an initial 55°C step for optimal RT activity followed by a higher temperature step for fast DNA amplification could increase sensitivity and/or speed.

The current design of RT-LAMP assay can be improved to include an IAC that is more representative of the SARS-CoV-2 virus. We currently use a DNA IAC, which has been used in several EUA assays³⁴, but the DNA control does not capture the potential failure of the lysis and cDNA conversion processes. We are developing a new IAC that uses RNA that can be packaged in a viral envelope. This modification may require re-development of excipients that preserve the RT-LAMP enzymes and the enveloped RNA.

POC tests like Harmony could provide rapid testing to enable the re-opening of businesses, schools, and borders; but the need for POC testing extends beyond the current emergency. As natural immunity and vaccination take hold, COVID-19 is likely to become endemic with seasonal outbreaks. Tests such as Harmony will be necessary not only to help identify individuals with SARS-CoV-2, but other respiratory pathogens such as influenza and RSV, as the symptoms of these are similar. Prior to the current SARS-CoV-2 pandemic, the yearly influenza infections caused nearly 30 million cases in the U.S. alone, with 400,000 hospital admission and 35,000 deaths(14, 15). Indeed, particularly when more typical patterns of respiratory infections occur, there will be a need for tests that can perform multiplex testing for individuals presenting respiratory

symptoms. Tests such as Harmony will become a common practice for confirming presence of viral infections that can be treated with antiviral therapies or that demand infection control, as well as helping to reduce inappropriate use of antibiotics. Further, test platforms developed for this pandemic could be more readily adapted to detect new diseases to increase readiness for future pandemics. The advances in awareness and acceptance of testing for infections at scale, and the technology platforms developed will have ongoing benefit for COVID-19 control, reducing harm from other endemic diseases, and fighting future pandemics.

MATERIALS AND METHODS

Preparation of synthetic RNA standards

MERS and SARS-CoV-1 plasmid standards (10006623 and 10006624, Integrated DNA Technologies, Coralville, IA) were amplified using M13 PCR primers to generate DNA templates tagged with T7 promoter sequences. DNA templates containing T7 promoter sequences were transcribed using Hi-T7 RNA polymerase (M0658, New England Biolabs, Lawley, MA) following the manufacturer's protocol, with 1 U/ μ L RNase inhibitor (N2611, Promega, Madison, WI). Following transcription, 0.01 U/ μ L DNase I (EN0521, ThermoFisher Scientific, Waltham, MA) was added to each reaction, and incubated at 37°C for 15 min. RNA transcripts were purified with Monarch RNA Cleanup Kits (T2040, New England Biolabs). RNA was quantified on a Qubit 4 fluorometer (Q32852, RNA HS Assay Kit, ThermoFisher Scientific, Waltham, MA), and checked for length and integrity by electrophoresis on TapeStation (5067 RNA screening tape, Agilent Technologies, Santa Clara, CA). RNA at 10¹⁰ copies/ μ L was stored in nuclease-

free water in single-use aliquots. SARS-CoV-2 RNA was prepared and quantified as previously described(16).

In-house polymerase production

Plasmid preparation: The chimeric polymerase (TFv1) was generated by assembly PCR using genomic material from *Thermus thermophilus* HB27 (ATCC BAA-163D-5™, Gaithersburg, MD) and a synthetic gene fragment derived from *Thermodesulfatator indicus* (*T. in*), (Integrated DNA Technologies, Coralville, IA), similar to the method previously described(17). Briefly, a 3' terminal fragment of the *T.th* HB27 **PolA** gene downstream of the finger domain was amplified by PCR using Phusion® Hot Start Flex DNA Polymerase (M0535L, New England Biolabs) with a modified forward primer containing sequence overlap with the *T. in polA* gene fragment (AA734-795, codon optimized for *E. coli*). A chimeric sequence was then generated by assembly PCR. A second *T.th* polymerase fragment, upstream of the finger domain and lacking the 3' to 5' exonuclease domain, was similarly amplified and appended to the 5' end of the chimeric fragment. The resultant sequence corresponds to a *T. th* DNA polymerase I fragment, orthologous to the Stoffel fragment of *T.aq.* polymerase, with a replacement in the finger domain derived from *T. in*. This gene was then cloned into the expression vector pET His6 TEV LIC cloning vector (2B-T) (a gift from Scott Gradia, Plasmid # 29666, Addgene, Watertown, MA) by ligation independent cloning(18). The annealed vector and insert were transformed into T7 Express lysY/lq Competent *E. coli* (C2987H, New

England Biolabs) per the manufacturer's protocol and plated on LB 100 μ g/mL carbenicillin plates. Colony plasmid inserts were verified by Sanger sequencing.

Protein expression and purification: 50mL starter culture in LB broth with 100 μ g/mL carbenicillin was inoculated with a sequence-verified colony and incubated at 30°C overnight in a shaking incubator at 200rpm. The culture was centrifuged at 5,000 RCF for 15 min and resuspended in an equivalent volume fresh LB carbenicillin media. Flasks of 500mL LB broth with 100 μ g/mL carbenicillin were inoculated with 5mL of resuspended starter culture and incubated at 37°C in a shaking incubator at 200rpm until OD₆₀₀ of 0.4-0.6 had been achieved. Cultures were induced with 0.4mg/mL of IPTG and incubated for two hours then harvested by centrifugation at 5,000 RCF for 15 min, decanted, and stored as cell pellets at -80°C until needed. Prior to purification, cell pellets were resuspended in lysis buffer (20mM sodium phosphate pH7.4, 80mM NaCl, 1%(v/v) Triton (648466 Sigma, St. Louis, MO), and 1mg/mL lysozyme from chicken egg white (Millipore Sigma, Burlington, MA)). Suspensions were heated at 37°C for 20 min, then lysed by ultrasonic disruption using a Fisherbrand™ Model 120 Sonic Dismembrator (Fisher Scientific, Waltham, MA) with two cycles of two second pulses at 50% amplitude for 2 min, while on ice. Lysates were heat-clarified at 65°C for 20 min, followed by centrifugation at 20,000 RCF. The supernatant was collected and diluted 1:6 with sample diluent buffer (20mM sodium phosphate, 1.8M NaCl, pH 7.4) and filtered with a 0.22 μ m syringe filter (25-342 Olympus, Genesee Scientific, San Diego, CA). The Purified clarified lysate was applied to a 1mL Hitrap Talon Crude column (Cytiva, Marlborough, MA) and enriched per manufacturers default recommended

specifications using an ÄKTA start system with the Frac30 fraction collector (Cytiva). Elution fractions were evaluated by A_{280} on a spectrophotometer (2200 Nanodrop, Fisher Scientific) using the proteins expected molecular weight (63.062kDa) and extinction coefficient ($60,850\text{m}^2/\text{mol}$) of the His-tagged polymerase. Elution fractions containing protein were diluted 1:10 with 10mM sodium phosphate and applied to a 1mL HiTrap® Heparin HP column and purified per the manufacturers recommended specifications except for a modified binding buffer (10mM sodium phosphate, and 30mM NaCl, at pH 7). Fractions containing protein were concentrated and buffer exchanged into storage buffer (17mM Tris-HCl, 167mM KCl, 17% glycerol pH7.5) using an Amicon® Ultra-2 30K device (Millipore Sigma), as recommended by the manufacturer. The recovered polymerase solution was then adjusted to 10mM Tris-HCl, 100mM KCl, 2mM DTT, 0.1% Triton X-100, 50% glycerol pH7.5. Polymerase purity was evaluated by SDS-PAGE and spectrophotometry throughout purification.

RT-LAMP assay design

Alignment of SARS-CoV-2 sequences (available on NCBI up to January 2020) were performed using Geneious (Auckland, New Zealand). Three sets of LAMP primers (Harmony NC1, Harmony NC2, and Harmony NC3) were developed targeting the 22,285 – 28,517 (232 bases), 28,528 – 28,471 (213 bases), and 29,135 – 29,416 (281 bases) regions of the SARS-CoV-2 nucleocapsid phosphoprotein gene. *In-silico* analysis showed that this primer design matched 100% of SARS-CoV-2 sequences and did not have cross-reactivity to MERS and SARS-CoV-1. We used two detection probes, each

comprised of a fluorescent probe and its complementary quencher probe. For detection of SARS-CoV-2 RNA amplification, we used FAM labeled detection probe and Iowa Black™ quencher probe. For detection of IAC amplification, we used TEX 615 labeled detection probe and BHQ2 labeled quencher probe. The primer and probe sequences are listed in **table S1**.

Lyophilization of RT-LAMP reagents

Reactions containing 2x primers (3.2mM each FIP/BIP primer, 0.4mM each loop primer, and 0.8mM each F3/B3 primer) and probes (0.4mM fluorophore probes and 0.6mM quencher probes) in the presence of 5% (v/v) mannitol (OPS Diagnostics EXMN 500-01), 10mM DTT, 1.4μg DNA polymerase, 1.2U/μL WarmStart RTx (M0380L, New England Biolabs), 2.8mM each dNTP in 10mM Tris-HCl, pH 7.4, 2U/μL RNase inhibitor (N2615, Promega), 50mU/μL thermostable inorganic pyrophosphatase (M0296L, New England Biolabs), and 2500 copies/μL IAC DNA were prepared in 20μL aliquots of 0.2mL reagent tubes and lyophilized. Packages were stored in desiccant packets at -20°C until use.

RT-LAMP reaction setup

Lyophilized reagents were resuspended in a final volume of 40μL containing 1x rehydration buffer (ThermoPol buffer (B9004S, New England Biolabs), 6mM MgSO₄, and 0.5%(v/v) Triton X-100). For analytical sensitivity experiments, 2μL SARS-CoV-2 RNA in water was added. For clinical testing, 10μL extracted RNA was added to each

RT-LAMP reaction. For testing the contrived specimens, 10 μ L sample (not extracted) was added to each RT-LAMP reaction. For usability testing, 40 μ L swab eluate was added directly to the RT-LAMP reaction using transfer pipettes or the in-house unified dispensers.

Analytical sensitivity of RT-LAMP assay

Analytical sensitivity of lyophilized RT-LAMP was tested using purified RNA (0-2x10⁶ copies/40 μ L reaction). We also tested the effect of salt, human DNA, and mucin on amplification of SARS-CoV-2 RNA and IAC DNA using simulated nasal matrix as previously described(19).

Preparation of contrived swab samples and kits for usability testing

For the transfer pipette workflow, each kit contained one swab (either positive or negative), a vial of buffer (10-500-25, Fisher Scientific, Waltham, MA) containing 400 μ L of 1x rehydration buffer (see section RT-LAMP reaction setup) and one lyophilized RT-LAMP reagent tube. For positive swabs, 10 μ L of 100 copies/ μ L SARS-CoV-2 DNA in 0.05% Triton X-100 was pipetted onto the swab (1804-PF, Puritan Medical) and dried at room temperature in the air-clean chamber (825-PCR/HEPA, Plas-Labs, Lansing, NI) for 4 hours before packaging in a sealed foil pouch with desiccant (S-8032, ULINE). The unified sampler was built from parts using an assembly jig (custom order from Hygiena, LLC, Camarillo, CA). The unified dispenser contained a hollow-shaft polyurethane swab attached to a bulb preloaded with 400 μ L 1x rehydration buffer and a separate tube with

a filter and drip dispenser friction fit inside the tube. Each custom sampler kit contained an in-house assembled dispenser unit and one lyophilized RT-LAMP reagent tube.

Usability study

The Harmony usability study was approved by the Institutional Review Board at the University of Washington (IRB#: STUDY00010884), and informed consent was provided by participants. HCWs ($n = 10$), including registered nurses, medical students, and dental students were enrolled in this study over the course of 1 month. All data was collected without identifiers. The primary objective of this study was to determine the accuracy and precision of the two sample transfer methods and test if variation in reproducibility impacted assay performance, and the secondary objective was to gather user input on the feasibility of each protocol. Participants first received an explanation of the test and its intended use, then watched an instructional video narrated by the facilitator to become familiar with the Harmony workflow (**video S1 and video S2**). After watching the video, participants were given a packet with a unique identifier that contained comprehensive instructions for the study procedures and survey questions. Study procedures were modeled on a coaching methodology(20), to allow natural progression of the procedures and to identify blocking steps in the instructions. The facilitator sat 10 feet from the participant and kept a record of verbal questions and comments from participants but did not provide answers to procedural questions unless the participant was could not proceed.

For part 1, data related to user preference on sample preparation, tube sizes, and their level of confidence in performing each task were collected anonymously as prompts within their study packet. Open-ended questions were extracted for themes in errors or issues and supplemented with comments and observations recorded by the facilitator. Samples were saved with deidentified labels, and after all participants had completed the study facilitators measured volumes in each receiving tube from part 1. The variances of the volume recovered by each method were tested for their equality using F-Test, and the P-value was reported. The means of the recovered volumes transferred by LPs or HCWs were compared using student's t-test (two-sided), and the P-value was reported.

For part 2, each participant followed instructions in the packet and on the integrated mobile phone to complete the entire Harmony workflow on four total contrived samples, two samples using the transfer pipette protocol and two using the unified sampler system. Kits were prepared as specified above.

The lab personnel ($n = 5$) portion of the user testing aimed to compare performance of the Harmony device by HCW to that of an "experienced" population, by replicating this process with four students and a researcher from the Lutz Lab. Survey questions were omitted for this group because these volunteers provided feedback throughout the development of Harmony and their views are included in the discussion of this

manuscript. The accuracies of tests performed by LPs and HCWs using each method were tested for their equality using Z-test (two-sided), and the P-value was reported.

Clinical nasal swab specimens from individuals presenting respiratory symptoms

A minimum sample size of 30 (each negative and positive) was required to achieve the confidence interval (CI) of 90 -100% (binomial cumulative distribution function) should all results be accurate. Our clinical specimen panel ($n = 110$) was previously used to evaluate other SARS-CoV-2 assays(16), but 9 specimens had insufficient volume for RNA extraction and were thus excluded from the study. These specimens were collected with informed consent as part of the Seattle Flu Study, approved by the Institutional Review Board at the University of Washington (IRB#: STUDY0006181). All participant specimens were de-identified prior to the transfer to the Lutz lab at the University of Washington for analysis. The remaining 101 specimens were evaluated using the newly-extracted RNA for analysis by RT-qPCR and Harmony system, to avoid variations introduced from different batches of extraction. RNA was extracted from clinical nasal swab specimens stored in viral transport media using the QIAamp Viral RNA Mini Kit (52906; Qiagen, Hilden, Germany). Prior to extraction, 100 μ L of each specimen was mixed with 40 μ L negative VTM to reach the suggested 140 μ L sample volume. RNA was extracted according to the manufacturer's protocol, eluted into 70 μ L of EB buffer, and stored at -80°C in single-use aliquots until amplification. 5 μ L of RNA was used for analysis in 20 μ L RT-qPCR using U.S. CDC N1, N2, or RP human control assays(9). 10 μ L of RNA was used for analysis in 40 μ L

lyophilized RT-LAMP in the Harmony workflow. Each sample was run in duplicate. Only samples with positive results from both Harmony replicates were reported as positive for SARS-CoV-2. Sensitivity ($TP/(TP + FN)$), specificity ($TN/(TN+FP)$), and concordance ($((TP+TN)/(TP+TN+FP+FN))$) were calculated and reported with 95%CI using binomial exact proportions.

XPRIZE blinded contrived sample panel

The blinded panel was assembled and distributed by HudsonAlpha Discovery (Huntsville, AL) for COVID-19 XPRIZE competition, and information on the samples was revealed to the authors only after results were submitted to XPRIZE. Samples were shipped to the Lutz Lab at the University of Washington. We reported results for contrived samples of chemically-inactivated SARS-CoV-2 viral particle (ZeptoMetrix Corporation, Buffalo, NY) spiked in phosphate-buffered saline (**1x PBS**) ($n = 29$), nasal specimens ($n = 20$), or saliva ($n = 20$) at 0 to 1000 copies/rxn. To affirm the identify of these specimens, we also performed RT-qPCR using CDC primer/probes after the samples had been tested using the Harmony system.

Housing

The housing combines the heater/reader assembly with the cell phone, while concealing the connector cables and presenting a sturdy and easy-to-use platform. The 3D CAD program, Fusion 360 (AutoDesk), was used for the design of housing. Prototypes were printed in PLA plastic (Overture™ Filament) using a Prusa i3 MK3s at 0.3mm layer

height. Power was supplied to both the heater/reader assembly and the phone simultaneously using the Charge-Plus USB-C (LAVA Computer Manufacturing Inc.) connector device and a standard fast-charging American cellphone charger. The LAVA device also facilitates the transfer of data between the devices.

Reader/heater device

The aluminum block to house the reagent tubes was custom-made by Bryan Willman according to the specifications in **fig. S3**. The circuit boards were designed using Autodesk Eagle (San Rafael, CA), and the assembly of the circuit boards are shown in **fig. S4** and **fig. S5**. The lid, latch, and upper and lower housing of the device were drawn using SolidWorks (Dassault Systèmes, Waltham, MA) and printed by Xometry - HP MultiJet Fusion 3D (Gaithersburg, MD). All parts were assembled as shown in **fig. S6**.

Real-time fluorescence signal analysis

After the user inserts the reaction tube into the device, the software adjusts the temperature to re-equilibrate the temperature of heat block back to the reaction temperature. During this initial period, the signals from the photo diodes (the blue LED (FAM signal for target amplification) and the yellow LED (TEX 615 signal for the IAC amplification) are collected but not used in analyses. Once the initial period is over, signal analysis commences. Each signal is analyzed in real time by comparing the current mean of the signal over the most recent 60s interval to the signal's behavior

some 360s in the past. Specifically, the current mean is compared to a bound consisting of the signal's mean over the period from 420s to 360s in the past plus a multiple (1.9x for target, 1.5x for IAC) of the signal's sample standard deviation over that same period. If the current mean exceeds the bound, and maintains that condition for at least 15s, the signal is said to have "indicated". If the signal from the target emission indicates, the presence of SARS-CoV-2 has been detected, and the software immediately reports "positive COVID-19". If the signal from the IAC emission indicates, successful amplification of the IAC has been detected. Then, if the target emission does not also indicate before the end of the run, then a "negative COVID-19" result is reported; this delay in reporting negative results provides the maximum opportunity to detect low concentrations of SARS-CoV-2 that may be present in samples. If neither the target emission nor the IAC emission is detected, "test failed" is reported.

REFERENCES AND NOTES

1. World Health Organization. (2020), vol. 2021.
2. World Health Organization. (2021), vol. 2021.
3. A. Sandford. (Euronews with AP, AFP, 2020).
4. G. Bonaccorsi, F. Pierri, M. Cinelli, A. Flori, A. Galeazzi, F. Porcelli, A. L. Schmidt, C. M. Valensise, A. Scala, W. Quattrocioni, F. Pammolli, Economic and social consequences of human mobility restrictions under COVID-19. *Proc Natl Acad Sci U S A* **117**, 15530-15535 (2020).

- 610 5. I. Y. Chu, P. Alam, H. J. Larson, L. Lin, Social consequences of mass quarantine
611 during epidemics: a systematic review with implications for the COVID-19
612 response. *J Travel Med* **27**, (2020).
- 613 6. J. Garre-Olmo, O. Turro-Garriga, R. Marti-Lluch, L. Zacarias-Pons, L. Alves-
614 Cabratosa, D. Serrano-Sarbosa, J. Vilalta-Franch, R. Ramos, G. Girona Healthy
615 Region Study, Changes in lifestyle resulting from confinement due to COVID-19
616 and depressive symptomatology: A cross-sectional a population-based study.
617 *Compr Psychiatry* **104**, 152214 (2021).
- 618 7. Centers for Diseases Control and Prevention. (2021), vol. 2021.
- 619 8. Y. Mori, K. Nagamine, N. Tomita, T. Notomi, Detection of loop-mediated
620 isothermal amplification reaction by turbidity derived from magnesium
621 pyrophosphate formation. *Biochem Biophys Res Commun* **289**, 150-154 (2001).
- 622 9. U.S. Centers for Disease Control and Prevention, "CDC 2019-Novel Coronavirus
623 (2019-nCoV) Real-Time RT-PCR Diagnostic Panel For Emergency Use Only,"
624 (2020).
- 625 10. P. Craw, W. Balachandran, Isothermal nucleic acid amplification technologies for
626 point-of-care diagnostics: a critical review. *Lab Chip* **12**, 2469-2486 (2012).
- 627 11. E. Gonzalez-Gonzalez, I. M. Lara-Mayorga, I. P. Rodriguez-Sanchez, Y. S.
628 Zhang, S. O. Martinez-Chapa, G. T. Santiago, M. M. Alvarez, Colorimetric loop-
629 mediated isothermal amplification (LAMP) for cost-effective and quantitative
630 detection of SARS-CoV-2: the change in color in LAMP-based assays

- quantitatively correlates with viral copy number. *Anal Methods* **13**, 169-178 (2021).
12. J. Rodriguez-Manzano, K. Malpartida-Cardenas, N. Moser, I. Pennisi, M. Cavuto, L. Miglietta, A. Moniri, R. Penn, G. Satta, P. Randell, F. Davies, F. Bolt, W. Barclay, A. Holmes, P. Georgiou, Handheld Point-of-Care System for Rapid Detection of SARS-CoV-2 Extracted RNA in under 20 min. *ACS Cent Sci* **7**, 307-317 (2021).
13. Chaibio. (Santa Clara, CA, 2021).
14. W. Putri, D. J. Muscatello, M. S. Stockwell, A. T. Newall, Economic burden of seasonal influenza in the United States. *Vaccine* **36**, 3960-3966 (2018).
15. Centers for Diseases Control and Prevention. (2012).
16. N. Panpradist, Q. Wang, P. S. Ruth, J. H. Kotnik, A. K. Oreskovic, A. Miller, S. W. A. Stewart, J. Vrana, P. D. Han, I. A. Beck, L. M. Starita, L. M. Frenkel, B. R. Lutz, Simpler and faster Covid-19 testing: Strategies to streamline SARS-CoV-2 molecular assays. *EBioMedicine* **64**, 103236 (2021).
17. N. Morant, University of Bath, (2015).
18. Addgene. (Watertown, Massachusetts, USA, 2021), vol. 2021.
19. N. Panpradist, B. J. Toley, X. Zhang, S. Byrnes, J. R. Buser, J. A. Englund, B. R. Lutz, Swab sample transfer for point-of-care diagnostics: characterization of swab types and manual agitation methods. *PLoS One* **9**, e105786 (2014).
20. R. Mack, J. B. Robinson, in *The Psychology of Expertise*, H. R.R., Ed. (Springer, New York, NY, 1992), pp. 245-268.

653 21. FPbase. (2021), vol. 2020.

654 22. Newport. (2021), vol. 2021.

655 23. Mousers Electronics. (2021), vol. 2021.

656 24. A. Alekseenko, D. Barrett, Y. Pareja-Sanchez, R. J. Howard, E. Strandback, H.
657 Ampah-Korsah, U. Rovsniak, S. Zuniga-Veliz, A. Klenov, J. Malloo, S. Ye, X. Liu,
658 B. Reinius, S. J. Elsasser, T. Nyman, G. Sandh, X. Yin, V. Pelechano, Direct
659 detection of SARS-CoV-2 using non-commercial RT-LAMP reagents on heat-
660 inactivated samples. *Sci Rep* **11**, 1820 (2021).

661 25. B. A. Rabe, C. Cepko, SARS-CoV-2 detection using isothermal amplification and
662 a rapid, inexpensive protocol for sample inactivation and purification. *Proc Natl*
663 *Acad Sci U S A* **117**, 24450-24458 (2020).

664 26. S. Wei, E. Kohl, A. Djandji, S. Morgan, S. Whittier, M. Mansukhani, E. Hod, M.
665 D'Alton, Y. Suh, Z. Williams, Direct diagnostic testing of SARS-CoV-2 without the
666 need for prior RNA extraction. *Sci Rep* **11**, 2402 (2021).

667 27. V. L. Dao Thi, K. Herbst, K. Boerner, M. Meurer, L. P. Kremer, D. Kirrmaier, A.
668 Freistaedter, D. Papagiannidis, C. Galmozzi, M. L. Stanifer, S. Boulant, S. Klein,
669 P. Chlanda, D. Khalid, I. Barreto Miranda, P. Schnitzler, H. G. Krausslich, M.
670 Knop, S. Anders, A colorimetric RT-LAMP assay and LAMP-sequencing for
671 detecting SARS-CoV-2 RNA in clinical samples. *Sci Transl Med* **12**, (2020).

672 28. L. Bokelmann, O. Nickel, T. Maricic, S. Paabo, M. Meyer, S. Borte, S.
673 Riesenberger, Point-of-care bulk testing for SARS-CoV-2 by combining

hybridization capture with improved colorimetric LAMP. *Nat Commun* **12**, 1467 (2021).

29. W. Yamazaki, Y. Matsumura, U. Thongchankaew-Seo, Y. Yamazaki, M. Nagao, Development of a point-of-care test to detect SARS-CoV-2 from saliva which combines a simple RNA extraction method with colorimetric reverse transcription loop-mediated isothermal amplification detection. *J Clin Virol* **136**, 104760 (2021).
30. J. Qian, S. A. Boswell, C. Chidley, Z. X. Lu, M. E. Pettit, B. L. Gaudio, J. M. Fajnzylber, R. T. Ingram, R. H. Ward, J. Z. Li, M. Springer, An enhanced isothermal amplification assay for viral detection. *Nat Commun* **11**, 5920 (2020).
31. M. Patchsung, K. Jantarug, A. Pattama, K. Aphicho, S. Suraritdechachai, P. Meesawat, K. Sappakhaw, N. Leelahakorn, T. Ruenkam, T. Wongsatit, N. Athipanyasilp, B. Eiamthong, B. Lakkanasirorat, T. Phoodokmai, N. Niljianskul, D. Pakotiprapha, S. Chanarat, A. Homchan, R. Tinikul, P. Kamutira, K. Phiwkaow, S. Soithongcharoen, C. Kantiwiriyanitch, V. Pongsupasa, D. Trisrivirat, J. Jaroensuk, T. Wongnate, S. Maenpuen, P. Chaiyen, S. Kamnerdnakta, J. Swangsri, S. Chuthapisith, Y. Sirivatanauksorn, C. Chaimayo, R. Sutthent, W. Kantakamalakul, J. Joung, A. Ladha, X. Jin, J. S. Gootenberg, O. Abudayyeh, F. Zhang, N. Horthongkham, C. Uttamapinant, Clinical validation of a Cas13-based assay for the detection of SARS-CoV-2 RNA. *Nat Biomed Eng* **4**, 1140-1149 (2020).

32. J. Joung, A. Ladha, M. Saito, M. Segel, R. Bruneau, M. W. Huang, N. G. Kim, X. Yu, J. Li, B. D. Walker, A. L. Greninger, K. R. Jerome, J. S. Gootenberg, O. O. Abudayyeh, F. Zhang, Point-of-care testing for COVID-19 using SHERLOCK diagnostics. *medRxiv*, (2020).
33. J. P. Broughton, X. Deng, G. Yu, C. L. Fasching, V. Servellita, J. Singh, X. Miao, J. A. Streithorst, A. Granados, A. Sotomayor-Gonzalez, K. Zorn, A. Gopez, E. Hsu, W. Gu, S. Miller, C. Y. Pan, H. Guevara, D. A. Wadford, J. S. Chen, C. Y. Chiu, CRISPR-Cas12-based detection of SARS-CoV-2. *Nat Biotechnol* **38**, 870-874 (2020).

ACKNOWLEDGMENTS

We thank Bryan Willman for his support in fabricating customized heat blocks, Jan Gray for his help with firmware development, Tom Blank for sharing the PCB fabrication resources, and Dr. Liz Sanocki for the useful feedback on the software user interface. We thank Shawna Cooper for sharing the graphics and content used in Audere's mobile app's flu@home, which influenced the design of the first few screens of the Harmony COVID-19 software. We thank Dr. Paul Yager and Steven Bennett for use of their optical filters during the initial stage of device development. We thank Jason Tauscher at the Washington Nanofabrication Facility for his assistance in cutting the plastic sheet filters. The Washington Nanofabrication Facility is a National Nanotechnology Coordinated Infrastructure (NNCI) site at the University of Washington with partial support from the National Science Foundation via awards NNCI-2025489 and NNCI-

1542101. We thank Seattle Flu Study researchers Dr. Jay Shendure and Dr. Jase Gehring for their helpful technical discussion. We thank Dr. Syamal Raychaudhuri, Dr. Frances Chu from Inbios International, and Dr. Gwong-Jen J. Chang for the stimulating discussion. We thank the Seattle Flu Study and the Seattle Coronavirus Assessment Network (SCAN) teams led by Principal Investigators: Helen Y. Chu, MD, MPH, Michael Boeckh, MD, PhD, Janet A. Englund, MD, Michael Famulare, PhD, Barry R. Lutz, PhD, Deborah A. Nickerson, PhD, Mark J. Rieder, PhD, Lea M. Starita, PhD, Mathew Thompson, MD, MPH, DPhil, Jay Shendure, MD, PhD and Trevor Bedford, PhD for providing specimens for testing. **Funding:** This work was supported by the Seattle Flu Study (funded by Gates Ventures) and the National Institutes of Health (R01AI140845; 5R61AI140460-03). We thank the Washington Entrepreneurial Research Evaluation and Commercialization HUB program (WE-REACH, U01 HL152401) and their matching grant partners for partly supporting reagents, supplies, and personnel. I.T.H. was supported in part by the National Institute of General Medical Sciences of the National Institutes of Health under Award Number T32GM008268. This work does not reflect the views of the funders. The funders were not involved in the design of the study, and funders do not have any ownership over the management and conduct of the study, the data, or the rights to publish. **Author contributions:** N.P. developed lyophilized RT-LAMP and outlined the overall plan for experiments and data visualization presented in this manuscript. R.G.A. developed a mobile software application for operating the device and analyzing results in real-time. M.R. developed the heater and fluorescent reader device. E.K. designed the RT-LAMP assay primers and probes and engineered

polymerase. E.K., I.T.H., and Q.W. synthesized polymerase and provided usability feedback during the development of the Harmony COVID-19 system. I.T.H. and Q.W. prepared *in vitro* RNA transcripts. J.H.K. and V.L. led the usability study. N.P., A.K.O., and C.B. tested clinical and contrived specimens. R.G.A., N.P., A.K.O., and J.H.K. analyzed the data. I.T.H. and Q.W. prepared the kits and blinded mock samples for usability testing. D.L. developed housing units to integrate the cell phone, sample preparation station, and the device. L.M.S. and P.D.H. designed and characterized the clinical specimen panel used in this study. M.T. oversaw the usability study. B.R.L. oversaw the development and evaluation of the Harmony COVID-19 system. All authors contributed to writing this manuscript and approved the final version. **Competing interest:** Provisional patent applications have been filed on several components of the Harmony COVID-19 system. N.P., R.G.A., M.R., E.K., Q.W., I.T.H., D.L., and B.R.L. are inventors on one or more provisional patent applications. N.P., R.G.A., M.R., E.K., Q.W., I.T.H., J.H.K., A.K.O., C.B., D.L., and B.R.L. anticipate forthcoming equity in a startup company that has licensed related technology and supports ongoing work in the B.R.L. laboratory at the University of Washington but played no role in the study. M.T. is on the Advisory Board for Visby Medical which produces tests for COVID-19 and has received reimbursement for this work. He has also received reimbursement for medical advice to Roche Molecular Diagnostics and Inflammatix. **Data and materials availability:** All data are presented in the main text and supplementary information. **Code availability:** Sequence alignment and primer design was performed using Geneious 8.1.9 (Auckland, New Zealand). CFX-Maestro (BioRad, Hercules, CA) was

761 used to analyze the RT-PCR results. TapeStation Software (Agilent, Santa Clara, CA)
 762 was used to visualize fragments of synthetic RNA. A provisional patent application was
 763 filed regarding aspects of the cell phone application. The data analysis algorithm is
 764 described in the manuscript.

765

766 **FIGURE CAPTIONS:**

767

A

B

C

D

Fig. 1 Harmony COVID-19 workflow, analytical performance, and interpretation.

(A) User workflow. A nasal swab is collected and eluted in the rehydration buffer using either a unified sampler-dispenser (shown) or a tube-and-bulb method (Fig. 6). The swab eluate is then transferred into the reaction tube to rehydrate lyophilized, pre-loaded RT-LAMP reagents. Users follow the instructions on the cell phone app to record

the sample identity, insert the tube in the test slot, and close the device. Four samples can be analyzed in parallel. (Photo credits: a1-a4, Mark Stone at the University of Washington and a5, Christopher Snyder at North Seattle College). **(B)** Examples of on-device analysis. Lyophilized reagents were rehydrated with the eluate from swabs spiked with 0, 200, 2000, or 20,000 copies of SARS-CoV-2 RNA, run on the device, and analyzed in real time. The samples with SARS-CoV-2 RNA at 2000, 200, and 20 copies/reaction (**rxn**) were reported as positive by software at 27 min, 30 min, and 41 min, respectively; after 60 min the NTC reaction was classified as negative. **(C)** Time-to-result. Individual detection time points ($n = 3$) are plotted. Variation in detection time increases as samples approach the detection limit. **(D)** Diagnostic algorithm. Harmony COVID-19 software calls samples as positive, negative, or indeterminate based on detection of the three regions in nucleocapsid gene (NC1 overlapping with the CDC N1 region, NC2, and NC3 overlapping with the CDC N2 region) and the engineered IAC sequence.

Fig. 2 Lyophilized RT-LAMP and its analytical sensitivity.

(A) To set up the assay, a sample in elution/rehydration buffer containing magnesium, Thermopol buffer, and detergents is added to rehydrate the lyophilized RT-LAMP reagents in a single step. **(B)** Real-time amplification signal of RT-LAMP reactions containing SARS-CoV-2 RNA (**NTC**; no template control), 10, 15, 20, 200, 2000, 2×10^4 , or 2×10^5 copies/40 μ L reaction (**rxn**). The left panel shows the average ($n = 3$) FAM signal detecting SARS-CoV-2. *Only 2/3 replicates of 10 copies/rxn amplified so the average was from duplicate reactions. The right panel shows TEX 615 signal detecting the internal amplification control (IAC). The IAC was detected in the NTC ($n = 3$) and in the single 10 copies/rxn that did not detect SARS-CoV-2. **(C)** Time to detection of SARS-CoV-2 and IAC targets reported by the real-time thermal cycler. Individual data points are plotted. Note that only 2/3 replicates of 10 copies/rxn amplified. IAC signals were properly detected in NTC ($n = 3$) and 1 of the 3 replicates of 10 copies/rxn undetected for SARS-CoV-2. **(D)** Evaluation of analytical sensitivity at 20 copies/rxn of SARS-CoV-2 RNA ($n = 40$), performed in two different runs ($n = 20$ each) with two different serially diluted RNA samples. Time-to-result of individual samples are plotted along with medians (middle lines of the boxes), interquartile ranges (edges of the boxes), ranges (the ends of whiskers), and outliers (data points outside the whiskers). **(E)** Impact of nasal matrix. SARS-CoV-2 RNA at 0, 40, or 100 copies/rxn was amplified in lyophilized RT-LAMP in the presence of simulated nasal matrix ($n = 3$). All data was

812 measured every 13s using FAM and TEX 615 channels, by a real-time thermal cyclers in
813 1h reactions at 63.3°C, and (C).

814

815

Fig.3 Overview of Harmony COVID-19 hardware.

819 (A) Plastic case cradles the cell phone (removable and connected by a spring cable),
 820 mounts the heater/reader device, and houses a Charge-Plus USB-C connector hub (not
 821 shown). A sample station holds the sample and reaction tube during processing and is
 822 detachable to allow cleaning. Weight: 266g housing, 104g heater/reader device (total
 823 <0.5kg). Dimensions: 205mm x 167.5mm x 125mm housing, 30mm x 120mm x 50mm
 824 heater/reader device. (B) The heater/reader device contains an aluminum block with
 825 wells for four samples (0.2mL reagent tubes) and is heated from the bottom by a
 826 resistive heater made from serpentine conductive traces on a printed circuit board. In
 827 the middle of the heater, a temperature sensor (Class B Platinum RTD) is used to
 828 measure the heat block temperature. Power to the heater is supplied by a MOSFET and
 829 modulated by varying the PWM duty cycle of the MOSFET. (C) Temperature profile
 830 after the lid opening throughout a 60-min assay run. The microcontroller
 831 (die) is measured to ensure the electronics are not overheated. (D) Lid heater, LED, and
 832 detector. A heater (85°C) on top of the tube prevents condensation on the tube lid. The
 833 lid heater is spring loaded to apply a preload force to the tubes to ensure good thermal
 834 contact and contains pass-through holes for excitation LEDs. (E) Spectral
 835 characteristics of the detection system. Blue and yellow LEDs provide excitation for
 836 FAM and TEX 615, respectively, and photodetectors on each side detect light passing
 837 through emission filters (spectral information from manufacturers(21-23)). The LEDs are
 838 operated independently (only one LED is illuminated at a given time). (F) Effect of
 839 thermostable inorganic pyrophosphatase (TiPP) on RT-LAMP signal in the Harmony

device. Normalized signal of negative control samples using the RT-LAMP formulation with and without TiPP ($n = 1$ each) at 64°C.

Fig. 4 Cell phone user interface and result reporting.

845 The Android Harmony application guides the user through steps to set up and run the
 846 test. **(A)** Intended use of the device. **(B)** Instructions for the user to prepare samples.
 847 **(B)** Selection of the number of samples. **(D)** Collection of sample identifiers by barcode,
 848 photo capture, or manual typing. **(E)** Instruction to insert the tube when the temperature
 849 of the device reaches the reaction temperature. **(F)** Real-time analysis during the test
 850 run. **(G)** Positive results are reported as soon as they are detected. **(H)** Negative and
 851 IND results are reported at the end of the run time.

852

853

Fig. 5 Harmony COVID-19 performance on clinical and blinded contrived specimens.

(A) Detection time called by Harmony COVID-19 software for the analysis of purified RNA extracted from human nasal specimens from 101 patients presenting respiratory symptoms. Stored specimens were re-tested by RT-qPCR using N1 and N2 CDC primers; results matched previous results in 30 previous positive specimens (positive for N1 and N2) and 68 previous negative specimens, with three previous positive

specimens classified as inconclusive (INC) because they were negative for either CDC N1 or CDC N2 by RT-qPCR. Samples are ranked from left to right by increasing SARS-CoV-2 concentrations as determined by RT-qPCR. Extracted RNA from specimens was analyzed by Harmony COVID-19 (10 μ L RNA in 40 μ L RT-LAMP dry reagents, $n = 2$). Samples were classified as positive by Harmony only if both replicates were detected. The classification results by Harmony were compared to RT-qPCR using CDC primers and probes (5 μ L RNA in 20 μ L RT-qPCR reaction). **(B)** Organisms in clinical samples collected from individuals presenting with respiratory symptoms were quantified using OpenArray, as previously described(16). High quantification cycle values indicate low concentrations. All these specimens were correctly identified as SARS-CoV-2 negative by Harmony. **(C)** Summary table for the sensitivity and specificity of the Harmony COVID-19 system using extracted RNA from clinical specimens (RT-qPCR INC samples excluded). **(D)** Blinded analysis of XPRIZE contrived sample panel. The assay was run for 90 min (30 min longer than **Fig. 5A** to accommodate inhibitors that may be present in these blinded samples). The table below shows the detectable samples over the total number of tested technical replicates. **(E)** Summary table for the sensitivity and specificity of the Harmony COVID-19 system on the XPRIZE contrived samples.

Fig. 6 Usability study of Harmony COVID-19 workflow including two sample transfer methods at the POC.

(A) Study design. (B) Illustration of the first sample transfer method including a volumetric transfer pipette (40 μ L), a nasal swab, and a pre-loaded buffer vial. **video S1** demonstrates this workflow. Briefly, swabs are manually rubbed on the side of the tube for 30 seconds, soaked for 1 minute, and removed. The pipette is then used to transfer the swab eluate to rehydrate the lyophilized RT-LAMP reaction tube. (C) Illustration of the second sample transfer method option using a unified dispenser, which integrates a swab, buffer container, and a dropper in a single device. The contained buffer is released to elute the swab sample by cracking a conduit, the user agitates the fluid by shaking the tube, and a dropper tip allows the use to transfer two drops of fluid to the reaction tube by squeezing the body of the device. **video S2** demonstrates this workflow. (D) Volume transferred by each method performed by healthcare workers (HCWs) or laboratory personnel (LPs). Individual volume measurements are plotted along with medians (middle lines of the boxes), interquartile ranges (edges of the boxes), and ranges (the ends of whiskers). **e** and **f**, Assay performance on blinded contrived swabs (one negative swab and one positive swab with SARS-CoV-2 DNA at 1,000 copies/swab, corresponding to 100 copies/rxn) by LPs ($n = 5$) and untrained HCWs ($n = 10$) using transfer pipettes, (E) or unified dispensers, (F) The usability experiment written protocol is available on protocol.io (<https://dx.doi.org/10.17504/protocols.io.bkvskw6e>) and as a visual demonstration in

video S1 and video S2. (G) Performance summary for the transfer pipette method. **(H)** Performance summary for the unified dispenser method.

TABLE CAPTIONS:

Table 1 Feature comparison between the Harmony COVID-19 system and other assays. **(A)** Comparison to the U.S. CDC assay. **(9)** Full cost calculations for reagents and devices are enclosed in **tables S3** and **table S4**.

Item	Standard RT-PCR assay (based on the U.S. CDC protocol)(9)	Covid-19 Harmony
SARS-CoV-2 primer detection	Separate reactions for N1 and N2	Multi-target redundancy amplifying three regions of the N gene in a single reaction tube
Process control	Human RP gene	Engineered internal control
LoD	1-3.2 copies/ μ L from extracted RNA	0.5 copies/ μ L (20 copies/40 μ L rxn) from synthetic RNA; 1 copies/ μ L (40 copies/40 μ L rxn) from extracted RNA from VTM; 2.5 copies/ μ L (100 copies/40 μ L rxn) from contrived nasal samples
Consumable cost	\$12.94 (\$6 extraction kit and \$6.94 for RT-PCR Taq Path/IDT probes)	\$8.00
Easy to use reagents	No	Yes
Batch size	24 specimens	Not required, run up to 4 samples
Equipment cost	> \$5000 qPCR machine	\$247 customized heater/reader
Time for sample-to- result	260 min	21-91 min

Hands-on: 120 min	Hands-on: 1 min for direct amplification
Wait-time: 120 min	Wait-time: 20-45 min for positive samples, but 90 min
Interpretation: 20 min	is required to confirm true negative

912 **(B) Comparison of Harmony COVID-19 to other isothermal assays for SARS-CoV-2.**

Published assays	Amplification chemistry	Multi-target redundancy in a single tube	Reported LoD	User-friendly devices with ready-to-use reagents and a closed system	Demonstration of use by HCWs
Harmony COVID-19 (This study)	RT-LAMP	Yes	0.5-2.5 copies/ μ L	Yes	Yes
A Alekseenko <i>et. al.</i> 2021(24)	RT-LAMP	No	10 copies/ μ L	No	No
BA Rabe and C Cepko(25)	RT-LAMP	No	50 copies/ μ L	No	No
S Wei, <i>et. al.</i> (26)	RT-LAMP	No	2.5 copies/ μ L	No	No
D Thi, <i>et. al.</i> (27)	RT-LAMP	No	5 copies/ μ L	No	No
L Bokelmann <i>et. al.</i> (28)	RT-LAMP	No	5-20 copies/ μ L	No	No
W Yamaziki <i>et. al.</i> (29)	RT-LAMP	Yes	25 copies/ μ L	No	No
J Qian <i>et. al.</i> (30)	RPA	Yes	0.5 copies/ μ L	No	No

M Patchsung <i>et. al.</i> (31)	SHERLOCK	No	2.1 copies/ μ L	No	No
J Joung <i>et. al.</i> (32)	SHERLOCK	No	2 copies/ μ L	No	No
J Broughton <i>et.al.</i> (33)	CRISPR cas12	No	10 copy/ μ L	No	No

SUPPLEMENTARY INFORMATION:

fig. S1 Effects of excipients on fresh RT-LAMP reactions tested on a commercial real-time thermal cycler.

fig. S2 Cross-reactivity of MERS and SARS in lyophilized RT-LAMP.

fig. S3 Heat block used in Harmony COVID-19 device.

fig. S4 The design of optical and heater circuit boards for the Harmony COVID-19 device.

fig. S5 Circuit board assembly and required parts

fig. S6 Device assembly

video S1 Demonstration of transfer pipette workflow incorporated with Harmony COVID-19.

video S2 Demonstration of the integrated sampling system workflow with Harmony COVID-19

table S1 Primers, probes, and control sequences.

table S2 Feedback from the HCWs

table S3 Detailed reagent and device cost at a production scale of 10,000 units.

table S4 Consumable costs per test

SUPPLEMENTARY INFORMATION

Trehalose can stabilize several enzymes, but we found that 2.5% (w/v) trehalose resulted in a slower reaction at 63°C, the optimal reaction temperature of the excipient-free RT-LAMP. An increase of reaction temperature to 65°C can speed up the trehalose-containing RT-LAMP reactions but resulted in unreliable detection at 20 RNA copies/reaction (**fig. S1**).

A

Target (copies/ reaction)	min to detection							
	65 °C		63 °C		59 °C		55 °C	
	SARS-CoV-2	IAC	SARS-CoV-2	IAC	SARS-CoV-2	IAC	SARS-CoV-2	IAC
0	N/A	27.3, 27.7	N/A	31.8, 30.7	N/A	44.7, 42.6	N/A	N/A
20	N/A	29.8, 27.5	N/A, 31.63	30.6, 32.1	N/A, 43.0	42.0, 42.8	N/A	N/A, 60.4
200	N/A	27.3, 27.4	N/A, 27.86	29.4, 39.5	N/A, 43.2	47.9, 40.4	N/A	N/A, 62.0
2000	24.7, 28.6	N/A	28.0, 28.2	30.3, 32.8	N/A, 51.0	43.3, 41.8	N/A	N/A, 60.3

B

% mannitol	Target (copies/reaction)	SARS-CoV-2 (min to detection)	IAC (min to detection)
0	0	N/A	33.0, 33.0
	20	25.0, 25.7	N/A
2.5	0	N/A	33.9, 34.0
	20	26.3, 26.3	N/A

fig. S1 Effects of excipients on fresh RT-LAMP reactions tested on a commercial real-time thermal cycler. (A) Time to detection of SARS-CoV-2 RNA and IAC targets in the RT-LAMP containing 2.5% (*w/v*) trehalose and 0-2000 copies of SARS-CoV-2 RNA/40 μ L reaction. N/A means undetectable. For each condition, two technical replicates were run at 4 different reaction temperatures (i.e., 55°C, 59°C, 63°C, and 65°C). **(B)** Time to detection of SARS-CoV-2 RNA and IAC targets in the RT-LAMP reactions containing 0 or 20 copies of SARS-CoV-2 RNA/reaction, with and without 2.5% (*w/v*) mannitol, at 63°C.

non-SARS-CoV-2 RNA (copies/reaction)	SARS-CoV-2 RNA (copies/reaction)	min to detection SARS-CoV-2	IAC
4000 copies SARS-CoV-1	0	N/A	29.0, 30
4000 copies SARS-CoV-1	100	23.5, 24.1	N/A
4000 copies MERS	0	N/A	29.8, 30.1
4000 copies MERS	100	24.7, 24.7	N/A

fig.S2 Cross-reactivity of MERS and SARS in lyophilized RT-LAMP.

Detection time (min) of SARS-CoV-2 RNA and IAC targets in the presence of 4000 copies of either SARS-CoV-1 RNA or MERS RNA in 40 μ L RT-LAMP reactions at 63°C. Two technical replicates were conducted in each condition. In the reactions without SARS-CoV-2 RNA, we did not observe amplification of FAM signal but observed IAC amplification indicating the amplification reaction was functioning properly in each replicate. 100 copies of SARS-CoV-2 RNA were amplified in each positive control replicate, indicating proper function of SARS-CoV-2 amplification in this master mix.

961

962

963 **fig. S3 Heat block used in Harmony COVID-19 device.** Drawing credit: Bryan

964 Willman.

965

fig. S4 The design of optical and heater circuit boards for the Harmony COVID-19 device. (A) LED board (B) the detector board

fig. S5 Circuit board assembly and required parts.

(A) LED board. (B) LED board assembled to the lid heater board. (C) Custom-cut filter pieces. (D) Detector board and the main board. Each device require the following parts: thick film resistors (RCL12251R20FKEG, RCL12252R20FKEG, CRCW08050000Z0EBC and RCL12254R70FKEG from Vishey; RN73C1J75RBTDF from TE Connectivity; FTS-105-03-F-DV from Samtec; CR0402AJW-472GAS from Bourns), thin film resistors (RN73H1ETTP3003B25 from KOA Speer; ERJ-2RKF22R0X from Panasonic), multilayer ceramic capacitors (C0402C104M4RACAUTO, KEMET;

0603YC105KAT2A, AVX), multilayer ceramic capacitor (GRM188R60J226MEA0D, Murata), USB connectors (10118193-0001LF and 10118192-0001LF, FCI / Amphenol), FFC & FPC connector (52745-1497, Molex), standard yellow LEDs (LY E63B-CBEA-26-1-Z, Osram Opto Semiconductor), standard blue LEDs (150141BS63140, Würth Elektronik), board-to-board & mezzanine connector (CLP-105-02-F-D and FLE-105-01-G-DV-K-TR, Samtec), FFC/FPC jumper cables (15166-0143, Molex), FFC & FFC connector (52559-1452, Molex), 2 headers & wire Housings (FTSH-105-01-F-DH, Samtec), MOSFET (FDMA410NZ, ON Semiconductor), 2 board-mount temperature sensors (PTS060301B100RP100 from Vishay), ARM Microcontrollers (CY8C5888LTI-LP097, Cypress Semiconductor), Photodiodes (SFH 2201, Opto Semiconductor).

fig. S6 Device assembly.

992 (A) Top part (B) Bottom part of the device are assembled using 10 screws, 2
 993 compression springs, and 2 torsion springs (McMaster-Carr, Elmhurst, IL) including:
 994 Torsion Spring 90 Degree Angle, Left-Hand Wound, 0.204" OD (9271K109), Torsion
 995 Spring 90 Degree Angle, Right-Hand Wound, 0.204" OD (9271K143), Black-Oxide Alloy
 996 Steel Socket Head Screw M3 x 0.5 mm Thread, 5 mm Long (91290A110), Black-Oxide
 997 Alloy Steel Socket Head Screw M3 x 0.5 mm Thread, 10 mm Long (91290A115), Black-
 998 Oxide Alloy Steel Socket Head Screw M3 x 0.5 mm Thread, 15 mm Long (91290A572),
 999 Black-Oxide Alloy Steel Socket Head Screw M3 x 0.5 mm Thread, 20 mm Long
 1000 (91290A123), Black-Oxide Alloy Steel Socket Head Screw M3 x 0.5 mm Thread, 25 mm
 1001 Long (91290A125), Black-Oxide Alloy Steel Socket Head Screw M3 x 0.5 mm Thread,
 1002 30 mm Long, Partially Threaded (91290A130), Music-Wire Steel Compression Springs
 1003 0.875" Long, 0.36" OD, 0.296" ID (9434K73).

1004

1005

1006 **video S1**

1007 Demonstration of transfer pipette workflow incorporated with Harmony COVID-19

1008 **video S2**

1009 Demonstration of the integrated sampling system workflow with Harmony COVID-19

1010 Both can be played from this slide presentation:

1011

1012

1013

1014 **table S1 Primers, probes, and control sequences.**

1015 For primers and probes F2/B2 sequences are underlined, non-template linker

1016 sequences are italicized, and adapter sequences are in bold.

1017

SARS-Cov-2 NC1	Sequence (5' to 3')
Primers	
NC1 FIP	CCACTGCGTTCTCCATTCTTTT <u>CCCCGCATTACGTTTGGT</u>
NC1 BIP	GCGATCAAAACAACGTCGGTTATTGCCATGTTGAGTGAGAGCG
NC1 LF	TGGTTACTGCCAGTTGAATCT
NC1 LB + T Adapter	ACCAACACCTCACATCACACATAATAGGTTTACCCAATAATACTG CGTCTTG
NC1 F3	TGGACCCCAAAATCAGCG
NC1 B3	ATCTGGACTGCTATTGGTGTTA
SARS-CoV-2 NC2	Sequence
Primer	
NC2 FIP	CAGCTTCTGGCCCAGTTCCTGTGGTGGTGACGGTAAATG
NC2 BIP	CTTCCCTATGGTGCTAACAAAGTCCAATGTGATCTTTTGGTGTATT CA
NC2 LF	GTAGTAGAAATACCATCTTGGACT
NC2 LB + T Adapter	ACCAACACCTCACATCACACATAATAATATGGGTTGCAACTGAG GGAG
NC2 F3	CTACTACCGAAGAGCTACCAG
NC2 B3	GCAGCATTGTTAGCAGGATTG
SARS-CoV-2 NC3	Sequence
Primers	
NC3 FIP	TGTGTAGGTCAACCACGTTCTGCTTCAGCGTTCTTCGGA

NC3 BIP	GTGCCATCAAATTGGATGACAAAG <u>GTTTTGTATGCGTCAATATGCT</u> <u>TATTCAG</u>
NC3 LF + T Adapter	ACCAACACCTCACATCACACATAATATCCATGCCAATGCGCGAC A
NC3 LB	CCAAATTTCAAAGATCAAGTCAT
NC3 F3	GACCAGGAACTAATCAGACAAG
NC3 B3	GCTTGAGTTTCATCAGCCTTC
IAC (NC1) primer	Sequence
IAC FL + C Adapter	ACCACACCTACCACCACTAATAACTAACTCCAGCCATCCTCACC ATC
Target UDP	Sequence
CoV UDP Probe	FITC – CCATCAGCACCAAGACTACCCACCTCGCCACCAA ACCAACACCT CACATCACACATAATA
CoV UDP Quencher	TTGGTGGCGAGGTGGGTAGTCTTGGTGCTGATGG– Iowa Black® FQ
Control UDP	Sequence
IAC UDP Probe	Tex615 – CCTGACCACTTCCGAACCCAACCACCTACGACAG ACCACACCTA CCACCACTAATAACTAA
IAC UDP Quencher	CTGTCGTAGGTGGTTGGGTTTCGGAAGTGGTCAGG – BHQ2
IAC Template	Sequence
IAC (NC1) ssDNA	AAT GGA CCC CAA AAT CAG CGA AAT GCA CCC CGC ATT ACG TTT GGT GGA CCC TCT GGA GTC AAT GGG TGG TGC CAG AAT GGA GAA CGC AGT GGG GCG CGA TCA AAA CAA CGT CGG CCC CAA GTT GAT

CTC CAG CCA TCC TCA CCA TCG TTC ACC GCT CTC ACT CAA CAT
GGC AAG AAT TAA CAC CAA TAG CAG TCC AGA TG

1018

1019

1020 **Sample transfer methods at the POC**

1021 Volumetric transfer pipettes are used in POC tests, but we have not identified a unified
1022 dispenser used in commercial pathogen tests. A unified dispenser integrating the swab
1023 and the buffer container in a single unit could reduce opportunities for sample mix-up
1024 when multiple samples are processed simultaneously. A simple workflow is crucial,
1025 especially in busy settings like clinics. However, in our hands, the in-house built
1026 dispenser led to variable dispensed fluid volumes and had a higher failure rate than
1027 those of the transfer pipette method. While the unified dispenser has many attractive
1028 features, we would not recommend using this in-house assembled dispenser unit until
1029 the method has been optimized to achieve a more accurately dispensed volume.

1030

1031 **table S2 Feedback from the HCWs**

1032 Among HCWs, 30% (3/10) reported problems dispensing the fluid using the unified
1033 system. However, 20% (2/10) of HCWs agreed that the unified dispenser offered an
1034 advantage in its similarity to other tools used in healthcare settings, and 30% (3/10)
1035 HCWs were concerned about contaminating the sample or the environment with the
1036 transfer pipette compared to the unified dispenser system. Only 1/10 (10%) HCW
1037 preferred the smallest tube (0.2mL) for either method, and 5/10 (50%) reported that
1038 larger (1.5mL) tubes were helpful for the unified system, while 40% (4/10) of HCWs said

receptacle size did not make a difference when using the transfer pipette, and none reported preference for tube size when using the unified dispenser system. HCWs reported higher confidence in correctly completing the second kit compared to the first kit of each method, indicating a similar learning curve.

(A) User preference for reaction tube size

Note: Cells with 0% response were left empty to aid data visualization.

Question	User response, n (%)				
	0.2 mL	0.5 mL	1.5 mL	0.5mL or 1.5mL	no difference
Which tube size were you most confident using with the unified dispenser swab?		4 (40%)	5 (50%)	1 (10%)	
Which tube size were you most confident using with the transfer pipette?	1 (10%)	4 (40%)	1 (10%)		4 (40%)

(B) User responses to survey questions during sample transfer usability testing.

Users were asked to provide their response on a Likert scale: (1) Not at all, (2) Slightly, (3) Somewhat, (4) Fairly, (5) Completely. Note: Phrasing of the questions below was slightly modified to be clear outside of the context of the written protocol/survey ([original survey here](#)). Cells with 0% response were left empty to aid data visualization.

1054

Question	User response, n (%)				
	(1)	(2)	(3)	(4)	(5)
How confident were you that you added the correct amount of liquid to each reaction tube using the unified dispenser system ?			3 (30%)	3 (30%)	4 (40%)
How confident were you that you completed the protocol correctly using the unified dispenser system with the first kit ?		2 (20%)	1 (10%)	4 (40%)	3 (30%)
How confident were you that you completed the protocol correctly using the unified dispenser system with the second kit ?				5(50%)	5 (50%)
How confident were you that you added the correct amount of liquid to each reaction tube using the transfer pipette ?			2 (20%)	2(20%)	6 (60%)
How confident were you that you completed the protocol correctly using the transfer pipette with the first kit ?			3 (30%)	4 (40%)	3 (30%)
How confident were you that you completed the protocol correctly				5 (50%)	5 (50%)

using the **transfer pipette** with the
second kit?

How confident were you that you 1 (10%) 4 (40%) 5 (50%)
completed the **on-screen**
instructions correctly?

1055
1056 **(C) User reported errors/challenges between two sample transfer methods.** Cells
1057 with no response were left empty to aid data visualization.
1058

Reported error	Sample transfer method	
	Number of occurrences for the unified dispensing system	Number of occurrences for the transfer pipette
The transfer method was hard to aim at the receiving reaction tube.	1	
Transfer method dripped/leaked before dispensing to the reaction tube.	2	
The materials were too small or difficult to handle.		3
It was hard to execute specific instructions due to unfamiliarity with the sample transfer device.	2	1
Challenging to dispense a consistent amount of fluid.	3	

Not confident liquid was dispensed

to the tube.

2

Concerned about contaminating

samples or the environment.

3

table S3 Detailed reagent and device cost at a production scale of 10,000 units

Item	Supplier	Cost/device (US\$)
Detector boards	Macrofab	\$7 each x 2 (\$14 total)
Main board	Macrofab	\$25 each
LED board	Macrofab	\$9 each
Lid Heater	Oshpark	\$7 each
Heat block	Bryan Willman	\$12 each
Housing	Xometry – HP MultiJet Fusion 3D printing, no volume discount	\$50 (costs here should be reduced for scale-up production)
Red filters	Newport	\$10 each, \$40 in total
Green filters	Newport	\$20 each, \$80 in total
Assembly	N/A	\$30 (half an hour assembly time at \$60/hour cost)
Total		\$267 per unit

table S4 Consumable costs per test

Item	Supplier	Costs per kit (US\$)
------	----------	----------------------

Sampling components (\$2 maximum)	Multiple sources	\$2
Desiccant	Multiple sources	\$0.5
PCR tube	Multiple sources	Negligible
DNA polymerase	Produced in-house	Negligible
Reverse transcriptase	New England Biolabs	\$3.5
Thermostable inorganic pyrophosphatase	New England Biolabs	\$0.23
Primers	Integrated DNA Technologies	\$0.03
Fluorescent probe/quenchers	Integrated DNA Technologies	\$0.24
Triton-X100	Sigma-Aldrich	Negligible
dNTPs	New England Biolabs	\$0.38
Mannitol	OPS Diagnostics	Negligible
DL-Dithiothreitol (DTT)	Promega	Negligible
Nuclease-free water	VWR	Negligible
1X Tris low-EDTA buffer	VWR	Negligible
Rnasin ribonuclease inhibitor	Promega	\$0.71
DNA internal amplification control	Integrated DNA Technologies	\$0.01
Cost of goods		\$8.00

1064

1065