1 Evaluation of commercial anti-SARS-CoV-2 antibody assays and comparison of
2 standardized titers in vaccinated healthcare workers.

3

4 Kahina Saker¹, Vanessa Escuret^{1,4}, Virginie Pitiot², Amélie Massardier-Pilonchéry², Stéphane
 5 Paul³, Bouchra Mokdad¹, Carole Langlois-Jacques⁵, Muriel Rabilloud⁵, David Goncalves⁶,
 6 Nicole Fabien⁶, Nicolas Guibert², Jean-Baptiste Fassier², Antonin Bal¹, Sophie Trouillet-
 7 Assant^{1,4}, Mary-Anne Trabaud¹.

8

9 ¹Laboratoire de Virologie, Institut des Agents Infectieux, Laboratoire associé au Centre
 10 National de Référence des virus des infections respiratoires, Hospices Civils de Lyon, IAI,
 11 Centre de Biologie Nord, Groupement Hospitalier Nord, F-69317, Lyon Cedex 04, France

12 ²Occupational Health and Medicine Department, Hospices Civils de Lyon, Lyon, France

13 ³Laboratory of Immunology and Immunomonitoring, CIC 1408 INSERM, GIMAP EA3064,
 14 University Hospital of Saint-Etienne, F-42055, Saint-Etienne cedex 2, France

15 ⁴CIRI- International Center of Research in Infectiology, INSERM U1111, CNRS UMR5308,
 16 ENS Lyon, Claude Bernard Lyon 1 University, F-69008, Lyon, France.

17 ⁵CNRS, UMR 5558, University of Lyon, Laboratoire de Biométrie et Biologie Evolutive,
 18 Equipe Biostatistique-Santé, F-69100, Villeurbanne, France

19 ⁶Immunology Department, Lyon Sud Hospital, Hospices Civils de Lyon, F-69495, Pierre-
 20 Bénite cedex, France

21 Corresponding author

22 Mary-Anne Trabaud, mary-anne.trabaud01@chu-lyon.fr

23 **Running title:** Comparison of SARS-CoV-2 antibody titers after vaccination

24 **Keywords:** SARS-CoV-2 antibodies; quantification; commercial assays; standardized titers;
 25 vaccination; health care workers

26 Abstract

27 With the availability of vaccines, commercial assays detecting anti-SARS-CoV-2 antibodies
28 (Ab) evolved towards quantitative assays directed to the spike glycoprotein or its receptor
29 binding domain (RBD). The main objective of the present study was to compare the Ab titers
30 obtained with quantitative commercial binding Ab assays, after 1 dose (convalescent
31 individuals) or 2 doses (naïve individuals) of vaccine, in healthcare workers (HCW).

32 Antibody titers were measured in 255 sera (from 150 HCW) with 5 quantitative
33 immunoassays (Abbott RBD IgG II quant, bioMérieux RBD IgG, DiaSorin Trimeric spike
34 IgG, Siemens Healthineers RBD IgG, Wantai RBD IgG). One qualitative total antibody anti
35 RBD detection assay (Wantai) was used to detect previous infection before vaccination. The
36 results are presented in binding Ab units (BAU)/mL after application, when possible, of a
37 conversion factor provided by the manufacturers and established from a World Health
38 Organization (WHO) internal standard.

39 There was a 100% seroconversion with all assays evaluated after two doses of vaccine. With
40 assays allowing BAU/ml correction, Ab titers were correlated (Pearson correlation
41 coefficient, ρ , range: 0.85-0.94). The titer differences varied by a mean of 10.6% between
42 Siemens and bioMérieux assays to 60.9% between Abbott and DiaSorin assays. These results
43 underline the importance of BAU conversion for the comparison of Ab titer obtained with the
44 different quantitative assays. However, significant differences persist, notably, between kits
45 detecting Ab against the different antigens.

46 A true standardization of the assays would be to include the International Standard in the
47 calibration of each assays to express the results in IU/mL.

48

49 **Introduction**

50 Since the end of 2020, severe acute respiratory syndrome coronavirus-2 (SARS-CoV-2)
 51 vaccines have become available worldwide with the aim of achieving herd immunity to
 52 control the pandemic. Vaccine immunity involves both cellular and humoral pathways.
 53 Cellular immunity is not easy to assess on a large scale, as is the neutralizing humoral
 54 response owing to requirement for a biosafety level 3 (BSL3) containment laboratory. The
 55 evaluation of vaccine effectiveness therefore mainly relies on high throughput serological
 56 tests to assess individual humoral immunity as well as monitoring SARS-CoV-2
 57 seroprevalence (1).

58 To effectively utilize measurements of binding antibodies (Ab) as indicators of vaccine
 59 effectiveness, several conditions must be met. First, binding Ab assays should be quantitative;
 60 second, titers should be consistent between different assays; third, binding Ab titers should
 61 correlate with neutralizing Ab titers; fourth, the minimum binding Ab titer associated with
 62 virus neutralization must be found; and fifth, the association between neutralizing Ab and
 63 vaccine protection must be demonstrated. It can be considered that the first and fifth
 64 conditions have been met given that commercial tests for the quantitative detection of binding
 65 Ab have been developed (2–8), and that the role of neutralizing Ab in the infection protection
 66 have been demonstrated in animals and humans (9-12). This is not the case for the other
 67 conditions; in particular, the second point is of importance for widespread evaluation of
 68 vaccines, but until now, Ab titers were often expressed as an index or unit with regard to an
 69 internal standard that differs between manufacturers. Recently, the World Health
 70 Organization (WHO) has developed an international standard (13) against which each
 71 supplier can standardize their assay, allowing comparability of titers between kits. The present
 72 study was conducted to evaluate the performance of commercial antibody assays in detecting
 73 vaccination-associated anti-SARS-CoV-2 Ab seroconversion; the main objective was to

74 compare Ab titers from quantitative assays after conversion of titers with the conversion
75 factor obtained using the WHO standard and provided by each manufacturer.

76

77 **Materials and Methods**

78 **Antibody binding assays**

79 Six CE-marked Ab binding assays, validated by each manufacturer, were investigated
80 according to the protocol recommended by each manufacturer (the characteristics of the
81 assays are summarized in Table 1). Five were quantitative: Siemens Healthineers (Erlangen,
82 Germany) Atellica® IM SARS-CoV-2 IgG (sCOVG; used in routine in our laboratory),
83 DiaSorin (Saluggia, Italy) Liaison® SARS-CoV-2 TrimericS IgG, bioMérieux (Marcy
84 l'Etoile, France) Vidas® SARS-CoV-2 IgG (clinically used for confirmation if necessary),
85 Abbott (Abbott park, IL, USA) Architect SARS-CoV2 IgG II Quant, and Wantai (Beijing,
86 China) SARS-CoV-2 IgG assays. The Wantai SARS-CoV-2 total antibody assay is qualitative
87 and was selected to detect a previous infection before vaccination based on its better
88 sensitivity in infected individuals compared to other commercial qualitative assays we have
89 evaluated in a previous study (14). The First International Standard developed by the WHO
90 (National Institute for Biological Standards and Control code: 20/136) corresponds to
91 lyophilized pooled plasma from patients who had been infected with SARS-CoV-2; after
92 reconstitution, the solution contains 1000 binding antibody units (BAU) per mL (13).
93 For conversion of titers obtained using the quantitative assays, the concentrations expressed in
94 arbitrary units per mL, or index according to the assay (Table 1), were converted to BAU/mL
95 using the conversion factors provided a posteriori (not included in the main procedure but as a
96 separate document – either by electronic or postal mail) by the manufacturer (with the
97 exception of the Wantai SARS-CoV-2 IgG assay for which the conversion factor was not
98 available and presented here only to compare the positivity rate between assays); these were

21.8 for the Siemens assay, 2.6 for the DiaSorin assay, 20.33 for the bioMérieux assay, and 0.142 for the Abbott assay (considering that 1 BAU/mL = conversion factor x AU/mL or index). Samples with results above the upper limit of quantification were tested again after dilution (1/5 when above 3270 BAU/mL for the Siemens assay, 1/20 when above 2080 BAU/mL for the DiaSorin assay, 1/20 when above 18 index for the bioMérieux assay, and 1/2 when above 5680 BAU/mL for the Abbott assay).

Samples from the study population

A prospective longitudinal cohort study was conducted at the laboratory associated with the national reference center for respiratory viruses (University Hospital of Lyon, France). Healthcare workers, excluding pregnant women (HCW; n=150) who were scheduled to receive 2 doses of Pfizer BioNtech vaccine (n=94; BNT162b2/BNT162b2; 78% female; median age 48.5 [range: 21-76] years) or 1 dose of AstraZeneca vaccine followed by 1 dose of Pfizer BioNtech vaccine (n=56; ChAdOx1/BNT162b2; 70% female; median age 33.5 [range: 21-55] years) were included. Blood samples were collected i) before the first dose of vaccine, ii) before the second injection of vaccine corresponding to 4 weeks after the first dose for participants vaccinated with 2 doses of Pfizer BioNtech vaccine (median delay of 28 [range: 21-37] days) or 12 weeks for those vaccinated with the AstraZeneca vaccine (median delay of 85 [range: 84-97] days), and iii) 4 weeks after the full vaccination. The pre-vaccination blood sample was used to document a previous SARS-CoV-2 infection. Among the participants, 26 who were previously infected with SARS-CoV-2 (convalescent group; 17.4%) had only one vaccine injection (Pfizer BioNtech, n=15 or Astra Zeneca, n=11); for these the second sample was omitted. Three participants were infected with SARS-CoV-2 between the 2 doses. Serum samples were prepared from 5 mL of whole blood collected in BD Vacutainer® Serum Separator Tubes II Advance (Beckon Dickinson Diagnostics). After collection, tubes were shaken gently and serum allowed to clot for a minimum 30 min at room

temperature to obtain total coagulation, followed by centrifugation at 2,000 g for 10 min. Serum removed from gel was stored at -80°C until serological assays were performed. Written informed consent was obtained from all participants; ethics approval was obtained from the regional review board for biomedical research in April 2020 (*Comité de Protection des Personnes Sud Méditerranée I*, Marseille, France; ID RCB 2020-A00932-37), and the study was registered on ClinicalTrials.gov (NCT04341142).

Statistical analyses

Results were expressed by the median and interquartile range [IQR]. Paired comparison between assays was performed using the Wilcoxon test. The correlation between concentrations obtained by each assay was investigated using Pearson correlation coefficients and 95% confidence interval (CI). To estimate proportional bias between two methods, Passing and Bablok regression was used and the regression line equation was calculated from the two data sets. The Bland-Altman method was used to measure the mean difference and 95% limit of agreement between log-transformed concentrations obtained with each assay. Statistical analyses were conducted using GraphPad Prism® software (version 8; GraphPad software, La Jolla, CA, USA). A p-value <0.05 was considered statistically significant.

Results

In the first part of the study, the performances of the six assays were compared to verify whether the ability to detect anti-SARS-CoV-2 antibodies of the Wantai total Ab assay, previously found as the most sensitive post-infection compared to other commercial qualitative assays (14), was similar after vaccination, and whether the sensitivity of qualitative and quantitative assays were also similar. The sera collected from patients scheduled to receive only Pfizer BioNtech vaccine (two doses 3 to 4 weeks apart, n=79) were used for this

evaluation. Four weeks after the first injection, the proportion of positive samples was over 90% for all assays except the Wantai assay detecting total antibodies (86.1%). Only one sample was negative with all assays. Six samples were negative with only one assay (5 with the Wantai total Ig assay, 1 with the Wantai IgG assay). Four samples were negative with 3 to 5 assays. In all of these cases the signals from positive assays were low. Four weeks after the second injection anti SARS CoV-2 antibodies were detected for all participants with all the assays (Table 1).

The second part of this study was to compare Ab titers after BAU/mL conversion; for this, only the assays adapted or developed for the quantification of Ab were tested. The Siemens, DiaSorin, bioMérieux and Abbott assays were compared using sera samples collected before the second injection of vaccine and those collected after full vaccination, and for which assays gave a positive quantitative result (255 samples). The median [IQR] values obtained were 744.9 [108.1; 2482] BAU/mL for the Siemens, 1240 [262.6; 3370] BAU/mL for the DiaSorin, 951.4 [142.5; 2314] BAU/mL for the bioMérieux, and 768.9 [100.8; 1916] BAU/mL for the Abbott assays; there was a significant difference in median titers between DiaSorin and Siemens ($p < 0.0001$), between DiaSorin and bioMérieux ($p < 0.0001$), as well as between Abbott and each assay ($p < 0.0001$). The difference in median titers between Siemens and bioMérieux assays was not significant (Figure 1). There was a strong correlation between assays (Table 2).

The slopes of the Passing and Bablock regression curves were calculated by pairwise comparison of each assay. The deviation from the perfect correlation was significantly greater between DiaSorin and each of the other assays than between each of these other assays. In most cases, there was a proportional difference between assays (Figure 2). In particular, DiaSorin assay values reached higher levels than the other methods; 11 samples had a value

above 10,000 BAU/ml while only three reached this threshold with the Siemens assay and none with the BioMérieux and Abbott assays.

According to the Bland-Altman method there was a mean [95%CI] difference in titers expressed as BAU concentrations of 50.9% [46.4%; 55.3%] between the Siemens and DiaSorin assays, 40.4% [35.5%; 45.3%] between bioMérieux and DiaSorin assays, 60.9% [55.9%; 66%] between Abbott and DiaSorin assays, 10.6% [5.8%; 15.4%] between Siemens and bioMérieux assays, 12.1% [16.2%; 8%] between Siemens and Abbott assays, and 22.3% [17.5%; 27.1%] between Abbott and bioMérieux assays (Figure 3).

Discussion

In the present cohort of vaccinated HCW, the performance of qualitative serological assays, developed for diagnostic purposes at the beginning of the pandemic, as well as those adapted or developed when the vaccine became available, were similar. However, assays developed for detecting a past infection are not useful to monitor vaccination effectiveness since they are not quantitative or could not be compared to others. The more important finding of the present study was that the quantitative assays, whose results could be standardized to BAU/ml, produce results that are correlated to each other. The Ab titers obtained with the DiaSorin assay, although correlated with those of the other assays, remained higher after conversion using the WHO standard. This may be related to difference in the antigens targeted, as the DiaSorin assay is the only one evaluated in this study, to use the trimeric spike protein (2). The WHO standard was obtained from the plasma of convalescent patients (13) and must therefore contain antibodies against numerous epitopes, thus the DiaSorin assay is probably capable of reacting with more antibodies than assays detecting only antibodies specific for RBD.

There are currently very few reports that have examined Ab binding assays with the use of the WHO standard (3, 7). Perkman et al. (7) compared four assays detecting binding Ab, including the Liaison anti SARS-CoV-2 trimeric S IgG from DiaSorin and the Architect anti SARS-CoV-2 RBD IgG from Abbott, used in the present study. In vaccinated individuals, after the first dose of vaccine, titers varied significantly between the assays, but they indicated that the recalculation in BAU/ml with the conversion factor given by the manufacturer did not solve error problems between tests. However, the assays compared were more different in their format than those investigated herein: total anti-RBD Ig versus anti-RBD IgG, anti-monomeric spike IgG or anti-trimeric spike IgG. Interestingly, comparison of Abbott and DiaSorin assays found, as was the case herein, higher titers for the DiaSorin assay. Later, Bradley et al. (3) performed linear regressions from sample dilutions of the WHO standard to determine a detection limit in international units (IU) per ml for each test; the authors confirmed the linearity of the Abbott anti-SARS-CoV-2 IgG II Quant assay over the analytical measurement interval but the conversion factor found seemed to be higher than indicated by the manufacturer (1 IU/ml gave 6,1 arbitrary units/ml while Abbott indicated that 1 IU correspond to 7.1 arbitrary units). Taken together, these data suggest that there are remaining differences between assays after conversion in BAU/ml and that this could be due to the incorrect adjustment of the correction factor by the manufacturers. The next step for a true harmonization would therefore be to use the international standard to calibrate each assay instead of applying a conversion factor to a result obtained with an assay previously calibrated with an internal standard (15).

A limitation of this study is the absence of specificity analysis. However, assay specificity analyses have been performed by manufacturers and independent groups (2, 4–6) showing specificity $\geq 99\%$ for all the quantitative assays. A point that may be also considered as a limitation is the choice to use the conversion factors obtained by the manufacturers without

testing the WHO standard ourselves; but the aim of the present study was to evaluate these assays under the conditions offered by the manufacturers to all their customers. In addition, not all commercial quantitative anti-SARS-CoV-2 Ab assays were evaluated, limiting the scope of the conclusions. Furthermore, neutralizing Ab were not investigated that could have helped determine whether anti-RBD or anti-spike assays are the most correlated with virus neutralization. However, before investigating this, harmonization of neutralizing Ab titers is also necessary to determine a common threshold from which vaccine protection could be predicted, allowing then to find the corresponding threshold with high throughput binding Ab assays. A study comparing different cell-based assays (with either live or pseudotyped viruses) to measure neutralization in vitro is rather reassuring, although differences were found according to the viruses used for pseudotyping (16). However, comparison of cell-based assays with surrogate virus neutralization tests (sVNT) that are based on ELISA, and measuring the competition of Ab and RBD for the binding to ACE, the cellular entry receptor of the virus, did not find good agreement; this is inconvenient, as these assays could be promising given that they have potential for large-scale.

In conclusion, the evaluated assays correlated well with each other but a difference in titers remained after adjustment to the same International Standard. Thus, the titer harmonization is not yet completely achieved, but it is better between assays detecting the same Ab against the same antigen than between assays with different targets.

Acknowledgements

This study was supported by Hospices Civils de Lyon and Fondation des Hospices Civils de Lyon. The respective suppliers kindly provided all the serological kits used in this study. We thank all the staff members of the occupational health and medicine department of the Hospices Civils de Lyon who contributed to the sample collection. We thank all Clinical

Research Associates, for their excellent work. We thank Karima Brahima and all the members of the clinical research and innovation department for their reactivity (DRCI, Hospices Civils de Lyon). Human biological samples and associated data were obtained from NeuroBioTec (CRB HCL, Lyon France, Biobank BB-0033-00046). We thank all the technicians from the virology laboratory of the Hospices Civils de Lyon who performed the assays on the automated platforms. Lastly, we thank all the healthcare workers for their participation in this clinical study, and Philip Robinson for critical reading of the manuscript and language revision.

References

1. Shi A. C., Ren P. SARS-CoV-2 Serology Testing: Progress and Challenges. 2021. J Immunol Methods 494 113060. <https://doi.org/10.1016/j.jim.2021.113060>
2. Bonelli F., Blocki F. A., Bunnell T., Chu E., De La O A., Grenache D. G., Marzucchi G., Montomoli E., Okoye L., Pallavicini L., Strevia V. A., Torelli A., Wagner A., Zanin D., Zierold C., Wassenberg J. J. Evaluation of the Automated LIAISON® SARS-CoV-2 TrimericS IgG Assay for the Detection of Circulating Antibodies. 2021. Clin Chem Lab Med 59 (8) 1463–1467. <https://doi.org/10.1515/cclm-2021-0023>.
3. Bradley B. T., Bryan A., Fink S. L., Goecker E. A., Roychoudhury P., Huang M.-L., Zhu H., Chaudhary A., Madarampalli B., Lu J. Y. C., Strand K., Whimbey E., Bryson-Cahn C., Schippers A., Mani N. S., Pepper G., Jerome K. R., Morishima C., Coombs R. W., Wener M., Cohen S., Greninger A. L. Anti-SARS-CoV-2 Antibody Levels Measured by the Advise Dx SARS-CoV-2 Assay Are Concordant with Previously Available Serologic Assays but Are Not Fully Predictive of Sterilizing Immunity. 2021. J Clin Microbiol JCM0098921. <https://doi.org/10.1128/JCM.00989-21>.

4. English E., Cook L. E., Piec I., Dervisevic S., Fraser W. D., John W. G. Performance of the Abbott SARS-CoV-2 IgG II Quantitative Antibody Assay Including the New Variants of Concern (VOC 202012/V1 (UK) and VOC 202012/V2 (South Africa)): And First Steps towards Global Harmonization of COVID-19 Antibody Methods. 2021. J Clin Microbiol JCM0028821 <https://doi.org/10.1128/JCM.00288-21>.
5. Irsara C., Egger A. E., Prokop W., Nairz M., Loacker L., Sahanic S., Pizzini A., Sonnweber T., Holzer B., Mayer W., Schennach H., Loeffler-Ragg J., Bellmann-Weiler R., Hartmann B., Tancevski I., Weiss G., Binder C. J., Anliker M., Griesmacher A., Hoermann G. Clinical Validation of the Siemens Quantitative SARS-CoV-2 Spike IgG Assay (SCOVG) Reveals Improved Sensitivity and a Good Correlation with Virus Neutralization Titers. 2021. Clin Chem Lab Med 59 (8) 1453–1462. <https://doi.org/10.1515/cclm-2021-0214>.
6. Narasimhan M., Mahimainathan L., Araj E., Clark A. E., Markantonis J., Green A., Xu J., SoRelle J. A., Alexis C., Fankhauser K., Parikh H., Wilkinson K., Reczek A., Kopplin N., Yekkaluri S., Balani J., Thomas A., Singal A. G., Sarode R., Muthukumar A. Clinical Evaluation of the Abbott Alinity SARS-CoV-2 Spike-Specific Quantitative IgG and IgM Assays among Infected Recovered and Vaccinated Groups. 2021. J Clin Microbiol 59 (7) e0038821. <https://doi.org/10.1128/JCM.00388-21>.
7. Perkmann T., Perkmann-Nagele N., Koller T., Mucher P., Radakovics A., Marculescu R., Wolzt M., Wagner O. F., Binder C. J., Haslacher H. Anti-Spike Protein Assays to Determine SARS-CoV-2 Antibody Levels: A Head-to-Head Comparison of Five Quantitative Assays. 2021. Microbiol Spectr 9 (1) e0024721. <https://doi.org/10.1128/Spectrum.00247-21>.
8. Di Meo A, Miller JJ, Fabros A, Brinc D, Hall V, Pinzon N, Ierullo M, Ku T, Ferreira VH, Kumar D, Pasic MD, Kulasingam V. 2021. Evaluation of Three anti-SARS-CoV-2

Serologic Immunoassays for Post-Vaccine Response. J Appl Lab Med.

<https://doi.org/10.1093/jalm/jfab087>.

9. Addetia A., Crawford K. H. D., Dingens A., Zhu H., Roychoudhury P., Huang M.-L.,

Jerome K. R., Bloom J. D., Greninger A. L. Neutralizing Antibodies Correlate with

Protection from SARS-CoV-2 in Humans during a Fishery Vessel Outbreak with a

High Attack Rate. 2020. J Clin Microbiol 58 (11) e02107-20.

<https://doi.org/10.1128/JCM.02107-20>.

10. Letizia A. G., Ge Y., Vangeti S., Goforth C., Weir D. L., Kuzmina N. A., Balinsky C.

A., Chen H. W., Ewing D., Soares-Schanoski A., George M.-C., Graham W. D., Jones

F., Bharaj P., Lizewski R. A., Lizewski S. E., Marayag J., Marjanovic N., Miller C. M.,

Mofsowitz S., Nair V. D., Nunez E., Parent D. M., Porter C. K., Santa Ana E.,

Schilling M., Stadlbauer D., Sugiharto V. A., Termini M., Sun P., Tracy R. P.,

Krammer F., Bukreyev A., Ramos I., Sealfon S. C. SARS-CoV-2 Seropositivity and

Subsequent Infection Risk in Healthy Young Adults: A Prospective Cohort Study.

2021. Lancet Respir Med 9 (7) 712–720.

[https://doi.org/10.1016/S2213-2600\(21\)00158-2](https://doi.org/10.1016/S2213-2600(21)00158-2)

11. McMahan K., Yu J., Mercado N. B., Loos C., Tostanoski L. H., Chandrashekar A., Liu

J., Peter L., Atyeo C., Zhu A., Bondzie E. A., Dagotto G., Gebre M. S., Jacob-Dolan

C., Li Z., Nampanya F., Patel S., Pessaint L., Van Ry A., Blade K., Yalley-Ogunro J.,

Cabus M., Brown R., Cook A., Teow E., Andersen H., Lewis M. G., Lauffenburger D.

A., Alter G., Barouch D. H. Correlates of Protection against SARS-CoV-2 in Rhesus

Macaques. 2021. Nature 590 (7847) 630–634.

<https://doi.org/10.1038/s41586-020-03041-6>

12. Rogers T. F., Zhao F., Huang D., Beutler N., Burns A., He W., Limbo O., Smith C.,

Song G., Woehl J., Yang L., Abbott R. K., Callaghan S., Garcia E., Hurtado J., Parren

- M., Peng L., Ramirez S., Ricketts J., Ricciardi M. J., Rawlings S. A., Wu N. C., Yuan
- M., Smith D. M., Nemazee D., Teijaro J. R., Voss J. E., Wilson I. A., Andrabi R.,
- Briney B., Landaïs E., Sok D., Jardine J. G., Burton D. R. Isolation of Potent SARS-
- CoV-2 Neutralizing Antibodies and Protection from Disease in a Small Animal Model.
2020. *Science* 369 (6506) 956–963. <https://doi.org/10.1126/science.abc7520>.
13. Mattiuzzo G., Bentley E. M., Hassall M., Routley S., Bernasconi V., Kristiansen P.,
- Harvala H., Roberts D., Semple G., Turtle L. C., Openshaw P. J., Baillie K.,
- Investigators C., Nissen-Meyer L. S. H., Brants A. B., Atkinson E., Rigsby P., Padley
- D., Almond N., Rose N. J., Page M. 2020. Establishment of the WHO International
- Standard and Reference Panel for Anti-SARS-CoV-2 Antibody. WHO/BS/2020.2403.
- <https://www.who.int/publications/m/item/WHO-BS-2020.2403>
14. Bal A., Pozzetto B., Traubad M.-A., Escuret V., Rabilloud M., Langlois-Jacques C.,
- Paul A., Guibert N., D'Aubarède-Frieh C., Massardier-Pilonchery A., Fabien N.,
- Goncalves D., Boibieux A., Morfin-Sherpa F., Pitiot V., Gueyffier F., Lina B., Fassier
- J.-B., Trouillet-Assant S., COVID SER Study Group. Evaluation of High-Throughput
- SARS-CoV-2 Serological Assays in a Longitudinal Cohort of Patients with Mild
- COVID-19: Clinical Sensitivity Specificity and Association with Virus Neutralization
- Test. 2021. *Clin Chem* 67 (5) 742–752. <https://doi.org/10.1093/clinchem/hvaa336>.
15. Kristiansen P. A., Page M., Bernasconi V., Mattiuzzo G., Dull P., Makar K., Plotkin S.,
- Knezevic I. WHO International Standard for Anti-SARS-CoV-2 Immunoglobulin.
2021. *The Lancet* 397 (10282) 1347–1348. [https://doi.org/10.1016/S0140-](https://doi.org/10.1016/S0140-6736(21)00527-4)
- [6736\(21\)00527-4](https://doi.org/10.1016/S0140-6736(21)00527-4).
16. Sholukh A. M., Fiore-Gartland A., Ford E. S., Miner M. D., Hou Y. J., Tse L. V.,
- Kaiser H., Zhu H., Lu J., Madarampalli B., Park A., Lempp F. A., St. Germain R.,
- Bossard E. L., Kee J. J., Diem K., Stuart A. B., Rupert P. B., Brock C., Buerger M.,

345 Doll M. K., Randhawa A. K., Stamatatos L., Strong R. K., McLaughlin C., Huang M.-
 346 L., Jerome K. R., Baric R. S., Montefiori D., Corey L. Evaluation of Cell-Based and
 347 Surrogate SARS-CoV-2 Neutralization Assays. 2021. J Clin Microbiol. JCM0052721.
 348 <https://doi.org/10.1128/JCM.00527-21>.
 349
 350

TABLE 1: Performance of six commercial anti SARS-CoV-2 antibody assays.

Manufacturer	Wantai	Wantai	Siemens Healthineers	DiaSorin	bioMérieux	Abbott
platform			Atellica® IM	Liaison®	Vidas®	Architect
SARS-CoV-2 Detected Ab	Total Ab	IgG	IgG	IgG	IgG	IgG
Assay type	ELISA	ELISA	CLIA	CLIA	ELFA	CMIA
Antigen	RBD	RBD	RBD	Trimeric Spike	RBD	RBD
Positive threshold	Index = 1	U/mL = 0	U/mL = 1	AU/mL = 13	Index = 1	AU/mL = 50
Conversion factor (WHO standard)	N/A	N/A	21.8	2.6	20.33	0.142
Positive threshold (BAU/mL)	N/A	N/A	21.8	33.8	20.33	7.1
Sensitivity (%)	94.5	N/A	96.41	99	96.6	100
Specificity (%)	100	N/A	99.90	100	99.9	99.9
Positive samples, n (%)						
4 weeks after first injection (n=79)	68 (86.1%)	77 (97.5%)	74 (93.6%)	76 (96.1%)	75 (94.9%)	79 (100%)

4 weeks after
full
vaccination 94 (100%) 94 (100%) 94 (100%) 94 (100%) 94 (100%) 92/92 *
(n=94) (100%)

Positivity was established according to manufacturers' instructions. Sensitivity and specificity data were those described in the instruction for utilization sheet from each manufacturer. Abbreviations: Ab: antibodies, Ig: immunoglobulin, ELISA: enzyme-linked immunosorbent assay, CMIA: chemiluminescence microparticle immunoassay CLIA: chemiluminescence immunoassay, ELFA: enzyme-linked fluorescent assay, N: number of samples, RBD: Receptor Binding Domain, CI: confidence interval.

* 2 samples did not remain in sufficient quantity to perform the Abbott assay

TABLE 2: Pearson correlation coefficient (ρ [IQR]) between each assay.

ASSAY	Abbott	bioMérieux	DiaSorin
Siemens	0.91 [0.89; 0.93]	0.85 [0.81; 0.88]	0.94 [0.92; 0.95]
DiaSorin	0.90 [0.87; 0.92]	0.86 [0.82; 0.89]	
bioMérieux	0.90 [0.87; 0.92]		

FIG 1: Comparison of anti SARS-CoV-2 antibodies concentration (BAU/mL) between all assays in sera collected from vaccinated subjects. The statistical difference was evaluated by Wilcoxon's test. Comparison of median titers between Siemens, DiaSorin, bioMérieux, and Abbott assays. BAU/mL: Binding Antibodies Unit/mL. ****p<0.0001. Data from patients scheduled to be vaccinated with 2 doses of Pfizer BioNtech vaccine (black) or with 1 dose of AstraZeneca vaccine followed by 1 dose of Pfizer BioNtech (blue) are presented.

FIG 2: Passing and Bablok regression analyses using the Siemens, DiaSorin, bioMérieux, and Abbott assays. A, The Siemens assay compared to the DiaSorin assay. B, bioMérieux compared to DiaSorin. C, Abbott compared to DiaSorin. D, Siemens compared to bioMérieux. E, Siemens compared to Abbott. F, Abbott compared to bioMérieux.

FIG 3: Bland-Altman plots comparing agreement between concentrations determined using the Siemens, DiaSorin, bioMérieux, and Abbott assays. A, The Siemens assay compared to the DiaSorin assay. B, bioMérieux compared to DiaSorin. C, Abbott compared to DiaSorin. D, Siemens compared to bioMérieux. E, Siemens compared to Abbott. F, Abbott compared to bioMérieux. The solid blue line represents the bias between assays, the dashed blue lines represent 95%CI.

A

Slope=0,69

B

Slope=0,66

C

Slope=0,54

D

Slope=0,95

E

Slope=0,80

F

Slope=1,16