Agile design and development of a high throughput cobas® SARS-CoV-2 RT-PCR diagnostic test

Chitra Manohar¹, Jingtao Sun¹, Peter Schlag¹, Chris Santini¹, Marcel Fontecha¹, Pirmin Lötcher², Carolin Bier², Kristina Goepfert², Dana Duncan¹, Gene Spier¹, Daniel Jarem^{1*}, Dmitriy Kosarikov¹

¹ Roche Molecular Systems, Inc., Pleasanton, United States

² Roche Diagnostics International AG, Rotkreuz, Switzerland

*Corresponding author

Global Medical and Scientific Affairs, Roche Molecular Diagnostics

4300 Hacienda Drive, Pleasanton, CA 94588 USA

Tel: +1 (908) 524-4705 Email: daniel.jarem@roche.com

Keywords: SARS-CoV-2, COVID-19, pandemic, RT-PCR, diagnostic testing, molecular assay design

Running Title: cobas SARS-CoV-2 test development

Abstract

Diagnostic testing is essential for management of the COVID-19 pandemic. An agile assay design methodology, optimized for the cobas® 6800/8800 system, was used to develop a dual-target, qualitative SARS-CoV-2 RT-PCR test using commercially available reagents and existing sample processing and thermocycling profiles. The limit of detection was 0.004 to 0.007 TCID₅₀/mL for USA-WA1/2020. Assay sensitivity was confirmed for SARS-CoV-2 variants Alpha, Beta, Gamma, Delta and Kappa. The coefficients of variation of the cycle threshold number (Ct) were between 1.1 and 2.2%. There was no difference in Ct using nasopharyngeal compared to oropharyngeal swabs in universal transport medium (UTM). A small increase in Ct was observed with specimens collected in cobas® PCR medium compared to UTM. In silico analysis indicated that the dual-target test is capable of detecting all >1,800,000 SARS-CoV-2 sequences in the GISAID database. Our agile assay design approach facilitated rapid development and deployment of this SARS-CoV-2 RT-PCR test.

1. Introduction

A novel coronavirus, SARS-CoV-2, is the causative agent of COVID-19, a complex and potentially lethal human disease (1) that has infected over 200 million individuals worldwide as of August 2021 (2). The COVID-19 pandemic has been associated with more than four million deaths (2) and enormous economic impact across the world. SARS-CoV-2 is transmitted person-to-person via respiratory secretions, causing fever, respiratory symptoms (cough, shortness of breath), and subsequent immune system dysregulation. The clinical presentation of COVID-19 can vary from asymptomatic infection to mild illness to fatal disease (3-5).

The *Coronaviridae* is a family of viruses that cause illness ranging from mild respiratory infection (human coronaviruses 229E, NL63, HKU1, and OC43) to more severe diseases such as Middle East Respiratory Syndrome (MERS-CoV) and Severe Acute Respiratory Syndrome (SARS-CoV) (6). SARS-CoV-2 coronavirus belongs to the *Sarbecovirus* sub-genus, which also includes SARS-CoV and other betacoronaviridae identified in bats (7, 8).

Diagnostic testing is an essential component of infection prevention, control and disease management. One of the most sensitive types of diagnostic test currently available is based on specific detection of viral nucleic acids. One commonly used technology platform for such tests is real-time reverse transcription (RT)-PCR, which involves binding of primers and probes to specific regions of the pathogen's genome. Rapid response to the need for testing for novel pathogens can be achieved by adaptation of existing automated instruments, well-established generic reagents and production facilities. The cobas® 6800/8800 system (Roche Molecular Systems) is a widely used platform that supports the detection of many different clinically important viruses and bacteria using real-time PCR (9).

Early in the pandemic only a few SARS-CoV-2 genomic sequences were available in public databases (e.g. GISAID or NCBI). When the cobas® SARS-CoV-2 test was designed in early 2020, little was known about what regions of the genome might be subject to sequence variation and/or recombination. Despite the paucity of knowledge about potential sequence variation, we designed a single well, dual-target assay to detect SARS-CoV-2-specific sequences using targets in the non-structural region of the ORF1a/b locus, and a conserved region in the structural envelope (E)-gene common to all sarbecoviruses, including SARS-CoV-2. The test was designed to meet the need for high throughput testing on the cobas 6800/8800 system, which performs fully automated sample preparation, real time RT-PCR reaction setup, target amplification and detection.

SARS-CoV-2 can evolve in response to external selection pressures. Strong but incomplete inhibition of replication, which might occur in an infected person with partial immunity, can result in the selection of SARS-CoV-2 variants that have higher replicative fitness than the wild-type virus in a population of susceptible hosts. Similarly, if a naturally occurring variant were to arise with increased ability to spread in an immunologically naïve population, it could out-compete the wild-type virus in a relatively short period of time. The emergence of several “variants of concern” (VOC) and “variants of interest” in many different locations of the world in recent months is therefore not unexpected, and has several important public health and clinical implications (10-16). Sequence variation in such variants has the potential to interfere with molecular diagnostic test performance.

2. Methods

2.1 Viruses

An isolate of SARS-CoV-2 from the first patient diagnosed with COVID-19 in the US (USA-WA1/2020, catalog number NR-52281, lot number 70033175, 2.8×10^5 TCID₅₀/mL) (17) was obtained from BEI Resources (Manassas, VA). Based on information provided in the Certificate of Analysis from the vendor, one TCID₅₀/mL is equal to 7393 genome equivalents (RNA copies) by droplet digital PCR™ (Bio-Rad®). An isolate of SARS-CoV-2 from a German patient (BetaCoV/Germany/BavPat1/2020, catalog number 026V-03883, 3.2×10^6 pfu/mL) was obtained from the European Virus Archive Global (Marseille, France). All experiments with replication competent SARS-CoV-2 were performed in a biocontainment level 3 facility in Switzerland. Virus stocks were diluted in a simulated matrix, consisting of human cells and mucin in Universal Viral Transport Medium (UTM, Copan Diagnostics, Murrieta, CA; <https://www.copanusa.com/wp-content/uploads/2020/02/UTM-Package-Insert.pdf>), that was shown to be equivalent to natural nasopharyngeal matrix in assay performance (data not shown). Virus stocks for variants of concern were obtained from BEI Resources (catalog numbers NR-54000, NR-54008, NR-54982, NR-55611, NR-55486, NR-55308, NR-55309, NR-55439, NR-55469, and NR-55654). For the wild-type strain, genomic RNA was used (NR-52499). Virus RNA concentration was determined by droplet digital PCR.

2.2 cobas SARS-CoV-2 assay design

The cobas 6800/8800 platform is an end-to-end system that includes hardware, software, reagents and consumables that performs automated nucleic acid testing. The platform is intended for moderate- to high-throughput laboratories, where a large number of test results are needed within short periods of time. Different tests that are performed on the cobas 6800/8800 system use the same generic reagents and share common sample processing and PCR profiles coupled with target-specific assay oligonucleotides and positive control. Therefore, cobas SARS-CoV-2

was developed using the automated, well established conditions for the cobas 6800/8800 system, sample preparation workflow and thermal cycling profile, allowing for simultaneous inclusion of other diagnostic tests designed for the cobas 6800/8800 platform (amplification/detection on the same PCR plate). All reagents were developed using synergies whenever possible, such as common raw material, manufacturing and use test (kit release) procedures, controls and PCR reaction master mix formulation.

To design primers and probes for the PCR assay, Agile Assay Design (AAD) software (Roche Molecular Systems) was used to select optimal oligonucleotide length and sequence based on the seven available SARS-CoV-2 sequences in the Global Initiative on Sharing All Influenza Data (GISAID, www.gisaid.org) (18). The designs took into consideration key PCR parameters that predict efficient assay performance using Roche master mix and reagents on the cobas 6800/8800 system. To evaluate inclusivity, the delay in cycle threshold (dCt) compared to perfectly matching primers was modeled. The algorithm was based on experiments performed with 20 perfectly matched primers at different locations and 268 corresponding, mis-matched primers containing one to six nucleotide mismatches, using both DNA and RNA templates to experimentally measure dCt. The models use thermodynamics parameters (free energies, dG) for DNA:DNA interactions (19). The melting temperatures for probes were calculated using Melting5 software (20) at 100 nM probe concentration, 1.7 mM Mg^{2+} and 50mM Na^{+}/K^{+} .

2.3 Inclusivity

Inclusivity analysis was performed using all available SARS-CoV-2 sequences in GISAID as of June 15, 2021 (n= 1,874,933). The predicted impact of each variant Target 1 and Target 2 primer and probe binding site sequence (six sites) was evaluated using the AAD software and quantitated as the predicted increase in Ct or probe melting temperature (T_m).

2.4 Sensitivity (Limit of Detection)

To determine the limit of detection (LoD), USA-WA1/2020 live virus was serially diluted in simulated clinical matrix. A total of seven concentrations, generated using 3-fold serial dilutions of the stock virus, were tested, with a total of 21 replicates per concentration and an additional 10 replicates of diluent only. LoD was determined by probit analysis based on the titer given by the supplier and dilution factor. The LoD was confirmed using a second virus isolate, BetaCoV/Munich/BavPat1/2020, similarly serially diluted.

2.5 Precision

Precision was assessed with a panel made using cultured SARS-CoV-2 (USA-WA1/2020, heat-inactivated) in simulated clinical matrix in UTM. SARS-CoV-2 virus stock material was serially diluted to generate a panel consisting of three concentration levels (weak, low and moderate positive) corresponding to approximately 0.3x, 1x and 3x the LoD, respectively. The samples were tested over 15 days, three reagent lots, on three instruments and by three operators. Each test day, two runs were performed per lot and per instrument, using three replicates per panel member per run. A total of 90 replicates per concentration level were tested over the course of the study. The 90 replicates were distributed across three reagent lots (30 replicates each) and three cobas® 6800/8800 Systems.

2.6 Matrix/Collection media equivalency

The relative performance of different specimen types or transport media was evaluated using cultured virus (USA-WA1/2020 strain), spiked into paired specimens from SARS-CoV-2 negative individuals with symptoms of an upper respiratory infection collected in 2018 (before the SARS-CoV-2 outbreak) and stored frozen at -80°C. The final virus concentration was

approximately 0.054 TCID₅₀/mL, or 1.5 times higher than the LoD. The relative performance of nasopharyngeal swab (NPS) and oropharyngeal swab (OPS) specimens was compared in universal viral transport medium (BD, Franklin Lakes, NJ; equivalent to UTM). NPS swabs were used for the comparison between UTM or the virus-inactivating cobas® PCR Media (CPM; Roche Molecular Systems) (21). A total of 21 replicates at 0.054 TCID₅₀/mL were tested. Similarly, detection of SARS-CoV-2 RNA in specimens collected using polyester woven or nylon flocked swabs in UTM or 0.9% saline physiological solution was evaluated using cultured virus (USA-WA1/2020 strain) spiked into matched nasal swab (NS) specimens from SARS-CoV-2 negative individuals. Three specimens were self-collected from each of 45 healthy donors: two using either a woven polyester or nylon flocked swab and placed in UTM, and one (from the other nostril) using a woven polyester swab placed in 0.9% physiological saline. A total of 17 replicates at 0.054 TCID₅₀/mL were tested.

3. Results

3.1 Assay Design

At the time when the cobas SARS-CoV-2 test was designed (January 2020), only seven SARS-CoV-2 sequences were available from the Global Initiative on Sharing All Influenza Data (GISAID). To guard against the possibility that viral sequence changes in the target site may negatively impact detection, two complementary strategies were employed. First, a dual target design was chosen, which dramatically increases the likelihood of viral sequence detection and PCR signal preservation from one of the two targets even if the other has sequence variations. Second, selection of target regions where sequences are conserved between virus species (e.g.

between SARS-CoV-2 and SARS-CoV) greatly enhances the chance that this sequence will remain conserved within species. We selected one target (Target 1) in the ORF1a/b coding region that is highly specific for SARS-CoV-2. The second target (Target 2) is located in a region of the E gene that is conserved among sarbecoviruses (Fig. 1). A proprietary software program, AAD (see Methods), was used to evaluate the seven available SARS-CoV-2 sequences and over 1000 sequences from other sarbecoviruses, including SARS-CoV, that were available from NCBI. Target 2 positivity can be interpreted unambiguously given the clinical and epidemiological context (i.e. with knowledge that SARS-CoV-2 is circulating but SARS-CoV is not).

3.2 Inclusivity

In June 2021, an updated *in silico* analysis of 1,874,933 SARS-CoV-2 sequences in the GISAID database was performed. Table 1 summarizes the predicted impact of sequence variation at primer or probe binding sites for variants represented in at least 0.02% (403 or more) of all SARS-CoV-2 sequences in the GISAID database (a complete listing of all variants including very infrequent sequences can be found in the Supplemental Data Table S1). A summary of the numbers of individual haplotypes (defined as a specific sequence including the four primer and two probe binding sites) is shown in Table 2. Overall, 98.56% of sequences have no changes in primer or probe binding sites at either target, and 1.4% have only a single change. One of these (variant 13, found in 0.025% of sequences), which has a single change in the probe binding site for Target 2 (Table 1), is predicted to reduce the reactivity of single probes or primers, and has been previously shown to be associated with failure of detection at Target 2 (22). An additional 74 (0.004%) sequences, bearing between two and eight nucleotide changes, were observed that are predicted to impact assay performance for target 1 (n=31) or target 2 (n=43; Table 2). No

variant reported in the GISAID database had changes in both target regions simultaneously. Therefore no impact on cobas SARS-CoV-2 test performance is anticipated, considering all the sequences available.

3.3 Sensitivity (Limit of Detection)

Assay sensitivity was determined by replicate testing of serial dilutions of USA-WA1/2020 virus. As shown in Table 3, the concentration level with observed test positivity rates $\geq 95\%$ were 0.009 and 0.003 TCID₅₀/mL for Target 1 and 2, respectively. The probit model predicted 95% test positivity rates at virus titers of 0.007 (95% CI: 0.005 – 0.036) and 0.004 (95% CI: 0.002 – 0.009) TCID₅₀/mL for Target 1 and 2, respectively. Based on an estimate of the number of RNA copies per TCID₅₀ in the virus stock used for this experiment, this corresponds to 52 copies/mL for Target 1 and 30 copies/mL for Target 2.

A second sensitivity study was performed using a different virus isolate (BetaCoV/Germany/BavPat1/2020) whose concentration is reported in plaque forming units (pfu) instead of TCID₅₀. The concentration levels with observed test positivity rates greater than or equal to 95% using this isolate were 0.011 pfu/mL for Target 1 and 0.004 pfu/mL for Target 2 (data not shown). Probit analysis predicted 95% test positivity rates at virus titers of 0.007 pfu/mL (95% CI: 0.005 – 0.023) for Target 1 and 0.004 pfu/mL (95% CI: 0.002 – 0.009) for Target 2.

Assay sensitivity for SARS-CoV-2 variants of concern alpha, beta, gamma, and delta (B.1.1.7, B.1.351, P.1, and B.1.617.2, respectively), variant of interest lambda (C.37), and several variants under monitoring (some of which were formerly variants of interest) was confirmed by testing 5 to 8 replicates of three different dilutions of virus stocks for each variant near the LoD determined above (50 to 250 copies/mL). For Target 1, at the two highest concentrations, all

results were positive for all variants (Table 4). At 50 copies/mL for Target 1, 62.5% (beta), 87.5% (wild-type, alpha and gamma) or 100% (delta, kappa) of results were positive. At all concentrations tested, all results were positive for Target 2.

3.4 Precision

Summary statistics for Ct values for the weak (~0.3x), low (~1.0x), and moderate (~3.0x) positive concentration levels by variance component are shown in Table 5. Coefficients of variation of less than 2.0% CV were observed for all variables and concentration levels. Slightly more variability was observed between reagent lots and in the within-run residual category. Precision values of 0.8% CV or less were observed between instruments, day-to-day, and run-to-run. Overall, the coefficients of variation ranged from 1.1 to 1.9% for Target 1, and from 1.1 to 2.2% for Target 2.

3.5 Matrix/Collection media equivalency

Nasopharyngeal swab (NPS) and oropharyngeal swab (OPS) specimens are suitable for use in the diagnosis of respiratory virus infections. To demonstrate matrix equivalency, cultured virus (USA-WA1/2020 strain) was spiked into paired OPS or NPS specimens from SARS-CoV-2 negative individuals to final concentrations of approximately 0.054 TCID₅₀/mL, or 1.5 times higher than the LoD. The cycle threshold number (Ct, inversely correlated with RNA quantity in the specimen) from replicate tests for each target is shown in Fig. 2. There was no statistically significant difference between mean Ct values in NPS vs OPS for either target (paired t-test *P* value >0.1).

Similar experiments were performed to compare the Ct values in specimens collected in different types of swabs and collection media. Ct values for specimens diluted in CPM were slightly

higher (difference in mean Ct of 0.4 to 0.7) than in Universal Transport Media (UTM) for both targets (paired t-test P value ≤ 0.002 ; Fig. 3A). Specimens collected using nylon flocked swabs in UTM yielded a minimal increase in mean Ct vs. polyester woven swabs in the same medium (difference 0.6 Ct, paired t-test P value 0.0065), while polyester woven swabs in saline yielded similar Ct values compared to the same type of swab in UTM (paired t-test P value > 0.4 ; Fig. 3B).

4. Discussion

Deployment of new diagnostic tests for emerging pathogens in a timely manner is an integral part of the public health response to pandemics, such as the one caused by infection with SARS-CoV-2. Several different factors can be leveraged to speed development and deployment of novel diagnostic assays at the start of a pandemic or any emerging infectious disease. When tests for new targets can be deployed using existing infrastructure central testing laboratories can capitalize on already deployed instruments and trained personnel. The cobas 6800/8800 platform represents such an opportunity, since real-time PCR primers and probes can be designed to work in the context of existing, well-characterized assay chemistry and conditions. The AAD software for primer/probe design is an effective tool that facilitates rapid assay development for new pathogens. In addition, when reference materials are available for preparation of contrived specimens in a variety of authentic and simulated clinical matrices, initial performance evaluations required for emergency use authorization can be completed quickly. The convergence of these features enabled the rapid development of this diagnostic test, which was the first such test granted emergency use authorization in the US in early 2020. The paucity of sequence data at the early stages of an outbreak is a significant challenge for molecular test development (23, 24). The cobas SARS-CoV-2 test primers and probes were

designed at a time when only seven genomic sequences were publicly available. However, sequence conservation may be predicted by considering data from related viruses for which more sequences have been characterized. A region of conserved sequence across different virus species is also likely to remain conserved within a species in the future.

Our approach included two different sets of primers and probes, one of which (Target 1) is specific for ORF1a/1b of SARS-CoV-2, while the other (Target 2) is intended to react with E-gene sequences of SARS-CoV-2, SARS-CoV, and other sarbecoviruses that infect bats. The genes in which the targets should be located were not pre-determined, but instead the AAD approach identified sites anywhere in the genome predicted to provide the best performance and desired level of sequence conservation. While Target 2 reactivity is expected for both SARS-CoV and SARS-CoV-2, virus positivity can be unambiguously established with knowledge of extant virus and disease prevalence, and in combination with the SARS-CoV-2 specific Target 1 result. Importantly, the use of two targets enables test accuracy even in the presence of sequence variation in one of the two target sequences. This has been demonstrated to occur in at least one case (22).

The limit of detection of the cobas SARS-CoV-2 test was determined to be between 0.004 and 0.007 TCID₅₀/mL or between 30 and 52 copies/mL for the particular virus stock used here. It should be noted that the relationship between TCID₅₀ (or plaque forming units) and number of RNA copies may differ between virus preparations; this may explain small differences in LoD reported in RNA copies/mL elsewhere (25). A lower LoD for the E-gene target compared to other targets has been reported previously (26), which is consistent with our results. We noted that at low virus input levels, Target 1 positivity is impacted more than Target 2, in spite of higher Ct values for Target 2.

Previous evaluations have provided conflicting results regarding the relative sensitivity of SARS-CoV-2 RT-PCR assays using NPS, OPS or other specimen types (27-34). Our results support the use of either NPS or OPS as the specimen type, since there was no difference in the ability to detect SARS-CoV-2 RNA spiked into either specimen at low levels. While NPS is viewed as the gold standard for many respiratory pathogens, OPS are easier to obtain and less intrusive for the patient. These findings should be confirmed with specimens from infected individuals in the clinic.

In the early phases of the COVID-19 pandemic, many clinicians experienced a shortage of recommended sample collection materials including media for specimen storage and shipping. Our data indicate equivalent sensitivity of the cobas SARS-CoV-2 test when specimens are stored in UTM, CPM or saline, as long as specimens are refrigerated (2-8 °C) and stored for 6 days or less. CPM has the added advantage of inactivating the infectivity of SARS-CoV-2, thus improving biosafety for specimen handling in the laboratory (21).

Several independent studies have reported on the performance of the cobas SARS-CoV-2 test, in comparison with laboratory-developed tests (LDTs) or other commercial assays. Generally, overall percent agreement values range from 95 to 99% (25, 26, 35-37), with more discordance observed in specimens with low viral loads (35, 36, 38).

In conclusion, the cobas SARS-CoV-2 is a robust, sensitive and specific test for qualitative diagnosis of infection by SARS-CoV-2 that can be performed using equipment and infrastructure already widely available in clinical reference laboratories globally. The rapid development and deployment of this test was made possible by the application of the AAD approach and early availability of sequence information and reference reagents.

Acknowledgements

This publication was supported by the European Virus Archive GLOBAL (EVA-GLOBAL) project that has received funding from the European Union’s Horizon 2020 research and innovation programme under grant agreement No 871029. Manuscript preparation services were provided by Data First Consulting (Sebastopol, CA).

Conflict of Interest

All authors except C.B., K.G. and P.L. are employees and stockholders of Roche Molecular Systems, Inc. C.B., K.G. and P.L. are employees of Roche Diagnostics International AG.

References

1. Zhu N, Zhang D, Wang W, Li X, Yang B, Song J, Zhao X, Huang B, Shi W, Lu R, Niu P, Zhan F, Ma X, Wang D, Xu W, Wu G, Gao GF, Tan W, China Novel Coronavirus I, Research T. 2020. A Novel Coronavirus from Patients with Pneumonia in China, 2019. *N Engl J Med* 382:727-733.
2. Dong E, Du H, Gardner L. 2020. An interactive web-based dashboard to track COVID-19 in real time. *Lancet Infect Dis* 20:533-534.
3. Guan WJ, Ni ZY, Hu Y, Liang WH, Ou CQ, He JX, Liu L, Shan H, Lei CL, Hui DSC, Du B, Li LJ, Zeng G, Yuen KY, Chen RC, Tang CL, Wang T, Chen PY, Xiang J, Li SY, Wang JL, Liang ZJ, Peng YX, Wei L, Liu Y, Hu YH, Peng P, Wang JM, Liu JY, Chen Z, Li G, Zheng ZJ, Qiu SQ, Luo J, Ye CJ, Zhu SY, Zhong NS, China Medical Treatment Expert Group for C. 2020. Clinical Characteristics of Coronavirus Disease 2019 in China. *N Engl J Med* 382:1708-1720.
4. Onder G, Rezza G, Brusaferro S. 2020. Case-Fatality Rate and Characteristics of Patients Dying in Relation to COVID-19 in Italy. *JAMA* 323:1775-1776.
5. Wang D, Hu B, Hu C, Zhu F, Liu X, Zhang J, Wang B, Xiang H, Cheng Z, Xiong Y, Zhao Y, Li Y, Wang X, Peng Z. 2020. Clinical Characteristics of 138 Hospitalized Patients With 2019 Novel Coronavirus-Infected Pneumonia in Wuhan, China. *JAMA* 323:1061-1069.
6. Bchetnia M, Girard C, Duchaine C, Laprise C. 2020. The outbreak of the novel severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2): A review of the current global status. *J Infect Public Health* 13:1601-1610.

7. Boni MF, Lemey P, Jiang X, Lam TT, Perry BW, Castoe TA, Rambaut A, Robertson DL. 2020. Evolutionary origins of the SARS-CoV-2 sarbecovirus lineage responsible for the COVID-19 pandemic. *Nat Microbiol* 5:1408-1417.
8. Lu R, Zhao X, Li J, Niu P, Yang B, Wu H, Wang W, Song H, Huang B, Zhu N, Bi Y, Ma X, Zhan F, Wang L, Hu T, Zhou H, Hu Z, Zhou W, Zhao L, Chen J, Meng Y, Wang J, Lin Y, Yuan J, Xie Z, Ma J, Liu WJ, Wang D, Xu W, Holmes EC, Gao GF, Wu G, Chen W, Shi W, Tan W. 2020. Genomic characterisation and epidemiology of 2019 novel coronavirus: implications for virus origins and receptor binding. *Lancet* 395:565-574.
9. Cobb B, Simon CO, Stramer SL, Body B, Mitchell PS, Reisch N, Stevens W, Carmona S, Katz L, Will S, Liesenfeld O. 2017. The cobas(R) 6800/8800 System: a new era of automation in molecular diagnostics. *Expert Rev Mol Diagn* 17:167-180.
10. Luring AS, Hodcroft EB. 2021. Genetic Variants of SARS-CoV-2-What Do They Mean? *JAMA* 325:529-531.
11. Tegally H, Wilkinson E, Giovanetti M, Iranzadeh A, Fonseca V, Giandhari J, Doolabh D, Pillay S, San EJ, Msomi N, Mlisana K, von Gottberg A, Walaza S, Allam M, Ismail A, Mohale T, Glass AJ, Engelbrecht S, Van Zyl G, Preiser W, Petruccione F, Sigal A, Hardie D, Marais G, Hsiao M, Korsman S, Davies MA, Tyers L, Mudau I, York D, Maslo C, Goedhals D, Abrahams S, Laguda-Akingba O, Alisoltani-Dehkordi A, Godzik A, Wibmer CK, Sewell BT, Lourenco J, Alcantara LCJ, Kosakovsky Pond SL, Weaver S, Martin D, Lessells RJ, Bhiman JN, Williamson C, de Oliveira T. 2021. Emergence of a SARS-CoV-2 variant of concern with mutations in spike glycoprotein. *Nature*.
12. Rambaut A, Loman N, Pybus O, Barclay W, Barrett J, Carabelli A, Connor T, Peacock T, Robertson DL, Volz E, COVID-19 Genomics Consortium UK Preliminary genomic

characterisation of an emergent SARS-CoV-2 lineage in the UK defined by a novel set of spike mutations. [Online.] <https://virological.org/t/preliminary-genomic-characterisation-of-an-emergent-sars-cov-2-lineage-in-the-uk-defined-by-a-novel-set-of-spike-mutations/563>.

Accessed March 17, 2021.

13. Eurosurveillance editorial team. 2021. Updated rapid risk assessment from ECDC on the risk related to the spread of new SARS-CoV-2 variants of concern in the EU/EEA - first update. Euro Surveill 26.

14. Galloway SE, Paul P, MacCannell DR, Johansson MA, Brooks JT, MacNeil A, Slayton RB, Tong S, Silk BJ, Armstrong GL, Biggerstaff M, Dugan VG. 2021. Emergence of SARS-CoV-2 B.1.1.7 Lineage - United States, December 29, 2020-January 12, 2021. MMWR Morb Mortal Wkly Rep 70:95-99.

15. Harvey WT, Carabelli AM, Jackson B, Gupta RK, Thomson EC, Harrison EM, Ludden C, Reeve R, Rambaut A, Consortium C-GU, Peacock SJ, Robertson DL. 2021. SARS-CoV-2 variants, spike mutations and immune escape. Nat Rev Microbiol 19:409-424.

16. Boehm E, Kronig I, Neher RA, Eckerle I, Vetter P, Kaiser L, Geneva Centre for Emerging Viral D. 2021. Novel SARS-CoV-2 variants: the pandemics within the pandemic. Clin Microbiol Infect.

17. Harcourt J, Tamin A, Lu X, Kamili S, Sakthivel SK, Murray J, Queen K, Tao Y, Paden CR, Zhang J, Li Y, Uehara A, Wang H, Goldsmith C, Bullock HA, Wang L, Whitaker B, Lynch B, Gautam R, Schindewolf C, Lokugamage KG, Scharon D, Plante JA, Mirchandani D, Widen SG, Narayanan K, Makino S, Ksiazek TG, Plante KS, Weaver SC, Lindstrom S, Tong S, Menachery VD, Thornburg NJ. 2020. Severe Acute Respiratory Syndrome Coronavirus 2 from Patient with Coronavirus Disease, United States. Emerg Infect Dis 26:1266-1273.

18. Shu Y, McCauley J. 2017. GISAID: Global initiative on sharing all influenza data - from vision to reality. *Euro Surveill* 22:30494.
19. SantaLucia J, Jr., Hicks D. 2004. The thermodynamics of DNA structural motifs. *Annu Rev Biophys Biomol Struct* 33:415-440.
20. Dumousseau M, Rodriguez N, Juty N, Le Novère N. 2012. MELTING, a flexible platform to predict the melting temperatures of nucleic acids. *BMC Bioinformatics* 13:101.
21. Norz D, Frontzek A, Eigner U, Oestereich L, Wichmann D, Kluge S, Fischer N, Aepfelbacher M, Pfefferle S, Lutgehetmann M. 2020. Pushing beyond specifications: Evaluation of linearity and clinical performance of the cobas 6800/8800 SARS-CoV-2 RT-PCR assay for reliable quantification in blood and other materials outside recommendations. *J Clin Virol* 132:104650.
22. Artesi M, Bontems S, Gobbels P, Franckh M, Maes P, Boreux R, Meex C, Melin P, Hayette MP, Bours V, Durkin K. 2020. A Recurrent Mutation at Position 26340 of SARS-CoV-2 Is Associated with Failure of the E Gene Quantitative Reverse Transcription-PCR Utilized in a Commercial Dual-Target Diagnostic Assay. *J Clin Microbiol* 58.
23. Damond F, Roquebert B, Benard A, Collin G, Miceli M, Yeni P, Brun-Vezinet F, Descamps D. 2007. Human immunodeficiency virus type 1 (HIV-1) plasma load discrepancies between the Roche COBAS AMPLICOR HIV-1 MONITOR Version 1.5 and the Roche COBAS AmpliPrep/COBAS TaqMan HIV-1 assays. *J Clin Microbiol* 45:3436-3438.
24. Damond F, Avettand-Fenoel V, Collin G, Roquebert B, Plantier JC, Ganon A, Sizmann D, Babel R, Glaubitz J, Chaix ML, Brun-Vezinet F, Descamps D, Rouzioux C. 2010. Evaluation of an upgraded version of the Roche Cobas AmpliPrep/Cobas TaqMan HIV-1 test for HIV-1 load quantification. *J Clin Microbiol* 48:1413-1416.

25. Dust K, Hedley A, Nichol K, Stein D, Adam H, Karlowsky JA, Bullard J, Van Caesele P, Alexander DC. 2020. Comparison of commercial assays and laboratory developed tests for detection of SARS-CoV-2. *J Virol Methods* 285:113970.
26. Pujadas E, Ibeh N, Hernandez MM, Waluszko A, Sidorenko T, Flores V, Shiffrin B, Chiu N, Young-Francois A, Nowak MD, Paniz-Mondolfi AE, Sordillo EM, Cordon-Cardo C, Houldsworth J, Gitman MR. 2020. Comparison of SARS-CoV-2 detection from nasopharyngeal swab samples by the Roche cobas 6800 SARS-CoV-2 test and a laboratory-developed real-time RT-PCR test. *J Med Virol* 92:1695-1698.
27. Wang H, Liu Q, Hu J, Zhou M, Yu MQ, Li KY, Xu D, Xiao Y, Yang JY, Lu YJ, Wang F, Yin P, Xu SY. 2020. Nasopharyngeal Swabs Are More Sensitive Than Oropharyngeal Swabs for COVID-19 Diagnosis and Monitoring the SARS-CoV-2 Load. *Front Med (Lausanne)* 7:334.
28. Wang X, Tan L, Wang X, Liu W, Lu Y, Cheng L, Sun Z. 2020. Comparison of nasopharyngeal and oropharyngeal swabs for SARS-CoV-2 detection in 353 patients received tests with both specimens simultaneously. *Int J Infect Dis* 94:107-109.
29. Rogers AA, Baumann RE, Borillo GA, Kagan RM, Batterman HJ, Galdzicka MM, Marlowe EM. 2020. Evaluation of Transport Media and Specimen Transport Conditions for the Detection of SARS-CoV-2 by Use of Real-Time Reverse Transcription-PCR. *J Clin Microbiol* 58.
30. Patel MR, Carroll D, Ussery E, Whitham H, Elkins CA, Noble-Wang J, Rasheed JK, Lu X, Lindstrom S, Bowen V, Waller J, Armstrong G, Gerber S, Brooks JT. 2020. Performance of oropharyngeal swab testing compared to nasopharyngeal swab testing for diagnosis of COVID-19 -United States, January-February 2020. *Clin Infect Dis*.

31. Mawaddah A, Gendeh HS, Lum SG, Marina MB. 2020. Upper respiratory tract sampling in COVID-19. *Malays J Pathol* 42:23-35.
32. Liu M, Li Q, Zhou J, Ai W, Zheng X, Zeng J, Liu Y, Xiang X, Guo R, Li X, Wu X, Xu H, Jiang L, Zhang H, Chen J, Tian L, Luo J, Luo C. 2020. Value of swab types and collection time on SARS-COV-2 detection using RT-PCR assay. *J Virol Methods* 286:113974.
33. Calame A, Mazza L, Renzoni A, Kaiser L, Schibler M. 2020. Sensitivity of nasopharyngeal, oropharyngeal, and nasal wash specimens for SARS-CoV-2 detection in the setting of sampling device shortage. *Eur J Clin Microbiol Infect Dis*.
34. Bwire GM, Majigo MV, Njiro BJ, Mawazo A. 2020. Detection profile of SARS-CoV-2 using RT-PCR in different types of clinical specimens: A systematic review and meta-analysis. *J Med Virol*.
35. Craney AR, Velu PD, Satlin MJ, Fauntleroy KA, Callan K, Robertson A, La Spina M, Lei B, Chen A, Alston T, Rozman A, Loda M, Rennert H, Cushing M, Westblade LF. 2020. Comparison of Two High-Throughput Reverse Transcription-PCR Systems for the Detection of Severe Acute Respiratory Syndrome Coronavirus 2. *J Clin Microbiol* 58.
36. Moran A, Beavis KG, Matushek SM, Ciaglia C, Francois N, Tesic V, Love N. 2020. Detection of SARS-CoV-2 by Use of the Cepheid Xpert Xpress SARS-CoV-2 and Roche cobas SARS-CoV-2 Assays. *J Clin Microbiol* 58.
37. Poljak M, Korva M, Knap Gasper N, Fujs Komlos K, Sagadin M, Ursic T, Avsic Zupanc T, Petrovec M. 2020. Clinical Evaluation of the cobas SARS-CoV-2 Test and a Diagnostic Platform Switch during 48 Hours in the Midst of the COVID-19 Pandemic. *J Clin Microbiol* 58.

38. Lowe CF, Matic N, Ritchie G, Lawson T, Stefanovic A, Champagne S, Leung V, Romney MG. 2020. Detection of low levels of SARS-CoV-2 RNA from nasopharyngeal swabs using three commercial molecular assays. J Clin Virol 128:104387.

Figure Legends

Figure 1. SARS-CoV-2 genome diagram and location of cobas assay target regions. ORF: open reading frame; S: spike; E: envelope; M: matrix; N: nucleocapsid. Forward and reverse primers are represented by arrows, and probes by black rectangles, for Target 1 (ORF1a/b) and 2 (E gene).

Figure 2. Specimen type equivalency. Individual results are plotted for each specimen type and target; horizontal bars represent the mean and standard deviation (SD) for each group. NPS: nasopharyngeal swab; OPS: oropharyngeal swab. Target 1 (filled circles) and Target 2 (open circles) are shown separately.

Figure 3. Specimen collection matrix equivalency. **A:** CPM vs. UTM. **B.** flocked or woven swabs in UTM and woven swabs in saline. Individual results are plotted for each matrix and target; horizontal bars represent the mean and standard deviation (SD) for each group. CPM: cobas PCR medium; UTM: universal transport medium. Target 1 (filled circles) and Target 2 (open circles) are shown separately.

Table 1 Analysis of cobas SARS-CoV-2 primer and probe binding site sequence variation.

	Frequency	Percentage	Target 1 (number of differences)				Target 2 (number of differences)				Assay Overall
			Forward	Probe	Reverse	Predicted Impact*	Forward	Probe	Reverse	Predicted Impact*	
Reference	1,847,992	98.56%	0	0	0	No	0	0	0	No	?
Variant 1	3895	0.21%	0	1	0	No	0	0	0	No	?
Variant 2	3717	0.20%	0	0	0	No	1	0	0	No	?
Variant 3	2395	0.13%	0	1	0	No	0	0	0	No	?
Variant 4	1383	0.074%	0	0	1	No	0	0	0	No	?
Variant 5	1185	0.063%	0	1	0	No	0	0	0	No	?
Variant 6	790	0.042%	0	0	0	No	1	0	0	No	?
Variant 7	771	0.041%	1	0	0	No	0	0	0	No	?
Variant 8	747	0.040%	0	0	0	No	0	0	1	No	?
Variant 9	606	0.032%	0	1	0	No	0	0	0	No	?
Variant 10	529	0.028%	0	1	0	No	0	0	0	No	?
Variant 11	496	0.027%	0	0	1	No	0	0	0	No	?
Variant 12	479	0.026%	0	1	0	No	0	0	0	No	?
Variant 13**	462	0.025%	0	0	0	No	0	1	0	Yes	?
Variant 14	431	0.023%	0	0	0	No	1	0	0	No	?
Variant 15	422	0.023%	0	0	0	No	1	0	0	No	?
Variant 16	403	0.022%	0	0	1	No	0	0	0	No	?

* “Predicted Impact” indicates the predicted combined impact of the sequence changes in the primers or probe binding site on assay signal. No: predicted Ct increase less than 5 cycles, probe Tm >65°C; Yes: predicted to potentially increase Ct >5 cycles, probe Tm <65°C, or the reported C-T mismatch near the probe 5’ end (22); the majority of cases with any predicted increase in Ct were 1-2 cycles. Variants present in more than 0.02% of sequences in GISAID as of June 15, 2021 are shown. For a complete listing of less

common variants, see Supplemental Material Table S1. ** This variant confirmed experimentally to reduce Target 2 (but not Target 1) reactivity (22).

Table 2. Summary of Assay Inclusivity.

Target	N differences in haplotype	N different haplotypes	Frequency	Count with impact	Frequency with impact	Total %	% with impact
1	0	238	1,857,440	0	0	99.07%	0%
	1	235	17,139	0	0	0.91%	0.0000%
	2	42	339	9	21	0.018%	0.0011%
	3	10	10	7	7	0.0005%	0.0004%
	4	2	2	1	1	0.0001%	0.0001%
	5	1	1	0	0	0.0001%	0%
	6	1	1	1	1	0.0001%	0.0001%
	7	1	1	1	1	0.0001%	0.0001%
2	0	243	1,865,363	0	0	99.49%	0%
	1	227	9,465	5	467	0.50%	0.0249%
	2	34	65	3	4	0.0035%	0.0002%
	3	4	4	3	3	0.0002%	0.0002%
	4	3	3	3	3	0.0002%	0.0002%
	5	3	3	3	3	0.0002%	0.0002%
	6	6	17	6	17	0.0009%	0.0009%
	7	6	9	6	9	0.0005%	0.0005%
	8	2	2	2	2	0.0001%	0.0001%
combined	9	2	2	2	2	0.0001%	0.0001%
	0	1	1,847,992	0	0	98.56%	0%
	1	365	26,367	0	0	1.41%	0%
	2	121	515	0	0	0.027%	0%
	3	16	18	0	0	0.0010%	0%
	4	5	5	0	0	0.0003%	0%
	5	4	4	0	0	0.0002%	0%
	6	7	18	0	0	0.0010%	0%
	7	6	7	0	0	0.0004%	0%
	8	3	5	0	0	0.0003%	0%
	9	2	2	0	0	0.0001%	0%

Table 3 Limit of detection

Concentration (TCID ₅₀ /mL)	RNA concentration (copies/mL)*	Total valid results	Test positivity (%)†		Mean Ct‡	
			Target 1	Target 2	Target 1	Target 2
0.084	621	21	100	100	31.0	33.0
0.028	207	21	100	100	31.8	34.1
0.009	67	21	100	100	32.7	35.2
0.003	22	21	38.1	100	33.5	36.4
0.001	7.4	21	0	52.4	n/a	37.9
0.0003	2.2	21	0	14.3	n/a	37.2
0.0001	0.7	21	0	9.5	n/a	38.5
0 (blank)	0	10	0	0	n/a	n/a

* Conversion from TCID₅₀/mL to RNA copies/mL based on information provided in the Certificate of Analysis from the vendor: one TCID₅₀/mL is equal to 7393 genome equivalents (RNA copies) by droplet digital PCR

† All replicates where Target 1 was positive were also positive for Target 2.

‡ Calculations only include positive results.

Table 4 Detection of variants of concern, variants of interest and variants under monitoring

Variant	Concentration (copies/mL)	N positive/N tested (Target 1)	N positive/N tested (Target 2)
Wildtype	250	8/8	8/8
	100	8/8	8/8
	50	7/8	8/8
Alpha (B.1.1.7)	250	8/8	8/8
	100	8/8	8/8
	50	7/8	8/8
Beta (B.1.351)	250	8/8	8/8
	100	8/8	8/8
	50	5/8	8/8
Gamma (P.1)	250	8/8	8/8
	100	8/8	8/8
	50	7/8	8/8
Delta (B.1.617.2)	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5
Kappa (B.1.617.1)	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5
Epsilon (B.1.427)	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5
Epsilon (B.1.429)	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5
Zeta (P.2)	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5
Lambda (C.37)	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5
R.1	250	5/5	5/5
	100	5/5	5/5
	50	5/5	5/5

Table 5 Assay precision

Target	Level (x LoD)	Hit rate	Mean Ct	Instrument to- Instrument		Lot-to-Lot		Day-to-Day		Run-to-Run		Within-Run (Residual)		Total	
				SD	CV%	SD	CV%	SD	CV%	SD	CV%	SD	CV%	SD	CV%
Target 1	~0.3x	10.0%	32.5	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.5	1.4	0.5	1.4
	~1.0x	91.1%	32.1	0.0	0.0	0.2	0.6	0.1	0.3	0.0	0.0	0.6	1.8	0.6	1.9
	~3.0x	100.0%	31.2	0.0	0.0	0.2	0.7	0.0	0.0	0.0	0.0	0.3	0.9	0.4	1.1
Target 2	~0.3x	34.4%	35.4	0.0	0.0	0.5	1.3	0.3	0.8	0.1	0.2	0.5	1.5	0.8	2.2
	~1.0x	93.3%	34.2	0.0	0.0	0.1	0.3	0.2	0.6	0.0	0.0	0.7	2.0	0.7	2.2
	~3.0x	100.0%	32.9	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.4	1.1	0.4	1.1

Figures

Figure 1 SARS-CoV-2 genome diagram and location of cobas assay target regions

Figure 2. Specimen type equivalency.

Figure 3. Specimen collection matrix equivalency.