Field clinical performance of SARS-CoV-2 point-of-care diagnostic tests: a living systematic review of trials up to 17th of August, 2021

Gabriel Hawthorne^{1,2}, Adam Harvey²

¹Cambridge University Hospitals NHS Foundation Trust, Cambridge, United Kingdom

²Diagnostics for the Real World EU, Chesterford Research Park, Saffron Walden, United Kingdom

Corresponding address: hawthornegab@gmail.com

Abstract

Point-of-care assays offer a decentralized and fast solution to the diagnosis of SARS-CoV-2, providing benefits for patients, healthcare workers and healthcare facilities. This technology has the potential to prevent outbreaks, enable fast adoption of potentially life-saving measures and improve hospital workflow. While reviews regarding the laboratory performance of those assays exist, a review focused on the real-life clinical performance and true point-of-care feasibility of those platforms is missing. Therefore, the objective of this study is to help clinicians, healthcare providers and organizations to understand the real-life performance of point-of-care assays, aiding in their implementation in decentralised, true point-of-care facilities, or inside hospitals. 1246 studies were screened in 3 databases and 87 studies were included, evaluating 27 antigen tests and 11 nucleic-acid amplification platforms deemed feasible for true point-of-care placement. We excluded studies that used processed samples, pre-selected populations, archived samples and laboratory-only evaluations and strongly favored prospective trial designs. We also investigated package inserts, instructions for use, comments on published studies and manufacturer's websites in order to assess feasibility of point-of-care placement and additional information of relevance to the end-user. Apart from performance in the form of sensitivity and specificity, we present information on time to results, hands-on time, kit storage, machine operating conditions and regulatory status. To the best of our knowledge, this is the first review to systematically compare point-of-care test performance in real-life clinical practice. We found the performance of tests in clinical practice to be markedly different from the manufacturers reported performance and laboratory-only evaluations in the majority of scenarios. Our findings may help in the decision-making process related to SARS-CoV-2 test in real-life clinical settings.

Rationale for the review

A review focused on the real-life clinical performance and point-of-care feasibility of SARS-CoV-2 diagnostic platforms is missing, impairing the ability of individuals, healthcare providers and test providers to make informed decisions.

Objective(s) or question(s) the review addresses:

The objective of this study is to help clinicians, healthcare providers and organizations to understand the real-life performance of point-of-care assays, aiding in their implementation in decentralised, true point-of-care facilities or in complex healthcare environments.

Introduction

In December 2019, SARS-CoV-2 was first reported in Wuhan, China, and a pandemic was declared by the World Health Organization (WHO) in March, 2020. Reverse transcription-quantitative polymerase chain reaction (RT-qPCR) is the gold standard for diagnosis of SARS-CoV-2 infection. However, this technique has disadvantages including the requirement of centralised facilities with specific equipment, the requirement of highly trained staff and a long turnaround time between sample collection to results, approaching over 48 hours in some scenarios¹. Driven by the need for diagnostic solutions during the pandemic, multiple new assays and platforms were developed, including multiple fast molecular and antigen tests. The FIND SARS-COV-2 DIAGNOSTIC PIPELINE², which collates an overview of commercially available SARS-CoV-2 tests in real time, listed 1152 results for diagnostic solutions up to 29th of October, 2021.

Point-of-care (POC) diagnostic platforms are defined as diagnostic systems that can deliver results near patients, without the need for centralized laboratories or diagnostic facilities³. These platforms tend to deliver fast results, often within 120 minutes, enabling rapid medical decisions and facilitating timely interventions. POC diagnostic assays are currently in use in health systems for different ends, from the bedside glucose test⁴ to the analysis of blood gases and electrolytes⁵. Besides being able to provide faster results, an important advantage of POC tests is to facilitate diagnosis in locations that previously could not have access to centralized laboratory diagnostic techniques. In the context of transmissible infectious diseases, some of these assays enable quick decisions regarding treatment and public health measures such as isolation of individuals and contact-tracing. Before the pandemic, POC tests were already in use for the diagnosis of conditions such as influenza-like illnesses in different settings, including accident and emergency departments in hospitals and outpatient clinics⁶. Other assays focus on the diagnosis of sexually transmitted diseases like HIV⁷ and Chlamydia trachomatis⁸. As a consequence of the SARS-CoV-2 pandemic and the need for more effective diagnostic solutions, not only centralized diagnostic solutions but also POC diagnostic assays had an unprecedented expansion, since time from sample collection to results is key to prevent further infections and to speed up patient triage in hospitals and healthcare settings overall.

However, POC tests can have limitations such as decreased sensitivity or specificity, increased costs, and lower throughput compared to centralised laboratory facilities and techniques such as real-time PCR. Some tests also require multiple manual steps in sample preparation or computers for their execution, which can make platforms too complex for true POC placement. Additionally, tests with poor performance can have multiple consequences. False negative results can cause inadequate placement of patient in hospitals (e.g, moving a infectious patient to a ‘green ward’), causing new outbreaks in an already diseased population, and also deem a community patient not infectious, thus increasing the risks of propagating infection to contacts. False positive results can inversely place patients in high-risk environments in hospitals (e.g, inside a ‘red ward’) and cause unnecessary isolation and economic impact in an outpatient setting. If a test is considered inaccurate and might require confirmation before results can be clinically acted on, this defeats the purpose of a rapid test.

The challenges of testing in real-life scenarios have also been explored. For instance, Micocci et al⁹ interviewed staff from English care homes and reported on the difficulties of COVID-19 testing in that setting. Isolation and testing procedures were found to be challenging, requiring reconfiguration of staffing and the environment. One of the conclusions of the study was that each POC test must be evaluated in the context the test is going to be conducted in, validating the need for in depth detail for the platform in question.

The regulatory aspect of novel tests in the context of a pandemic has been complex, and the extent of this process was likely unprecedented. Due to the urgent need for testing solutions, many abbreviated validation studies were accepted by the scientific community and by regulatory agencies with initial or partial evaluations, and few assays had their field performance assessed before being released to the market. Despite showing good accuracy in internal laboratory validations, multiple POC tests for SARS-CoV-2 diagnosis had a lower-than-expected performance once released for clinical use¹⁰. This topic was the subject of political and legal debate¹¹ and resulted in some previously approved tests being later withdrawn from the market¹². There are many examples of disparities between manufacturer claims based on laboratory-only evaluations and data from clinical trials. For instance, the platform ID Now (Abbott) claimed a sensitivity of 100% and a specificity of 100% on the product's package insert¹³, but clinical evaluations showed a sensitivity below 75%¹⁴ or below 50% in some cases¹⁵. Another example is the study conducted by Jokela et al¹⁶, where the Mobidiag Novodiag platform had 93.4% (100/107) sensitivity in archived samples but only identified 60% (3/5) positive samples in a clinical setting.

POC diagnosis plays an important role in SARS-CoV-2 management. Faster diagnosis speeds up isolation measures and therefore the prevention of new outbreaks. In the same way, faster confirmation of SARS-CoV-2 absence helps avoiding unnecessary isolation for individuals and their contact groups, providing social and economic benefit. The workflow of patients inside a hospital can be greatly facilitated by using tests that are reliable and provide a fast result, aiding in the placement of patients inside “red” or “green” wards and preventing SARS-CoV-2 nosocomial spread while freeing up rooms and improving the capacity of emergency departments. For example, in the study conducted by Collier et al using the SAMBA platform¹⁷, mean length of stay on COVID-19 “holding” (or “amber”) wards was reduced by nearly 30h using the POC test. Additionally, timely interventions like the use of dexamethasone in patients requiring respiratory support¹⁸ or the use of interleukin-6 receptor antagonists in critically ill patients¹⁹ benefit from a fast diagnostic modality, especially considering the waiting-time for a centralised test result.

In this review, we address the POC tests for SARS-CoV-2 diagnosis, that are divided in two large categories. The first category is of the nucleic acid amplification tests (NAAT), also called molecular tests. These tests use some sort of nucleic acid amplification technique (polymerase chain reaction or a similar technique), and therefore require platforms for test conduction. The second category concerns antigen tests, which directly target SARS-CoV-2 surface antigens. These tests are usually attached to a lateral flow detection cassette and more often than not have no associated platform. We specifically focus on clinical trials clarifying the performance of those tests in real life settings. Our criteria was rigidly tailored to exclude samples that were pre-selected and testing conducted in laboratory conditions and/or with frozen samples. The objective of this study is to help clinicians, healthcare providers and organizations to understand field performance of POC assays, aiding in their safe implementation in decentralised, true POC facilities or in more complex healthcare environments.

Other reviews of POC assays targeting SARS-CoV-2 are available and use different inclusion criteria. Dinnes et al²⁰ published a review that includes 64 studies for 16 antigen platforms and 5 molecular assays, collected up to September 2020. This study attempted to divide reports between symptomatic and asymptomatic individuals, and also included laboratory-only evaluations and retrospective studies, which diverges from our goal. Yoon et al²¹ conducted a study using the FDA Emergency Use Authorization (FDA-EUA) POC tests up to August, 2020, in which 26 studies were analyzed. Hayer et al²² reviewed antigen tests but only included assays not needing a separate reader. Other POC diagnostic technologies that have been in use in the management of COVID-19 patients, like the use of POC ultrasound for patient follow-up after diagnosis, do not fit the scope of this work.

1. Methodology

This systematic review was conducted following Preferred Reporting Items for Systematic Reviews and Meta-Analyses (PRISMA) guidelines.

After debate between the authors, a broad search phrase was defined with the intention to capture a large number of studies, given the novelty of the field. Filtering results by date using the abovementioned algorithm was not necessary, since the name SARS-CoV-2 was coined after the identification of the pathogen in 2019.

2.1 Search strategy

Algorithm searched on 17/08/2021	Source	Results
“((point-of-care) AND (SARS-CoV-2)) AND (performance) OR ((point-of-care) AND (SARS-CoV-2)) AND (evaluation)”	MedRxiv	645
	MEDLINE	506
	BioRxiv	95
	Cross-references	1

2.2 Exclusion criteria

Exclusion criteria
-Study unrelated to POC testing;
-Platform not deemed feasible for the POC (e.g requires the use of centralised equipment);
-Test main focus other than diagnosis of acute infection (e.g antibody assays);
-Study not a clinical trial (e.g proof of concepts, validations, studies conducted with frozen samples, test conducted in laboratory conditions);
-Samples collected from patients processed before testing;
-Population pre-selected (e.g, only known positives were enrolled);

2.3 Selection process

2.31 Defining a POC assay

The definition of a POC diagnostic assay is not straightforward, since there are no rigid criteria for reference. The defining factor is being able to provide a diagnostic solution near the patient, thus removing the need for a centralized testing facility. While some platforms are totally mobile, being able to

move to the patient's location and clearly defined as POC, others require the use of an energy source, centralised computers, tablets or other accessories. Other assays require multiple sample preparation steps before test conduction (such as the need for RNA extraction, the use of a centrifuge, heating blocks or multiple pipetting steps), even though the final testing step can theoretically be conducted near the patient; this is the case of most LAMP platforms. In addition, requirements for isolation and measures for preventing infection spread further complicate this definition during the SARS-CoV-2 pandemic. We therefore recognize the definition of a POC test or platform remains subjective. Aligned with our objectives, in this review, we considered an outpatient community setting as our reference point. Therefore, assays that use reagents requiring processing in a central laboratory or facility were excluded from our analysis. The rationale for that is that these platforms could not be operational without such facilities in the vicinity and therefore could not be implemented in the community in a truly POC fashion. Based on the information available in publications and manufacturer's manuals, assays that are estimated to be technically capable of implementation in lower complexity settings are included. These platforms will have their potential limitations/considerations (e.g, the requirement for cold chain storage or the need for a desktop computer) described in the comments of the results table.

2.32 Defining target trials and comments regarding the criteria

Because of the disruption and urgency caused by the COVID-19 the pandemic, many studies have not followed rigid clinical trial designs, and a high level of heterogeneity between methods and designs is acknowledged. Issues such as need for self-isolation, patient discomfort upon repeated swabbing, multi-platform evaluations and insufficient number of positive samples in low-incidence scenarios affect the feasibility of the studies and need to be taken into consideration when conducting a review. For instance, the study of Tu et al²³ evaluating the ID NOW platform (Abbott) had an original design to enroll 200 positive patients, but a prevalence drop made the study incompletable. Given the heterogeneity of the trials, finding a rigid, unifying criteria for inclusion was difficult and would defeat the purpose of this review.

In this work, as mentioned, our goal is reviewing the efficacy of tests in real life conditions. Therefore, we included studies that evaluated tests in patients in a true POC fashion and excluded laboratory-only or in silico evaluations. As a consequence, we have excluded from our analysis all proof-of-concepts and studies that used spiked samples. We have also excluded studies that used solely pre-tested frozen samples, as these conditions are vastly different from conditions in the field. As mentioned before, the selection of frozen samples may exclude samples with inhibitors, invalid or borderline results, and with low viral-loads, favoring samples with low Ct values.

Naturally, many trials used cooled or frozen samples at some point, especially when considering multi-platform evaluations. We tended to consider time from sample collection to testing in our selection criteria; while it was difficult to decide on a clear cut-off time, samples that were stored for a brief period of time to allow testing with a POC platform were accepted. We recognize this likely does not have the same value as fresh POC testing, it is often a necessary accommodation for validation studies where a comparator assay is used. As an example, Lephart et al collected samples from 88 patients (13 of which were known positive), stored at 4 °C and tested within 24h; this study was included in our table. On the other hand, studies using samples that were part of frozen panels tested in retrospect, often weeks after sample collection, were excluded. As an example of our criteria, we did not include the work by Corman et al²⁴ who conducted a comparison of seven SARS-CoV-2 antigen tests available in Europe because processed samples were used and only negative swabs were collected from patients. We favored prospective studies.

Given the heterogeneity of designs and the conditions for studies, we debated between authors before inclusion when methods were not clear. For instance, Cerutti et al conducted a trial with 330 patients including a minority of frozen samples ($n = 13$); this trial was considered for our study²⁵. Conversely, a prospective study by Courtellemont et al²⁶ was not included as known positive patients were pre-selected to enroll; therefore, operators knew the status of the patients beforehand. A similar situation was found in the trial by Ghofrani et al²⁷, who conducted a comparison with known positive patients and selected eligible samples; the methodology of this study was unclear and the sensitivity, reported as 96.7%, was much higher than usually reported in literature for antigen assays.

The vast majority of studies used nasopharyngeal samples, although a few studies used nasal samples only. Studies evaluating POC assays using saliva samples were reported^{28,29} and usually show poor performance. For instance, Basso et al³⁰ found a sensitivity of only 13% testing saliva with antigen tests; similarly, the performance of saliva samples was inferior in the study by Agulló et al³¹ evaluating the Panbio assay. We therefore reported the results for either nasopharyngeal or nasal swabs and (as per manufacturer's advice) when they were part of an assessment with multiple sample types. When collection methods were compared (e.g, sensitivity of self-swab against healthcare collection^{32,33}), we reported results obtained by the healthcare staff worker.

We also excluded studies where POC assays were compared to other POC assays, as we considered no gold-standard was included. An example is the study conducted by Basu et al³⁴, where Abbott ID Now COVID-19 was compared to the Cepheid Xpert Xpress SARS-CoV-2 without a PCR standard; the same can be said of the study conducted by Agarwal et al, which compared the Standard Q COVID-19 antigen test to the TrueNat POC platform³⁵. It is important to mention that the reliability of the gold-standard was questioned in some studies³⁶. Ideally, a reference standard would be built using a combination of criteria, including more than one assay¹⁵, cross-checked clinical information, radiological evidence and other laboratory information (e.g antibody production, viral markers or inflammatory makers), but this is understandably complex and unfeasible in many circumstances.

We also carefully considered the population type in the studies. Naturally, testing known-positive patients presents a bias. Due to challenges imposed by factors such as lockdowns, the urgency for results, different prevalence levels, differences in viral load at different stages of disease and the size of the trials, some studies tested known positive populations in order to obtain statistically significant data for sensitivity. If a study was conducted solely with known positive patients, it was excluded. On the other hand, studies that complemented a prospective evaluation by randomly testing positive populations were accepted. As an example, Basso et al tested antigen assays in 139 selected inpatients (this population had a 60% positivity rate) and 96 outpatients prospectively (3% positivity rate); this study was included. Some studies tested exclusively in a paediatric population, and were also included³⁷.

Taking these factors into consideration, trials were assessed on the overall level of heterogeneity in their methodology. In the study conducted by Osterman et al³⁸ evaluating 2 antigen assays, there was a significant variability between 2 sites as samples were collected in different time-frames (site 1 from March to October 2020 and site 2 between November-December 2020) and some samples in site 1 had different storage methods, with some being frozen for days before testing. Ultimately, we decided to include this study. Other studies like Marti et al³⁹ were excluded due to a high level of heterogeneity in their methods, using both a POC and a centralised PCR as their standard and using different populations, including a population of known positive individuals. We attempted to include detailed explanations of the reasons for inclusion or exclusion of individual trials in appendix 1 of this review. We encourage authors to contact us for clarifications and potential corrections in the next version of this living review.

2.33 Technical information on platforms

We investigated package inserts, instructions for use, comments on published studies and manufacturer's websites in order to assess feasibility of POC placement and additional information that may be relevant to the end user. Apart from sensitivity and specificity, we included time to results, hands-on time, kit storage, machine operating conditions and regulatory status. We also made comments on testing requirements and additional details that were deemed relevant. We opted to use publicly available information, such as the instructions for use in the FDA website¹³. When that information was unavailable, we attempted to obtain the package insert by contacting the manufacturers or examining public information published by hospitals, government entities and other third-party institutions using the platform.

2. Results

Figure 1 – identification, screening, eligibility and inclusion flowsheet

3.1 Excluded studies and studies debated between authors

Comments on individual studies and their criteria for exclusion can be found in Appendix 1.

3.2 Comments on reference standards, sensitivity, specificity, positive predictive value (PPV) and negative predictive value (NPV)

POC tests have to be compared against a selected gold-standard in order to properly evaluate performance. This standard is real-time PCR in the vast majority of the studies (as mentioned above, some studies used another POC assay as a comparison and were excluded from our analysis). It is important to understand that the performance numbers reflect values against a reference standard, which is not always necessarily better or more accurate than the object of study. Most studies used a third platform as a tie-breaker in this context and the results of this full analysis were considered whenever

appropriate. As mentioned previously, if a clear reference standard was used based antibodies and/or clinical and radiological evidence, this was also taken into consideration.

Positive predictive value (PPV) and negative predictive value (NPV) are useful to understand how much a result can be trusted given the prevalence of an infectious disease in a given time. Assays are always trialed settings with an estimated prevalence at a moment in time; in performance studies, the reference method provides that value. PPV and NPV change accordingly depending on the prevalence in the setting. For instance, the study of an antigen test by Peña et al⁴⁰ reported a sensitivity of 69.86%, a specificity of 99.61%, but a PPV of 94.44% and a NPV of 97.22% as the prevalence in that setting was 8.64%. Here, we avoided using PPV and NPV projections whenever possible and aimed to report the provided ‘sensitivity’ and ‘specificity’ values, even though we recognize these values are intertwined. The decision to not use PPV and NPV projections was made because (1) in rapidly contagious infections like SARS-CoV-2, an accurate real-time monitoring of prevalence parameters is difficult, in contrast to what is found for infectious agents with a clearly defined and predictable epidemiology; therefore the prevalence of SARS-CoV-2 infection in a given setting may change rapidly considering outbreaks, lockdowns, and new variants. (2) An accurate real time monitoring of SARS-CoV-2 regional prevalence is challenging even for developed countries, resulting in prevalence values that are often retrospective. (3) An assay’s performance can be distorted using different prevalence levels. Finally, (4) an analysis including multiple PPV and NPV projections would make this review more speculative and less practical. We strongly suggest that the referenced trials are read in full for further information and clarification of performance as the term ‘sensitivity’ and ‘specificity’, as reported in this review, are always relative to prevalence in the particular setting of each study.

3.3 Included studies

We present the included studies in tables below. NAAT are grouped separately from antigen tests and are presented with an additional column discussing testing requirements. While most manufacturers used viral RNA copies/ml in a serial dilution to assess the assay limit of detection, some manufacturers used plaque forming units (PFU) instead of viral RNA copies/ml. We presented the limit of detection in copies/ml if both information were available, according to manufacturer’s instructions for use. The limit of detection of the different assays was converted into a copies/ml format when possible (for instance, if this value was given in copies/uL). We also did not include a claimed limit of detection for antigen assays.

We have not presented results for different populations (eg, symptomatic vs asymptomatic, children vs adults); therefore, the values presented represent the average for all individuals tested as provided in the studies. For example, the study by Ford et al⁴¹ presented separate values for children and adults, but we have reported these values as an overall.

Some assays had little technical information available, despite our best efforts to obtain package inserts or instructions for use. This was especially true for novel antigen tests. We have therefore included information to the best of our knowledge and indicated the information we could not obtain as “not available (N/A)” in the table.

Tables are divided between assays that diagnose of SARS-CoV-2 alone and multiplex assays (e.g, SARS-CoV-2 + FluA/FluB).

3.31 SARS-CoV-2 studies

SARS-CoV-2 POC NAAT WITH FIELD CLINICAL EVALUATIONS (PART 1)				
	Cobas Liat (Roche)	CovidNudge (DNA Nudge)	Cue COVID-19 (Cue)	ID Now (Abbott)
Testing requirements	Requires strict temperature control for kit storage (near-POC)	Requires disposable scissors (not provided by manufacturer), and an iPad	Direct testing wand into cartridge. Results on a mobile smart device ⁴²	Direct swab into cartridge is possible
Sensitivity	94.5% (35/37) ⁴³ 100% (162/162) ⁴⁴	94% (67/71) ⁴⁵	91.7% (22/24) ⁴⁶	48% (12/25) ¹⁵ 74.73% (139/186) ¹⁴ 86.4% (19/22) ⁴⁷ 91% (30/33) ⁴⁸ 91.3% (21/23) ²³ 95.1% (158/166) ⁴³ 98% (151/154) ⁴⁹
Specificity	97.4% (190/195) ⁴⁴ 97.7% (476/487) ⁴³	100% (315/315) ⁴⁵	98.4% (239/243) ⁴⁶	96.9% (506/520) ⁴³ 97.5% (235/241) ⁴⁹ 99.3% (1474/1484) ⁴⁷ 99.4% (336/338) ¹⁴ 100% (63/63) ¹⁵ 100% (151/151) ⁴⁸ 100% (762/762) ²³
Claimed limit of detection	12 copies/ml ¹³	5000 copies/ml ⁵⁰	1300 copies/ml ¹³	125 copies/ml ¹³
Time to results	~20 min ¹³	75 min ⁵⁰	25 min post pre-heating (total time not provided) ¹³	13 min or less ¹³
Hands on time	1 min ¹³	1 min ⁵⁰	1 min ¹³	2 min ¹³
Kit storage	2-8°C ¹³	25°C or less ⁵⁰	15-30°C ¹³	2-30°C ¹³
Machine operating conditions	15-30°C ¹³	16-30°C ⁵⁰	15-30°C ¹³	15-30°C ¹³
Regulatory status	FDA EUA, CE ¹³	N/A	FDA EUA, CE ¹³	FDA EUA, CE ¹³
Additional Details	Invalid specimens on Liat testing were excluded (n = 3) ⁴³ . FDA Issued risk of false positives from leaking tubes ⁵¹	Invalid rate of 5.7% ⁴⁵ . Report (03/2021) from the UK's government DHSC shows a sensitivity of 82.1% and a specificity of 99.1% after evaluation in 11 sites (details unpublished) ⁵²	8.6% invalid/cancelled results (25/292) ⁴⁶	-

N/A = not available

SARS-CoV-2 POC NAAT WITH FIELD CLINICAL EVALUATIONS (PART 2)			
	Novodiag (Mobidiag)	POCKIT (HORIBA UK)	QIAstat-Dx (Qiagen)
Testing requirements	Requires strict temperature control for kit storage	Potentially requires a biological safety cabinet or hood ⁵³	Requires strict temperature control for kit storage
Sensitivity	60% (3/5) ¹⁶	85.7% (6/7*) ⁵⁴ 90.2% (65/72) ⁴³	100% (23/23) ⁵⁵
Specificity	99.7% (355/356) ¹⁶	98.7% (246/249*) ⁵⁴ 99.6% (2051/2059) ⁴³	90% (18/20) ⁵⁵
Claimed limit of detection	313 copies/ml ⁵⁶	60.000 copies/ml ⁵²	500 copies/ml ¹³
Time to results	80 min ⁵⁷	85 min ⁵⁸	67 min ¹³
Hands on time	1 min ⁵⁶	'Minimal' (details not provided) ⁵⁸	1 min ¹³
Kit storage	18-25°C ⁵⁶	10-40°C ⁵⁸	15-25°C ¹³
Machine operating conditions	15-30°C ⁵⁶	15-35°C ⁵⁸	15-30°C ⁵⁹
Regulatory status	CE ⁵⁷	CE ⁵⁸	FDA, CE ¹³
Additional Details	-	-	Multi-pathogen assay, tests 22 pathogens ¹³

SARS-CoV-2 POC NAAT WITH FIELD CLINICAL EVALUATIONS (PART 3)				
	RapiPREP/RapiPRO (MicrosensDx)	SAMBA-II (Diagnostics for the Real World)	VitaPCR (Credo)	Xpert Xpress (Cepheid)
Testing requirements	Details of testing requirements are unclear	Requires a tablet for test conduction	Requires strict temperature control for kit storage	Requires a computer for the delivery of results ¹³ .
Sensitivity	80% (8/10) ⁶⁰	96.9% (31/32) ¹⁷	90% (126/140) ⁶¹ 99.3% (155/156) ³⁶	100% (25/25) ¹⁵ 100% (13/13) ⁴⁷
Specificity	73% (8/11) ⁶⁰	100% (117/117) ¹⁷	94.7% (358/378) ³⁶ 99% (98/99) ⁶¹	97% (60/62) ¹⁵ 99.6% (488/490) ⁴⁷
Claimed limit of detection	N/A	150 copies/ml ⁶²	2730 copies/ml ⁶³	0.0200 PFU/mL (plaque forming units) ¹³
Time to results	25 min on average ⁶⁰	71-86 min ⁶²	20 min ⁶³	30-45 min ⁶⁴
Hands on time	N/A	1 min ⁶²	1 min ⁶³	1 min ⁶⁴
Kit storage	N/A	2-37°C ⁶²	15-30°C ⁶³	2-28°C ¹³
Machine operating conditions	N/A	10-38°C ⁶²	10-38°C ⁶³	15-30°C ¹³
Regulatory status	CE ⁶⁵	CE ⁶²	CE ⁶³	FDA EUA, CE ⁶⁴
Additional Details	Test consists of extraction followed by RT-LAMP; details and true POC feasibility unclear ⁶⁶	Testing buffer inactivates virus	-	Gene E is not specific for SARS-CoV-2; samples exclusively positive for E gene need retesting ¹³

N/A = not available

SARS-CoV-2 ANTIGEN POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 1)			
	AAZ-LMB COVID-VIRO® (AAZ)	BINAXnow (Abbott)	PANBIO (Abbott)
Sensitivity	84.1% (116/138) ⁶⁷	57.7% (15/26) ⁶⁸ 74% (57/77)* ⁶⁹ 77.4% (226/292) ⁷⁰ 79.4 (267/334)* ⁴¹	38.6% (39/101) ⁷¹ , 48.1% (38/79) ⁷² 46.67% (7/15) ⁷³ , 54% (258/477*) ⁷⁴ 56.8% ^A , 57% (76/132) ³¹ , 57.1% (48/84) ⁷⁵ , 60.5% (118/195) ⁷⁶ , 63.5% (21/33)* ⁷⁷ , 66% (58/88*) ³⁰ , 66% (79/119) ⁷⁸ 66.4% (148/223) ⁷⁹ , 70.5% (24/34) ⁸⁰ , 71.4% (100/140) ⁸¹ , 73% (27/37) ⁴¹ , 73.3% (44/60) ⁸² , 74% (186/250) ⁸³ , 77.7% (14/18) ³⁷ , 79.6% (43/54) ⁸⁴ , 80.8% (240/297) ⁴¹ , 80.8% ^{A,85} 82.1% (602/733) ⁸⁶ , 84% (59/70*) ⁸⁷ , 85.5% (106/124) ⁸⁸ , 86.1% (101/122) ⁶⁷ , 86.7% (39/45) ⁸⁹ , 86.8% (92/106) ⁹⁰ , 90.5% (325/359) ⁹¹
Specificity	100% (186/186) ⁶⁷	99% (218/220)* ⁶⁹ 99.4% (2002/2016*) ⁷⁰ 99.9% (1774/1776*) ⁴¹ 100% (845/845) ⁶⁸	98% (56/57*) ⁸⁷ , 98% (581/592*) ⁷⁴ , 98.8% (592/599) ⁹¹ , 99% (145/146*) ³⁰ , 99.1% (3420/3450) ⁸⁶ , 99.5% (184/185) ⁷¹ , 99.61% (3146/3158) ⁷³ , 99.8% ^A 99.8% (519/520) ³¹ , 99.8% (1220/1222) ⁸¹ , 99.9% (1000/1001) ⁹⁰ , 99.9% (3738/3741) ⁸³ 99.9% (1586/1588) ⁴¹ , 100% (217/217) ⁷² , 100% (368/368) ⁷⁹ , 100% (369/369*) ⁸⁴ , 100% (422/422) ³⁷ , 100% (195/195) ⁸² , 100% (709/709) ⁷⁶ , 100% (149/149) ⁶⁷ , 100% (411/411) ⁸⁸ , 100% (1017/1017*) ⁷⁵ , 100% (323/323) ⁸⁰ , 100% (658/658) ⁷⁷ , 100% (188/188) ⁴¹ , 100% (388/388) ⁸⁹ , 100% (703/703) ⁷⁸ , N/A ^{A,85}
Time to results	15 min ⁹²	15 min ¹³	15-20 min ⁹³
Hands on time	3 min ⁹²	1 min ¹³	5 min ⁹³
Kit storage	2-30 °C ⁹²	2-30°C ¹³	2-30°C ⁹³
Machine operating conditions	Nil (cassette) ⁹²	Nil (antigen card) ¹³	Nil (cassette) ⁹³
Regulatory status	CE ⁹²	FDA ¹³	CE/WHO EUL ⁹³
Additional Details	-	-	Reagents must be brought to room temperature 30min before use ⁹³

A = Details could not be extracted from paper

N/A = not available

*numbers were calculated based on information on paper but not explicitly provided by authors.

SARS-CoV-2 ANTIGEN POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 2)					
	CareStart COVID-19 Antigen (Access Bio)	BD Veritor (BD)	2019-CoV Ag Fluorescence Rapid Test Kit (Bioeasy)	NowCheck COVID-19 Ag- RDT (BioNote)	COVID-19 antigen Respi- Strip (Coris)
Sensitivity	49% (25/51) ⁹⁴ 57.7% (135/234) ⁹⁵	63.1% (41/65) ⁹⁶ 66.4% (77/116) ⁹⁷ 80.7% (112/140) ^{*98} 87.7% ^{A85}	66.7% (10/15) ⁹⁹	84.5% (128/152) ^{*100}	37.1% (13/35) ¹⁰¹ 50% (4/8) ⁹⁹ 50% (47/94) ¹⁰² 80.0% ^{A,85}
Specificity	98.3% (1243/1264) ^{*95} 99.5% (577/580) ⁹⁴	98.8% (1253/1268) ⁹⁷ 99.4% (330/332) ⁹⁸ 99.7% (742/744) ⁹⁶ N/A ^{A85}	93.1% (663/712) ⁹⁹	94.4% (820/845) ^{*100}	95.8% (392/409) ⁹⁹ 100% (328/328) ¹⁰¹ 100% (44/44) ¹⁰² N/A ^{A,85}
Time to results	10 min ¹⁰³	15-20min ¹³	10 min ¹⁰⁴	15~30 min ¹⁰⁵	15-30 min ¹⁰⁶
Hands on time	1-2 min (estimated)	2-3 min (estimated)	2-3 min (estimated)	1-2 min (estimated)	1-2 min (estimated)
Kit storage	1-30°C ¹⁰³	2-30°C ¹³	N/A	2-30°C ¹⁰⁵	4-30°C ¹⁰⁶
Machine operating conditions	Nil (Ag cassette)	2-30°C ¹³	N/A	Nil (Ag cassette)	Nil (Ag cassette)
Regulatory status	FDA ¹⁰³	FDA ¹³	CE ¹⁰⁴	CE ¹⁰⁵	CE ¹⁰⁶
Additional Details	-	Requires analyser device to read results ¹⁰⁷	Requires analyser device to read results ¹⁰⁴	-	-

A = Details could not be extracted from paper

N/A = not available

*numbers were calculated based on information on paper but not explicitly provided by authors.

SARS-CoV-2 ANTIGEN POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 3)					
	QuickNavi™ COVID19 (Denka)	ESPLINE rapid test (Fujirebio)	SARS-CoV-2 Antigen (Innova)	SARS-CoV-2 Antigen Rapid Test Kit (Lepu Medical)	LumiraDx SARS- CoV-2 (LumiraDx)
Sensitivity	80.3% (151/188)* ¹⁰⁸ 86.7% (77/91) ¹⁰⁹	48% (42/88*) ³⁰	40.0% (28/70) ¹¹⁰ 86.4% (242/280) ¹¹¹	45.5% (46/101) ⁷¹	82.2% (120/146) ¹¹² 90.3% (269/298) ¹¹³ 97.5% (39/40) ¹¹⁴
Specificity	100% (1081/1081) ¹⁰⁹ 100% (997/997) ¹⁰⁸	100% (146/146*) ³⁰	95.1 (426/448)% ¹¹¹ 99.9% (5431/5434) ¹¹⁰	89.2% (165/185) ⁷¹	92.1% (561/609) ¹¹³ 97.7% (380/389*) ¹¹⁴ 99.3% (611/615) ¹¹²
Time to results	15 min ¹¹⁵	30 min ¹¹⁶	20-30 min ¹¹⁷	15 min ¹¹⁸	12 min ¹³
Hands on time	1-2 min (estimated)	>5 min ¹¹⁶	2 min ¹¹⁷	>5 min ¹¹⁸	2 min ¹³
Kit storage	N/A	2-30°C ¹¹⁶	2-30°C ¹¹⁷	4-30°C ¹¹⁸	2-30°C ¹³
Machine operating conditions	Nil (Ag cassette)	Nil (Ag cassette)	Nil (Ag cassette)	Nil (Ag card)	15-30°C ¹³
Regulatory status	N/A	CE ¹¹⁶	CE ¹¹⁷	CE ¹¹⁸	CE, FDA ¹³
Additional Details	-	-	-	Recalled from market due to likely risk of false results ¹¹⁹	Requires an instrument for the delivery of results

N/A = not available

*numbers were calculated based on information on paper but not explicitly provided by authors.

SARS-CoV-2 ANTIGEN POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 4)					
	MEDsan® SARS-Cov-2 (MEDsan)	NADAL® COVID-19 (Nal vonminden)	Quick Chaser® Auto SARS- CoV-2 (Mizuho Medy)	PCL COVID-19 Ag rapid (PCL)	INDICAID® COVID-19 (Phase Scientific)
Sensitivity	36.51% (23/63) ⁷³	56.52% (13/23) ⁷³	74.7% (62/83) ¹²⁰	38.9% (14/36) ¹²¹	85.3% (64/75) ¹²²
Specificity	99.62% (1010/1014) ⁷³	100% (783/783) ⁷³	99.8% (1316/1318) ¹²⁰	83% (38/46*) ¹²¹	94.8% (260/274) ¹²²
Time to results	15-20 min ¹²³	N/A	15 min ¹²⁰	10 min ¹²⁴	20 min ¹²²
Hands on time	2 min ¹²³	N/A	N/A	N/A	1-2 min (estimated)
Kit storage	2-30°C ¹²³	N/A	N/A	N/A	2-30°C ¹⁰³
Machine operating conditions	Nil (Ag cassette)	N/A	N/A	N/A	Nil (Ag cassette)
Regulatory status	CE ¹²³	N/A	N/A	N/A	FDA ¹⁰³
Additional Details	-	-	Requires dedicated reader ¹²⁰	-	-

N/A = not available

SARS-CoV-2 ANTIGEN POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 5)				
	Sofia SARS-CoV-2 Antigen Assay (Quidel)	BIOCREDIT COVID-19 Ag (RapiGEN)	SARS-CoV-2 Rapid Antigen Test (Roche)	Romed Coronavirus Ag Rapid Test (Romed)
Sensitivity	57.1% (52/91) ¹²⁵	56.4% (44/78) ¹²⁶	43.6% (44/101) ⁷¹ 50.3% (224/445) ³⁸ 72.5% (109/149) ¹²⁷ 80.3% (171/213) ¹²⁸	73.3% (220/300)* ¹²⁹
Specificity	99.3% (1282/1291) ¹²⁵	100% (28/28) ¹²⁶	96.2% (178/185) ⁷¹ 97.7% (377/386) ³⁸ 99.4% (171/172) ¹²⁷ 99.1% (325/328) ¹²⁸	99.8% (599/600)* ¹²⁹
Time to results	15 min ¹³⁰	15-30 min ¹³¹	15-30 min ¹³²	15 min ¹³³
Hands on time	2-3 min (estimated)	>5 min ¹³¹	2 min ¹³²	1-2 min (estimated)
Kit storage	15-35°C ¹³⁰	1-40°C ¹³¹	2-30°C ¹³²	N/A (package insert not in English)
Machine operating conditions	15-30°C ¹³⁴	Nil (test device) ¹³¹	Nil (Ag cassette) ¹³²	Nil (Ag cassette) ¹³³
Regulatory status	FDA, CE ¹³⁰	CE ¹³¹	CE ¹³²	N/A
Additional Details	Requires reading instrument	-	-	-

N/A = not available

*numbers were calculated based on information on paper but not explicitly provided by authors.

SARS-CoV-2 ANTIGEN POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 6)				
	STANDARD Q COVID-19 Ag Test (SD Biosensor/Roche)	RapidTest^a SARS-CoV-2 (Sekisui Medical)	CLINITEST[®] Rapid COVID-19 Antigen Test (Siemens)	SGTI-flex COVID-19 Ag (Sugentech)
Sensitivity	28.6% (2/7) ¹³⁵ 37.84 (14/37) ¹³⁶ 42.86% (90/210) ¹³⁷ 45.4% (173/381) ³⁸ 60.4% ^A 62.3% (139/223) ⁷⁹ 64% (27/42) ¹³⁸ 69.86% (51/73) ⁴⁰ 70.6 % (77/109) ²⁵ 76.6% (36/47) ⁹⁹ 78.3% ^{A,85} 79.5% (31/39) ³³ 84.9% (158/186) ¹³⁹ 85% (34/40) ³² 89% (170/191) ⁸⁸ 92.9% (104/122) ⁶⁷	78.4% (58/74) ¹⁴⁰	51.5% (52/101) ⁷¹ 73.5% (78/106)* ¹⁴¹	52.6% (41/78) ¹²⁶
Specificity	81% (123/151) ¹³⁸ 97.8% (352/360) ³⁸ 98.2% (110/112) ¹³⁵ 99.1% (105/106) ³² 99.3% (1207/1216) ⁹⁹ 99.5% (366/368) ⁷⁹ 99.5% (780/784) ¹³⁹ 99.6% (249/250) ³³ 99.61% (766/769) ⁴⁰ 99.7% (337/338) ⁸⁸ 99.7% ^A 99.89% (1816/1818) ¹³⁷ 100% (221/221) ²⁵ 100% (221/221) ⁶⁷ 100% (1231/1231) ¹³⁶ N/A ^{A,85}	99.2% (1003/1028) ¹⁴⁰	98.4% (182/185) ⁷¹ 100% (164/164)* ¹⁴¹	96.4% (26/27) ¹²⁶
Time to results	15-30 min ¹⁴²	10 min ¹⁴⁰	15 min ¹⁴³	15 min ¹⁴⁴
Hands on time	2 min ¹⁴²	1-2 min (estimated)	>5min ¹⁴³	1-2 min (estimated)
Kit storage	2-30°C ¹⁴²	N/A	2-30°C ¹⁴³	N/A
Machine operating conditions	Nil (Test device) ¹⁴²	Nil (Test cassette) ¹⁴⁰	Nil (Ag cassette) ¹⁴³	Nil (Ag cassette) ¹⁴⁴
Regulatory status	CE ¹⁴²	N/A	CE ¹⁴³	N/A
Additional Details	-	Can be operated with automatic reader	-	-

A = Details could not be extracted from paper

3.31 Multiplex studies

Further adjustments to inclusion criteria had to be made to consider multiplex assays studies, taking into consideration factors like the drop in prevalence of some pathogens during the coronavirus pandemic, the different prevalence of pathogens across the globe and, consequently, the difficulty in getting enough numbers for a prospective analysis of pathogens other than SARS-CoV-2.

SARS-CoV-2 ANTIGEN MULTIPLEX POC TESTS WITH FIELD CLINICAL EVALUATIONS (PART 4)	
	BD Veritor SARS-CoV-2 and Flu A+B (BD)
Sensitivity	<u>SARS-CoV-2</u> 86.6% (52/60) ¹⁴⁶ <u>Influenza A</u> 100% (40/40 – archived samples) ¹⁴⁶ <u>Influenza B</u> 100% (35/35 – archived samples) ¹⁴⁶
Specificity	<u>SARS-CoV-2</u> 99.5% (216/217) ¹⁴⁶ <u>Influenza A</u> 100% (75/75 – archived samples) ¹⁴⁶ <u>Influenza B</u> 100% (80/80 – archived samples) ¹⁴⁶
Time to results	~15 min ¹⁰³
Hands on time	1-2 min (estimated)
Kit storage	2-30°C ¹⁰³
Machine operating conditions	15–30°C ¹⁰³
Regulatory status	FDA ¹⁰³ , CE ¹⁴⁷
Additional Details	Triplex targeting SARS-CoV-2 + Flu A + Flu B. SARS-CoV-2 tested prospectively. Influenza tested using archived Influenza samples, as study was conducted off Flu season. Requires reading analyzer. Can run batches of samples.

3 Discussion

It is well known that results from laboratory-only evaluations often differ from results of external clinical studies. Usually, evaluating an intervention in patients in real-life scenarios has an increased level of complexity. In the particular situation of SARS-CoV-2 testing, there are many possible reasons for those discrepancies. The first group of differences stems from factors related to the patients themselves, such as tolerance to swabbing and the presence of inhibitors in the nasopharynx secretions. There has been debate regarding the reliability of viral load and Ct values¹⁴⁸, suggesting that viral load may greatly vary not only with the stage of the disease but also depending on the quality of swab collection and elements like dilution of swabs in a buffer and the RNA extraction process. Therefore, it appears to be important for a test to have a low limit of detection regardless of the ‘average’ viral load in a group of patients in order to have a reliable diagnostic performance. The presence of inhibitors on samples is also relevant, as many assays describe interactions between food, beverages and medications and their amplification chemistry in their instructions for use or package inserts¹³. Once a sample is already known positive (with a known Ct value or an estimated viral load) and has been selected in a panel for a comparison against other tests, this risk of poor-swabbing technique and inhibitors becomes non-existent, and therefore laboratory-only evaluations may report higher sensitivity than what would be expected in a clinical setting.

Another group of differences is related to the feasibility of the workflow of the proposed testing platform in a true POC context, including the technique of swab collection and the expertise needed to conduct testing (such as sample preparation, instrument operation, and cleaning). Many tests that claim the possibility of use in a true POC setting would find resistance to adoption due to technical complications, as it is the case of most Loop-mediated isothermal amplification (LAMP) assays (e.g., Yamazaki et al¹⁴⁹). There are also factors related to the particularities of the pandemic scenario and the precautions needed to prevent cross-infections, including the desirable use of viral inactivation methods in samples and the suitability of amplification techniques that have the potential to release amplicons in a clinical setting, thus possibly increasing false-positive results. Furthermore, there are differences between laboratory evaluations and clinical settings; pre-selection of samples and repeating experiments is possible in a laboratory setting, but much harder inside a clinical workflow. In addition, the risk of false-positives in hospital environments is elevated, especially in areas of high movement and turnover like emergency departments, outpatient settings and clinical wards.

We identified 27 antigen tests and 11 NAAT with clinical trials that meet our defined criteria. These 38 assays were covered by 87 studies, with some studies covering more than one test. Considering the high heterogeneity of methods and outcomes between studies, and also the unbalanced number of studies per test, we opted not to conduct a meta-analysis in this study. We decided against providing an ‘average’ performance for tests as this would likely be misleading and would potentially downplay methodological discrepancies between trials.

NAAT, on average, takes longer to provide results and requires more equipment for test operation compared to antigen tests. However, NAAT results appear to be more reliable in clinical practice. Applying selection criteria specifically targeted at prospective studies, we noticed important differences between performance reported by manufacturers, laboratory evaluations and performance in field conditions. While this is true for both NAAT and antigen assays, the discrepancies were more marked in the antigen group. Healthcare facilities, individuals and test providers must be aware of the field performance of the tests before deciding on their implementation. We hope this systematic review can help making informed decisions regarding SARS-CoV-2 testing.

The accuracy of diagnostic tests is affected by numerous factors, including days since symptom onset, individual viral load, quality of sample collection, site of sample collection (nasopharyngeal, nasal only,

pharyngeal only, saliva only) and test modality (nucleic acid amplification or antigen). As previously mentioned, all studies included were compared to PCR or qPCR assays, considered the gold standard for SARS-CoV-2 diagnosis. Results suggest a strong tendency of antigen tests to be less accurate than NAAT in field clinical trials. This finding is aligned with findings of other reviews²⁰.

Importantly, we noticed differences in performance within a given test. For instance, the sensitivity of the antigen assay STANDARD Q COVID-19 Ag Test (SD Biosensor) ranged from 28% to 92%, and the sensitivity of the antigen test PANBIO (Abbott) ranged from 38 to 90%. The explanations for the phenomenon of high variability in results, particularly in sensitivity, are likely multifactorial, including differences in SARS-CoV-2 prevalence between studies and differences in sample size and methodology. One such factor may be testing kits transporting and storage, as most antigen tests need to be stored between 5 and 30C°. Haage et al¹⁵⁰ assessed 11 antigen tests and identified reductions of up to 10 fold in sensitivity for 46% of the assays after 10 minutes outside the ideal temperature range; this number grew to 73% if the exposure lasted three weeks. Pollock et al⁷⁰ reported similar findings evaluating the BinaxNOW test. This finding is significant and may partially explain false negative results, considering that many regions have oscillations in temperature outside the target range and some factors like stock storage and transportation are beyond the end-user control. In this context, Abdul-Mumin et al¹³⁸ reported suboptimal results for both sensitivity and specificity using the STANDARD Q COVID-19 test (64% and 81%, respectively); authors reported the difficulty of maintaining the tests within the ideal temperature range near patients given the temperatures in Ghana, despite following the manufacturer's storage conditions. Therefore, for optimal results, it is important to consider not only the performance of the tests but also context they are being used in.

Point-mutations generating changes in the SARS-CoV-2 nucleocapsid protein structures can also play a role in suboptimal performance; Bourassa et al uncovered a 1000-fold loss in sensitivity for the Sofia Antigen test (Quidel) which was associated with the D399N mutation¹⁵¹. Importantly, the high variability range in results may reflect a publishing bias, as high variability was usually seen in assays with the largest number of published studies.

Additionally, we found some evidence that the sensitivity of antigen tests increases if they are used within the first days of symptoms, but this is still significantly inferior to the average performance of NAAT. For instance, Bulilete et al reported that the sensitivity of the Panbio assay improved from 71.4% to 77.2% if the test was conducted in the first 5 days of symptoms⁸¹. However, using time from symptoms to choose between test modality has several limitations. The definition of being symptomatic and date of first symptom can be subjective, depending on factors such as threshold of perception and memory, which not only may vary between individuals but may also be unreliable. This is especially important in settings such as care-homes, accident and emergency rooms and for populations such as children and cognitively impaired individuals. Furthermore, respiratory compromise is usually a late clinical presentation of COVID and therefore it is reasonable to expect a significant portion of patients to present late to healthcare services, when symptoms worsen. Not only memory of symptom onset may be weakened by then, but a significant portion of frail individuals may also present with delirium. At that stage, timely interventions such as the use of dexamethasone in patients requiring respiratory support will require confirmation of SARS-CoV-2 presence¹⁸, which can be challenging if tests only perform well for a small period of time. In addition, there is evidence that immunosuppressed individuals, who are more likely to have complications of SARS-CoV-2 infection, may have different viral load and shedding patterns, making the use of time criteria for diagnosis even more complicated¹⁵².

Some studies have also provided different performance values for symptomatic and asymptomatic individuals, but a division between those two groups of individuals has similar problems with symptom

recognition. Also, a significant portion of asymptomatic individuals are in fact pre-symptomatic and will develop symptoms in the future, but may already be in the shedding phase; this becomes even more important if the individual has had a high-risk contact.

It appears that around the 8th day after the beginning of symptoms, individuals are not capable of transmitting SARS-CoV-2¹⁵³. It has also been proposed that antigen tests or Ct values of PCR tests could be used as a cut-off for infectivity. This approach assumes that individuals with a negative antigen test or a high Ct value would hardly have a viral load capable of kickstart infection in another individual. While information such as Ct values can be informative, there may be several caveats to the use of Ct values for clinical decision making. Factors such as the quality of swabbing and the tolerance of the patient can add intrinsic variability to estimations of viral load (and thus Ct values), as well as issues with transportation of samples and limitations of the extraction and amplification systems¹⁴⁸. As Binnicker discusses, even assays using the same gene target may have significant variations in Ct value¹⁵⁴, and factors such as the type of sample and swab also need to be considered. Additionally, the volume of buffer the swab is diluted into and the transport media itself can also affect viral load, and it is likely that the quantities of virus in the nasopharynx surfaces are uneven, varying between different regions of the tract and depending on factors such as the ingestion of food and drinks and time. As Basso et al has demonstrated, repeat testing of the same sample can result in different Ct values¹⁵⁵, and this variability has been reported in clinical practice as well¹⁵⁶. It is also unclear which Ct value would be used as a cut-off for infectiousness, as values as different as 24¹⁵⁷ and 34¹⁵⁸ have been proposed in the literature.

Some studies also showcased the implications of using tests with suboptimal specificity in settings of low prevalence. Hoehl et al¹⁵⁹ used an antigen test for self-collected home testing of teachers, with the goal to prevent clusters of infections. Out of a population of 602 individuals, 5 were confirmed positive but 16 false positive results were recorded. The same concern was voiced by Kriemler et al¹⁶⁰ when using antigen tests to assess the point-prevalence of acute SARS-CoV-2 infections in school children. In a study by Colavita et al, of 73,634 individuals tested at international airports, 1176 were reported antigen positive but only 34.3% were confirmed positive. A post-implementation assessment of two antigen tests used for screening of asymptomatic staff (n=71,847) conducted by Kanji et al¹⁶¹ had similar findings.

Regarding kit storage, most platforms will require the use of refrigerated or air-conditioned facilities given the average upper storage limit was 30°C. Some of them deserve mention for requiring strict temperature control, particularly the Cobas Liat (2-8°C), the QIAstat-DX (15-25°C) and Novodiag (18-25°C). Only 5 assays have a published kit storage temperature above 30°C: SAMBA-II (NAAT), VitaPCR (NAAT), HORIBA (NAAT), Sofia SARS-CoV-2 (antigen) and BIOCREREDIT (antigen).

Time to results was highly variable between NAAT, ranging from ~13 minutes (ID Now) to ~85 minutes (HORIBA, SAMBA-II). Antigen tests take an average of 15 minutes for results, not surpassing 30 minutes. Hands-on time, used in this context as the time needed to prepare the test before a run (prepare samples, load machine, configure test conduction) was usually around 1 to 2 minutes, and rarely over 5 minutes across all platforms.

4.2 Other assays

Assays other than NAAT and antigen tests have also been used for COVID-19 diagnosis or SARS-CoV-2 identification. We found a few studies using the FebriDx device (Lumos diagnostics), which captures Myxovirus resistance protein A (MxA - a marker of interferon-induced antiviral host response) and C reactive protein (a well-known and widely used inflammatory marker in medical practice). In one study, the assay had a sensitivity of 93% and a specificity of 86%¹⁶² (with an estimated prevalence of 48% in the studied population). There are other studies available regarding this assay^{163,164,165} but a comprehensive

analysis of this platform goes beyond the purpose of our review. As the markers are commonly elevated for a range of pathogens, the test has a low specificity and has limited use in settings with low prevalence.

Few tests had a satisfactory number of clinical studies, and in many situations the number of individuals enrolled was suboptimal. Further research and reviews of this topic are encouraged.

4.3 Limitations

One of the main limitations of this review is the selection criteria. Considering the high heterogeneity of methods and outcomes between studies, finding a clear-cut unified exclusion criteria was not possible. We debated between authors when in doubt, but a level of subjectivity was inevitable.

For the same reasons of heterogeneity, we opted not to conduct a meta-analysis in this study. Authors have decided against providing an ‘average’ performance for platforms as this would likely be misleading and would potentially downplay the method discrepancies in the trials.

5. Other information

5.1 Registration

This review was registered in the International prospective register of systematic reviews (PROSPERO) with registration number CRD42021260694. A protocol for this study can be assessed online.

5.2 Conflicts of interest

GHH and AH are employed by Diagnostics of the Real World, who owns the SAMBA-II platform mentioned in this open systematic review.

Appendix 1 – comments on individual studies and exclusions

In this section, we comment on criteria for inclusion or exclusion of individual studies. As the criteria for exclusion can often be similar, not all excluded studies will be mentioned. We aim to provide commentary especially when inclusion or exclusion was debated between authors and to give examples of our exclusion pathway.

Multiple studies were excluded in our final screening as tests were conducted using spiked or synthetic samples. The study by Zheng et al¹⁶⁶ evaluating an unnamed lateral flow dipstick assay is an example. Another example is the work of Tanida et al¹⁶⁷, evaluating the ARIES SARS-CoV-2 Assay; this study used patient samples with defined copy numbers and synthetically spiked samples. The study by Shelite et al¹⁶⁸ was not included as it was conducted with processed, archived samples. Details of the platform and its true feasibility for POC placement are unclear.

As the case of most LAMP assays, the saliva LAMP assay proposed by Yamazaki et al¹⁴⁹ was not deemed to be feasible in a true POC scenario as it requires RNA extraction, a heat block and apparently multiple pipetting steps, and therefore a reasonable level of expertise. The assay was also tested with pre-selected samples. The study of Desai et al¹⁶⁹ evaluating the Atila iAMP test and the OptiGene Direct Plus RT-LAMP was not included as platforms were also not considered feasible for POC implementation. The study by Yoshikawa et al¹⁷⁰ and Casati et al¹⁷¹ are also examples of this exclusion criteria. The study by Nörz et al¹⁷² evaluating the Elecsys® SARS-CoV-2 Antigen assay was not deemed feasible in a POC environment given the modules necessary for test conduction are centralised equipment with a large footprint.

Similarly, the study conducted by Peto et al¹⁷³ regarding the LamPORE platform was not included in the analysis as the platform was not deemed to be feasible for POC implementation. RNA needs to be extracted and primers have to be added and incubated with a thermocycler, followed by multiple manual steps. The trial also was conducted with pre-selected, frozen samples. For similar reasons, the study by Singha et al¹⁷⁴ using a glucose meter to detect SARS-CoV-2 was not included, as the test requires a centrifuge, a magnet, and incubation in water baths. This evaluation was also conducted with known positive samples only.

The study by Caffry et al¹⁷⁵ using the Q-POC device was not included as the study was conducted with known-positive/negative frozen samples. Another example is the study by Krause et al¹⁷⁶ regarding the ID now platform, conducted with frozen, pre-processed samples. The study by Hagbom et al¹⁷⁷ was also not included given samples were collected from known-positive patients. Similarly, the study by Muthamia et al¹⁷⁸ evaluating the BD Veritor platform was not included as the analysis was conducted with known-positive patients. Nordgren et al¹⁷⁹ was excluded given test was conducted with pre-selected, known positive samples. Onyilagha et al¹⁸⁰ evaluation of the Biomeme SARS-CoV-2 assay and the Precision Biomonitoring TripleLock SARS-CoV-2 assay was not included given evaluations were made with known positive samples. A platform was also used for extraction, though this was considered to be feasible for POC use. Stokes et al¹⁸¹ evaluation of the PANBIO antigen test was not included as it only recruited known positive individuals.

The study by Renzoni et al¹⁸² evaluating the Visby Medical RT-PCR Portable Device was not included given it was conducted retrospectively, with frozen samples. Agulló et al evaluated nasopharyngeal, nasal only, and saliva samples against nasopharyngeal samples in the Cobas z 480 Analyzer (Roche). Because of this division, sample size ended up being small and heterogeneous. The concordance for positive results was 57.3% for nasopharyngeal samples and as low as 23.1% in saliva; we included the results of

the nasopharyngeal testing in our table. Studies conducted with frozen samples are available for this assay¹⁸³, but as mentioned, were not included in our review.

Basso et al³⁰ tested antigen assays in both saliva and NPS in a mixed population (139 inpatients, 96 outpatients), providing individual figures of sensitivity and specificity for the NPS samples. Since the comparator gold-standard was also tested in saliva and NPS, the ultimate reference standard was unclear. In the case of antigen tests, the detailed number of individuals to give the figures for sensitivity and specificity were not provided in the study or in the supplementary material to the best of our knowledge. We decided to include this study in our table with a commentary pointing towards the fact that the number of individuals used to make the figures was an estimation made for practical purposes and was not provided in the original paper; therefore, it may reflect slightly different patient numbers.

Halfon et al¹⁸⁴ study was not included as samples were pre-selected based on symptom onset and Ct value. Similarly, the study conducted by Jääskeläinen et al¹⁸⁵ evaluating 3 antigen platforms was not included due to the use of pre-selected, known-positive frozen samples. The study conducted by Blairon et al¹⁸⁶ was excluded due to the same reason. Kweon et al¹⁸⁷ also conducted an evaluation of 2 antigen platforms where positive samples were obtained by testing known-positive individuals and pre-selected before the evaluation, and thus was not included in our results.

An unnamed antigen test by R-Biopharm was also evaluated by Toptan et al¹⁸⁸ but this study was not included given that testing was conducted in archived samples only. In line the same criteria, the evaluation of the SIENNA™ COVID-19 Antigen test by Bouassa et al¹⁸⁹ was not included as the comparison was made in the laboratory with frozen samples. The study by Mitchell and George¹⁹⁰ evaluating the ID NOW assay and the study by Assennato et al¹⁹¹ evaluating the SAMBA-II platform were excluded for the same reasons. The study by Stevens et al¹⁹² evaluating the Cepheid Xpert Xpress SARS-CoV-2 assay was also not included given the analysis was made on frozen, pre-selected samples. Consequently, the study by Young et al¹⁹³ assessing BD Veritor and the Sofia 2 SARS Antigen test was excluded as samples were shipped frozen and evaluations happened in a laboratory setting. Miscio et al¹⁹⁴ evaluated the bKIT Virus Finder COVID-19™ (Hyris Ltd); this study used frozen samples which were further manipulated, and therefore was not included in our list of trials. Another example is the study by Panpradist et al¹⁹⁵ evaluating the Harmony COVID-19 platform. The study by Koskinen et al¹⁹⁶ on the MariPOC platform (ArcDia) was not included as it was conducted retrospectively with frozen samples.

Some studies were excluded due to methodology. Hoehl et al¹⁵⁹ study using the RIDA® QUICK SARS-CoV-2 (R-Biopharm) was not included as a minimal number of samples were tested with a confirmation method, and thus false negative results could not be determined. The study by Mlcochova et al¹⁹⁷ was not included in our table because (1) the methodology used frozen pre-selected samples, (2) antibodies were evaluated together with NAAT, making the selected time-frame for analysis questionable as NAAT was used to test samples up to 28 days after symptoms, (3) the criteria used for the reference standard was not entirely clear and (4) the number of samples was small (n=45) and divided between different assays. For similar reasons, the study by Veyrenche¹⁹⁸ et al was not included in our list. A study by Micocci et al⁹ evaluating the LumiraDx antigen test in care homes was not included due to methodology individuals were tested three times a week, twice a week with antigen tests but only once a week with a RT-PCR standard.

After discussion between authors, the study by Olearo et al¹⁹⁹ that evaluated 4 antigen tests was not included in the table for having openly deviated from the manufacturers recommended sample matrix/handling instructions and presenting a sensitivity between 49.4-54.9%, which is on the lower side of what is expected for antigen tests.

The methodology in the study by Hogan et al²⁰⁰ evaluating the Mesa Accula assay (now the Thermo Scientific™ Accula™ SARS-CoV-2 Test after acquisition by Thermo Fisher) was not totally clear, since there is no mention of frozen samples or time to test after sample collection. However, it appears that samples were pre-selected (N=100) and tested in a laboratory after being tested by a centralised PCR assay. We therefore believe that this study, which showed a sensitivity of 68% and a specificity of 100%, is unlikely to accurately reflect results in the field.

A different methodology was used by Rastawicki et al¹²¹ evaluating the PCL COVID-19 Ag rapid fluorescent immunoassay (FIA); 4 swabs were collected in the course of 2 days and antibodies were also evaluated. We considered the comparison between RT-PCR and antigen test straightforward enough for the trial to be included in our table, despite the low number of patients enrolled.

The study by Regev-Yochay²⁰¹ could not be included as multiple antigen platforms were classified as a whole, and individual data for individual assays was not available.

The study by Smithgall et al²⁰² regarding the Cepheid Xpert Xpress and Abbott ID Now was not included as only remnant patient samples were tested; therefore, the platforms were not evaluated in a proper clinical environment. Loeffelholz et al²⁰³ study evaluating the Xpert Xpress SARS-CoV-2 was excluded as all but one site tested the platform with remnant frozen samples. The study by Wolters et al²⁰⁴ on the same platform was also excluded as it only evaluated diluted and processed sample panels in a laboratory setting.

Jokela et al¹⁶ has made two different evaluations of the Novodiag by Mobidiag where, by our understanding, an initial phase was a laboratory evaluation and the second phase was a prospective clinical trial. However, there was a major drop in prevalence in the second phase of the study, which only enabled collection of 5 positive samples in a population of 362 individuals. We included the second phase of this trial in our table despite the low number of positive samples.

Moeren et al⁹⁸ study evaluating the BD Veritor antigen test had a mixed design, where 352 symptomatic adults were evaluated prospectively and known-positive individuals (n = 123) were added to the pool, visiting them at home within 72h of their RT-PCR positive result. Because the assay was tested in a true POC fashion and this was necessary to obtain statistical significance, we decided to include this study in our table. We also reported the specificity considering the 2 false-positive results obtained by the analyser, not considering the eye readings.

The study by Berke et al on the Visby Medical COVID-19 test using pooling strategies in a school setting was not a performance study, as negative pools were not tested against a gold-standard. Therefore, it was not included in our study.

The study by Cassuto et al²⁰⁵ regarding the COVID-□ VIRO® antigen assay was not included given weak-positive samples were excluded from analysis.

The study by Kernéis et al²⁰⁶ was not included as we could not assess the true performance of the antigen test against NAAT. Authors report a positivity rate of 129/1443 for NAAT tests and 89/1115 for antigen tests, but further details are unclear. A reference standard that included saliva samples amplified with NAAT was used later, which further complicated our understanding.

The study by Lamb et al²⁰⁷ was not included as no sensitivity analysis was conducted; only positive individuals were offered a confirmatory PCR test. The study by LeGoff et al²⁰⁸ was not included; the goal was to evaluate a saliva LAMP test, and despite using an antigen test, not all RT-PCR samples had paired antigen samples.

Linder et al²⁰⁹ report on three antigen-detecting tests had no detailed methodology, and therefore was not included. The study by Perkins et al²¹⁰ was not included given only 37 antigen tests were conducted for 853 individuals, thus making a performance analysis not adequate.

The study by Oshiro et al²¹¹ where an antigen test was developed had an unclear methodology, but followed a proof-of-concept model and to the best of our knowledge was not tested in a true POC fashion. Therefore, it was not included. The study by Papadakis et al²¹² was excluded for similar reasons; the evaluation without extracted RNA also used a high proportion of frozen samples compared to fresh samples. The study by Wu et al²¹³ evaluating a newly developed cross-priming isothermal amplification kit (named Kit A) is another example.

The study by Wen et al²¹⁴ evaluating the Xpert Xpress (Cepheid) platform was not included as it was conducted with pre-selected, known-positive frozen samples; the study mentions a small number of tests on fresh specimens (of which 6 were positive) but in addition to the proportionally small number, it is unclear if those specimens were integrated in the final analysis²¹⁴.

Pilarowski et al⁶⁸ has changed the criteria for evaluation of a positive result in the middle of the trial, later recalculating their data. This apparently diverged from manufacturer's instructions for use but was advised by experts by the manufacturer's research staff, and therefore we have included this study.

The study by Šterbenc et al²¹⁵ evaluating the Rapid Antigen Test (Roche) was not included given its methodology was not optimised for an evaluation of performance. Only two positive samples were found in the small population (n = 36), tested three times a week for 2 weeks (191 samples); one of them was considered the tail capture of RNA (the individual was positive for IgG previously) and the other was missed by the antigen test, which would give a sensitivity of 0%.

The study conducted by Tulloch et al²¹⁶ evaluating the Innova SARS-CoV-2 antigen test was not optimised to be a performance study; only 828 of the 1638 samples had matching PCR tests. We therefore excluded this study as the gold-standard was not employed systematically.

Osmanodja et al¹⁴⁵ have trialed a test that was generically called 'Novel SARS-CoV-2 Antigen-Detecting Rapid Diagnostic Test', reporting a sensitivity of 88.6% and a specificity of 99.7%. Because we could not find further details on this platform, we have opted to not include it in the comparison tables.

References

1. Weekly statistics for NHS Test and Trace (England): 24 June to 30 June 2021. *GOV.UK* <https://www.gov.uk/government/publications/weekly-statistics-for-nhs-test-and-trace-england-24-june-to-30-june-2021>.
2. Test directory. *FIND* <https://www.finddx.org/test-directory/>.
3. Kozel, T. R. & Burnham-Marusich, A. R. Point-of-Care Testing for Infectious Diseases: Past, Present, and Future. *J Clin Microbiol* **55**, 2313–2320 (2017).
4. Rebel, A., Rice, M. A. & Fahy, B. G. Accuracy of point-of-care glucose measurements. *J Diabetes Sci Technol* **6**, 396–411 (2012).
5. Allardet-Servent, J. *et al.* Point-of-Care Versus Central Laboratory Measurements of Hemoglobin, Hematocrit, Glucose, Bicarbonate and Electrolytes: A Prospective Observational Study in Critically Ill Patients. *PLoS One* **12**, (2017).
6. Mitamura, K. *et al.* Clinical evaluation of ID NOW influenza A & B 2, a rapid influenza virus detection kit using isothermal nucleic acid amplification technology - A comparison with currently available tests. *J Infect Chemother* **26**, 216–221 (2020).
7. Ritchie, A. V. *et al.* SAMBA HIV Semiquantitative Test, a New Point-of-Care Viral-Load-Monitoring Assay for Resource-Limited Settings. *Journal of Clinical Microbiology* **52**, 3377–3383 (2014).
8. Harding-Esch, E. M. *et al.* A 30-Min Nucleic Acid Amplification Point-of-Care Test for Genital Chlamydia trachomatis Infection in Women: A Prospective, Multi-center Study of Diagnostic Accuracy. *EBioMedicine* **28**, 120–127 (2018).
9. Micocci, M. *et al.* Point of Care Testing using rapid automated Antigen Testing for SARS-COV-2 in Care Homes – an exploratory safety, usability and diagnostic agreement evaluation. 2021.04.22.21255948 <https://www.medrxiv.org/content/10.1101/2021.04.22.21255948v2> (2021) doi:10.1101/2021.04.22.21255948.
10. Commissioner, O. of the. Coronavirus (COVID-19) Update: FDA Informs Public About Possible Accuracy Concerns with Abbott ID NOW Point-of-Care Test. *FDA* <https://www.fda.gov/news-events/press-announcements/coronavirus-covid-19-update-fda-informs-public-about-possible-accuracy-concerns-abbott-id-now-point> (2020).
11. McGinley, L. & Abutaleb, Y. Trump administration bars FDA from regulating some laboratory tests, including for coronavirus. *Washington Post*.
12. Health, C. for D. and R. Removal Lists of Tests that Should No Longer Be Used and/or Distributed for COVID-19: FAQs on Testing for SARS-CoV-2. *FDA* (2020).
13. Health, C. for D. and R. In Vitro Diagnostics EUAs - Molecular Diagnostic Tests for SARS-CoV-2. *FDA* (2021).
14. Harrington, A. *et al.* Comparison of Abbott ID Now and Abbott m2000 Methods for the Detection of SARS-CoV-2 from Nasopharyngeal and Nasal Swabs from Symptomatic Patients. *J Clin Microbiol* **58**, (2020).

15. Lephart, P. R. *et al.* Comparative study of four SARS-CoV-2 Nucleic Acid Amplification Test (NAAT) platforms demonstrates that ID NOW performance is impaired substantially by patient and specimen type. *Diagnostic Microbiology and Infectious Disease* **99**, 115200 (2021).
16. Jokela, P. *et al.* SARS-CoV-2 sample-to-answer nucleic acid testing in a tertiary care emergency department: evaluation and utility. *Journal of Clinical Virology* **131**, 104614 (2020).
17. Collier, D. A. *et al.* Point of Care Nucleic Acid Testing for SARS-CoV-2 in Hospitalized Patients: A Clinical Validation Trial and Implementation Study. *CR Med* **1**, (2020).
18. Dexamethasone in Hospitalized Patients with Covid-19. *New England Journal of Medicine* **384**, 693–704 (2021).
19. Interleukin-6 Receptor Antagonists in Critically Ill Patients with Covid-19. *New England Journal of Medicine* **384**, 1491–1502 (2021).
20. Dinnes, J. *et al.* Rapid, point-of-care antigen and molecular-based tests for diagnosis of SARS-CoV-2 infection. *Cochrane Database of Systematic Reviews* (2021) doi:10.1002/14651858.CD013705.pub2.
21. Yoon, S. H. *et al.* Point-of-care testing for the detection of SARS-CoV-2: a systematic review and meta-analysis. *Eur Rev Med Pharmacol Sci* **25**, 503–517 (2021).
22. Hayer, J., Kasapic, D. & Zemmrich, C. Real-world clinical performance of commercial SARS-CoV-2 rapid antigen tests in suspected COVID-19: A systematic meta-analysis of available data as per November 20, 2020. *International Journal of Infectious Diseases* **0**, (2021).
23. Tu, Y.-P., Iqbal, J. & O'Leary, T. Evaluation of ID NOW and RT-PCR for Detection of SARS-CoV-2 in an Ambulatory Population. *medRxiv* 2020.12.07.20245225 (2020) doi:10.1101/2020.12.07.20245225.
24. Corman, V. M. *et al.* Comparison of seven commercial SARS-CoV-2 rapid point-of-care antigen tests: a single-centre laboratory evaluation study. *The Lancet Microbe* **2**, e311–e319 (2021).
25. Cerutti, F. *et al.* Urgent need of rapid tests for SARS CoV-2 antigen detection: Evaluation of the SD-Biosensor antigen test for SARS-CoV-2. *J Clin Virol* **132**, 104654 (2020).
26. Courtellemont, L. *et al.* High performance of a novel antigen detection test on nasopharyngeal specimens for diagnosing SARS-CoV-2 infection. *J Med Virol* 10.1002/jmv.26896 (2021) doi:10.1002/jmv.26896.
27. Ghofrani, M. *et al.* Performance characteristics of the ID NOW COVID-19 assay: A regional health care system experience. *medRxiv* 2020.06.03.20116327 (2020) doi:10.1101/2020.06.03.20116327.
28. Chen, J. H.-K. *et al.* Evaluating the use of posterior oropharyngeal saliva in a point-of-care assay for the detection of SARS-CoV-2. *Emerg Microbes Infect* **9**, 1356–1359 (2020).
29. SoRelle, J. A. *et al.* Saliva for use with a point of care assay for the rapid diagnosis of COVID-19. *Clin Chim Acta* **510**, 685–686 (2020).
30. Basso, D. *et al.* Salivary SARS-CoV-2 antigen rapid detection: A prospective cohort study. *Clinica Chimica Acta* **517**, 54–59 (2021).

31. Agulló, V. *et al.* Evaluation of the rapid antigen test Panbio COVID-19 in saliva and nasal swabs in a population-based point-of-care study. *J Infect* **82**, 186–230 (2021).
32. Lindner, A. K. *et al.* SARS-CoV-2 patient self-testing with an antigen-detecting rapid test: a head-to-head comparison with professional testing. *medRxiv* 2021.01.06.20249009 (2021) doi:10.1101/2021.01.06.20249009.
33. Lindner, A. K. *et al.* Head-to-head comparison of SARS-CoV-2 antigen-detecting rapid test with self-collected nasal swab versus professional-collected nasopharyngeal swab. *European Respiratory Journal* **57**, (2021).
34. Basu, A. *et al.* Performance of Abbott ID Now COVID-19 Rapid Nucleic Acid Amplification Test Using Nasopharyngeal Swabs Transported in Viral Transport Media and Dry Nasal Swabs in a New York City Academic Institution. *Journal of Clinical Microbiology* **58**, (2020).
35. Agarwal, J., Das, A., Pandey, P., Sen, M. & Garg, J. “David vs. Goliath”: A simple antigen detection test with potential to change diagnostic strategy for SARS-CoV-2. *The Journal of Infection in Developing Countries* **15**, 904–909 (2021).
36. Fournier, P.-E. *et al.* Contribution of VitaPCR SARS-CoV-2 to the emergency diagnosis of COVID-19. *J Clin Virol* **133**, 104682 (2020).
37. González-Donapetry, P. *et al.* Think of the Children: Evaluation of SARS-CoV-2 Rapid Antigen Test in Pediatric Population. *The Pediatric Infectious Disease Journal* **40**, 385–388 (2021).
38. Osterman, A. *et al.* Evaluation of two rapid antigen tests to detect SARS-CoV-2 in a hospital setting. *Med Microbiol Immunol* **210**, 65–72 (2021).
39. Martí, J. L. G. *et al.* Differences in detected viral loads guide use of SARS-CoV-2 antigen-detection assays towards symptomatic college students and children. *medRxiv* 2021.01.28.21250365 (2021) doi:10.1101/2021.01.28.21250365.
40. Peña, M. *et al.* Performance of SARS-CoV-2 rapid antigen test compared with real-time RT-PCR in asymptomatic individuals. *medRxiv* 2021.02.12.21251643 (2021) doi:10.1101/2021.02.12.21251643.
41. Ford, L. *et al.* *Characteristics of children and antigen test performance at a SARS-CoV-2 community testing site.* 2021.07.06.21259792
<https://www.medrxiv.org/content/10.1101/2021.07.06.21259792v1> (2021)
doi:10.1101/2021.07.06.21259792.
42. Real-time COVID-19 testing | Cue. <https://www.cuehealth.com/products/how-cue-detects-covid-19/>.
43. Mahmoud, S. A. *et al.* *Evaluation of seven different rapid methods for nucleic acid detection of SARS-COV-2 virus.* 2021.04.15.21255533
<https://www.medrxiv.org/content/10.1101/2021.04.15.21255533v1> (2021)
doi:10.1101/2021.04.15.21255533.
44. Hansen, G. *et al.* Clinical Performance of the Point-of-Care cobas Liat for Detection of SARS-CoV-2 in 20 Minutes: a Multicenter Study. *J Clin Microbiol* **59**, (2021).

45. Gibani, M. M. *et al.* Assessing a novel, lab-free, point-of-care test for SARS-CoV-2 (CovidNudge): a diagnostic accuracy study. *The Lancet Microbe* **1**, e300–e307 (2020).
46. Donato, L. J. *et al.* Evaluation of the Cue Health point-of-care COVID-19 (SARS-CoV-2 nucleic acid amplification) test at a community drive through collection center. *Diagn Microbiol Infect Dis* **100**, 115307 (2021).
47. Leong, K. W. *et al.* Excellent negative predictive value (99.8%) of two rapid molecular COVID-19 tests compared to conventional RT-PCR for SARS-CoV-2 (COVID-19) in 2,011 tests performed in a single centre. 2021.06.20.21258392 <https://www.medrxiv.org/content/10.1101/2021.06.20.21258392v1> (2021) doi:10.1101/2021.06.20.21258392.
48. Cradic, K. *et al.* Clinical Evaluation and Utilization of Multiple Molecular In Vitro Diagnostic Assays for the Detection of SARS-CoV-2. *American Journal of Clinical Pathology* **154**, 201–207 (2020).
49. Van, J.-C. N. *et al.* Prospective evaluation of ID NOW COVID-19 assay used as point-of-care test in an Emergency Department. 2021.03.29.21253909 <https://www.medrxiv.org/content/10.1101/2021.03.29.21253909v1> (2021) doi:10.1101/2021.03.29.21253909.
50. DnaNudge: CovidNudge: rapid, lab-free COVID-19 test. <https://www.dnanudge.com/en/COVID-Nudge>.
51. Health, C. for D. and R. Potential for False Results with Roche Molecular Systems, Inc. cobas SARS-CoV-2 & Influenza Test for use on cobas Liat System-Letter to Clinical Laboratory Staff, Point-of-Care Facility Staff, and Health Care Providers. *FDA* (2021).
52. Coronavirus COVID-19 serology and viral detection tests: technical validation reports. *GOV.UK* <https://www.gov.uk/government/publications/coronavirus-covid-19-serology-and-viral-detection-tests-technical-validation-reports>.
53. Coronavirus COVID-19 serology and viral detection tests: technical validation reports. *GOV.UK* <https://www.gov.uk/government/publications/coronavirus-covid-19-serology-and-viral-detection-tests-technical-validation-reports>.
54. Micocci, M. *et al.* Is point-of-care testing feasible and safe in care homes in England? An exploratory usability and accuracy evaluation of a point-of-care polymerase chain reaction test for SARS-CoV-2. *Age Ageing* **50**, 1464–1472 (2021).
55. Visseaux, B. *et al.* Evaluation of the QIAstat-Dx Respiratory SARS-CoV-2 Panel, the First Rapid Multiplex PCR Commercial Assay for SARS-CoV-2 Detection. *J Clin Microbiol* **58**, (2020).
56. MOBIDIAG. Instructions for Use for the Novodiag® System. (2020).
57. Mobidiag - Molecular diagnostics of coronavirus infection. *Mobidiag* <https://mobidiag.com/products/coronavirus/>.
58. Horiba Diagnostics. POKKIT Central. <https://www.horiba.com/uk/medical/products/point-of-care/pockit-central/pockit-central-details/pockit-central-46105/>.

59. QIAstat-Dx Analyzer 1.0 User Manual - QIAGEN.
<https://www.qiagen.com/us/resources/resourcedetail?id=658139de-0b59-4801-a1fd-92d4cd75da3f&lang=en>.
60. Österdahl, M. F. *et al.* Detecting SARS-CoV-2 at point of care: preliminary data comparing loop-mediated isothermal amplification (LAMP) to polymerase chain reaction (PCR). *BMC Infectious Diseases* **20**, 783 (2020).
61. Fitoussi, F., Dupont, R., Tonen-Wolyec, S. & Bélec, L. Performances of the VitaPCR™ SARS-CoV-2 Assay during the second wave of the COVID-19 epidemic in France. *Journal of Medical Virology* **n/a**,.
62. Diagnostics for the Real World. SAMBA II SARS-CoV-2 Test Instructions for use. (2021).
63. VitaPCRTM Platform. <https://www.mediconire.com/vitapcrtm-platform>.
64. Xpert® Xpress SARS-CoV-2 has received FDA Emergency Use Authorization.
<https://www.cepheid.com/coronavirus>.
65. Member Press Release - MICROSENSDX RapiPRO™ LAMP for SARS-CoV-2 RECEIVES CE MARK APPROVAL. BIVDA - The British In Vitro Diagnostics Association <https://www.bivda.org.uk/News-Events/BIVDA-Press-Releases/ArticleID/529/Member-Press-Release-MICROSENSDX-RapiPRO™-LAMP-for-SARS-CoV-2-RECEIVES-CE-MARK-APPROVAL>.
66. Products Archive. *MicrosensDx* <https://www.microsensdx.com/product/>.
67. Schwob, J. M. *et al.* Antigen rapid tests, nasopharyngeal PCR and saliva PCR to detect SARS-CoV-2: a prospective comparative clinical trial. *medRxiv* 2020.11.23.20237057 (2020)
doi:10.1101/2020.11.23.20237057.
68. Pilarowski, G. *et al.* Performance Characteristics of a Rapid Severe Acute Respiratory Syndrome Coronavirus 2 Antigen Detection Assay at a Public Plaza Testing Site in San Francisco. *The Journal of Infectious Diseases* **223**, 1139–1144 (2021).
69. Frediani, J. K. *et al.* Multidisciplinary assessment of the Abbott BinaxNOW SARS-CoV-2 point-of-care antigen test in the context of emerging viral variants and self-administration. *Sci Rep* **11**, 14604 (2021).
70. Pollock, N. R. *et al.* Performance and Implementation Evaluation of the Abbott BinaxNOW Rapid Antigen Test in a High-Throughput Drive-Through Community Testing Site in Massachusetts. *J Clin Microbiol* **59**, (2021).
71. Baro, B. *et al.* Performance characteristics of five antigen-detecting rapid diagnostic test (Ag-RDT) for SARS-CoV-2 asymptomatic infection: a head-to-head benchmark comparison. *medRxiv* 2021.02.11.21251553 (2021) doi:10.1101/2021.02.11.21251553.
72. Torres, I., Poujois, S., Albert, E., Colomina, J. & Navarro, D. Evaluation of a rapid antigen test (Panbio™ COVID-19 Ag rapid test device) for SARS-CoV-2 detection in asymptomatic close contacts of COVID-19 patients. *Clinical Microbiology and Infection* **27**, 636.e1-636.e4 (2021).
73. Wagenhäuser, I. *et al.* Clinical performance evaluation of SARS-CoV-2 rapid antigen testing in point of care usage in comparison to RT-qPCR. *EBioMedicine* **69**, (2021).

74. Thirion-Romero, I. *et al.* *Evaluation of a Rapid Antigen Test for Sars-Cov-2 in Symptomatic Patients and Their Contacts: A Multicenter Study*. 2021.05.24.21257020
<https://www.medrxiv.org/content/10.1101/2021.05.24.21257020v1> (2021)
doi:10.1101/2021.05.24.21257020.
75. Bruins, M. J., Santos, C. O. dos, Spoelman-Lunsche, M., Bos-Kromhout, M. I. van den & Debast, S. B. *Evaluation of the Panbio™ rapid antigen test for COVID-19 diagnosis in symptomatic health care workers*. 2021.06.21.21259234 <https://www.medrxiv.org/content/10.1101/2021.06.21.21259234v1> (2021) doi:10.1101/2021.06.21.21259234.
76. Masiá, M. *et al.* Nasopharyngeal Panbio COVID-19 Antigen Performed at Point-of-Care Has a High Sensitivity in Symptomatic and Asymptomatic Patients With Higher Risk for Transmission and Older Age. *Open Forum Infectious Diseases* **8**, (2021).
77. Ferté, T. *et al.* Accuracy of COVID-19 rapid antigenic tests compared to RT-PCR in a student population: The StudyCov study. *J Clin Virol* **141**, 104878 (2021).
78. L'Huillier, A. G. *et al.* *Diagnostic accuracy of SARS-CoV-2 rapid antigen detection testing in symptomatic and asymptomatic children in the clinical setting*. 2021.04.15.21255577
<https://www.medrxiv.org/content/10.1101/2021.04.15.21255577v1> (2021)
doi:10.1101/2021.04.15.21255577.
79. Drevinek, P. *et al.* The sensitivity of SARS-CoV-2 antigen tests in the view of large-scale testing. *medRxiv* 2020.11.23.20237198 (2020) doi:10.1101/2020.11.23.20237198.
80. Carbonell-Sahuquillo, S. *et al.* Evaluation of a rapid antigen detection test (Panbio™ COVID-19 Ag Rapid Test Device) as a point-of-care diagnostic tool for COVID-19 in a pediatric emergency department. *Journal of Medical Virology* **93**, 6803–6807 (2021).
81. Bulilete, O. *et al.* Evaluation of the Panbio™ rapid antigen test for SARS-CoV-2 in primary health care centers and test sites. *medRxiv* 2020.11.13.20231316 (2020) doi:10.1101/2020.11.13.20231316.
82. Linares, M. *et al.* Panbio antigen rapid test is reliable to diagnose SARS-CoV-2 infection in the first 7 days after the onset of symptoms. *J Clin Virol* **133**, 104659 (2020).
83. Landaas, E. T. *et al.* Diagnostic performance of a SARS-CoV-2 rapid antigen test in a large, Norwegian cohort. *J Clin Virol* **137**, 104789 (2021).
84. Albert, E. *et al.* Field evaluation of a rapid antigen test (Panbio™ COVID-19 Ag Rapid Test Device) for COVID-19 diagnosis in primary healthcare centres. *Clin Microbiol Infect* **27**, 472.e7-472.e10 (2021).
85. Yin, N. *et al.* SARS-CoV-2 Diagnostic Tests: Algorithm and Field Evaluation From the Near Patient Testing to the Automated Diagnostic Platform. *Front Med (Lausanne)* **8**, 650581 (2021).
86. Abdulrahman, A. *et al.* Comparison of SARS-COV-2 nasal antigen test to nasopharyngeal RT-PCR in mildly symptomatic patients. *medRxiv* 2020.11.10.20228973 (2020)
doi:10.1101/2020.11.10.20228973.

87. Faico-Filho, K. S. *et al.* Evaluation of the Panbio™ COVID-19 Ag Rapid Test at an Emergency Room in a Hospital in São Paulo, Brazil. *medRxiv* 2021.03.15.21253313 (2021) doi:10.1101/2021.03.15.21253313.
88. Berger, A. *et al.* Diagnostic accuracy of two commercial SARS-CoV-2 Antigen-detecting rapid tests at the point of care in community-based testing centers. *medRxiv* 2020.11.20.20235341 (2020) doi:10.1101/2020.11.20.20235341.
89. Kolwijck, E. *et al.* Validation and implementation of the Panbio COVID-19 Ag rapid test for the diagnosis of SARS-CoV-2 infection in symptomatic hospital healthcare workers. *Infect Prev Pract* **3**, 100142 (2021).
90. Krüger, L. J. *et al.* The Abbott PanBio WHO emergency use listed, rapid, antigen-detecting point-of-care diagnostic test for SARS-CoV-2-Evaluation of the accuracy and ease-of-use. *PLoS One* **16**, e0247918 (2021).
91. Merino, P. *et al.* Multicenter evaluation of the Panbio™ COVID-19 rapid antigen-detection test for the diagnosis of SARS-CoV-2 infection. *Clinical Microbiology and Infection* **27**, 758–761 (2021).
92. COVID-VIRO | COVID-19 AAZ. <https://www.covid19aaz.com/en/rapide-antigenic-test/> (2020).
93. Abbott. Panbio™ COVID-19 Ag Rapid Test Device. (2021).
94. Suliman, S. *et al.* Evaluation of the Access Bio CareStart™ rapid SARS-CoV-2 antigen test in asymptomatic individuals tested at a community mass-testing program in Western Massachusetts. 2021.06.17.21259109 <https://www.medrxiv.org/content/10.1101/2021.06.17.21259109v1> (2021) doi:10.1101/2021.06.17.21259109.
95. Pollock, N. R. *et al.* Performance and Operational Evaluation of the Access Bio CareStart Rapid Antigen Test in a High-Throughput Drive-Through Community Testing Site in Massachusetts. *Open Forum Infectious Diseases* **8**, (2021).
96. Bonde, J. *et al.* Clinical validation of point-of-care SARS-COV-2 BD Veritor antigen test by a single throat swab for rapid COVID-19 status on hospital patients predominantly without overt COVID symptoms. 2021.04.12.21255299 <https://www.medrxiv.org/content/10.1101/2021.04.12.21255299v1> (2021) doi:10.1101/2021.04.12.21255299.
97. Kilic, A., Hiestand, B. & Palavecino, E. Evaluation of Performance of the BD Veritor SARS-CoV-2 Chromatographic Immunoassay Test in Patients with Symptoms of COVID-19. *J Clin Microbiol* **59**, e00260-21 (2021).
98. Moeren, N. V. der *et al.* PERFORMANCE EVALUATION OF A SARS-COV-2 RAPID ANTIGENTEST: TEST PERFORMANCE IN THE COMMUNITY IN THE NETHERLANDS. *medRxiv* 2020.10.19.20215202 (2020) doi:10.1101/2020.10.19.20215202.
99. Krüger, L. J. *et al.* Evaluation of the accuracy, ease of use and limit of detection of novel, rapid, antigen-detecting point-of-care diagnostics for SARS-CoV-2. *medRxiv* 2020.10.01.20203836 (2020) doi:10.1101/2020.10.01.20203836.

100. Onsongo, S. N. *et al.* Field performance of NowCheck rapid antigen test for SARS-CoV-2 in Kisumu County, western Kenya. 2021.08.12.21261462
<https://www.medrxiv.org/content/10.1101/2021.08.12.21261462v1> (2021)
doi:10.1101/2021.08.12.21261462.
101. Cheung, N. N. *et al.* On-site rapid molecular testing, mobile sampling teams and eHealth to support primary care physicians during the COVID-19 pandemic. *medRxiv* 2020.11.20.20234898 (2020)
doi:10.1101/2020.11.20.20234898.
102. Lambert-Niclot, S. *et al.* Evaluation of a Rapid Diagnostic Assay for Detection of SARS-CoV-2 Antigen in Nasopharyngeal Swabs. *J Clin Microbiol* **58**, (2020).
103. Health, C. for D. and R. In Vitro Diagnostics EUAs - Antigen Diagnostic Tests for SARS-CoV-2. *FDA* (2021).
104. 2019-Novel Coronavirus (2019-nCoV) Antigen Rapid Test Kit (FIA). *Bioeasy*
<http://en.bioeasy.com/?p=1038> (2020).
105. NowCheck COVID-19 Ag. 바이오토티
http://bionote.co.kr:/eng/board/now_check/board_view.asp?num=2022.
106. CORIS. COVID-19 Ag Respi-Strip. (2020).
107. COVID-19 Rapid Antigen Test | BD Veritor™ Plus System. *BD Veritor*
<https://bdveritor.bd.com/en-us/rapid-antigen-testing/covid-19>.
108. Kiyasu, Y. *et al.* Prospective analytical performance evaluation of the QuickNavi™-COVID19 Ag for asymptomatic individuals. 2021.04.01.21254813
<https://www.medrxiv.org/content/10.1101/2021.04.01.21254813v1> (2021)
doi:10.1101/2021.04.01.21254813.
109. Takeuchi, Y. *et al.* The evaluation of a newly developed antigen test (QuickNavi™-COVID19 Ag) for SARS-CoV-2: A prospective observational study in Japan. *Journal of Infection and Chemotherapy* **27**, 890–894 (2021).
110. García-Fiñana, M. *et al.* Performance of the Innova SARS-CoV-2 antigen rapid lateral flow test in the Liverpool asymptomatic testing pilot: population based cohort study. *BMJ* **374**, n1637 (2021).
111. Houston, H., Gupta-Wright, A., Toke-Bjølgerud, E., Biggin-Lamming, J. & John, L. Diagnostic accuracy and utility of SARS-CoV-2 antigen lateral flow assays in medical admissions with possible COVID-19. *Journal of Hospital Infection* **110**, 203–205 (2021).
112. Krüger, L. J. *et al.* Evaluation of accuracy, exclusivity, limit-of-detection and ease-of-use of LumiraDx™: An antigen-detecting point-of-care device for SARS-CoV-2. *Infection* (2021)
doi:10.1007/s15010-021-01681-y.
113. Bianco, G. *et al.* Evaluation of an antigen-based test for hospital point-of-care diagnosis of SARS-CoV-2 infection. *J Clin Virol* **139**, 104838 (2021).

114. Drain, P. K. *et al.* A Rapid, High-Sensitivity SARS-CoV-2 Nucleocapsid Immunoassay to Aid Diagnosis of Acute COVID-19 at the Point of Care: A Clinical Performance Study. *Infect Dis Ther* **10**, 753–761 (2021).
115. Quick NaviTM-COVID 19 Ag Rapid Diagnostic Test in Japan Now Enables Test Sample Collection from Front of Nasal Cavity | News Releases. *Otsuka Pharmaceutical Co., Ltd.*
https://www.otsuka.co.jp/en/company/newsreleases/2020/20201005_1.html.
116. Fujirebio. ESPLINE SARS-CoV-2. (2021).
117. Innova. SARS-CoV-2 Antigen Rapid Qualitative Test Instructions for Use. (2020).
118. Lepu Medical. SARS-CoV-2 Antigen Rapid Test Kit (Colloidal Gold Immunochromatography). (2020).
119. Health, C. for D. and R. Lepu Medical Technology Recalls SARS-CoV-2 Antigen Rapid Test Kit and Leccurate SARS-CoV-2 Antibody Rapid Test Kit (Colloidal Gold Immunochromatography) due to Risk of False Results. *FDA* (2021).
120. Kurihara, Y. *et al.* The evaluation of a novel digital immunochromatographic assay with silver amplification to detect SARS-CoV-2. 2021.05.06.21256738
<https://www.medrxiv.org/content/10.1101/2021.05.06.21256738v1> (2021)
doi:10.1101/2021.05.06.21256738.
121. Rastawicki, W., Gierczyński, R., Juszczak, G., Mitura, K. & Henry, B. M. Evaluation of PCL rapid point of care antigen test for detection of SARS-CoV-2 in nasopharyngeal swabs. *J Med Virol* **93**, 1920–1922 (2021).
122. Chiu, R. Y. T. *et al.* Evaluation of the INDICAID™ COVID-19 Rapid Antigen Test in symptomatic populations and asymptomatic community testing. 2021.05.26.21257063
<https://www.medrxiv.org/content/10.1101/2021.05.26.21257063v1> (2021)
doi:10.1101/2021.05.26.21257063.
123. MEDsan. MEDsan SARS-CoV-2 Antigen Rapid Test. (2021).
124. ApS, A., Lora Grafisk. PCL Covid19 Ag Gold Saliva. *Vitrex* <https://vitrexmedical.com/covid-19-products/covid-19-test-kits/pcl-covid-19-ag-gold-saliva/>.
125. Bornemann, L. *et al.* Real-life evaluation of the Sofia SARS-CoV-2 antigen assay in a large tertiary care hospital. *J Clin Virol* **140**, 104854 (2021).
126. Shidlovskaya, E. V. *et al.* The Value of Rapid Antigen Tests to Identify Carriers of Viable SARS-CoV-2. *medRxiv* 2021.03.10.21252667 (2021) doi:10.1101/2021.03.10.21252667.
127. Salvagno, G. L., Gianfilippi, G., Bragantini, D., Henry, B. M. & Lippi, G. Clinical assessment of the Roche SARS-CoV-2 rapid antigen test. *Diagnosis* **8**, 322–326 (2021).
128. Thell, R. *et al.* Evaluation of a novel, rapid antigen detection test for the diagnosis of SARS-CoV-2. 2021.04.22.21255637 <https://www.medrxiv.org/content/10.1101/2021.04.22.21255637v1> (2021)
doi:10.1101/2021.04.22.21255637.

129. Koeleman, J. G. M., Brand, H., de Man, S. J. & Ong, D. S. Y. Clinical evaluation of rapid point-of-care antigen tests for diagnosis of SARS-CoV-2 infection. *Eur J Clin Microbiol Infect Dis* **40**, 1975–1981 (2021).
130. Sofia SARS Antigen FIA | Quidel. <https://www.quidel.com/immunoassays/rapid-sars-tests/sofia-sars-antigen-fia>.
131. Biocredit. COVID-19 Ag. (2020).
132. Roche SARS-CoV 2 Rapid Antigen Test x 25 (For COVID-19). *UK Medisave UK View* <https://www.medisave.co.uk/roche-sars-cov-2-rapid-anti-gen-test-kit.html>.
133. BV, G. Romed SARS-CoV-2 Antigen Rapid Test Cassette (Swab), COVID-RAPID-AG25. *Stentor 6.5 Demo* <https://www.romed.nl/en/products/corona-rapid-test/corona-rapid-test/sars-cov-2-antigen-rapid-test/romed-sars-cov-2-antigen-rapid-test-cassette-swab-covid-rapid-ag25/groups/g+c+a+nr+view>.
134. Sofia 2 Fluorescent Immunoassay Analyzer | Quidel. <https://www.quidel.com/immunoassays/sofia-tests-kits/sofia-2-analyzer>.
135. Caruana, G. *et al.* The dark side of SARS-CoV-2 rapid antigen testing: screening asymptomatic patients. *New Microbes and New Infections* **42**, 100899 (2021).
136. Dankova, Z. *et al.* Comparison of SARS-CoV-2 Detection by Rapid Antigen and by Three Commercial RT-qPCR Tests: A Study from Martin University Hospital in Slovakia. *Int J Environ Res Public Health* **18**, 7037 (2021).
137. Korenkov, M. *et al.* *Assessment of SARS-CoV-2 infectivity by a Rapid Antigen Detection Test*. 2021.03.30.21254624 <https://www.medrxiv.org/content/10.1101/2021.03.30.21254624v2> (2021) doi:10.1101/2021.03.30.21254624.
138. Abdul-Mumin, A. *et al.* *Field evaluation of specificity and sensitivity of a standard SARS-CoV-2 antigen rapid diagnostic test: A prospective study at a teaching hospital in Northern Ghana*. 2021.06.03.21258300 <https://www.medrxiv.org/content/10.1101/2021.06.03.21258300v2> (2021) doi:10.1101/2021.06.03.21258300.
139. Igliński, Z. *et al.* Clinical Evaluation of Roche SD Biosensor Rapid Antigen Test for SARS-CoV-2 in Municipal Health Service Testing Site, the Netherlands - Volume 27, Number 5—May 2021 - Emerging Infectious Diseases journal - CDC. doi:10.3201/eid2705.204688.
140. Suzuki, H. *et al.* *Diagnostic performance of a novel digital immunoassay (RapidTesta SARS-CoV-2): a prospective observational study with 1,127 nasopharyngeal samples*. 2021.07.26.21261162 <https://www.medrxiv.org/content/10.1101/2021.07.26.21261162v2> (2021) doi:10.1101/2021.07.26.21261162.
141. Torres, I. *et al.* Point-of-care evaluation of a rapid antigen test (CLINITEST® Rapid COVID-19 Antigen Test) for diagnosis of SARS-CoV-2 infection in symptomatic and asymptomatic individuals. *Journal of Infection* **82**, e11–e12 (2021).
142. SD Biosensor, Inc. STANDARD Q COVID-19 Ag Test. (2020).
143. SIEMENS. CLINITEST® Rapid COVID-19 Antigen Test. (2021).

144. Sugentech. <http://www.sugentech.com/>.
145. Osmanodja, B. *et al.* Accuracy of a Novel SARS-CoV-2 Antigen-Detecting Rapid Diagnostic Test from Standardized Self-Collected Anterior Nasal Swabs. *J Clin Med* **10**, 2099 (2021).
146. Christensen, K., Ren, H., Chen, S., Cooper, C. K. & Young, S. *Clinical evaluation of BD Veritor™ SARS-CoV-2 and Flu A+B Assay for point-of-care (POC) System*. 2021.05.04.21256323 <https://www.medrxiv.org/content/10.1101/2021.05.04.21256323v1> (2021) doi:10.1101/2021.05.04.21256323.
147. BD Announces CE Mark of Portable, Rapid Point-of-Care Antigen Test to Detect SARS-CoV-2 in 15 minutes - BD. <https://investors.bd.com/news-releases/news-release-details/bd-announces-ce-mark-portable-rapid-point-care-antigen-test>.
148. Dahdouh, E., Lázaro-Perona, F., Romero-Gómez, M. P., Mingorance, J. & García-Rodríguez, J. Ct values from SARS-CoV-2 diagnostic PCR assays should not be used as direct estimates of viral load. *J Infect* **82**, 414–451 (2021).
149. Yamazaki, W., Matsumura, Y., Thongchankaew-Seo, U., Yamazaki, Y. & Nagao, M. Development of a point-of-care test to detect SARS-CoV-2 from saliva which combines a simple RNA extraction method with colorimetric reverse transcription loop-mediated isothermal amplification detection. *Journal of Clinical Virology* **136**, 104760 (2021).
150. Haage, V. *et al.* Impaired performance of SARS-CoV-2 antigen-detecting rapid tests at elevated temperatures. <http://medrxiv.org/lookup/doi/10.1101/2021.01.06.21249314> (2021) doi:10.1101/2021.01.06.21249314.
151. Bourassa, L. *et al.* A SARS-CoV-2 Nucleocapsid Variant that Affects Antigen Test Performance. 2021.05.05.21256527 <https://www.medrxiv.org/content/10.1101/2021.05.05.21256527v1> (2021) doi:10.1101/2021.05.05.21256527.
152. Niyonkuru, M. *et al.* Prolonged viral shedding of SARS-CoV-2 in two immunocompromised patients, a case report. *BMC Infectious Diseases* **21**, 743 (2021).
153. Walsh, K. A. *et al.* SARS-CoV-2 detection, viral load and infectivity over the course of an infection. *J. Infect.* **81**, 357–371 (2020).
154. Binnicker, M. J. Challenges and Controversies to Testing for COVID-19. *J Clin Microbiol* **58**, e01695-20 (2020).
155. Basso, D. *et al.* SARS-CoV-2 RNA identification in nasopharyngeal swabs: issues in pre-analytics. *Clinical Chemistry and Laboratory Medicine (CCLM)* **58**, 1579–1586 (2020).
156. Li, Y. *et al.* Stability issues of RT-PCR testing of SARS-CoV-2 for hospitalized patients clinically diagnosed with COVID-19. *Journal of Medical Virology* **92**, 903–908 (2020).
157. Bullard, J. *et al.* Predicting Infectious Severe Acute Respiratory Syndrome Coronavirus 2 From Diagnostic Samples. *Clin Infect Dis* **71**, 2663–2666 (2020).

158. La Scola, B. *et al.* Viral RNA load as determined by cell culture as a management tool for discharge of SARS-CoV-2 patients from infectious disease wards. *Eur J Clin Microbiol Infect Dis* **39**, 1059–1061 (2020).
159. Hoehl, S. *et al.* At-home self-testing of teachers with a SARS-CoV-2 rapid antigen test to reduce potential transmissions in schools: Results of the SAFE School Hesse Study. *medRxiv* 2020.12.04.20243410 (2020) doi:10.1101/2020.12.04.20243410.
160. Kriemler, S. *et al.* Surveillance of Acute SARS-CoV-2 Infections in School Children and Point-Prevalence During a Time of High Community Transmission in Switzerland. *Front. Pediatr.* **9**, (2021).
161. Kanji, J. N. *et al.* Multicenter Postimplementation Assessment of the Positive Predictive Value of SARS-CoV-2 Antigen-Based Point-of-Care Tests Used for Screening of Asymptomatic Continuing Care Staff. *J Clin Microbiol* **59**, e0141121 (2021).
162. Clark, T. W. *et al.* Diagnostic accuracy of a host response point-of-care test in patients with suspected COVID-19. *medRxiv* 2020.05.27.20114512 (2020) doi:10.1101/2020.05.27.20114512.
163. Houston, H. *et al.* Use of the FebrIDx point-of-care assay as part of a triage algorithm for medical admissions with possible COVID-19. *medRxiv* 2021.01.05.21249154 (2021) doi:10.1101/2021.01.05.21249154.
164. Ismail, S. A. *et al.* Horses for courses? Assessing the potential value of a surrogate, point-of-care test for SARS-CoV-2 epidemic control. *Influenza and Other Respiratory Viruses* **15**, 3–6 (2021).
165. Karim, N. *et al.* Utility of the FebrIDx point-of-care test for rapid triage and identification of possible coronavirus disease 2019 (COVID-19). *International Journal of Clinical Practice* **75**, e13702 (2021).
166. Zheng, Y.-Z. *et al.* Reverse Transcription Recombinase-Aided Amplification Assay With Lateral Flow Dipstick Assay for Rapid Detection of 2019 Novel Coronavirus. *Front. Cell. Infect. Microbiol.* **11**, (2021).
167. TANIDA, K. *et al.* Evaluation of the automated cartridge-based ARIES SARS-CoV-2 Assay (RUO) against automated Cepheid Xpert Xpress SARS-CoV-2 PCR as gold standard. *Eur J Microbiol Immunol (Bp)* **10**, 156–164 (2020).
168. Shelite, T. R., Uscanga-Palomeque, A. C., Castellanos-Gonzalez, A., Melby, P. C. & Travi, B. L. Isothermal recombinase polymerase amplification-lateral flow detection of SARS-CoV-2, the etiological agent of COVID-19. *J Virol Methods* **296**, 114227 (2021).
169. Desai, K. T. *et al.* Multi-site clinical validation of Isothermal Amplification based SARS-COV-2 detection assays using different sampling strategies. 2021.07.01.21259879 <https://www.medrxiv.org/content/10.1101/2021.07.01.21259879v1> (2021) doi:10.1101/2021.07.01.21259879.
170. Yoshikawa, R. *et al.* Development and evaluation of a rapid and simple diagnostic assay for COVID-19 based on loop-mediated isothermal amplification. *PLOS Neglected Tropical Diseases* **14**, e0008855 (2020).

171. Casati, B. *et al.* ADESSO: a rapid, adaptable and sensitive Cas13-based COVID-19 diagnostic platform. 2021.06.17.21258371 <https://www.medrxiv.org/content/10.1101/2021.06.17.21258371v2> (2021) doi:10.1101/2021.06.17.21258371.
172. Nörz, D. *et al.* Multicenter Evaluation of a Fully Automated High-Throughput SARS-CoV-2 Antigen Immunoassay. *Infect Dis Ther* (2021) doi:10.1007/s40121-021-00510-x.
173. Peto, L. *et al.* Diagnosis of SARS-CoV-2 Infection with LamPORE, a High-Throughput Platform Combining Loop-Mediated Isothermal Amplification and Nanopore Sequencing. *J Clin Microbiol* **59**, (2021).
174. Singh, N. K. *et al.* Hitting the diagnostic sweet spot: Point-of-care SARS-CoV-2 salivary antigen testing with an off-the-shelf glucometer. *Biosens Bioelectron* **180**, 113111 (2021).
175. Caffry, J. *et al.* The QuantuMDx Q-POCTM SARS-CoV-2 RT-PCR assay for rapid detection of COVID-19 at point-of-care: preliminary evaluation of a novel technology. 2021.07.12.21260119 <https://www.medrxiv.org/content/10.1101/2021.07.12.21260119v1> (2021) doi:10.1101/2021.07.12.21260119.
176. Krause, E. *et al.* Sensitive on-site detection of SARS-CoV-2 by ID NOW COVID-19. 2021.04.18.21255688 <https://www.medrxiv.org/content/10.1101/2021.04.18.21255688v1> (2021) doi:10.1101/2021.04.18.21255688.
177. Hagbom, M. *et al.* Evaluation of SARS-CoV-2 rapid antigen diagnostic tests for saliva samples. 2021.05.14.21257100 <https://www.medrxiv.org/content/10.1101/2021.05.14.21257100v1> (2021) doi:10.1101/2021.05.14.21257100.
178. Muthamia, E. *et al.* ASSESSMENT OF PERFORMANCE AND IMPLEMENTATION CHARACTERISTICS OF RAPID POINT OF CARE SARS-CoV-2 ANTIGEN TESTING. 2021.06.03.21258290 <https://www.medrxiv.org/content/10.1101/2021.06.03.21258290v2> (2021) doi:10.1101/2021.06.03.21258290.
179. Nordgren, J. *et al.* SARS-CoV-2 rapid antigen test: High sensitivity to detect infectious virus. *J Clin Virol* **140**, 104846 (2021).
180. Onyilagha, C. *et al.* Evaluation of mobile real-time polymerase chain reaction tests for the detection of severe acute respiratory syndrome coronavirus 2. *Sci Rep* **11**, 9387 (2021).
181. Stokes, W. *et al.* Clinical performance of the Abbott Panbio with nasopharyngeal, throat, and saliva swabs among symptomatic individuals with COVID-19. *Eur J Clin Microbiol Infect Dis* **40**, 1721–1726 (2021).
182. Renzoni, A. *et al.* Analytical Evaluation of Visby Medical RT-PCR Portable Device for Rapid Detection of SARS-CoV-2. *Diagnostics (Basel)* **11**, 813 (2021).
183. Alemany, A. *et al.* Analytical and clinical performance of the panbio COVID-19 antigen-detecting rapid diagnostic test. *Journal of Infection* **82**, 186–230 (2021).

184. Halfon, P. *et al.* An optimized stepwise algorithm combining rapid antigen and RT-qPCR for screening of COVID-19 patients. *medRxiv* 2021.01.13.21249254 (2021)
doi:10.1101/2021.01.13.21249254.
185. Jääskeläinen, A. E. *et al.* Evaluation of three rapid lateral flow antigen detection tests for the diagnosis of SARS-CoV-2 infection. *J Clin Virol* **137**, 104785 (2021).
186. Blairon, L. *et al.* Efficacy comparison of three rapid antigen tests for SARS-CoV-2 and how viral load impact their performance. *J Med Virol* **93**, 5783–5788 (2021).
187. Kweon, O. J. *et al.* Evaluation of rapid SARS-CoV-2 antigen tests, AFIAS COVID-19 Ag and ichroma COVID-19 Ag, with serial nasopharyngeal specimens from COVID-19 patients. *PLOS ONE* **16**, e0249972 (2021).
188. Toptan, T. *et al.* Evaluation of a SARS-CoV-2 rapid antigen test: Potential to help reduce community spread? *J Clin Virol* **135**, 104713 (2021).
189. Bouassa, R.-S. M., Veyer, D., Péré, H. & Bélec, L. Analytical performances of the point-of-care SIENNA™ COVID-19 Antigen Rapid Test for the detection of SARS-CoV-2 nucleocapsid protein in nasopharyngeal swabs: A prospective evaluation during the COVID-19 second wave in France. *International Journal of Infectious Diseases* **106**, 8–12 (2021).
190. Mitchell, S. L. & George, K. St. Evaluation of the COVID19 ID NOW EUA assay. *J Clin Virol* **128**, 104429 (2020).
191. Assennato, S. M. *et al.* Performance Evaluation of the SAMBA II SARS-CoV-2 Test for Point-of-Care Detection of SARS-CoV-2. *J Clin Microbiol* **59**, (2020).
192. Stevens, B. *et al.* Comparison of a Point-of-Care Assay and a High-Complexity Assay for Detection of SARS-CoV-2 RNA. *The Journal of Applied Laboratory Medicine* **5**, 1307–1312 (2020).
193. Young, S. *et al.* Clinical Evaluation of BD Veritor SARS-CoV-2 Point-of-Care Test Performance Compared to PCR-Based Testing and versus the Sofia 2 SARS Antigen Point-of-Care Test. *J Clin Microbiol* **59**, (2020).
194. Miscio, L. *et al.* Evaluation of the diagnostic accuracy of a new point-of-care rapid test for SARS-CoV-2 virus detection. *J Transl Med* **18**, (2020).
195. Panpradist, N. *et al.* *Harmony COVID-19: a ready-to-use kit, low-cost detector, and smartphone app for point-of-care SARS-CoV-2 RNA detection*. 2021.08.12.21261875
<https://www.medrxiv.org/content/10.1101/2021.08.12.21261875v2> (2021)
doi:10.1101/2021.08.12.21261875.
196. Koskinen, J. M. *et al.* Clinical validation of automated and rapid mariPOC SARS-CoV-2 antigen test. *Sci Rep* **11**, 20363 (2021).
197. Mlcochova, P. *et al.* Combined Point-of-Care Nucleic Acid and Antibody Testing for SARS-CoV-2 following Emergence of D614G Spike Variant. *CR Med* **1**, (2020).
198. Veyrenche, N. *et al.* Diagnosis value of SARS-CoV-2 antigen/antibody combined testing using rapid diagnostic tests at hospital admission. *Journal of Medical Virology* **93**, 3069–3076 (2021).

199. Olearo, F. *et al.* Handling and accuracy of four rapid antigen tests for the diagnosis of SARS-CoV-2 compared to RT-qPCR. *J Clin Virol* **137**, 104782 (2021).
200. Hogan, C. A. *et al.* Comparison of the Accula SARS-CoV-2 Test with a Laboratory-Developed Assay for Detection of SARS-CoV-2 RNA in Clinical Nasopharyngeal Specimens. *Journal of Clinical Microbiology* **58**, (2020).
201. Regev-Yochay, G. *et al.* Real World Performance of SARS-CoV-2 Antigen Rapid Diagnostic Tests in Various Clinical Settings. *medRxiv* 2021.03.02.21252400 (2021) doi:10.1101/2021.03.02.21252400.
202. Smithgall, M. C., Scherberkova, I., Whittier, S. & Green, D. A. Comparison of Cepheid Xpert Xpress and Abbott ID Now to Roche cobas for the Rapid Detection of SARS-CoV-2. *J Clin Virol* **128**, 104428 (2020).
203. Loeffelholz, M. J. *et al.* Multicenter Evaluation of the Cepheid Xpert Xpress SARS-CoV-2 Test. *J Clin Microbiol* **58**, (2020).
204. Wolters, F. *et al.* Multi-center evaluation of cepheid xpert® xpress SARS-CoV-2 point-of-care test during the SARS-CoV-2 pandemic. *Journal of Clinical Virology* **128**, 104426 (2020).
205. Cassuto, N. G. *et al.* Evaluation of a SARS-CoV-2 antigen-detecting rapid diagnostic test as a self-test: Diagnostic performance and usability. *J Med Virol* 10.1002/jmv.27249 (2021) doi:10.1002/jmv.27249.
206. Kernéis, S. *et al.* Accuracy of antigen and nucleic acid amplification testing on saliva and naopharyngeal samples for detection of SARS-CoV-2 in ambulatory care. 2021.04.08.21255144 <https://www.medrxiv.org/content/10.1101/2021.04.08.21255144v1> (2021) doi:10.1101/2021.04.08.21255144.
207. Lamb, G. *et al.* Real-world evaluation of COVID-19 lateral flow device (LFD) mass-testing in healthcare workers at a London hospital; a prospective cohort analysis. *Journal of Infection* **83**, 452–457 (2021).
208. LeGoff, J. *et al.* Evaluation of saliva molecular point of care for detection of SARS-CoV-2 in ambulatory care. 2021.06.12.21258811 <https://www.medrxiv.org/content/10.1101/2021.06.12.21258811v2> (2021) doi:10.1101/2021.06.12.21258811.
209. Lindner, A. K. *et al.* SARS-CoV-2 variant of concern B.1.1.7: diagnostic accuracy of three antigen-detecting rapid tests. 2021.06.15.21258502 <https://www.medrxiv.org/content/10.1101/2021.06.15.21258502v1> (2021) doi:10.1101/2021.06.15.21258502.
210. Perkins, T. A. *et al.* Performance of three molecular tests for SARS-CoV-2 on a university campus estimated jointly with Bayesian latent class modeling. 2021.07.31.21261425 <https://www.medrxiv.org/content/10.1101/2021.07.31.21261425v1> (2021) doi:10.1101/2021.07.31.21261425.
211. Oshiro, S. *et al.* Development of an immunochromatographic kit to detect severe acute respiratory syndrome coronavirus 2. *J Virol Methods* **294**, 114183 (2021).

212. Papadakis, G. *et al.* Portable real-time colorimetric LAMP-device for rapid quantitative detection of nucleic acids in crude samples. 2020.07.22.215251
<https://www.biorxiv.org/content/10.1101/2020.07.22.215251v2> (2021)
doi:10.1101/2020.07.22.215251.
213. Wu, S. *et al.* Comparative evaluation of six nucleic acid amplification kits for SARS-CoV-2 RNA detection. *Ann Clin Microbiol Antimicrob* **20**, 38 (2021).
214. Wen, D. *et al.* Sample-to-Answer and Routine Real-Time RT-PCR. *J Mol Diagn* **23**, 665–670 (2021).
215. Šterbenc, A. *et al.* Usefulness of rapid antigen testing for SARS-CoV-2 screening of healthcare workers: a pilot study. *Clin Exp Med* 1–4 (2021) doi:10.1007/s10238-021-00722-y.
216. Tulloch, J. S. P. *et al.* Enhanced lateral flow testing strategies in care homes are associated with poor adherence and were insufficient to prevent COVID-19 outbreaks: results from a mixed methods implementation study. *Age and Ageing* (2021) doi:10.1093/ageing/afab162.

Identification

Articles identified
through MedRxiv
search (n= 645)

Articles identified
through PubMed
search (n= 506)

Articles identified
through BioRxiv
search (n= 95)

Screening

Articles screened
(n= 1246)

Articles excluded: duplicates, unrelated topic
(i.e, seroprevalence, reviews, risk scores,
prognostic scores, imaging scores), test does
not focus on acute diagnosis (e.g antibody
assays) or uses imaging techniques (e.g, point-
of-care ultrasound)

Articles added
through cross-
references (n=1)

Eligibility

Articles screened
for full reading (n= 487)

Articles excluded: Not POC, laboratory-only
evaluations, selected population, assay not
compared to standard (PCR).
Debate between authors for articles with high
heterogenicity in methods

Inclusion

Articles included in
systematic review
(n= 87)