A prospective diagnostic evaluation of accuracy of self-taken and healthcare worker-taken swabs for rapid COVID-19 testing

Authors: Helen R. Savage¹(0000-0003-2592-9257), Lorna Finch², (0000-0002-8167-8766) Richard Body³(0000-0001-9089-8130), Rachel L. Watkins²(0000-0001-5754-3469), LSTM Diagnostics group², CONDOR steering group, Gail Hayward⁴(0000-0003-0852-627X), Eloïse Cook³(0000-0003-1598-4703), Ana I Cubas-Atienzar² (0000-0002-9604-124X), Luis E. Cuevas¹ (0000-0002-6581-0587), Peter MacPherson^{1,5,6} (0000-0002-0329-9613), Emily R. Adams² (0000-0002-0816-2835).

¹ Department of Clinical Sciences, Liverpool School of Tropical Medicine and Hygiene, Liverpool, L3 5QA, United Kingdom, Helen Savage PhD candidate, Luis E. Cuevas Professor International Public Health and Epidemiology, Peter MacPherson Reader and Wellcome Trust Fellow. ² Centre for Drugs and Diagnostics, Liverpool School of Tropical Medicine and Hygiene, Liverpool, L3 5QA, United Kingdom³, Lorna Finch Post-doctoral Research Associate, Rachel L. Watkins Research Technician, Ana I Cubas-Atienzar Post-doctoral Research Associate, Emily Adams Reader. ³ Manchester University NHS Foundation Trust, Research and Innovation, Manchester, M13 9WL, United Kingdom, Richard Body Professor of Emergency Medicine, Eloise Cook Project Manager. ⁴ Nuffield Department of Primary Care Health Sciences, University of Oxford, Oxford, OX2 6GG, UK, Gail Hayward Associate Professor⁵ Malawi-Liverpool-Wellcome Trust Clinical Research Programme, Blantyre, Malawi. ⁶ Clinical Research Department, London School of Hygiene and Tropical Medicine, UK, Peter MacPherson Reader and Wellcome Fellow

Corresponding author: Emily Adams, Email: Emily.adams@lstmed.ac.uk Phone number: +44151 7053196

Authors' contributions: The FALCON C-19 study, in which this work is nested was designed and is run by RB, EC and the CONDOR steering group. The study was conceived by HRS, ERA and LEC. The study design was developed by HRS, LF, and ERA. Recruitment and data collection was conducted by HRS, RLW, LF and LSTM Diagnostics Group. Data analysis and interpretation were conducted by HRS, ERA, PMP and LEC. The initial manuscript was prepared by HRS. All authors edited and approved the final manuscript. The corresponding author attests that all listed authors meet authorship criteria and that no others meeting the criteria have been omitted. This work is guaranteed by HS and ERA. HS and ERA affirm that this manuscript is an honest, accurate, and transparent account of the study being reported; that no important aspects of the study have been omitted; and that any discrepancies from the study as planned (and, if relevant, registered) have been explained.

The Corresponding Author has the right to grant on behalf of all authors and does grant on behalf of all authors, a worldwide license to the Publishers and its licensees in perpetuity, in all forms, formats and media (whether known now or created in the future), to i) publish, reproduce, distribute, display and store the Contribution, ii) translate the Contribution into other languages, create adaptations, reprints, include within collections and create summaries, extracts and/or, abstracts of the Contribution, iii) create any other derivative work(s) based on the Contribution, iv) to exploit all subsidiary rights in the Contribution, v) the inclusion of electronic links from the Contribution to third party material where-ever it may be located; and, vi) license any third party to do any or all of the above. All authors have completed the ICMJE uniform disclosure form at <http://www.icmje.org/disclosure-of-interest/> and declare: This study received funding

from the UK Research Council through a PhD scholarship from the MRC Doctoral Training Partnership to HRS. PM is funded by Wellcome (200901/Z/16/Z), Wellcome Trust award ‘Detecting and Excluding Coronavirus disease 2019 (COVID-19) at the Point of Need’ (220764/Z/20/Z), FALCON C-19 study was funded by National Institute for Health Research (NIHR), Asthma UK and the British Lung Foundation; and is supported by the global alliance for diagnostics (FIND). ERA contributed to this study design and analysis in her role as PhD supervisor to HRS at LSTM, she is also Director of Epidemics and NTDs at Mologic Ltd a UK diagnostics company who provided the RDTs for this study under joint Wellcome funding. The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Patient and public involvement statement:

The CONDOR platform (COVID-19 National Diagnostic Research and evaluation platform) has two PPI representatives on its steering committee meeting fortnightly. They contribute to the prioritisation of research within the platform, support and lead on developing new research priorities and co-produce plain language summaries and infographics detailing study findings. Within the nested FALCON C-19 study at LSTM results will be disseminated to the public within Liverpool by social media statements, press releases and attendance at open science events.

Keywords: Covid-19, diagnosis, self-testing, health worker-taken, Lateral flow assays, adults

Word count: Abstract, 325; Manuscript, 2328

Summary:

What is already known on this topic?

- Rapid diagnostic tests (RDTs) for SARS-CoV-2 Ag are a cheaper point-of-care alternative to RT-PCR for diagnosing COVID-19 disease.
- The accuracy of tests can vary dependent on sampling technique, test processing and reading of results.

What this study adds?

- Self-taken throat-nasal swabs for RDTs can be used by symptomatic adults to give reliable results to diagnose SARS-CoV-2.
- Self-sampling can be implemented with little training and no assistance.

Abstract (317 words)

Objectives:

To compare self-taken and healthcare worker (HCW)-taken throat/nasal swabs to perform rapid diagnostic tests (RDT) for SARS-CoV-2, and how these compare to RT-PCR. We hypothesised that self-taken samples are non-inferior for use with RDTs and in clinical and research settings could have substantial individual and public health benefit.

Design

A prospective diagnostic accuracy evaluation as part of the 'Facilitating Accelerated Clinical Evaluation of Novel Diagnostic Tests for COVID -19 (FALCON C-19)'. workstream C (undifferentiated community testing)'.¹

Setting

NHS Test and Trace drive-through community PCR testing site (Liverpool, UK).

Participants

Eligible participants 18 years or older with symptoms of COVID-19. 250 participants recruited; one withdrew before analysis.

Sampling

Self-administered throat/nasal swab for the Covios® RDT, a trained HCW taken throat/nasal sample for PCR and HCW comparison throat/nasal swab for RDT.

Main outcome measures

Sensitivity, specificity, and positive and negative predictive values (PPV, NPV) were calculated; comparisons between self-taken and HCW-taken samples used McNemar's test.

Results

Seventy-five participants (75/249, 30.1%) were positive by RT-PCR. RDTs with self-taken swabs had a sensitivity of 90.5% (67/74, 95% CI: 83.9-97.2), compared to 78.4% (58/74, 95% CI: 69.0-87.8) for HCW-taken swabs (absolute difference 12.2%, 95% CI: 4.7-19.6, $p=0.003$). Specificity for self-taken swabs was 99.4% (173/174, 95% CI: 98.3-100.0), versus 98.9% (172/174, 95% CI: 97.3-100.0) for HCW-taken swabs (absolute difference 0.6%, 95% CI: 0.5-1.7, $p=0.317$). The PPV of self-taken RDTs (98.5%, 67/68, 95% CI: 95.7-100.0) and HCW-taken RDTs (96.7%, 58/60, 95% CI: 92.1-100.0) were not significantly different ($p=0.262$). However, the NPV of self-taken swab RDTs was significantly higher (96.1%, 173/180, 95% CI: 93.2-98.9) than HCW-taken RDTs (91.5%, 172/188, 95% CI: 87.5-95.5, $p=0.003$).

Conclusion

Self-taken swabs for COVID-19 testing offer substantial individual benefits in terms of convenience, accuracy, and reduced risk of transmitting infection. Our results demonstrate that self-taken throat/nasal samples can be used by lay individuals as part of rapid testing programmes for symptomatic adults.

Trial Registration

IRAS ID:28422, clinical trial ID: NCT04408170

Introduction

The Severe Acute Coronavirus 2 (SARS-CoV-2) is a novel pathogen causing Coronavirus Disease-19 (COVID-19) that emerged in December 2019 and spread quickly around the globe before being declared a pandemic on 20th March 2020. Confirmation of SARS-CoV-2 infection is recommended by real-time polymerase chain reaction (RT-PCR) testing, however this requires well-resourced laboratory facilities, which are not available in many settings (1). Given the need to rapidly upscale testing, rapid diagnostic tests (RDTs) were developed to detect SARS-CoV-2 antigen(s) (Ag), which can be used at point of care without a laboratory infrastructure. Guidance from the World Health Organization (WHO) recommends using RDTs in settings with trained health workers to facilitate collecting samples and processing tests (1). Currently, large-scale self-testing for SARS-CoV-2 is conducted in schools, workplaces, and homes in the U.K.; however concerns have been raised over the accuracy of these tests and the risk of missing infected individuals (2). Many tests designed and regulated for “Professional Use Only” have been implemented for self-testing use, but little accuracy information exists for self-swabbing and interpretation.

Previous work has suggested that RDTs achieve higher sensitivity when performed by laboratory scientists (sensitivity of the Innova lateral flow test 70%, 95% CI 63-76%) than by healthcare workers (sensitivity 79%, 95% CI 72-84%)(3). Sensitivity can be substantially affected by the quality of the sample and swabbing technique (4). A small number of studies have compared self-taken to healthcare worker (HCW) taken swabs for RT-PCR, with a high degree of concordance (5,6). To the best of our knowledge, no studies have compared self-taken and HCW-taken samples with identical swab types and identical RDTs rather than comparing alternative sampling strategies.

We therefore set out to compare the sensitivity and specificity of self-taken and HCW-taken throat/nasal swabs to perform a RDT for SARS-CoV-2, and how these sampling approaches perform compared to the RT-PCR. We hypothesised that, if self-taken samples are accurate for use with RDTs in clinical and research settings, this could have substantial individual and public health benefit.

Methods

We conducted a prospective diagnostic accuracy evaluation to compare self-taken and HCW-taken throat/nasal swabs RDTs with a standard HCW-taken throat/nasal swab tested using RT-PCR. Participants were recruited as part of the ‘Facilitating Accelerated Clinical Evaluation of Novel Diagnostic Tests for COVID -19 (FALCON), workstream C (undifferentiated community testing)’(7), which aims to evaluate the diagnostic accuracy of commercially supplied in-vitro diagnostic (IVD) tests for SARS-CoV-2 infection.

Participants were recruited consecutively when presenting at the Liverpool John Lennon Airport (Liverpool, UK) drive through community PCR testing site, a National Health Service (NHS) Test and Trace site for the general population with symptoms of COVID-19, defined as a high temperature, continuous cough or change in sense of taste or smell. People presenting for testing were assessed for eligibility in their vehicle and received a patient information sheet prior to offer of NHS test. Eligible participants were 18 years or older who verbally confirmed they currently had symptoms of COVID-19. If multiple occupants were in the vehicle, they were each assessed for eligibility. Informed verbal consent was taken and recorded by a researcher on site. Demographic and symptom data were recorded electronically. We excluded anyone under 18 years, or who did not state they currently had COVID-19 symptoms, or if they did not consent to

participate. We did not record people who declined to participate. As people presented to the testing centre in vehicles, we were unable to distinguish between drivers or relatives, and people presenting for testing services.

A HCW in personal protective equipment (PPE, in line with local guidelines) passed a self-collection kit for each participant into the vehicle. Each kit contained a short instruction sheet taken from the manufacturer protocol (see Appendix), a tissue, a swab, and collection tube. The participant self-administered their throat/nasal swab without further advice, observation, or supervision. Once complete the participant signaled the research team and a trained healthcare worker took a throat/nasal sample using a COPAN mini UTM (universal transport medium) kit 1ml for PCR from one side and a comparison swab for RDT from the other. All swabs taken for RDT were randomly numbered so that laboratory staff performing the RDTs were blinded to sample collection method and could not identify paired samples.

Samples were transported to the Liverpool School of Tropical Medicine (LSTM) by trained research staff (in accordance with the requirements for Category B substance UN3373(8)) where they were processed and tested in a category 3 laboratory within 3 hours of sampling. UTM samples for PCR were aliquoted and frozen at -80° C. The Covios® COVID-19 Antigen Rapid Diagnostic test, which detects the SARS-CoV-2 nucleoprotein, was used for testing all RDT samples(9). This test is CE marked and manufactured in the UK by Global Access Diagnostics (legal manufacturer Mologic) (patients 1-100 LOT: CALFD-102-1, patient 101-250 LOT: CALFD-130-1). RDTs were run according to the manufacturer's instructions and read and graded (1-10 from a visual card) by two trained researchers. If there was disagreement between the two readers, a third reader was requested.

RNA was extracted from batched UTM samples using the QiAamp96 Virus Qiacube HT kit and RT-PCRs were run following manufacturer's instruction using TaqPath COVID-19 RT-PCR on QuantStudio 5 (ThermoFisher). RT-PCR reactions volumes were made in 20 µl. Reverse transcription step was performed at 53°C for 10 minutes and this was followed by an activation step of 2 minutes at 95°C, then PCR was carried for 40 amplification cycles at 95°C for 3 minutes and 60°C for 30 seconds. Fluoresce was recorded in the FAM, VIC, ABY and JUN channels for the ORF1ab, N, S and MS2 targets respectively.

Sample size and statistical analysis

Sample size was calculated using an alpha of 0.05, anticipated prevalence of 20% (based on UK PHE positivity of all PCR tests carried out (10)), minimum test sensitivity of 80%, specificity 99%, and precision interval of 10% (11). These gave a planned sample size of 308. We described participant characteristics using summary statistics and compared self-swab sampling RDT results with RT-PCR results. RDTs were graded 0-10, with 0 representing a negative result, and 1-10 positive results based on a visual scale using the manufacturer's reference card. RT-PCR results with cycle threshold (CT) values <40 were considered positive and CT values ≥40 were graded as negative. We calculated sensitivity and specificity, and positive and negative predictive values (PPV and NPV), all with binomial exact 95% confidence intervals (CI) in R v4.1.1 (R Foundation for Statistical Computing, Vienna). Paired results were compared between self-taken and HCW-taken samples using McNemar's test. Indeterminate RDT results were recorded but excluded from further analysis. Indeterminate RT-PCR results were repeated twice, and the repeat test result was used for analysis. RT-PCR data were classified according to the mean RT-PCR CT threshold values (<20, 20-24.9, 25-29.9, 30-34.9 and ≥35).

Recruitment started on 31st March 2021 and 100 participants had been recruited until the 21st May 2021, when local COVID-19 prevalence declined (positivity testing rate 0.4%) giving very small numbers of positive cases. Recruitment was temporarily halted until July 2021 when prevalence increased (positivity testing rate 12.9%) and a further 150 participants were recruited. Recruitment ended on 9th August 2021.

Ethical approval

Ethical approval was obtained from the National Research Ethics Service (reference 20/WA/0169) and the Health Research Authority (IRAS ID:28422, clinical trial ID: NCT04408170).

Results

Two hundred and fifty participants were recruited between the 31st March 2021 and the 9th August 2021. One participant withdrew after recruitment and did not wish data or samples to be included, leaving 249 participants for the analysis. The mean age of participants was 40 years (range 18-82, interquartile range [IQR] 30.0-50.0 years), 104 (41.7%) were male and 216 white British (86.7%) (Table 1). One hundred and eighty (72.3%) had received at least one vaccine dose against SARS-CoV-2 and of these 113 had received a second dose. The time interval since vaccination and the vaccine brand were not available. The most common symptoms reported were cough (174, 69.9%), fever (78, 31.3%), sore throat (71, 28.5%) and headache (53, 21.3%); all participants were symptomatic. The median duration since symptom onset was 2 days (IQR 1-3 days).

Seventy-five participants (75/249, 30.1%) tested positive by RT-PCR. The mean age of RT-PCR positive participants was 37.6 years (range 18-70, IQR 24.5-50.0), 42 (56.0%) were male and 67 were white British (89.3%) (Table 1). Fifty-two (69.3%) of the 75 RT-PCR-positive participants had received a first COVID-19 vaccine dose and 32 (42.7%) a second dose. Since symptom onset, the median duration in days was 2 (IQR: 1-3) and the most commonly reported symptoms were cough (53, 70.3%), fever (31, 41.3%), headache (22, 29.3%), sore throat (19, 25.3%), loss of smell (17, 22.7%) and loss of taste (16, 21.3%).

Overall, self-taken throat/nasal RDTs were positive in 68/249 (27.3%, 95% CI: 21.9-33.3) participants, one (0.4%) was indeterminate and 180 (72.3%) negative. HCW-taken throat/nasal RDTs were positive in 61/249 (24.5%, 95% CI: 19.3-30.3) participants, none was indeterminate and 188 (75.5%) were negative. The participant with the indeterminate RDT was excluded from further analysis.

RDT kits using a self-taken swab had a sensitivity of 90.5% (67/74, 95% CI: 83.9-97.2) and specificity of 99.4% (173/174, 95% CI: 98.3-100.0) when compared to the reference standard (Table 2). HCW-taken RDTs had a sensitivity of 78.4% (58/74, 95% CI 69.0-87.8) and specificity of 98.9% (172/174, 95% CI: 97.3-100.0) compared to the reference standard. The difference in sensitivity was 12.2% (95% CI: 4.7-19.6, $p=0.003$), the difference in specificity was 0.6% (95% CI: 0.5-1.7, $p=0.317$). Of the self-HCW RDT pairs, 238/248 (96.0%) agreed, and 10/248 (4.0%) were discordant. Of the discordant pairs, on nine occasions the self-taken swab RDT was read as positive, while the HCW-taken swab was read negative (Table 3). Nine of these pairs were RT-PCR positive and one negative. The discordant pair that was RT-PCR negative, the HCW-taken swab RDT was positive, while the self-taken RDT was negative.

The PPV of self-taken RDTs (98.5%, 67/68, 95% CI: 95.7-100.0) and HCW-taken RDTs (96.7%, 58/60, 95% CI 92.1-100.0) were not significantly different ($p=0.262$). However, the NPV of self-taken swab RDTs was significantly higher (96.1%, 173/180, 95% CI: 93.2-98.9) than HCW-taken RDTs (91.5%, 172/188, 95% CI 87.5-95.5, $p=0.003$). Full data table available in Appendix.

Sensitivity of the RDTs varied by CT values (Table 4 and Figure 2). Self-taken and HCW-taken samples with CT values <20 had 100% (32/32, CI: 89.1-100.0) sensitivity; samples with CT values 20-24.9 had 91.7% (22/24, 95% CI: 73.0-99.0) for self-taken and 83.3% (20/24, 95% CI: 62.6-95.3) for HCW-taken RDTs. At CT values between 25 and 29.9, RDTs sensitivity was 80.0% (12/15, 95% CI: 51.9-95.7) for self-taken and 40.0% (6/15, 95% CI: 16.3-67.7) for HCW-taken swabs, while at CT values 30-34.9 both self-taken and HCW-taken swabs had sensitivity of 33.0% (1/3, 95% CI: 0.8-90.6). Sensitivity for samples with CT values ≥ 35 was 0% (0/1, 95% CI: 0.0-97.5).

Discussion

This study found that the sensitivity of self-taken swabs for the detection of SARS-CoV-2 antigen was higher (90.5%) than using HCW-taken swabs (78.3%), with similar specificity. No RT-PCR-positive results from HCW-taken swabs were missed by self-taken swabs and the PPV and NPV for both methods were over 90%.

Current WHO guidance for implementing RDTs indicates swabbing to collect samples should be conducted by trained professionals (1). Using self-sampling for testing would reduce the workload of health workers and increase the ability of services to test patients in both clinical and research settings where trained workers are not available. These results show self-taken

throat-nasal samples with only written and pictorial instructions can be used by the general public for RDTs and is not likely to reduce the sensitivity of testing, which could widen access.

Within this study, self-taken swabs had higher sensitivity than HCW swabs for RDTs in a general population setting in the UK. The self-swabbing technique was not monitored for quality, no participants failed to take the swabs and no assistance was given so the results could be extrapolated to other non-supervised settings. The high concordance of self- and HCW-taken results has been reported from studies comparing self- and HCW-taken swabs for PCR testing and with also studies looking at alternative swab types (nasopharyngeal, nasal only) for RDTs (6,12–14).

The limitations of this study are that the sampling order was not randomized as the HCW swab for RDT was taken after the swab for the RT-PCR, and only from one nostril; this may lower sensitivity or participants may experience sampling fatigue. However, previous studies have shown that repeated sampling from one nostril does not impact RT-PCR sensitivity or CT values (15). Participation was voluntary so people who were less confident to take their own sample may not have agreed to take part; it is also likely participants may have done previous COVID-19 tests, and so have experience of self-sampling.

By focusing on differences in the sampling process (thus removing issues of running and interpretation of RDTs), we have shown that the sampling quality is unlikely to explain the differences in test accuracy, and that if individuals self-take samples for RDTs, the results can be as accurate as professionally taken swabs. All RDT-positive tests in the UK are currently confirmed with follow-on PCR tests which, given the high prevalence of circulating infection, are unlikely to result in many false positives and the high workloads can lead to laboratory

errors. Recent test discrepancies in South-West England, which reported large numbers of RDT-positive but PCR-negative results, were due to incorrect PCR results, with correct RDTs results (16). Our study suggest that the public and healthcare professionals should trust RDT-positive tests from self-taken samples in symptomatic individuals.

In conclusion, self-taken swabs for COVID-19 testing offer substantial individual benefits in terms of convenience, accuracy, and reduced risk of transmitting infection during swabbing. Our results demonstrate that, with no training, self-taken throat/nasal samples can be used by lay individuals as part of rapid testing programmes for symptomatic adults. Self-testing has the potential to widen access to early diagnosis for COVID-19 in clinical services and outreach settings where the lack of trained healthcare workers restricts access to testing.

Acknowledgements

We acknowledge the support of the National Institute for Health Research (NIHR) Clinical Research Network, which supports delivery of the FALCON study. The views expressed in this article are those of the authors and not necessarily those of the NIHR, or the Department of Health and Social Care.

LSTM diagnostics group: Kate Buist, Karina Clerkin, Dr Thomas Edwards, Dr Susan Gould, Caitlin Greenland-Bews, Konstantina Kontogianni, Laryssa Mashenko, Caitlin R Thompson, Jahanara Wardale, Christopher T Williams and Dominic Wooding.

Condor steering group: Dr A. Joy Allen, Dr Julian Braybrook, Professor Peter Buckle, Professor Paul Dark, Dr Kerrie Davis, Professor Adam Gordon, Ms Anna Halstead, Dr Charlotte Harden, Dr Colette Inkson, Ms Naoko Jones, Dr William Jones, Professor Dan Lasserson, Dr Joseph

Lee, Dr Clare Lendrem, Dr Andrew Lewington, Mx Mary Logan, Dr Massimo Micocci, Dr Brian Nicholson, Professor Rafael Perera-Salazar, Mr Graham Prestwich, Dr D. Ashley Price, Dr Charles Reynard, Dr Beverley Riley, Professor AJ Simpson (Professor Simpson is an NIHR Senior Investigator), Dr Valerie Tate, Dr Philip Turner, Professor Mark Wilcox, Dr Melody Zhifang.

References

1. World Health Organization (WHO). SARS-CoV-2 antigen-detecting rapid diagnostic tests - An implementation guide [Internet]. Geneva; 2020. Available from: <http://apps.who.int>
2. Torjesen I. Covid-19: How the UK is using lateral flow tests in the pandemic. *BMJ*. 2021;372:1–3.
3. Peto T, Affron D, Afrough B, Agasu A, Ainsworth M, Allanson A, et al. COVID-19: Rapid antigen detection for SARS-CoV-2 by lateral flow assay: A national systematic evaluation of sensitivity and specificity for mass-testing. *EClinicalMedicine*. 2021;36:0–6.
4. García-Fiñana M, Hughes DM, Cheyne CP, Burnside G, Stockbridge M, Fowler TA, et al. Performance of the Innova SARS-CoV-2 antigen rapid lateral flow test in the Liverpool asymptomatic testing pilot: Population based cohort study. *BMJ*. 2021;374(July).
5. Altamirano J, Govindarajan P, Blomkalns AL, Kushner LE, Stevens BA, Pinsky BA, et al. Assessment of Sensitivity and Specificity of Patient-Collected Lower Nasal Specimens for Sudden Acute Respiratory Syndrome Coronavirus 2 Testing. *JAMA Netw Open*. 2020;3(6):10–3.

6. Tu Y-P, Jennings R, Hart B, Cangelosi GA, Wood RC, Wehber K, et al. Swabs Collected by Patients or Health Care Workers for SARS-CoV-2 Testing. *N Engl J Med*. 2020;383(5):494–6.
7. NIH U.S. National Library of Medicine. Facilitating Accelerated Clinical Validation Of Novel Diagnostics for COVID-19 (FALCON-C19) (FALCON-C19) [Internet]. 2021 [cited 2021 Nov 17]. Available from: <https://clinicaltrials.gov/ct2/show/NCT04408170>
8. Department for Transport. Packaging and transport requirements for patient samples – UN3373 [Internet]. United Kingdom. 2020 [cited 2021 Nov 16]. Available from: <https://www.gov.uk/government/publications/packaging-and-transport-requirements-for-patient-samples-un3373/packaging-and-transport-requirements-for-patient-samples-un3373>
9. Cubas-Atienzar AI, Bell F, Byrne RL, Buist K, Clark DJ, Coccozza M, et al. Accuracy of the Mologic COVID-19 rapid antigen test: a prospective multi-centre analytical and clinical evaluation. *Wellcome Open Res*. 2021;6(May):132.
10. Liverpool City Council Corporate Intelligence team. Liverpool COVID-19 report □: Produced on 21 January 2021. Liverpool; 2021.
11. Fenn Buderer NM. Statistical methodology: I. Incorporating the prevalence of disease into the sample size calculation for sensitivity and specificity. *Acad Emerg Med*. 1996;3(9):895–900.
12. Nikolai O, Rohardt C, Tobian F, Junge A, Corman VM, Jones TC, et al. Anterior nasal versus nasal mid-turbinate sampling for a SARS-CoV-2 antigen-detecting rapid test: does

- localisation or professional collection matter? medRxiv [Internet]. 2021;pre-print:2021.02.09.21251274-2021.02.09.21251274. Available from: <http://medrxiv.org/content/early/2021/02/16/2021.02.09.21251274.abstract>
13. Lindner AK, Nikolai O, Rohardt C, Burock S, Hülso C, Bölke A, et al. Head-to-head comparison of SARS-CoV-2 antigen-detecting rapid test with professional-collected nasal versus nasopharyngeal swab. *Eur Respir J*. 2021;57(5):3–6.
 14. Lindner AK, Nikolai O, Rohardt C, Kausch F, Wintel M, Bernhard J, et al. Diagnostic accuracy and feasibility of patient self-testing with a SARS-CoV-2 antigen-detecting rapid test. 2020;(January).
 15. Levy JM, Frediani JK, Tyburski EA, Wood A, Figueroa J, Kempker RR, et al. Impact of repeated nasal sampling on detection and quantification of SARS-CoV-2. *Sci Rep* [Internet]. 2021;11(1):1–5. Available from: <https://doi.org/10.1038/s41598-021-94547-0>
 16. Torjesen I. Covid-19: PCR testing is suspended at private laboratory after high rate of false negatives [Internet]. Vol. 375, *BMJ*. 2021. Available from: <https://www.bmj.com/content/375/bmj.n2535>

List of tables

Table 1: Demographic data.

Table 2: Sensitivity and Specificity results

Table 3: Table showing results for participants with discrepant RDT results

Table 4: Sensitivity by RT-PCR range (including indeterminate result)

List of figures

Figure 1: Correlation between self and healthcare worker graded RDT result by PCR cycle threshold.

Figure 2: A graph to show the sensitivity of the Covios® SARS-CoV-2 RDT by mean CT range and cumulative sensitivities.

Appendix:

Table of all PCR positive participants and their characteristics.

A box plot comparing self-taken and healthcare worker taken swabs tested by Covios® RDT by mean PCR CT value.

Self-taken and HCW taken Covios® Ag RDT results by RT-PCR CT range.

Covios® leaflet from study

Table 1: Characteristics of participants.

	All (N, %)	RT-PCR positive (N, %)
All	249	75
Age in years, mean, (range, IQR)	40 (18-82, 30.0-50.0)	37.6 (18-70, 24.5-50.0)
Male	104 (41.7)	42 (56.0)
Median symptom duration (days), range, IQR	2.0 (0-33, 1-3)	2.0 (0-32, 1-3)
Shortness of breath	7 (2.8)	2 (2.7)
Cough	174 (69.9)	53 (70.7)
Fever	78 (31.3)	31 (41.3)
Chest pain	11 (4.4)	7 (9.3)
Sore throat	71 (28.5)	19 (25.3)
Confusion	0 (0)	0 (0)
Rash	1 (0.4)	1 (1.3)
Loss of smell	27 (10.8)	17 (22.7)
Loss of taste	25 (10.0)	16 (21.3)
Abdominal pain	6 (2.4)	1 (1.3)
Vomiting	7 (2.8)	3 (4.0)
Diarrhoea	11 (4.4)	3 (4.0)
Headache	53 (21.3)	22 (29.3)
Tiredness/Fatigue	10 (4.0)	5 (6.7)
Tight chest	1 (0.4)	0 (0)
Other	70 (28.1)	31 (41.3)
White British	216 (86.7)	67 (89.3)
Irish	10 (4.0)	2 (2.7)
Other white	7 (2.8)	1 (1.3)
Indian	2 (0.8)	0 (0.0)
Mixed ethnic group	9 (3.6)	3 (4.0)
Other ethnic group	5 (2.0)	2 (2.7)
Vaccinated 1 st dose	180 (72.3)	52 (69.3)
Vaccinated 2 nd dose	113 (45.4)	32 (42.7)

Table 2: Sensitivity and Specificity of self- and healthcare worker-taken swabs for COVID-19 rapid diagnostic testing

	Comparison to RT-PCR							
	Sensitivity (%)	95% CI	Specificity (%)	95% CI	PPV (%)	95% CI	NPV (%)	95% CI
Self-taken RDT	90.5	83.9-97.2	99.4	98.3-100.0	98.5	95.7-100.0	96.1	93.3-98.9
HCW RDT	78.4	69.0-87.8	98.9	97.3-100.0	96.7	92.1-100.0	91.5	87.5-95.5

* CI = Confidence intervals, PPV = Positive predictive value, NPV = Negative predictive value

Table 3: Table showing results for participants with discrepant RDT results

Participant	RT-PCR result (mean Ct value)	Self-taken RDT result (Reader 1/Reader 2)	HCW-taken RDT result (Reader 1/Reader 2)
26	Positive (29.03)	4/4	0/0
38	Negative (>40)	0/0	2/2
106	Positive (21.68)	3/3	0/0
131	Positive (27.99)	3/3	0/0
140	Positive (25.44)	3/3	0/0
169	Positive (28.09)	2/2	0/0
189	Positive (20.35)	8/8	0/0
195	Positive (22.96)	3/3	0/0
231	Positive (27.33)	6/6	0/0
249	Positive (27.86)	5/5	0/0

Table 4: Sensitivity of self- and healthcare worker-taken swab for rapid diagnostic testing by RT-PCR CT ranges.

	RT-PCR CT range				
	<20	20-24.9	25-29.9	30-34.9	≥35
Self-taken RDT					
Positive	32	22	12	1	0
Negative	0	1	3	2	1
Indeterminate	0	1	0	0	0
Sensitivity	100.0%	91.7%	80.0%	33.3%	0%
95% CI	89.1-100.0%	73.0-99.0%	51.9-95.7%	0.8-90.6%	0.0-97.5%
Cumulative sensitivity	100.0%	96.4%	93.0%	90.5%	89.3%
95% CI	89.1-100.0%	87.7-99.6%	84.3-97.7%	81.5-96.1%	80.1-95.3%
HCW RDT					
Positive	32	20	6	1	0
Negative	0	4	9	2	1
Indeterminate	0	0	0	0	0
Sensitivity	100%	83.3%	40.0%	33.3%	0.0%
95% CI	89.1-100.0%	62.6-95.3%	16.3-67.7%	0.8-90.6%	0.0-97.5%
Cumulative sensitivity	100.0%	92.9%	81.7%	79.7%	78.7%
95% CI	89.1-100.0%	82.7-98.0%	70.7-89.9%	68.8-88.2%	67.6-87.3%

Sensitivity (95% confidence interval)

Healthworker-taken

Healthworker-taken cumulative

Self-taken

Self-taken cumulative

Mean SARS-CoV-2 PCR cycle threshold value