TITLE: Comparative evaluation of rapid isothermal amplification and antigen assays for screening testing of SARS-CoV-2

AUTHORS: Nol Salcedo^{1*}, Brena F. Sena^{1*}, Xiyang Qu^{1a}, Bobby Brooke Herrera^{1,2}

AFFILIATIONS: ¹E25Bio, Inc., Cambridge, MA, USA

²Department of Immunology and Infectious Diseases, Harvard T.H. Chan School of Public Health, Boston, MA, USA

^aCurrent Address: Flagship Pioneering 75, Cambridge, MA 02139, USA

*Equal contribution

#Corresponding authors: BBH, bbherrera@e25bio.com

ABSTRACT

Human transmission of SARS-CoV-2 and emergent variants of concern has continued to occur globally, despite mass vaccination campaigns. Public health strategies to reduce virus spread should therefore rely, in part, on frequent screening with rapid, inexpensive, and sensitive tests. We evaluated two digitally integrated rapid tests and assessed their performance using stored nasal swab specimens collected from individuals with or without COVID-19. An isothermal amplification assay combined with a lateral flow test had a limit of detection of 10 RNA copies per reaction, and a positive percent agreement (PPA)/negative percent agreement (NPA) during the asymptomatic and symptomatic phases of 100%/100% and 95.83/100%, respectively. Comparatively, an antigen-based lateral flow test, had a limit of detection of 30,000 copies, and a PPA/NPA during the asymptomatic and symptomatic phases of 82.86%/98.68% and 91.67/100%, respectively. Both the isothermal amplification and antigen-based lateral flow tests had optimized detection of SARS-CoV-2 during the peak period of transmission; however, the antigen-based test had reduced sensitivity in clinical samples with qPCR Ct values greater than 29.8. Low-cost, high-throughput screening enabled by isothermal amplification or antigen-based techniques have value for outbreak control.

INTRODUCTION

Alongside widespread vaccine campaigns, strategies continue to be implemented to reduce the human transmission of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) (1-4). Testing, in particular, has played an important role throughout the coronavirus disease 2019 (COVID-19) pandemic in detecting the virus and emergent variants of concern, enabling responses at the national, community, and individual levels (5, 6). However, most testing occurs in centralized settings that utilize quantitative polymerase chain reaction (qPCR) assays (7). While these molecular techniques can detect minute amounts of viral RNA and therefore most appropriate for clinical diagnosis, they cannot be scaled to meet demands for extensive public health surveillance or frequent screening of individuals, especially in resource limited settings. Inexpensive, accurate tests that can be self-administered or performed at the point-of-care, and provide actionable results, will further facilitate outbreak suppression.

Diagnostic testing for COVID-19 focuses on establishing the presence or absence of SARS-CoV-2 in symptomatic or asymptomatic individuals (8). In general, healthcare professionals collect respiratory specimens via nasopharyngeal swabs or use less invasive approaches such as anterior nares swabs or saliva collection (9, 10). The respiratory specimens then are processed by centralized high-complexity laboratories with specialized equipment using qPCR assays with results being reported within 24 to 48 hours. In some regions, bottlenecks in laboratory-based testing have led to turnaround times exceeding several days, diminishing the efficacy of this approach to prevent ongoing transmission.

Surveillance testing estimates the infection dynamics of SARS-CoV-2 in representative sample sets, with the primary goal to inform public health policy. Molecular-based techniques that are highly sensitive and specific, and report qPCR cycle thresholds (Ct) and viral loads, are typically used for surveillance testing (11-13). An emerging approach involves surveillance of SARS-CoV-2 via wastewater monitoring using qPCR assays (14, 15). While surveillance testing is performed in centralized settings, the resulting information can be used to monitor epidemic trajectory in specific communities and allow for real-time evaluation and/or implementation of mitigation programs.

Identification of individuals who are likely infectious with screening testing is one of the most effective, but underused, strategies to limit the ongoing transmission of SARS-CoV-2 (16). In approximately 20-40% of COVID-19 cases, the infection remains asymptomatic, and symptomatic disease is preceded by a pre-symptomatic incubation period (17). Yet, pre-symptomatic and asymptomatic cases contribute significantly to SARS-CoV-2 spread, challenging our ability to contain outbreaks (17).

Breaks in transmission chains can be most effectively achieved when screening testing is applied frequently and serially using self-administered rapid tests (18-20). Antigen-based tests, which utilize combinations of monoclonal antibodies and nanoparticles to detect viral proteins, do not require instruments or skilled operators; as of December 2021, 10 antigen tests for SARS-CoV-2 have been approved for at-home use in the United States (21). Although antigen tests have lower analytical sensitivity and specificity compared to qPCR assays, they have increased ability to detect SARS-CoV-2

during the acute phase of COVID-19 when an infected individual is most likely to transmit the virus (22-24).

Moreover, isothermal amplification technologies offer the simplicity and speed of antigen tests but have higher sensitivity and specificity (25-27). One of the most promising isothermal amplification technologies is recombinase polymerase amplification (RPA) (28, 29). In RPA, double stranded DNA denaturation and strand invasion is achieved by a cocktail of enzymes including recombinases, single-stranded binding proteins, and DNA polymerases; typically, this occurs by multiple heat cycles in PCR (30). RPA has added benefits over other isothermal amplification technologies (i.e., loop-mediated isothermal amplification, LAMP or CRISPR) as reactions occur at ambient temperatures (37-42°C), in shorter time periods, and results that can be visualized on a lateral flow test. One of three isothermal amplification technologies currently available in the United States for at-home detection of SARS-CoV-2 utilizes reverse transcription RPA (RT-RPA), and has been shown to detect the virus in nasal swab specimens with as low as twenty genome copies (31, 32). Given their robust sensitivity and specificity, RT-RPA assays are optimized to detect SARS-CoV-2 during the peak period of transmission in individuals with pre-symptomatic, symptomatic, and/or asymptomatic infections (28, 33, 34).

Here, we performed a comparative evaluation of a RT-RPA assay and an antigen test. Using previously characterized nasal swab dilution specimens, we assessed the analytical sensitivity of the two tests. We show that the RT-RPA assay allows for detection of SARS-CoV-2 down to 10 RNA copies per reaction compared to folds higher with the antigen test. We then calculated the positive percent agreement (PPA, or sensitivity) and negative percent agreement (NPA, or specificity) using stored, unextracted nasal swab

specimens collected from individuals with or without COVID-19. We demonstrate that the RT-RPA assay has increased sensitivity in nasal swab specimens, particularly in qPCR Ct values greater than 29.8, regardless of if the sample was collected during the asymptomatic or symptomatic phases. Supporting the innovation, manufacturing, approval, and distribution of isothermal amplification screening tests will enable more effective control of infectious disease outbreaks.

MATERIALS AND METHODS

Clinical samples

The nasal swab dilution panel was provided by the non-profit PATH (www.path.org). Nasal swab dilutions were prepared from human nasal swab eluate discards from suspected COVID-19 patients, collected within seven days post-symptoms onset. A single swab eluate positive for SARS-CoV-2 by qPCR was diluted into a single nasal eluate negative for SARS-CoV-2 by qPCR. For the dilution specimens with lower than 5,000 RNA copies, known quantities of RNA (ATCC, Manassas, VA) were spiked into nasal eluates negative for SARS-CoV-2. Dilution specimens were de-identified, coded, and then aliquoted and frozen at -80°C. Aliquots were thawed and characterized by qPCR as previously described (24). The primary studies under which the samples were collected received ethical clearance from the PATH Institutional Review Board (IRB) (approval number 0004244).

Additionally, nasal swab specimens were collected from a cohort of suspected COVID-19 patients with or without symptoms at a point-of-care site (POC nasal swab

specimens); for individuals with symptoms, specimen were collected within the first 3 days of symptoms onset. The nasal swabs were mixed in tubes containing 1X PBS (MilliporeSigma, Burlington, MA). Aliquots were de-identified, coded, and then frozen at -80°C. The study under which the samples were collected received ethical clearance from the Advarra, Inc. IRB (approval number Pro00044496).

qPCR

200 µl of the POC nasal swab specimens were used for extraction with the MagMAX Viral/Pathogen II Nucleic Acid Isolation Kit (ThermoFisher Scientific, Waltham, MA) on an epMotion 5075 (Eppendorf, Hamburg, Germany) liquid handler. Nucleic acids were eluted in 50 µl; 2 µl were used for qPCR confirmation using the GoTaq Probe 1-Step RT-qPCR System (Promega, Madison, WI) on a QuantStudio 7 Flex Real-Time PCR Instrument (ThermoFisher Scientific, Waltham, MA). The SARS-CoV-2 (2019-nCoV) CDC qPCR Probe Assay was used to detect the human RNaseP gene and two viral targets 2019-nCoV_N1 and 2019-nCoV_N2 (Integrated DNA Technologies, Coralville, IA).

RT-RPA assay

Prior to isothermal amplification testing, the nasal swab dilution specimens or POC nasal swab specimens were lysed at 95°C using a heat block (Southern Labware, Cumming, GA) for 3 minutes. Isothermal amplification reactions were conducted using AmpliFast enzymes and buffer (E25Bio, Inc., Cambridge, MA), 1 µl RNase H (5U/µl; ThermoFisher Scientific, Waltham, MA), 0.5 µl SuperScript IV RT (200 U/µl;

ThermoFisher Scientific, Waltham, MA), 0.5 µl of SARS-CoV-2 nucleocapsid (N) forward and reverse primers (300 nM final concentration), and 2 µl input template (nasal swab dilution specimen or POC nasal swab specimen). This mix was activated by addition of 1 µl of magnesium acetate (14 nM final concentration), MilliporeSigma, Burlington, MA) followed by thorough mixing. Reactions were incubated at 38°C for 20 minutes. A hybridization mix was prepared by combining 1 µl SARS-Cov-2 N biotinylated probe (0.167 nM final concentration) with 19 µl Tris pH 8 (10 mM). 20 µl of the hybridization mix was added to each reaction, and samples were heated to 95°C for 3 minutes followed by a cooling step at room temperature for 3 minutes. 40 µl of buffer (Pocket Diagnostic, York, United Kingdom) were added to each reaction, then the mixture was applied to the PCRD nucleic acid lateral flow test (Pocket Diagnostic, York, United Kingdom) and allowed to react for 10 minutes. Interactions of the immobilized test and control line antibodies with amplified nucleic acids and the nanoparticle conjugate produced visible bands, indicating whether a test was positive or negative.

Antigen test

Rapid antigen tests (E25Bio, Inc., Cambridge, MA) contain a monoclonal antibody and a nanoparticle conjugate that detect SARS-CoV-2 N. 100 µl of the nasal swab dilution specimens or POC nasal swab specimens were applied to the antigen test and allowed to react for 15 minutes. Interactions of the immobilized test and control line antibodies with antigen and the nanoparticle conjugate produced visible bands, indicating whether a test was positive or negative.

184 Image analysis

185 Results from the isothermal amplification lateral flow tests and antigen-based
186 lateral flow tests were captured via the Passport App (currently available through Apple,
187 Inc.'s TestFlight; E25Bio, Inc., Cambridge, MA). The images were machine-read and
188 processed to quantify test results. The average pixel intensity was quantified at the test
189 line, control line, and background areas. The background-subtracted test line signal was
190 then normalized to the background-subtracted control line signal and the final test signal
191 was expressed as percent of control. The Passport App only stores images and
192 identifiable test results locally on the user's mobile device, and the individual can share
193 the results with whomever, whenever they choose.

194

195 Statistics

196 GraphPad Prism 9.0 (San Diego, CA) was used to analyze and report the
197 performance of the isothermal amplification and antigen tests compared to qPCR. The
198 sensitivity was defined as the fraction of total qPCR confirmed positive samples that are
199 true positives according to the test. The specificity was defined as the fraction of total
200 qPCR confirmed negative samples that are true negatives according to the test. Where
201 appropriate, test signals were plotted using symbol and line graphs according to
202 asymptomatic or symptomatic infection status and qPCR Ct thresholds.

203

204 Data availability

205 All relevant data have been included in the manuscript. We will provide any
206 additional data upon request.

RESULTS

In the RT-RPA assay, viral RNA is first copied to cDNA by reverse transcriptase, then degraded by RNase H. The cDNA product is amplified by RPA using a forward and a FAM-labeled reverse pair of primers specific to the target sequence. The amplified nucleic acid target is denatured and hybridized to a biotinylated probe. Dual FAM-labeled and biotin-labeled products are then detected on a lateral flow test that contain nanoparticles and detection molecules (i.e., anti-FAM antibody and streptavidin) specific for FAM and biotin (Fig. 1A). In the antigen test, the interaction of antibodies and nanoparticles with protein targets produces detectable bands (Fig. 1B). Both the RT-RPA assay and the antigen test used in this study target SARS-CoV-2 N. To reduce errors in user-based interpretation, we used a mobile phone application to machine read and quantify the RT-RPA and antigen test results (Fig. 1C). Mobile phone image processing allowed test users to obtain an objective analysis of their results, despite varied use conditions, and share data in real-time.

We evaluated the analytical sensitivity of the RT-RPA assay and the antigen test using well-characterized nasal swab dilution specimens. The dilution specimens contained SARS-CoV-2 RNA copies ranging from 1 (Ct value 39.6) to 200,000 (Ct value 25.2). Consistent with expectations from qPCR, the RT-RPA assay yielded positive results with an input of 10 RNA copies per reaction (Ct value 37.3) (Fig. 2A-B). The antigen test reproducibly had detectable results with dilution specimens between 40,000 and 30,000 copies (Ct values 28.3 and 29.2, respectively) of SARS-CoV-2 (Fig. 2C-D).

The RT-RPA assay had a detection limit several orders of magnitude lower than the antigen test.

To further evaluate the RT-RPA and antigen tests, we compared their sensitivity and specificity using stored, unextracted nasal swab specimens collected from individuals with or without COVID-19. A total of 114 nasal swab specimens were negative and 59 were positive for SARS-CoV-2 by qPCR. Of the 114 negative specimens, 76 were collected from asymptomatic cases and 24 were collected from symptomatic cases. Of the 59 SARS-CoV-2 positive specimens, 35 and 24 were collected from asymptomatic or symptomatic cases, respectively. All 114 negative specimens were negative by the RT-RPA assay, regardless of symptoms, corresponding to a 100% specificity (Fig. 3, Table 1-2). Only 1 of 76 negative specimen from asymptomatic cases was positive by the antigen test, corresponding to a 98.68% specificity (Fig. 3A, Table 1). All negative specimens from 24 symptomatic cases were negative by the antigen test (100% specificity) (Fig. 3B, Table 2). These results confirmed a low false positive rate for the RT-RPA assay and antigen test.

All 35 SARS-CoV-2 positive specimens from asymptomatic cases were positive by the RT-RPA assay, corresponding to a 100% sensitivity (Fig. 3A, Table 1). Of the 24 positive specimens from the symptomatic phase, only 1 tested negative (95.83% sensitivity) (Fig. 3B, Table 2). In contrast, the antigen test detected 29 out of 35 (82.86% sensitivity) asymptomatic SARS-CoV-2 positives and 22 out of 24 (91.67% sensitivity) symptomatic positives (Fig. 3, Tables 1-2). Of note, the sensitivity of the antigen test decreased significantly in Ct values greater than 30.1, while the sensitivity of the RT-RPA assay was maintained (Fig. 3). Altogether, these results demonstrated that the true

positive rate of the RT-RPA assay was much higher than the antigen test especially during the asymptomatic phase and particularly in specimens with higher Ct values.

DISCUSSION

One of the most promising strategies aimed at SARS-CoV-2 outbreak suppression is the screening of infectious individuals. This type of testing requires frequent and serial testing of large populations that can be self-administered or performed at the point-of-care in high-transmission settings (i.e., schools, workplaces, etc.). The primary goal of screening testing is to achieve population-wide effects by breaking transmission chains through detection of cases, especially during the pre-symptomatic or asymptomatic phases.

Modeling studies have demonstrated that frequent rapid testing of large populations, even with varied test accuracies, can help achieve herd effects thereby suppressing transmission of SARS-CoV-2 (18, 20, 35). In Slovakia, ~80% of the population was screened for COVID-19 using antigen tests (36). In a 2-week period, 50,000 cases were identified, and along with other public health measures (i.e., wearing masks, quarantining, etc.), the incidence was reduced by 82%. Further, at-home antigen testing was performed twice per week in a coworking environment in Cambridge, MA over a 6-month period (19). In the case of a positive test, an individual would undergo a 10-day quarantine prior to returning to the workplace. Twice-weekly testing identified 15 individuals infected with SARS-CoV-2, with a test sensitivity of 96.2% on days 0-3 of symptoms. This testing strategy allowed the activities of the coworking sites to continue

without pause. To our knowledge, screening testing using isothermal amplification techniques has not been extensively evaluated.

In this study, we performed a comparative evaluation of a RT-RPA assay and an antigen test for SARS-CoV-2. We tested the analytical sensitivity using a nasal swab dilution panel. The RT-RPA assay had a detection limit far lower than the antigen test. We then analyzed the performance of the tests using qPCR characterized nasal swab specimens collected from individuals with or without COVID-19. The RT-RPA assay had a high sensitivity (>95%) and specificity (100%) in specimens from asymptomatic or symptomatic cases. In contrast, the sensitivity of the antigen test during the asymptomatic phase was much lower at 82.86%, and especially with specimens that had Ct values greater than 30. A likely explanation is that during the asymptomatic phase SARS-CoV-2 viremia has not peaked, resulting in reduced viral antigens in respiratory specimens. Additionally, Ct values >30 typically appear later in the course of SARS-CoV-2 infection (i.e., 7 days after exposure), when virus is being eliminated by the immune system, clearing antigen levels. In support of our hypotheses, the sensitivity of the antigen test increased to >90% in nasal swab specimens collected from symptomatic cases within 3 days of symptoms onset.

Future work should broaden the evaluation of isothermal amplification and antigen assays to additional settings, sample types, and disease states (i.e., pre-symptomatic phase). Performance testing on prospectively collected samples will further corroborate preliminary findings. As the RT-RPA assay used in this study uses an inexpensive water bath and heat block, there is a need to for these types of assays to perform reactions with consumer-designed hardware that would allow for at-home or point-of-care testing.

Optimizing reaction mixes can also help reduce the temperatures and time required for test processing. Finally, the mobile phone application used in this study lessens user error by interpreting the results via pre-designed algorithms. Additional open-source, low-cost methods for data capture and reporting are warranted.

Public health screening requires rapid, inexpensive, and sensitive tests that can be scaled for frequent and serial testing in large numbers. Antigen tests and upcoming isothermal amplification assays fit these needs and could be scaled to millions of tests per day. Despite being shown to be highly effective at detecting infectious individuals, there are only a handful of rapid tests currently available for self-administration or at-home use in the United States. Even with approvals, these manufacturers have been unable to meet the scale and demand, leaving individuals without access to these valuable inexpensive, rapid testing options. The support of manufacturing, rapid approval processes, and distribution of screening tests will help control COVID-19 outbreaks.

ACKNOWLEDGEMENTS

We thank the staff of Cataldo Ambulance Service, notably David, Karen, and Pippa, for their help with patient recruitment and sample collection. We thank Professor Phyllis Kanki for critical reading of the manuscript.

COMPETING INTERESTS

This work was supported by E25Bio, Inc. NS, BFS, and BBH are employed by E25Bio, Inc., a company that develops rapid diagnostic tests for infectious diseases. At

the time of the study, XQ was employed by E25Bio, Inc. BBH is an inventor on a U.S. Provisional Patent Application (63/189,502).

REFERENCES

1. El Sahly HM, Baden LR, Essink B, Doblecki-Lewis S, Martin JM, Anderson EJ, et al. Efficacy of the mRNA-1273 SARS-CoV-2 Vaccine at Completion of Blinded Phase. *N Engl J Med*. 2021;385(19):1774-85.
2. Barda N, Dagan N, Ben-Shlomo Y, Kepten E, Waxman J, Ohana R, et al. Safety of the BNT162b2 mRNA Covid-19 Vaccine in a Nationwide Setting. *N Engl J Med*. 2021;385(12):1078-90.
3. Sadoff J, Gray G, Vandebosch A, Cardenas V, Shukarev G, Grinsztejn B, et al. Safety and Efficacy of Single-Dose Ad26.COV2.S Vaccine against Covid-19. *N Engl J Med*. 2021;384(23):2187-201.
4. CDC. Mitigation measures for COVID-19 in households and markets in non-US low-resource settings. 2021.
5. CDC. Testing Strategies for SARS-CoV-2. 2021.
6. Erster O, Beth-Din A, Asraf H, Levy V, Kabat A, Mannasse B, et al. SPECIFIC DETECTION OF SARS-COV-2 B.1.1.529 (OMICRON) VARIANT BY FOUR RT-qPCR DIFFERENTIAL ASSAYS. *medRxiv*. 2021:2021.12.07.21267293.
7. Binnicker MJ. Challenges and Controversies to Testing for COVID-19. *J Clin Microbiol*. 2020;58(11).
8. Vandenberg O, Martiny D, Rochas O, van Belkum A, Kozlakidis Z. Considerations for diagnostic COVID-19 tests. *Nat Rev Microbiol*. 2021;19(3):171-83.

- 342 9. Bwire GM, Majigo MV, Njiro BJ, Mawazo A. Detection profile of SARS-CoV-2 using
343 RT-PCR in different types of clinical specimens: A systematic review and meta-analysis.
344 J Med Virol. 2021;93(2):719-25.
- 345 10. LeBlanc JJ, Heinstein C, MacDonald J, Pettipas J, Hatchette TF, Patriquin G. A
346 combined oropharyngeal/nares swab is a suitable alternative to nasopharyngeal swabs
347 for the detection of SARS-CoV-2. J Clin Virol. 2020;128:104442.
- 348 11. Kahn R, Kennedy-Shaffer L, Grad YH, Robins JM, Lipsitch M. Potential Biases
349 Arising From Epidemic Dynamics in Observational Seroprotection Studies. Am J
350 Epidemiol. 2021;190(2):328-35.
- 351 12. Hay JA, Kennedy-Shaffer L, Kanjilal S, Lennon NJ, Gabriel SB, Lipsitch M, et al.
352 Estimating epidemiologic dynamics from cross-sectional viral load distributions. Science.
353 2021;373(6552).
- 354 13. Tom MR, Mina MJ. To Interpret the SARS-CoV-2 Test, Consider the Cycle
355 Threshold Value. Clin Infect Dis. 2020;71(16):2252-4.
- 356 14. Peccia J, Zulli A, Brackney DE, Grubaugh ND, Kaplan EH, Casanovas-Massana
357 A, et al. Measurement of SARS-CoV-2 RNA in wastewater tracks community infection
358 dynamics. Nat Biotechnol. 2020;38(10):1164-7.
- 359 15. Randazzo W, Truchado P, Cuevas-Ferrando E, Simon P, Allende A, Sanchez G.
360 SARS-CoV-2 RNA in wastewater anticipated COVID-19 occurrence in a low prevalence
361 area. Water Res. 2020;181:115942.
- 362 16. Mina MJ, Parker R, Larremore DB. Rethinking Covid-19 Test Sensitivity - A
363 Strategy for Containment. N Engl J Med. 2020;383(22):e120.

- 364 17. He X, Lau EHY, Wu P, Deng X, Wang J, Hao X, et al. Temporal dynamics in viral
365 shedding and transmissibility of COVID-19. *Nat Med*. 2020;26(5):672-5.
- 366 18. Larremore DB, Wilder B, Lester E, Shehata S, Burke JM, Hay JA, et al. Test
367 sensitivity is secondary to frequency and turnaround time for COVID-19 screening. *Sci*
368 *Adv*. 2021;7(1).
- 369 19. Harmon A, Chang C, Salcedo N, Sena B, Herrera BB, Bosch I, et al. Validation of
370 an At-Home Direct Antigen Rapid Test for COVID-19. *JAMA Netw Open*.
371 2021;4(8):e2126931.
- 372 20. Nash B, Badea A, Reddy A, Bosch M, Salcedo N, Gomez A. Validating and
373 Modeling the Impact of High-Frequency Rapid Antigen Screening on Covid-19 Spread
374 and Outcomes. *J Clin Trials*. 2021;11:483.
- 375 21. U.S. FDA. In Vitro Diagnostics EUAs - Antigen Diagnostic Tests for SARS-CoV-2.
- 376 22. Corman VM, Haage VC, Bleicker T, Schmidt ML, Muhlemann B, Zuchowski M, et
377 al. Comparison of seven commercial SARS-CoV-2 rapid point-of-care antigen tests: a
378 single-centre laboratory evaluation study. *Lancet Microbe*. 2021;2(7):e311-e9.
- 379 23. Pickering S, Batra R, Merrick B, Snell LB, Nebbia G, Douthwaite S, et al.
380 Comparative performance of SARS-CoV-2 lateral flow antigen tests and association with
381 detection of infectious virus in clinical specimens: a single-centre laboratory evaluation
382 study. *Lancet Microbe*. 2021;2(9):e461-e71.
- 383 24. Salcedo N, Harmon A, Herrera BB. Pooling of Samples for SARS-CoV-2 Detection
384 Using a Rapid Antigen Test. *Front Trop Dis*. 2021;2:707865.

- 385 25. Ganguli A, Mostafa A, Berger J, Aydin MY, Sun F, Ramirez SAS, et al. Rapid
386 isothermal amplification and portable detection system for SARS-CoV-2. Proc Natl Acad
387 Sci U S A. 2020;117(37):22727-35.
- 388 26. Baek YH, Um J, Antigua KJC, Park JH, Kim Y, Oh S, et al. Development of a
389 reverse transcription-loop-mediated isothermal amplification as a rapid early-detection
390 method for novel SARS-CoV-2. Emerg Microbes Infect. 2020;9(1):998-1007.
- 391 27. Rabe BA, Cepko C. SARS-CoV-2 detection using isothermal amplification and a
392 rapid, inexpensive protocol for sample inactivation and purification. Proc Natl Acad Sci U
393 S A. 2020;117(39):24450-8.
- 394 28. Qian J, Boswell SA, Chidley C, Lu ZX, Pettit ME, Gaudio BL, et al. An enhanced
395 isothermal amplification assay for viral detection. Nat Commun. 2020;11(1):5920.
- 396 29. Mancuso CP, Lu ZX, Qian J, Boswell SA, Springer M. A Semi-Quantitative
397 Isothermal Diagnostic Assay Utilizing Competitive Amplification. Anal Chem.
398 2021;93(27):9541-8.
- 399 30. Piepenburg O, Williams CH, Stemple DL, Armes NA. DNA detection using
400 recombination proteins. PLoS Biol. 2006;4(7):e204.
- 401 31. U.S. FDA. In Vitro Diagnostics EUAs - Molecular Diagnostic Tests for SARS-CoV-
402 2.
- 403 32. Donato LJ, Trivedi VA, Stransky AM, Misra A, Pritt BS, Binnicker MJ, et al.
404 Evaluation of the Cue Health point-of-care COVID-19 (SARS-CoV-2 nucleic acid
405 amplification) test at a community drive through collection center. Diagn Microbiol Infect
406 Dis. 2021;100(1):115307.

33. Ptasinska A, Whalley C, Bosworth A, Poxon C, Bryer C, Machin N, et al. Diagnostic accuracy of loop-mediated isothermal amplification coupled to nanopore sequencing (LamPORE) for the detection of SARS-CoV-2 infection at scale in symptomatic and asymptomatic populations. Clin Microbiol Infect. 2021;27(9):1348 e1- e7.
34. Dong Y, Wu X, Li S, Lu R, Li Y, Wan Z, et al. Comparative evaluation of 19 reverse transcription loop-mediated isothermal amplification assays for detection of SARS-CoV-2. Sci Rep. 2021;11(1):2936.
35. Paltiel AD, Zheng A, Walensky RP. Assessment of SARS-CoV-2 Screening Strategies to Permit the Safe Reopening of College Campuses in the United States. JAMA Netw Open. 2020;3(7):e2016818.
36. Pavelka M, Van-Zandvoort K, Abbott S, Sherratt K, Majdan M, group CC-w, et al. The impact of population-wide rapid antigen testing on SARS-CoV-2 prevalence in Slovakia. Science. 2021;372(6542):635-41.

FIGURES

Figure 1. Schematic of RT-RPA assay versus antigen-based test. A) In RT-RTPA, viral RNA is copied to cDNA by reverse transcriptase, then degraded by RNase H. Using a forward and a FAM-labeled reverse pair of primers specific to a target sequence, the cDNA product is amplified by RPA, then denatured and hybridized to a biotinylated probe. FAM-labeled and biotin-labeled products are detected on a lateral flow strip using molecules specific for FAM and biotin and nanoparticles. B) In an antigen test, protein targets are detected by a lateral flow strip using protein-specific antibodies and nanoparticles. C) A mobile phone application was used to image capture, machine-read,

and quantify test results. The average pixel intensity is quantified at the test line, control line, and background areas. The background-subtracted test line signal is then normalized to the background-subtracted control line and expressed at % of control.

Figure 2. Analytical sensitivity of the RT-RPA assay and the antigen test using nasal swab dilution specimens. A) Lateral flow strips for the RT-RPA reactions with dilution specimens containing RNA copies ranging from 0-1,000. B) Plot from the RT-RPA assay results quantified by the mobile phone application. The x-axis corresponds to dilutions specimens with known input copies of SARS-CoV-2 RNA. The y-axis corresponds to background subtracted test signal normalized to the control line for each lateral flow strip. Test results less than 10% of control are considered negative results, which is indicated by the black dashed line. C) Lateral flow strips for the antigen tests with dilution specimens

containing RNA copies ranging from 0-200,000 D) Plot from the antigen tests results quantified by the mobile phone application. The x-axis corresponds to dilutions specimens with known input copies of SARS-CoV-2 RNA. The y-axis corresponds to background subtracted test signal normalized to the control line for each lateral flow strip. Test results less than 10% of control are considered negative results, which is indicated by the black dashed line.

Figure 3. Clinical performance of the RT-RPA assay and the antigen test using nasal swab specimens collected from individuals with or without COVID-19. A) Comparative evaluation of the RT-RPA assay (purple) and the antigen test (blue) using nasal swab specimens from asymptomatic cases. Comparative performance between the tests was plotted according to qPCR positive (Ct values between <20 to <40) and negative results. B) Comparative evaluation of the RT-RPA assay (purple) and the antigen test (blue) using nasal swab specimens from symptomatic cases. Comparative performance between the tests was plotted according to qPCR positive (Ct values between <20 to <40) and negative results.

TABLES

Table 1. Comparative performance of the RT-RPA assay and the antigen test against qPCR in asymptomatic cases.

Asymptomatic Phase								
		qPCR					95% CI	
		+	-	Total	PPA	100.00%	90.00%	100.00%
RT-RPA	+	35	0	35	NPA	100.00%	95.26%	100.00%
	-	0	76	76	PPV	100.00%		
Total		35	76	111	NPV	100.00%		
		qPCR					95% CI	
		+	-	Total	PPA	82.86%	66.35%	93.44%
Antigen	+	29	1	30	NPA	98.68%	92.89%	99.97%
	-	6	75	81	PPV	96.67%	80.45%	99.51%
Total		35	76	111	NPV	92.59%	85.78%	96.28%

Table 2. Comparative performance of the RT-RPA assay and the antigen test against qPCR in symptomatic cases.

Symptomatic Phase								
		qPCR					95% CI	
		+	-	Total	PPA	95.83%	78.88%	99.89%
RT-RPA	+	23	0	23	NPA	100.00%	90.75%	100.00%
	-	1	38	39	PPV	100.00%		
Total		24	38	62	NPV	97.44%	84.80%	99.62%
		qPCR					95% CI	
		+	-	Total	PPA	91.67%	73.00%	98.97%
Antigen	+	22	0	22	NPA	100.00%	90.75%	100.00%
	-	2	38	40	PPV	100.00%		
Total		24	38	62	NPV	95.00%	83.45%	98.62%