Adherence of SARS-CoV-2 delta variant to surgical mask and N95 respirators

Ana C. Lorenzo-Leal¹, Selvarani Vimalanathan^{1,2}, Horacio Bach^{1,*}

¹Faculty of Medicine, Division of Infectious Diseases, University of British Columbia, Vancouver, BC, Canada

² Pathology & Laboratory Medicine, University of British Columbia, Vancouver, BC, Canada.

*Corresponding author: 2660 Oak Street, University of British Columbia, Vancouver, BC, V6H 3Z6, Canada. Phone: 1-604-875-4111 x 62107. Email: hbach@mail.ubc.ca

Abstract

The use of facial protection, including masks and respirators, has been adopted globally due to the COVID-19 pandemic. These products have been demonstrated to be effective in reducing the transmission of the virus. To determine whether or not the virus adheres to masks and respirators, we dissected four respirators and one surgical mask into layers. These individual layers were contaminated with the SARS-CoV-2 delta variant, and its release by vortexing was performed. Samples were used to infect Vero cells, and a plaque assay was used to determine to evaluate the adherence of the virus. Results showed that a cumulative log reduction of the layers reduced the load of the virus six-folds. Our study confirms the effectiveness of facial protection in reducing the transmission and or infection of the virus.

Keywords: surgical masks, respirators, SARS-CoV-2, COVID-19, adherence

1. Introduction

The pandemic of the novel coronavirus SARS-CoV-2, the causative agent of COVID-19, was initially identified in Wuhan, China, and has spread worldwide. According to the World Health Organization (WHO), as of November 20th, 2021, there have been 5,127,696 deaths and 255,324,963 confirmed cases of COVID-19 [1].

COVID-19 disease seems to be spreading through airborne transmission via aerosol, and as it is known that respiratory infections mainly occur when breathing, speaking, sneezing, coughing, or talking occurs. These transmissions are facilitated by droplets (> 5 to $10\ \mu\text{m}$) and/or aerosols ($\leq 5\ \mu\text{m}$) containing the virus SARS-CoV-2 (ref). Aerosol's accumulation, also known as the "gas cloud," could remain infectious for hours in indoor spaces. Thus, the implementation of masks or respirators has a tremendous effect on reducing aerosol transmission [2–5]. This physical protection is in use and has been adopted globally.

There are different types of masks or respirators with various levels of protection, which can be reusable or disposable. Generally, masks do not fit as tightly as respirators. For example, N95 respirators are classified by the U.S. National Institute for Occupational Safety and Health (NIOSH) as air filtration, filtering at least 95% of airborne particles. N95 respirators and surgical masks are examples of the most common disposable masks, while reusable are mainly industrial masks with half or full facepiece respirators [5].

According to the Centers for Disease Control and Prevention (CDC) and the NIOSH, surgical masks are approved by the Food and Drug Administration (FDA) and intended to use as protection to the wearer against multiple hazardous environments, such as those created by sprays, dangerous fluids, splashes, and droplets. However, masks would not always confer respiratory

protection because of their lower reliable level of filtration against smaller particles that could be inhaled, such as airborne particles.

Surgical masks contain mainly three-layer structures with different functions. For example, the outer layer is repelling water, the middle layer functions as a trap of particles and prevents the penetration of microorganisms, and the inner layer absorbs moisture [2,6].

On the other hand, N95 respirators are defined by NIOSH as a filtration system that retains at least 95% of different airborne sizes particles (from aerosols to large droplets), reducing the exposure to the user [7]. Respirators may contain extra layers compared to surgical masks. For example, N95 respirators have between three to five layers, comprising an inner and outer layer, a filter layer, and in some cases, a support layer [6].

The penetration or filtration mechanism of aerosols containing SARS-CoV-2 in masks has been studied and reported in different studies [2,3,6,8,9]. Moreover, contamination with SARS-CoV-2 on the masks' outer surface has also been reported [10]. Still, a detailed investigation addressing the question of whether or not the virus remains adhered to the layers has not been reported. Then, this study aimed to evaluate the adherence of the SARS-CoV-2 delta variant on different layers of a surgical mask and N95 respirators.

2. Methods

2.1. Virus and cell lines

SARS-CoV-2 delta variant (B.1.617.2) was provided by the British Columbian Centre for Disease Control Public Health Laboratory, Vancouver, Canada. Vero E6 cells (ATCC CCL-81) were used to replicate the virus and infections. Cells were grown in MEM (Invitrogen) supplemented with 5% fetal calf serum (Invitrogen), pyruvate, and non-essential amino acids. All

the incubations were performed in an incubator at 37°C supplemented with 5% CO₂. All the work performed in this study was conducted in level 3 at the Facility for Infectious Disease and Epidemic Research (University of British Columbia, Vancouver, BC, Canada).

Viral stocks were prepared by infecting 80% confluent Vero E6 cells with 1 mL of the virus at a 5x10⁵ PFU/mL concentration. After 72 h, the supernatant was collected and cleared by centrifugation at 3200 × g for 10 min. The supernatant was aliquoted and kept at -80°C until further use. The titer of the virus was determined to be 5.9x10⁵ pfu/mL by plaque assay.

A plaque assay was performed overnight by dispensing 1.5x10⁵ Vero cells into 12-well plates. The next day, the 80-100% confluency was verified, and cells were infected with the virus for one h. Then, the virus was washed away, and the monolayer was overlaid with 1.5 mL of 2 x MEM supplemented with 2% cellulose (Sigma) (v/v 1:1)

2.2. Mask or respirators preparation

Five different masks or respirators (Table 1) were used in this study. Pieces of 1 cm² were cut, separated into individual layers, decontaminated by submerging the pieces into 70% ethanol for 10 min, and dry overnight at 37 °C inside a decontaminated Biosafety Containment Level 2. The decontaminated layers were stored in empty sterile 1.5 mL tubes until further use.

Table 1. Characteristics of the masks and respirators used in this study

Brand	Layer	Material	Reference
1 North N95 7130 (Lot No. 9E001) Closed-cell vinyl foam	Outer Filter Support Inner	Polypropylene Polypropylene composite Polyester Polyester	[11]

2 North N95 5130 (Lot No. 9G004) Open-cell foam seal	Outer	Polyester	[11]
	Filter	Polypropylene composite	
	Inner	Polyester	
Three 3-Ply surgical mask	Outer	Fluid resistant	[12]
	Filter	Polypropylene	
	Inner	Absorbent	
4 3M 8210 N95 (Lot No. A20047) Standard respirator	Outer	Polyester	[13]
	Filter	Polypropylene	
	Inner	Polyester	
5 3M 1860 N95 (Lot No. B16047) Surgical respirator	Outer	Polyester	[13]
	Filter	Polypropylene	
	Inner	Polyester

Figure 1. Mask or respirators layers. Description of the mask and respirator layers used in this study. Created by Biorender.com

2.3. Contamination of layers

The layers were exposed to 2×10^3 viruses (in 10 μ L MEM). After 10 min (time to dry the drop), the layers were submerged in 300 μ L of OptiMEM (Invitrogen) in a 1.5 mL screw-top tube

and vortexed vigorously for 1 min every. The samples were subjected to 10-fold dilutions using OptiMEM in 1.5 mL tubes.

2.4. Plaque assay

5.0×10^5 Vero E6 cells in complete MEM were seeded into 12-well cell culture plates (Thermo Scientific) and incubated overnight at 37 °C supplemented with 5% CO₂. Cells were then checked for confluency, and the medium was removed from the wells. The serial dilutions (300 µL) were added to the corresponded wells, and the plates were incubated as explained above. The plates were placed back in the incubator for an hour to allow infection. The plates were gently rocked every 15 min. Then, the virus and the medium were removed from the wells, and an overlay of 2% cellulose and 2x complete MEM (as detailed above) was gently added to each well. Plates were incubated for 72 h, fixed for 30 min with 4% *p*-formaldehyde (PFA, J.T. Baker) in PBS. After removing the PFA, the cells were stained with 250 µL of 1% crystal violet (Sigma), dissolved in 20% methanol (Fisher) for 3,0 min, washed twice with water, dried, and the number of plaques was counted and recorded.

3. Results and discussion

The adherence of SARS-CoV-2 to different types of buccal protection showed that all of the layers could retain the virus at different extents. In addition, adherent viruses were not released upon vortexing the samples, suggesting that they are tightly bound to the layers once they are deposited and dried on their surface.

Analysis of the results showed that an accumulation of the log reductions of the layers ranged between 6.15-8.45 (Figure 2). In addition, respirator **2** showed superior retention of the

virus with an increase of approximately 1-2 log reduction (Figure 2B) compared to the rest of the tested products.

Figure 2. Viral titers of SARS-CoV-2 deposited onto layers. The virus was deposited on the layers, and after drying, non-adherent viruses were released from the layers by vortexing and according to the Method section. (A) Respirator 7130, (B) Respirator 5130, (C) Surgical masks, (D) Respirator 8210, and (E) Respirator 1860. Shown is the mean of at least three independent experiments $\pm$ SD. The number located in the lower panel is the cumulative log reduction of the layers. The arrangement of the layers is from the outer layer (a) to the inner layer (c), except for **one** that has four layers (d).

People use medical masks as protection from sprays and splashes of fluids by sick people. Although extended usage or reuse of masks has become common, especially during the present pandemics, this could lead to an infection of the person wearing the mask due to the possible presence of pathogens in different layers of the masks [10].

Masks materials (Table 1) are mainly nonwoven fabric made by a mass of fibers (such as polypropylene, polyester, and polyurethane) bonded by different means such as mechanical, heat or chemical. The efficiency of particle filtration (permeability), bacterial or virus filtration, differential pressure, flammability, flexibility, and resistance to fluids are some of the performance characteristics for materials used in masks [2,3,6]. Regarding permeability, when resistance to fluid material is used, aerosols are likely to travel around the mask following contours seeking low resistance areas or leaks, rather than penetration of the filter material [3].

Between the different layers in a mask or respirator, the most important one is the filter layer, which is mainly produced through a melt-blown process, where a polymer (such as polyethylene) is exposed to blows of high-velocity air, forming a web shape of filaments. This filter layer (commonly used in N95 respirators) has an electrostatic charge; therefore, the polymer filtration capability for small particles increases because of electrostatic adsorption [2,14]. In our study, the respirators contained at least one layer of polypropylene, suggesting that an electrostatic layer increased the retention of the virus.

A model of viral retention based on the electrostatic charges on the surface of the virus and the modified polypropylene (electret) might explain one of the reasons for the log reduction of the virus. In the case of the virus, the Spike protein surrounds the viral particle. Assuming that other components of the virus are not contributing to the surface charge of the virus (lipids and nucleocapsid are part of the core of the virus [15], we can attribute the net surface charge of the virus to the Spike protein. This protein looks to have a particular disorder [16] probably because of its interaction with the membrane proteins [17]. Thus, the electrostatic interaction of the virus and the electret is based on the charge of the Spike protein [18]. One of the main variables regarding the isoelectric point of the Spike protein is the large number of ionizable amino acids

that contribute to the net charge (e.g., C, D, E, H, K, R, and Y) [15]. Then, the pH will play an important role in the fluctuation of the surface charge at which the virus.

The isoelectric point of the Spike protein is 6.2 [19]. Although this value is lower than the physiological pH (7.0), most Spike protein is negatively charged [15], except the receptor-binding domain. This receptor is located in the apical portion of the spike protein, and it has been calculated to remain positively charged through a range of pH values [20]. Interestingly, the positive charge of the receptor-binding domain has increased in the SARS-CoV-2 variants [20] because of the mutation observed in this region.

On the other hand, studies reported negative surface charges on the polypropylene electrets, varying between -200 to -1350 (V) [21,22]. Thus, a strong electrostatic attraction occurs between the receptor-binding domain of the Spike protein and the polypropylene electret, contributing to the retention of the viruses on the layers.

4. Conclusions

In this study, we report the adherence of the SARS-CoV-2 delta variant to different buccal protection (mask and respirators). We hypothesize that a strong electrostatic binding occurs, and as a result, the cumulative viral log reduction is over 6. Our results also showed that masks and respirators efficiently reduce infections either from the infected (wearer) or the recipient person.

Acknowledgement: We thank Dr. Mel Krajden from the British Columbian Centre for Disease Control Public Health Laboratory, Vancouver, Canada for the access to the SARS-CoV-2 delta variant.

Author Contributions: H.B. conceptualized, reviewed, edited the manuscript, and procured financing acquisition. A.L. investigated, wrote, reviewed, and edited the manuscript. S.V. investigated and reviewed the manuscript. All authors have read and agreed to the published version of the manuscript.

Funding: This work was supported by the Canadian Institutes of Health Research (CIHR) grant Canadian 2019 Novel Coronavirus (COVID-19) Rapid Research Funding Opportunity Therapeutics to H.B (No. OV3- 170627).

Conflicts of Interest: The authors have nothing to disclose.

References

1. World Health Organization WHO Coronavirus (COVID-19) Dashboard Available online: <https://covid19.who.int/> (accessed on 11 September 2021).
2. Chua, M.H.; Cheng, W.; Goh, S.S.; Kong, J.; Li, B.; Lim, J.Y.C.; Mao, L.; Wang, S.; Xue, K.; Yang, L.; et al. Face Masks in the New COVID-19 Normal: Materials, Testing, and Perspectives. *Research* **2020**, *2020*, 1–40, doi:10.34133/2020/7286735.
3. Hill, W.C.; Hull, M.S.; MacCuspie, R.I. Testing of Commercial Masks and Respirators and Cotton Mask Insert Materials Using SARS-CoV-2 Virion-Sized Particulates: Comparison of Ideal Aerosol Filtration Efficiency versus Fitted Filtration Efficiency. *Nano Lett.* **2020**, *20*, 7642–7647, doi:10.1021/acs.nanolett.0c03182.
4. Howard, J.; Huang, A.; Li, Z.; Tufekci, Z.; Zdiamal, V.; van der westhuizen, H.-M.; von Delft, A.; Price, A.; Fridman, L.; Tang, L.-H.; et al. An Evidence Review of Face Masks against COVID-19. **2021**, *118*, 1–12, doi:<https://doi.org/10.1073/pnas.2014564118>.
5. Prather, K.A.; Wang, C.C.; Schooley, R.T. Reducing Transmission of SARS-CoV-2. *Science* **2020**, *368*, 1422–1424, doi:10.1126/science.abc6197.
6. Tcharkhtchi, A.; Abbasnezhad, N.; Zarbini Seydani, M.; Zirak, N.; Farzaneh, S.; Shirinbayan, M. An Overview of Filtration Efficiency through the Masks: Mechanisms of

- the Aerosols Penetration. *Bioact. Mater.* **2021**, *6*, 106–122, doi:10.1016/j.bioactmat.2020.08.002.
7. CDC Healthcare Workers Available online: <https://www.cdc.gov/coronavirus/2019-ncov/hcp/respirator-use-faq.html> (accessed on 3 November 2021).
8. Konda, A.; Prakash, A.; Moss, G.A.; Schmoldt, M.; Grant, G.D.; Guha, S. Aerosol Filtration Efficiency of Common Fabrics Used in Respiratory Cloth Masks. *ACS Nano* **2020**, *14*, 6339–6347, doi:10.1021/acsnano.0c03252.
9. Lustig, S.R.; Biswakarma, J.J.H.; Rana, D.; Tilford, S.H.; Hu, W.; Su, M.; Rosenblatt, M.S. Effectiveness of Common Fabrics to Block Aqueous Aerosols of Virus-like Nanoparticles. *ACS Nano* **2020**, *14*, 7651–7658, doi:10.1021/acsnano.0c03972.
10. Chughtai, A.A.; Stelzer-Braid, S.; Rawlinson, W.; Pontivivo, G.; Wang, Q.; Pan, Y.; Zhang, D.; Zhang, Y.; Li, L.; MacIntyre, C.R. Contamination by Respiratory Viruses on Outer Surface of Medical Masks Used by Hospital Healthcare Workers. *BMC Infect. Dis.* **2019**, *19*, 491, doi:10.1186/s12879-019-4109-x.
11. Datasheet: Technical Data Sheet Technical Data 7130n95 8140p95 8150p100 North Filtering Facepiece Available online: <http://www.industrycortex.com/datasheets/profile/1259774979/technical-data-sheet-technical-data-7130n95-8140p95-8150p100-nort> (accessed on 3 November 2021).
12. Chin, A.W.H.; Chu, J.T.S.; Perera, M.R.A.; Hui, K.P.Y.; Yen, H.-L.; Chan, M.C.W.; Peiris, M.; Poon, L.L.M. Stability of SARS-CoV-2 in Different Environmental Conditions. *micro* **2020**, *1*, e10, doi:https://doi.org/10.1016/S2666-5247(20)30003-3.
13. 3M™ Particulate Respirator 8210, N95 160 EA/Case Available online: https://www.3m.com/3M/en_US/p/d/v000585997/ (accessed on 3 November 2021).
14. Campos, R.K.; Jin, J.; Rafael, G.H.; Zhao, M.; Liao, L.; Simmons, G.; Chu, S.; Weaver, S.C.; Chiu, W.; Cui, Y. Decontamination of SARS-CoV-2 and Other RNA Viruses from N95 Level Meltblown Polypropylene Fabric Using Heat under Different Humidities. *ACS Nano* **2020**, *14*, 14017–14025, doi:10.1021/acsnano.0c06565.
15. Javidpour, L.; Božič, A.; Naji, A.; Podgornik, R. Electrostatic Interactions between the SARS-CoV-2 Virus and a Charged Electret Fibre. *Soft Matter* **2021**, *17*, 4296–4303, doi:10.1039/D1SM00232E.
16. Beniac, D.R.; Andonov, A.; Grudeski, E.; Booth, T.F. Architecture of the SARS Coronavirus Prefusion Spike. *Nat. Struct. Mol. Biol.* **2006**, *13*, 751–752, doi:10.1038/nsmb1123.
17. Gorbalenya, A.E.; Baker, S.C.; Baric, R.S.; de Groot, R.J.; Drosten, C.; Gulyaeva, A.A.; Haagmans, B.L.; Lauber, C.; Leontovich, A.M.; Neuman, B.W.; et al. The Species Severe Acute Respiratory Syndrome-Related Coronavirus: Classifying 2019-NCoV and Naming It SARS-CoV-2. *Nat. Microbiol.* **2020**, *5*, 536–544, doi:10.1038/s41564-020-0695-z.
18. Heffron, J.; Mayer, B.K. Virus Isoelectric Point Estimation: Theories and Methods. *Appl. Environ. Microbiol.* **2021**, *87*, e02319-20, doi:10.1128/AEM.02319-20.
19. Scheller, C.; Krebs, F.; Minkner, R.; Astner, I.; Gil-Moles, M.; Wätzig, H. Physicochemical Properties of SARS-CoV-2 for Drug Targeting, Virus Inactivation and Attenuation, Vaccine Formulation and Quality Control. *Electrophoresis* **2020**, *41*, 1137–1151, doi:https://doi.org/10.1002/elps.202000121.
20. Pawłowski, P.H. Additional Positive Electric Residues in the Crucial Spike Glycoprotein S Regions of the New SARS-CoV-2 Variants. *Infect. Drug Resist.* **2021**, *14*, 5099–5105, doi:10.2147/IDR.S342068.

21. Tsai, P.P.; Schreuder-Gibson, H.; Gibson, P. Different Electrostatic Methods for Making Electret Filters. *J. Electrostat.* **2002**, *54*, 333–341, doi:10.1016/S0304-3886(01)00160-7.
22. Wang, J.; Rychkov, D.; Gerhard, R. Chemical Modification with Orthophosphoric Acid Enhances Surface-Charge Stability on Polypropylene Electrets. *Appl. Phys. Lett.* **2017**, *110*, 192901, doi:10.1063/1.4983348.