Limit of Detection for Rapid Antigen Testing of the SARS-CoV-2 Omicron Variant

Sydney Stanley¹, Donald J. Hamel¹, Ian D. Wolf¹, Stefan Riedel^{2,3}, Sanjucta Dutta², Annie Cheng², James E. Kirby^{*#2,3}, Phyllis J. Kanki^{*#1}

¹Department of Immunology and Infectious Diseases, Harvard T.H. Chan School of Public Health, Boston, Massachusetts, USA

²Beth Israel Deaconess Medical Center, Boston, Massachusetts, USA

³Harvard Medical School, Boston, Massachusetts, USA

*Co-Senior Authors

Address for correspondence:

Phyllis J. Kanki, DVM, PhD

651 Huntington Avenue

FXB Building, Room 405B

Boston, Massachusetts 02115

617.432.1267

pkanki@hsph.harvard.edu

James E. Kirby, MD

Beth Israel Deaconess Medical Center

330 Brookline Avenue – YA309

Boston MA 02215

617-667-3648

jekirby@bidmc.harvard.edu

Keywords: COVID-19, SARS-CoV-2, antigen test, Omicron, limit of detection, analytical sensitivity

Abstract

There has been debate in the literature about the ability of antigen tests to detect the SARS-CoV-2 Omicron variant including indication on the US Food and Drug administration website that antigen tests may have lower sensitivity for the Omicron variant without provision of data or the potential scale of the issue (see <https://www.fda.gov/medical-devices/coronavirus-covid-19-and-medical-devices/sars-cov-2-viral-mutations-impact-covid-19-tests-omicronvariantimpact>, accessed 1/27/2022). Here we determined the limit of detection (LoD) for the Omicron variant compared with the WA1 strain used for LoD studies described in the Instructions for Use for all Emergency Use Authorization (EUA)-approved antigen tests. Using live virus (to avoid artifactual findings potentially obtained with gamma-irradiated or heat-killed virus) quantified by plaque forming units (PFU), we examined the analytical sensitivity of three antigen tests widely used in the United States: the Abbott Binax Now, the AccessBio CareStart, and LumiraDx antigen tests. We found that the 95% detection threshold (LoD) for antigen tests was at least as good for Omicron as for the WA1 strain. Furthermore, the relationship of genome copies to plaque forming units for Omicron and WA1 overlap. Therefore, the LoD equivalency also applies if the quantitative comparator is genome copies determined from live virus preparations. Taken together, our data support the continued ability of the antigen tests examined to detect the Omicron variant.

To bolster COVID-19 pandemic mitigation efforts, the U.S. Food and Drug Administration (FDA) issued Emergency Use Authorization (EUA) for easy-to-use rapid antigen tests instrumental for diagnosis and surveillance of SARS-CoV-2 infection (1-2). Unlike sensitive molecular tests that detect multiple SARS-CoV-2 genes, antigen tests target a singular yet genetically-conserved nucleocapsid viral protein (3-6). As the pandemic continues, some hypothesized that new SARS-CoV-2 variants might compromise antigen test performance. This concern heightened with the spread of Omicron, the B.1.1.529 variant of concern (VoC) that caused 99.5% of SARS-CoV-2 infections in the United States early 2022 (7-8). Beyond the striking 36 amino acid mutations in the spike protein, Omicron also harbors P13L, Δ 31-33, R203K, and G204R nucleocapsid mutations (9). The limit of detection (LoD) of many FDA EUA antigen tests were established with gamma-irradiated or heat-inactivated preparations of the USA WA1/2020 (WA1) reference strain (13) lacking nucleocapsid mutations. This includes at-home lateral flow tests like the BinaxNOW COVID-19 Ag Card (Abbott Diagnostics Scarborough, Inc., Scarborough, ME) and the CareStart COVID-19 Antigen Home Test (Access Bio, Inc., Somerset, NJ), and the LumiraDx SARS-CoV-2 Ag Test (LumiraDx UK Ltd., Alloa, Great Britain), a microfluidic immunofluorescence assay for clinical laboratory testing (10-12). In the present study, we used cultured plaque-titered live Omicron and WA1 virus to assess differences in the LoD with the Binax, CareStart, and LumiraDx tests.

The WA1 (13) and Omicron lh01 (NCBI accession OL719310) virus were titrated with standard plaque (13) and calibrated RT-qPCR (14) assays. Ten-fold serial dilutions in PBS ranging from 2.5×10^4 to 2.5 plaque forming units (PFU)/mL were applied to swabs in 50 μ L volumes and tested in triplicate according to manufacturer instructions (10-12). Binax and CareStart kits contained all required consumables; iClean foam swabs (Supera CY-FS742, Houston, TX) were

used with the LumiraDx test. After identifying the lowest 10-fold dilution with three replicate positive tests, we iteratively tested 3-fold dilutions around this concentration until identifying the lowest dilution (the LoD) in which at least 19 of 20 replicates ($\geq 95\%$) were positive.

The LumiraDx LoD for both Omicron and WA1 was 2.5×10^2 PFU/mL (12.5 PFU/swab or 1.0×10^6 genome copies (gc)/swab) (Fig. 1). The Binax LoD was 8.3×10^1 PFU/mL (4.2 PFU/swab, 3.4×10^5 gc/swab) and 2.5×10^2 PFU/mL (12.5 PFU/swab, 1.0×10^6 gc/swab) for Omicron and WA1, respectively. The CareStart LoD was 2.8×10^3 PFU/mL (1.4×10^2 PFU/swab, 1.1×10^7 gc/swab) and 8.3×10^3 PFU/mL (4.2×10^2 PFU/swab, 3.5×10^7 gc/swab) for Omicron and WA1, respectively. The nearly identical relationship of PFU to genome copies for each variant indicates that the Omicron variant mutations do not change underlying diagnostic relationships and parameters (Figure 2). Our use of live virus, analyte volume, and swab type may explain the slight discrepancy with the manufacturers' LoDs. Our findings are consistent with similar investigations, but these studies fell short of the FDA's EUA requirement of 20 LoD replicates or included tests unavailable in the United States (15-17). In all, we demonstrate that the rapid antigen tests evaluated detect Omicron effectively, allaying concerns on the impact of the nucleocapsid mutations. Rapid antigen tests remain critical public health tools towards reducing SARS-CoV-2 variant transmission.

Figure 1. Limit of detection of the antigen tests. Limit of detection (LoD) in PFU/mL determined in our analysis (bars). Dotted lines reference the manufacturer reported LoD in respective Instructions for Use (IFU) documents (10-12), converted from TCID₅₀/mL to PFU/mL by multiplying the TCID₅₀/mL by 0.7, a standard conversion based on the Poisson distribution: LumiraDx (32 TCID₅₀/mL, 2.2×10^1 PFU/mL); Binax (140 TCID₅₀/mL, 9.8×10^1 PFU/mL), CareStart (800 TCID₅₀/mL, 5.6×10^2 PFU/mL).

Figure 2. Correlation of PFU/mL and viral load in genome copies/mL. Stocks of each strain was serially diluted 10-fold in PBS and analyzed by PFU (13) and calibrated RT-qPCR assays (14). Both axes are on a Log10 scale.

References

1. The U.S. Food and Drug Administration. In Vitro Diagnostics EUAs - Antigen Diagnostic Tests for SARS-CoV-2. Available at: <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas-antigen-diagnostic-tests-sars-cov-2>. Accessed 24 January 2022.
2. The Centers for Disease Control and Prevention. Overview of Testing for SARS-CoV-2, the virus that causes COVID-19. Available at: <https://www.cdc.gov/coronavirus/2019-ncov/hcp/testing-overview.html>. Accessed 24 January 2022.
3. Jegerlehner S, Suter-Riniker F, Jent P, Bittel P, Nagler M. 2021. Diagnostic accuracy of a SARS-CoV-2 rapid antigen test in real-life clinical settings. *Int J Infect Dis* 109:118-122.
4. Allan-Blitz LT, Klausner JD. 2021. A Real-World Comparison of SARS-CoV-2 Rapid Antigen Testing versus PCR Testing in Florida. *J Clin Microbiol* 59(10):e0110721.
5. Pray IW, Ford L, Cole D, Lee C, Bigouette JP, Abedi GR, Bushman D, Delahoy MJ, Currie D, Cherney B, Kirby M, Fajardo G, Caudill M, Langolf K, Kahrs J, Kelly P, Pitts C, Lim A, Aulik N, Tamin A, Harcourt JL, Queen K, Zhang J, Whitaker B, Browne H, Medrzycki M, Shewmaker P, Folster J, Bankamp B, Bowen MD, Thornburg NJ, Goffard K, Limbago B, Bateman A, Tate JE, Gieryn D, Kirking HL, Westergaard R, Killerby M, CDC COVID-19 Surge Laboratory Group. 2021. Performance of an Antigen-Based Test for Asymptomatic and Symptomatic SARS-CoV-2 Testing at Two University Campuses - Wisconsin, September-October 2020. *MMWR Morb Mortal Wkly Rep*. 2021 Jan 1 69(5152):1642-1647.

6. Grifoni A, Sidney J, Zhang Y, Scheuermann RH, Peters B, Sette A. 2020. A Sequence Homology and Bioinformatic Approach Can Predict Candidate Targets for Immune Responses to SARS-CoV-2. *Cell Host Microbe* 27(4):671-680.e2.
7. The World Health Organization. Classification of Omicron (B.1.1.529): SARS-CoV-2 Variant of Concern. Available at: [https://www.who.int/news/item/26-11-2021-classification-of-omicron-\(b.1.1.529\)-sars-cov-2-variant-of-concern](https://www.who.int/news/item/26-11-2021-classification-of-omicron-(b.1.1.529)-sars-cov-2-variant-of-concern). Accessed 21 January 2022.
8. The Centers for Disease Control and Prevention. Monitoring Variant Proportions. Available at: <https://covid.cdc.gov/covid-data-tracker/#variant-proportions>. Accessed 21 January 2022.
9. Nextstrain. Genomic epidemiology of novel coronavirus - Global subsampling. Available at: <https://nextstrain.org/ncov/gisaid/global>. Accessed 21 January 2022.
10. SARS-CoV-2 Antigen (Ag) Product Insert – LumiraDx. Available at: <https://www.lumiradx.com/assets/pdfs/covid-19-antigen-test/sars-cov-2-antigen-product-insert/sars-cov-2-ag-test-strip-product-insert-eua.pdf?v=6>. Accessed 21 January 2022.
11. BinaxNOW COVID-19 Ag Card – Instructions for Use. Available at: <https://www.fda.gov/media/141570/download>. Accessed 21 January 2022.
12. CareStart COVID-19 Antigen test – Instructions for Use. Available at: <https://www.fda.gov/media/142919/download>. Accessed 21 January 2022.
13. Harcourt J, Tamin A, Lu X, Kamili S, Sakthivel SK, Murray J, Queen K, Tao Y, Paden CR, Zhang J, Li Y, Uehara A, Wang H, Goldsmith C, Bullock HA, Wang L, Whitaker B, Lynch B, Gautam R, Schindewolf C, Lokugamage KG, Scharton D, Plante JA, Mirchandani D, Widen SG, Narayanan K, Makino S, Ksiazek TG, Plante KS, Weaver SC,

- Lindstrom S, Tong S, Menachery VD, Thornburg NJ. 2020. Severe Acute Respiratory Syndrome Coronavirus 2 from Patient with Coronavirus Disease, United States. *Emerg Infect Dis* 26(6):1266-1273.
14. Kirby JE, Riedel S, Dutta S, Arnaout R, Cheng A, Ditelberg S, Donald J. Hamel, Charlotte A. Chang, Phyllis J. Kanki. 2021. SARS-CoV-2 Antigen Tests Predict Infectivity Based on Viral Culture: Comparison of Antigen, PCR Viral Load, and Viral Culture Testing on a Large Sample Cohort. *medRxiv* 2021 <https://doi.org/10.1101/2021.12.22.21268274>.
15. Regan J, Flynn JP, Choudhary MC, Uddin R, Lemieux J, Boucau J, Bhattacharyya RP, Barczak AK, Li JZ , Siedner MJ. 2021. Detection of the omicron variant virus with the Abbott BinaxNow SARS-CoV-2 Rapid Antigen Assay. *medRxiv* <https://doi.org/10.1101/2021.12.22.21268219>.
16. Bekliz M, Perez-Rodriguez F, Puhach O, Adea K, Melancia SM, Baggio S, Corvaglia AR, Jacquerioz-Bausch F, Alvarez C, Essaidi-Laziosi M, Escadafal C, Kaiser L, Eckerle I. 2022. Sensitivity of SARS-CoV-2 antigen-detecting rapid tests for Omicron variant. *medRxiv* <https://doi.org/10.1101/2021.12.18.21268018>.
17. Deerrain J, Druce J, Tran T, Batty M, Yoga Y, Fennell M, Dwyer DE, Kok J, Williamson DA. 2021. Assessment of the analytical sensitivity of ten lateral flow devices against the SARS-CoV-2 omicron variant. *J Clin Microbiol* jcm-02479.