

Suppl. 02

Title:

The use of a new type of training system "HeartGo[®]", which is controlled by heart rate, for patients with chronic Heart failure on the pedelec.

Authors:

Erik B. Friedrich², Herbert Löllgen³, Helmut Röder⁴, Wolfgang Baltes⁵, Oliver Adam⁶, Martin Schlickel⁷, Günter Hennersdorf¹

2. methodology

2.1 Objectives and significance of the study:

Research question: Is a heart rate based training system using pedelecs in heart group participants with heart failure suitable to demonstrate effects on safety of the training process, acceptance of the system, clinical and prognostic data and training?

In a prospective pilot study with 10 patients from Saarland heart groups diagnosed with heart failure NYHA II to III, the novel system HeartGo[®], which was tested for practicability by preliminary examinations (1), was used. Not only parameters of the training process were measured, but also climatic parameters, which were intended to provide information about the course of health and thus about the prognosis of the disease.

2.2 Statistical methodology

To statistically validate the results, we used the Student's t-test for paired samples and applied a probability of error of 5%, (p-value of \leq or $>$ 0.05).

The value groups were compared at the beginning, in some cases halfway through the term, i.e. after one year, and at the end of the term. The study design is therefore prospective and comparative without control group. It is therefore neither randomized nor blinded.

2.3. study design:

10 volunteers with chronic heart failure were recruited with the following requirements (Table 1):

1. duration 2 years
2. 6 months of belonging to a heart group,
3. stable clinical condition,
4. stable medication (see Table 4)
5. NYHA II - III
- 6th LVEF \leq 50% (HfmrEF)
7. \approx 0.5 watts/kg power
8. training 1/week 60 - 150 min
9. training heart rate 60-70% of maximum heart rate
10. summer and winter mode
11. distances 2x10 to 2x25 (20 - 50) km

The test persons were provided with the bikes in fully functional condition. The presented study is a pilot study, as the strictly defined selection locally hardly allowed for higher numbers of test persons and was carried out with new, previously untested technology. The study was to be followed by further, preferably multi-center studies for verification purposes.

2.4 Training parameters:

Training duration, training distance, frequency behaviour, BORG queries, relation pedalling power to motor power, training blood pressure (by means of commercially available, calibrated wrist device, OMRON).

The Borg scale (2) was used as an orientation scale between 6 and 20 for the sensation of effort.

The training phases during the period of use were divided into two winter phases (indoor training) and two summer phases (outdoor training).

During the four phases (winter 2017/2018, summer 2018, winter 2018/2019, summer 2019) the power and target frequencies were adjusted. It started with 60%, from half time on 70% of the maximum frequency HFmax or the calculated training frequency was set. At a roughly constant speed of 20 km/h, the distances were increased and thus the training duration increased.

The four phases were structured as follows (Table 2):

Phase	Speed km/h	Distance km	Duration min
I	20	20 (2x10)	60
II	20	30 (2x15)	90
III	20	40 (2x20)	120
IV	≥20	50 (2x25)	≈150

The chest belt sensors continuously recorded heart rate (HF) and ECG signal "real time" (standard). For this purpose, the test person was given a target heart rate, which usually corresponded to a previously determined training heart rate (3).

The app offers the possibility of selecting four driving modes:

1. free ride (free selection of assistance via control element)
2. tour (support automatically determined by the system by terrain rise or fall: (GPS altitude))
3. outdoor training (summer mode)
4. stationary ergometer training (winter mode)

Modes 3 and 4 were used for the study. The winter training took place in a hall with jacked-up wheels.

The modes have the control algorithms of the RF-assist coupling. The reference parameter for the control was the pre-set target frequency, which has to be maintained by a combination of pedalling and motor power.

The setting of the target frequency was based on the respective maximum heart rate (HF-max) of the test person, which is either determined ergometrically or calculated according to the formula method of Tanaka (3). It was usually between 60 (at the beginning) and 70% of the HF-max (middle and end).

However, the target frequency had to be adjusted due to individual tolerance (e.g. beta-blocker medication) in order to avoid overloading the pedal power. This resulted in an adjustment of about 10% below the calculated target frequency (IF).

The control process causes the heart rate to fluctuate around the target rate depending on the load and the pedal assistance to increase as the HR increases. Correspondingly, the pedal power increases out of phase when the HF drops below the IF. This is the typical example of an intact control system.

2.5 Clinical parameters:

At the beginning (start), halfway through the study (middle) and at the end of the study (end), 3 separate clinical examinations are performed:

ECG, blood pressure, bicycle ergometry

Echocardiography

Long term ECG

Lab data:

Blood count, electrolytes, kidney values, CRP, HbA1c, NT-Pro-BNP, cholesterol, LDL-, HDL-cholesterol, triglycerides.

The pedelec was adapted by HeartGo[®] with deep entry, 250 watt rear wheel drive and a battery (450 Wh). The motor allows the activation of up to 10 levels of assistance and two levels of recuperation (power recovery).

2.6 Electronic equipment:

Pedal power and assistance monitor (Swiss Go Drive[®]) were connected to an Android app developed by HeartGo[®] on a standard smartphone. The units are linked to the subject's bicycle and heart rate via Bluetooth technology. The test person wears a chest strap sensor.

2.7 Data management

The following data were processed:

Real-time ECG via sensor,

Pedaling power (watts), engine power (watts), speed, (km/h)

Terrain profile, support levels (10)

The data was available to the test person on the smartphone display during the test. After completion of the training, they were transferred to the secure portal of the manufacturer and were then available for evaluation and information to the trainer admin or the trainer and also to the individual test persons.

2.8. acceptance questionnaire:

We used a questionnaire system (Well-being or Wellness-ess index) which was introduced and tested by Kolip (4). It is divided into three positive (answers are correct) and three negative indicators (answers are not correct) and was asked at the end of the study about the state of health before and after the end of the study (Table 6; Suppl.4)

2.9 Ethics Committee: The study was accepted by the Ethics Committee of the Medical Association of the Saarland on April 18, 2018:

"With reference to §2 of the statute of the Ethics Commission at the Medical Association of the Saarland, there are no objections to the implementation of the intended research project".

References:

1. Emrich, A, Leonhardt, F, Werth, D, Loos,
Proactive Assistance for Device-Integrated Training Adjustment in
Outdoor Rehabilitation Training
Extended Abstract im Rahmen der TAR 2015 /Technically Assisted
Rehabilitation)
2. BORG G,
Psychophysical bases of perceived exertion.
Medicine and Science in Sports and Exercise; 1982,14 (5), 377–381
3. Tanaka, H, Monahan K, Seals, DR;
Age-predicted maximal heart rate revisited.
JACC 2001, 37, 153–156
4. Kolip, P., Schmidt, B.
Fragebogen zur Erfassung körperlichen Wohlbefindens (FEW 16)
Zeitschrift für Gesundheitspsychologie, 1999, 7, 77-87